

New-Indy Oxnard, LLC  
ROSEMOUNT CEMS SOURCE TEST - March 15, 2021

3/15/2021 1:14 1.4

*data*

Stack NOx	Stack CO	Stack O2	Duct Burner Gas Flow	Turbine Gas Flow	Steam Injection	Ammonia Injection	Turbine Load
15-Mar-21 12:45:00	13.56	15.28	0.05	250.98	1.00	24.29	21.68
15-Mar-21 12:46:00	13.56	15.26	0.05	250.98	1.00	24.31	21.62
15-Mar-21 12:47:00	13.56	15.26	0.05	250.98	1.00	24.33	21.63
15-Mar-21 12:48:00	13.56	15.26	0.05	250.98	1.00	24.34	21.68
15-Mar-21 12:49:00	13.56	15.26	0.05	250.98	1.00	24.38	21.68
15-Mar-21 12:50:00	13.56	15.26	0.05	250.98	1.00	24.37	21.68
15-Mar-21 12:51:00	13.56	15.26	0.05	250.98	1.00	24.40	21.57
15-Mar-21 12:52:00	13.56	15.26	0.05	250.98	1.00	24.30	21.64
15-Mar-21 12:53:00	13.56	15.26	0.05	250.98	1.00	24.30	21.71
15-Mar-21 12:54:00	13.56	15.26	0.05	250.98	1.00	24.35	21.68
15-Mar-21 12:55:00	13.56	15.26	0.05	250.98	1.00	24.45	21.63
15-Mar-21 12:56:00	13.56	15.26	0.05	250.98	1.00	24.35	21.70
15-Mar-21 12:57:00	13.56	15.26	0.05	250.98	1.00	24.42	21.73
15-Mar-21 12:58:00	13.56	15.26	0.05	250.98	1.00	24.55	21.69
15-Mar-21 12:59:00	13.56	15.26	0.05	250.98	1.00	24.62	21.73
15-Mar-21 13:00:00	13.56	15.26	0.05	250.98	1.00	24.63	21.66
15-Mar-21 13:01:00	13.56	15.26	0.05	250.98	1.00	24.64	21.63
15-Mar-21 13:02:00	13.56	15.26	0.05	250.98	1.00	24.59	21.56
15-Mar-21 13:03:00	13.56	15.26	0.05	250.98	1.00	24.59	21.56
15-Mar-21 13:04:00	13.56	15.26	0.05	250.98	1.00	24.54	21.56
15-Mar-21 13:05:00	13.56	15.26	0.05	250.98	1.00	24.56	21.81
15-Mar-21 13:06:00	13.56	15.26	0.05	250.98	1.00	24.59	21.50
15-Mar-21 13:07:00	13.56	15.26	0.05	250.98	1.00	24.53	21.56
15-Mar-21 13:08:00	13.56	15.26	0.05	250.98	1.00	24.60	21.81
15-Mar-21 13:09:00	13.56	15.26	0.05	250.98	1.00	24.51	21.62
15-Mar-21 13:10:00	13.56	15.26	0.05	250.98	1.00	24.57	21.62
15-Mar-21 13:11:00	13.56	15.26	0.05	250.98	1.00	24.57	21.73
15-Mar-21 13:12:00	13.56	15.26	0.05	250.98	1.00	24.51	21.61
15-Mar-21 13:13:00	13.56	15.26	0.05	250.98	1.00	24.53	21.66
15-Mar-21 13:14:00	13.56	15.26	0.05	250.98	1.00	24.56	21.61
15-Mar-21 13:15:00	13.56	15.26	0.05	250.98	1.00	24.52	21.59
15-Mar-21 13:16:00	13.56	15.26	0.05	250.98	1.00	24.54	21.55
15-Mar-21 13:17:00	13.56	15.26	0.05	250.98	1.00	24.54	21.56
<b>Average Stack NOx, ppm (15% O2)</b>	<b>13.56</b>	<b>15.26</b>	<b>0.05</b>	<b>250.98</b>	<b>1.00</b>	<b>24.48</b>	<b>23.63</b>
<b>Average Stack CO, ppm (15% O2)</b>	<b>13.56</b>	<b>15.26</b>	<b>0.05</b>	<b>250.98</b>	<b>1.00</b>	<b>24.48</b>	<b>23.63</b>

*AVKlogu*  
7-8-9  
7-8-9  
7-8-9

*AVKlogu*  
7-8-9  
7-8-9  
7-8-9

*AVKlogu*  
7-8-9  
7-8-9  
7-8-9

*AVKlogu*  
7-8-9  
7-8-9  
7-8-9

*AVKlogu*  
7-8-9  
7-8-9  
7-8-9

*AVKlogu*  
7-8-9  
7-8-9  
7-8-9

*AVKlogu*  
7-8-9  
7-8-9  
7-8-9

*AVKlogu*  
7-8-9  
7-8-9  
7-8-9

*AVKlogu*  
7-8-9  
7-8-9  
7-8-9

*Average runs 7-8-9*  
*4192.74*  
*dsctm*

*251.03 x 1000 / 60 = 4183.8*  
*dsctm*

New-Indy Onnaird, LLC  
ROSEMOUNT GEMS SOURCE TEST - March 15, 2021

3/15/2021 13:18 3/15/2021 13:18 1.4

*Notes*

Stack NOx	Stack CO	Stack O2	Duct Burner Gas Flow	Turbine Gas Flow	Steam Injection	Ammonia Injection	Turbine Load
15-Mar-21 13:18:00	13.56	15.26	0.05	250.98	1.91	24.47	23.60
15-Mar-21 13:19:00	13.56	15.26	0.05	250.55	1.91	24.47	23.60
15-Mar-21 13:20:00	13.56	15.26	0.05	250.98	1.91	24.47	23.60
15-Mar-21 13:21:00	13.56	15.26	0.05	250.98	1.91	24.47	23.60
15-Mar-21 13:22:00	13.56	15.26	0.05	250.98	1.91	24.47	23.60
15-Mar-21 13:23:00	13.56	15.26	0.05	250.98	1.91	24.47	23.60
15-Mar-21 13:24:00	13.56	15.26	0.05	250.98	1.91	24.47	23.60
15-Mar-21 13:25:00	13.56	15.26	0.05	250.98	1.91	24.47	23.60
15-Mar-21 13:26:00	13.56	15.26	0.05	250.98	1.91	24.47	23.60
15-Mar-21 13:27:00	13.56	15.26	0.05	250.98	1.91	24.47	23.60
15-Mar-21 13:28:00	13.56	15.26	0.05	250.98	1.91	24.47	23.60
15-Mar-21 13:29:00	13.56	15.26	0.05	250.98	1.91	24.47	23.60
15-Mar-21 13:30:00	13.56	15.26	0.05	250.98	1.91	24.47	23.60
15-Mar-21 13:31:00	13.56	15.26	0.05	250.98	1.91	24.47	23.60
15-Mar-21 13:32:00	13.56	15.26	0.05	250.98	1.91	24.47	23.60
15-Mar-21 13:33:00	13.56	15.26	0.05	250.98	1.91	24.47	23.60
15-Mar-21 13:34:00	13.56	15.26	0.05	250.98	1.91	24.47	23.60
15-Mar-21 13:35:00	13.56	15.26	0.05	250.98	1.91	24.47	23.60
15-Mar-21 13:36:00	13.56	15.26	0.05	250.98	1.91	24.47	23.60
15-Mar-21 13:37:00	13.56	15.26	0.05	250.98	1.91	24.47	23.60
15-Mar-21 13:38:00	13.56	15.26	0.05	250.98	1.91	24.47	23.60
15-Mar-21 13:39:00	13.56	15.26	0.05	250.98	1.91	24.47	23.60
15-Mar-21 13:40:00	13.56	15.26	0.05	250.98	1.91	24.47	23.60
15-Mar-21 13:41:00	13.56	15.26	0.05	250.98	1.91	24.47	23.60
15-Mar-21 13:42:00	13.56	15.26	0.05	250.98	1.91	24.47	23.60
15-Mar-21 13:43:00	13.56	15.26	0.05	250.98	1.91	24.47	23.60
15-Mar-21 13:44:00	13.56	15.26	0.05	250.98	1.91	24.47	23.60
15-Mar-21 13:45:00	13.56	15.26	0.05	250.98	1.91	24.47	23.60
15-Mar-21 13:46:00	13.56	15.26	0.05	250.98	1.91	24.47	23.60
15-Mar-21 13:47:00	13.56	15.26	0.05	250.98	1.91	24.47	23.60
15-Mar-21 13:48:00	13.56	15.26	0.05	250.98	1.91	24.47	23.60
15-Mar-21 13:49:00	13.56	15.26	0.05	250.98	1.91	24.47	23.60
15-Mar-21 13:50:00	13.56	15.26	0.05	250.98	1.91	24.47	23.60
<b>Average Stack NOx, ppm (15% O2)</b>	<b>13.56</b>	<b>15.26</b>	<b>0.05</b>	<b>250.98</b>	<b>1.91</b>	<b>24.47</b>	<b>23.60</b>
<b>Average Stack CO, ppm (15% O2)</b>	<b>13.56</b>	<b>15.26</b>	<b>0.05</b>	<b>250.98</b>	<b>1.91</b>	<b>24.47</b>	<b>23.60</b>
<b>Average Duct Burner Gas Flow (MSCFH)</b>	<b>0.05</b>	<b>15.26</b>	<b>0.05</b>	<b>250.98</b>	<b>1.91</b>	<b>24.47</b>	<b>23.60</b>
<b>Average Turbine Gas Flow (MSCFH)</b>	<b>250.98</b>	<b>15.26</b>	<b>0.05</b>	<b>250.98</b>	<b>1.91</b>	<b>24.47</b>	<b>23.60</b>
<b>Average Steam Injection Rate (lb/hr)</b>	<b>1.91</b>	<b>15.26</b>	<b>0.05</b>	<b>250.98</b>	<b>1.91</b>	<b>24.47</b>	<b>23.60</b>
<b>Average Ammonia Injection (lb/hr)</b>	<b>24.47</b>	<b>15.26</b>	<b>0.05</b>	<b>250.98</b>	<b>1.91</b>	<b>24.47</b>	<b>23.60</b>
<b>Ave. Turbine Load (MWh)</b>	<b>23.54</b>	<b>15.26</b>	<b>0.05</b>	<b>250.98</b>	<b>1.91</b>	<b>24.47</b>	<b>23.60</b>

250.72 x 1000 / 60 = 4178.4  
dgsfm

Rate 9

New-Indy Onnard, LLC  
ROSEMOUNT CEMS SOURCE TEST - March 15, 2021

3/15/2021 13:11 3/15/2021 13:11

Stack NOx	Stack CO	Stack O2	Duct Burner Gas Flow	Turbine Gas Flow	Stream Injection	Ammonia Injection	Turbine Load
10.69	13.58	15.26	0.05	250.89	1.90	24.49	23.29
10.73	13.56	15.26	0.05	250.96	1.90	24.32	23.48
10.63	13.56	15.26	0.05	250.82	1.91	24.48	23.55
10.58	13.56	15.26	0.05	248.33	1.90	24.45	23.56
10.78	13.56	15.26	0.05	248.91	1.90	24.52	23.68
10.70	13.56	15.26	0.05	250.70	1.91	24.62	23.59
10.81	13.56	15.26	0.05	250.88	1.91	24.64	23.59
10.67	13.56	15.26	0.05	250.88	1.81	24.79	23.59
10.73	13.56	15.26	0.05	250.88	1.81	24.79	23.59
10.65	13.56	15.26	0.05	250.98	1.80	24.60	23.66
10.56	13.56	15.26	0.05	250.62	1.91	24.60	23.62
10.48	13.76	15.26	0.05	248.33	1.91	24.43	23.41
10.44	14.45	15.23	0.05	246.14	1.91	24.32	23.46
10.46	15.27	15.20	0.05	246.22	1.90	24.28	23.44
10.44	18.18	15.17	0.31	248.51	1.91	24.28	23.44
10.46	17.09	15.14	2.00	250.70	1.91	24.30	23.45
10.48	21.00	15.06	3.01	250.96	1.92	24.36	23.45
10.45	19.36	15.05	5.23	250.82	1.89	24.36	23.45
10.48	18.10	15.06	6.48	250.82	1.92	24.36	23.45
10.69	17.53	15.06	6.76	249.33	1.91	24.27	23.53
10.82	17.11	15.06	7.01	248.51	1.91	24.29	23.53
10.69	17.07	15.06	6.75	250.88	1.90	24.29	23.53
10.69	17.07	15.06	6.48	250.88	1.90	24.29	23.53
10.74	17.07	15.06	6.48	250.82	1.92	24.27	23.53
10.59	17.07	15.06	6.76	249.33	1.91	24.27	23.53
10.57	17.07	15.06	7.01	248.51	1.91	24.29	23.53
10.55	17.07	15.06	6.75	250.88	1.90	24.29	23.53
10.55	17.07	15.06	6.48	250.88	1.90	24.29	23.53
10.60	17.07	15.06	6.75	250.88	1.90	24.29	23.53
10.64	17.07	15.06	7.48	250.88	1.89	24.29	23.53
10.55	17.07	15.06	6.92	250.96	1.90	24.29	23.53
Average Stack NOx, ppm (15% O2)	Average Stack CO, ppm (15% O2)	Average Stack O2 (%)	Average Duct Burner Gas Flow (MSCFH)	Average Turbine Gas Flow (MSCFH)	Average Stream Injection Rate (lb/hr)	Average Ammonia Injection (lb/hr)	Ave. Turbine Load (MWh)
10.62	15.77	15.15	3.20	249.74	1.91	24.46	23.57

257.94 x 1000 / 60 = 4215.76  
desfm

EPA METHOD 20 MEASUREMENT  
SYSTEM PERFORMANCE TIME

RESPONSE TIME

DATE OF TEST: 3-15-21

PROBE LOCATION: Stack

ANALYZER TYPE & MODEL #

NOx: API 200 EM

CO: API 300 EM

O2: SERVOMEX 1400

Span Gas Concentration/Analyzer Full Scale Setting:

NOx; Gas, ppmv: 20.2

Full Scale setting, ppmv 25

CO; Gas, ppmv: 40.0

Full Scale setting, ppmv 50

O2; Gas, %: 20.13

Full Scale setting, % 25

UPSCALE TIME - Analyzer + Sampling System:

Trial #	Nox	CO	O2
1	<u>58</u>	<u>55</u>	<u>47</u> seconds
2	<u>59</u>	<u>55</u>	<u>47</u> seconds
3	<u>58</u>	<u>54</u>	<u>47</u> seconds
Average Response	<u>59</u>	<u>55</u>	<u>47</u> seconds

DOWNSCALE - Analyzer + Sampling System:

Trial #	Nox	CO	O2
1	<u>57</u>	<u>53</u>	<u>45</u> seconds
2	<u>58</u>	<u>54</u>	<u>45</u> seconds
3	<u>58</u>	<u>53</u>	<u>45</u> seconds
Average Response	<u>58</u>	<u>54</u>	<u>45</u> seconds

Slower Average 59 55 47 seconds

<b>DESCRIPTION OF METHOD(S) USED</b>
--------------------------------------


## CARB Method 100 – Sample Train Assembly

### Probe:

1. AIRx Testing Services, Inc. uses a stainless steel sampling probe.
2. The probe is connected to the sample conditioner using a heated (if necessary) Teflon sampling line.

### Sample Conditioner (Condenser System):

1. The sample conditioner consists of a moisture knock-out bottle immersed in an ice bath.
2. All parts of the conditioner exposed to the sample are either glass, stainless steel, or Teflon.
3. The sample conditioner is setup so that the sample gas is not bubbled through the condensate.
4. A temperature gauge is used to determine the temperature of the condenser outlet.
5. Capable of reducing sample gas temperature to 15 °C (60 °F).



### Manifold Panel & Strip Chart Recorder:

1. The sample gas exiting the condenser is transported to the instrument manifold panel.
2. The sample manifold directs the sample gas through to the various analyzers.
3. All components of the system that contact the sample are either stainless steel, Teflon, or glass.
4. The data is continuously recorded on a Sollec 10" strip chart recorder with a 1% resolution.

### Calibration System:

1. The system is composed of the Analyzer Calibration and the System BIAS Check.
2. Certified Calibration gases are supplied to the analyzers for calibration purposes.
3. The selection of zero or span calibration gases or sample gas is accomplished by operation of selector knobs located on the main flow control panel.

**Nitrogen Oxides:** – Chemiluminescent Analyzer: Based on the chemiluminescent reaction of NO and ozone to form NO<sub>2</sub> in an excited state. Light emission is monitored through an optical filter by a high sensitivity photomultiplier tube, the output of which is electronically processed so it is linearly proportional to the NO concentration. The output is in units of ppmv.

**Carbon Monoxide** – Non-Dispersive Infrared (NDIR) Analyzer: Radiation from an infrared source is passed through a narrow band-pass filter and a multiple optical pass sample cell where absorption by the sample gas occurs. The infrared radiation exits the sample cell and falls on a solid state infrared detector. The output is in units of ppmv.

**Oxygen** – Electrochemical Analyzer: Oxygen in the flow gas sample diffuses through a Teflon membrane and is reduced on the surface of the cathode. A corresponding oxidation occurs at the anode and an electric current proportional to the concentration of oxygen is produced. The output is in units of percent O<sub>2</sub> by volume.

**Carbon Dioxide** – Non Dispersive Infrared (NDIR) Analyzer: The instrument measures the differential in infrared energy absorbed from energy beams passed through a reference cell (containing a gas selected to have minimal absorption of infrared energy in the wavelength absorbed by CO<sub>2</sub>) and a sample cell through which the sample gas flows continuously. The output is in units of percent CO<sub>2</sub> by volume.

# EPA Method 20

## NO<sub>x</sub>, SO<sub>2</sub>, and Diluent (O<sub>2</sub> or CO<sub>2</sub>) Emissions from Stationary Gas Turbines

**Principle:** A sample of an exhaust gas stream is continuously extracted, conditioned, and conveyed to instrumental analyzers for the determination of NO<sub>x</sub> and diluent content. During each NO<sub>x</sub> and diluent determination, a separate measurement of SO<sub>2</sub> emissions is made, using Method 6 or its equivalent. The diluent determination is used to adjust the NO<sub>x</sub> and SO<sub>2</sub> concentration to a reference condition.

### Measurement System: Performance Test Procedures

**Calibration check:** The difference between the known concentration of the CalGas and the gas concentration exhibited by the gas analyzer when the CalGas is introduced directly to the analyzer.

- 1) Calibrate the analyzer with zero and mid-level (45-55% of range) CalGas
- 2) Introduce low-level (20-30% of range) and high-level (80-90% of range) CalGas

**Performance Spec:** The linear curve determined from the zero and mid-level CalGas responses must predict the actual response of the low-level (NO<sub>x</sub> only) and high-level (NO<sub>x</sub> and O<sub>2</sub>/CO<sub>2</sub>) gases within 2% of the range.

**Sampling system BIAS:** The difference between the gas concentrations exhibited by the measurement system when CalGas is introduced at the sampling probe tip filter and when the same CalGas is introduced directly to the analyzer.

**Performance Spec:** less than 5% of the range for the zero, mid-range, and high-range CalGas.

**Zero Drift:** The difference in the measurement system responses at a zero concentration level during the initial calibration, and final calibration check after a test. No adjustment to the measurement system is allowed at that point.

**Performance Spec:** less than 2% of the range

**Calibration Drift:** The difference in the measurement system responses at a mid-range concentration level during the initial calibration, and final calibration check after a test. No adjustment is allowed at that point.

**Performance Spec:** less than 2% of the range

**Response time:** Time required by data recorder to display 95% of a step change in gas concentration.

- 1) Introduce zero gas into the system until all readings are stable.
- 2) Switch to monitor stack effluent until a stable reading can be obtained. Record upscale response time.
- 3) Introduce high-level CalGas into the system until all readings are stable.
- 4) Switch to monitor stack effluent until a stable reading can be obtained. Record downscale response time.
- 5) Repeat the previous 4 steps three times

**Performance Spec:** No greater than 30 seconds (or change of < 5% of the measured average concentration for 2 min)

**Interference response:** The output response of the measurement system to a component in the sample gas, other than the gas component being measured. Introduce the following gases, 500±50 ppm CO, 200±20 ppm SO<sub>2</sub>, 10±1 ppm CO<sub>2</sub>, 20.9±1 ppm O<sub>2</sub>, and determine the total interference output response of the system to these components. Conduct the interference response check on each analyzer prior to its first use in the field.

**Performance Spec:** less than 2% of the range

### Emission Measurement Test Procedures

- Select sampling site as follows:
  1. Upstream of the point of introduction of dilution air into the duct;
  2. At least 2 stack diameters (or 5 feet whichever is less) upstream of stack outlet.

- Conduct preliminary O<sub>2</sub>/CO<sub>2</sub> traverse to determine the sampling points of lowest O<sub>2</sub> or highest CO<sub>2</sub> concentration. Conduct this test at the turbine's lowest operating load mentioned in the test protocol.
  - Select minimum # of points:
    1. 8 for stacks with cross-sectional areas < 1.5 m<sup>2</sup> (16.1 ft<sup>2</sup>)
    2. 8 + 1 additional sample point per each 0.2m<sup>2</sup> (2.2ft<sup>2</sup>) of areas for stacks whose areas 16.1ft<sup>2</sup> < A < 107.6ft<sup>2</sup>
    3. 49 (48 if circular stacks) for stacks whose cross-sectional areas are > 10.0 m<sup>2</sup> (107.6 ft<sup>2</sup>)
    4. Use Method 1 to locate the traverse points
  - Conduct preliminary diluent measurement while the turbine is operating at the lowest percent of peak load:
    1. Position probe at first traverse point and begin sampling for at least 1 minute + response time
    2. Do the same for each traverse point, recording the diluent concentration each time
    3. Select the 8 points at which the lowest O<sub>2</sub> (or highest CO<sub>2</sub>) concentrations were obtained
    4. Sample at each of these selected points during each run at the different turbine load conditions
  - **Three test runs at each load condition specified in the test protocol constitutes a complete test**
- Cleaning of sample train: Flush probe, lines and sample conditioner with DW, than acetone. Dry with filtered dry air.
  - Allow Continuous Analyzers to warm up
  - Sampling system preparation: assemble sample train as shown in Fig 100.1-1, 100.1-2 and 100.1-3. Leak check the vacuum side of the assembly to a minimum of 20" of Hg (gauge). The sampling system should hold 20" of Hg vacuum for 5 minutes with less than 1" Hg loss. Check the pressure side of the system with liquid soap solution and correct any leaks.
  - Allowable modifications: probe heating element can be eliminated if stack is at or below ambient temperature and condensation is not observed. Pitot tube can be eliminated if only concentration measurements are required.
  - Calibrate analyzers and data recorders: introduce CalGases directly to the instruments and make all necessary adjustments to calibrate the analyzer and data recorder. Adjust system components to achieve manufacturer's recommended sampling rates.
  - Instrument response time: establish during semi-annual certification
  - Sampling system BIAS check: Mandatory
    1. Backflash gas through the probe as necessary to prevent particulate buildup
    2. Zero, and either mid-range or high-range (whichever is closest to effluent concentration)
    3. Introduce upscale CalGas and record concentration on a form similar to 100.1-5
    4. Introduce zero CalGas and record concentration
    5. Make no adjustments to the system except those necessary to achieve the correct flow rate.
    6. If invalid calibration is exhibited (> 5% of the range), take corrective action and repeat check.
    7. If adjustment to the analyzer is required, first repeat the analyzer calibration error check, then repeat the sampling system BIAS check.
  - NO<sub>2</sub> to NO conversion check: EPA Method 20 or gas mixture of NO<sub>2</sub> in air



### Emission Test Procedure:

- Traverse stack to determine presence of stratification
- Single-point gas sampling is acceptable if gas composition is homogenous (<10% variation)
- Determine moisture content and velocity pressures or Mass flow rate may be obtained by stoichiometric and gas composition relations
- Chart recorder label: turn on strip chart recorder and label the chart as to pollutant, source, range, calibration cylinder ID number, certified expiration date, zero and upper range calibration settings, chart speeds, date, time, operator.
- Sample probe traverse and minimum sampling time:
  1. Insert probe in stack
  2. determine if single point sampling is acceptable
  3. if traverse is required, leave the probe at each traverse point for at least the system response time + 1 minute.
  4. minimum sampling time of 60 minute is recommended. See District Rules and Regulations and permit conditions for special requirements.
  5. When test duration exceeds one hour, conduct zero and span checks every 2 hours. Adjust settings as necessary, mark strip charts and record in log books.
- Zero and Calibration DRIFT Tests:
  1. Immediately preceding and following each run, or if adjustments are necessary during the run, repeat the sampling system BIAS procedure. Make no adjustments to the system until after the DRIFT checks are completed. Record the information on a form similar to Figure 100.1-5.
  2. If run is invalid (sampling system BIAS specs exceeded), repeat entire procedure before repeating run.
  3. If both the zero and upscale calibration values are within the sampling system BIAS specs, then use the average of the initial and final BIAS check values to calculate the gas concentration for the run.
  4. If the zero or upscale calibration DRIFT exceeds the DRIFT limits, repeat entire procedure before conducting additional runs.
- Post Run Leak Check:

<b>INSTRUMENT INFORMATION</b>
-------------------------------

## SOURCE EMISSION INSTRUMENTATION LIST

### OXIDES OF NITROGEN

<b>Unit No. - 1:</b> Manufacturer: API Model No.: 200 EH Serial No.: 233 Method: Chemiluminescence Range (ppmv) 0-5000	<b>Unit No. - 6:</b> Manufacturer: API Model No.: 200 A Serial No.: 1013 Method: Chemiluminescence Range (ppmv) 0-50
<b>Unit No. - 2:</b> Manufacturer: API Model No.: 200 EH Serial No.: 234 Method: Chemiluminescence Range (ppmv) 0-5000	<b>Unit No. - 7:</b> Manufacturer: Thermo Environmental (TECO) Model No.: 10AR Serial No.: 25559-221 Method: Chemiluminescence Range (ppmv) 0-2.5, 10, 25, 100, 250, 1000, 2500, 10000
<b>Unit No. - 3:</b> Manufacturer: API Model No.: 200 EH Serial No.: 109 Method: Chemiluminescence Range (ppmv) 0-5000	<b>Unit No. - 8:</b> Manufacturer: Thermo Environmental (TECO) Model No.: 10AR Serial No.: 38586-258 Method: Chemiluminescence Range (ppmv) 0-2.5, 10, 25, 100, 250, 1000, 2500, 10000
<b>Unit No. - 4:</b> Manufacturer: API Model No.: 200 EH Serial No.: 442 Method: Chemiluminescence Range (ppmv) 0-5000	
<b>Unit No. - 5:</b> Manufacturer: API Model No.: 200 EH Serial No.: 441 Method: Chemiluminescence Range (ppmv) 0-5000	



## SOURCE EMISSION INSTRUMENTATION LIST

### OXYGEN

<p><b><u>Unit No. - 5:</u></b>  <b>Manufacturer:</b> California Analytical Instruments  <b>Model No.:</b> 110P  <b>Serial No.:</b> T02034  <b>Method:</b> Paramagnetic  <b>Range (%):</b> 0-5, 10, 25</p>	<p><b><u>Unit No. - 13:</u></b>  <b>Manufacturer:</b> Servomex  <b>Model No.:</b> 1400  <b>Serial No.:</b> X1420/B707  <b>Method:</b> Paramagnetic  <b>Range (%):</b> 0-25</p>
<p><b><u>Unit No. - 7:</u></b>  <b>Manufacturer:</b> Teledyne  <b>Model No.:</b> 320-AX  <b>Serial No.:</b> 108742  <b>Method:</b> Electrochemical  <b>Range (%):</b> 0-5, 10, 25</p>	
<p><b><u>Unit No. - 9:</u></b>  <b>Manufacturer:</b> Servomex  <b>Model No.:</b> 1400  <b>Serial No.:</b> 01420/B701/730  <b>Method:</b> Paramagnetic  <b>Range (%):</b> 0-25, 100</p>	
<p><b><u>Unit No. - 10:</u></b>  <b>Manufacturer:</b> Servomex  <b>Model No.:</b> 1400  <b>Serial No.:</b> 01420/B308  <b>Method:</b> Paramagnetic  <b>Range (%):</b> 0-25</p>	
<p><b><u>Unit No. - 11:</u></b>  <b>Manufacturer:</b> Teledyne  <b>Model No.:</b> 320-A  <b>Serial No.:</b> 111211  <b>Method:</b> Electrochemical  <b>Range (%):</b> 0-5, 10, 25</p>	
<p><b><u>Unit No. - 12:</u></b>  <b>Manufacturer:</b> Servomex  <b>Model No.:</b> 1400  <b>Serial No.:</b> 01420/B7103  <b>Method:</b> Paramagnetic  <b>Range (%):</b> 0-25, 100</p>	



## SOURCE EMISSION INSTRUMENTATION LIST

### CARBON MONOXIDE


<b>Unit No. - 1:</b> Manufacturer: API Model No.: 300 EM Serial No.: 239 Method: NDIR/GFC Range (ppmv) 0-5, 10, 25, 50, 100, 250, 500, 1000, 2500, 5000	<b>Unit No. - 2:</b> Manufacturer: API Model No.: 300 EM Serial No.: 240 Method: NDIR/GFC Range (ppmv) 0-5, 10, 25, 50, 100, 250, 500, 1000, 2500, 5000
<b>Unit No. - 3:</b> Manufacturer: API Model No.: 300 EM Serial No.: 104 Method: NDIR/GFC Range (ppmv) 0-5, 10, 25, 50, 100, 250, 500, 1000, 2500, 5000	<b>Unit No - 5:</b> Manufacturer: Thermo Environmental (TECO) Model No.: 48H Serial No.: 25184-219 Method: NDIR/GFC Range (ppmv) 0-50, 100, 200, 500, 1000, 2000, 5000, 10000, 20000, 50000
<b>Unit No. - 6:</b> Manufacturer: Thermo Environmental (TECO) Model No.: 48H Serial No.: 29031-233 Method: NDIR/GFC Range (ppmv) 0-50, 100, 200, 500, 1000, 2000, 5000, 10000, 20000, 50000	<b>Unit No. - 7:</b> Manufacturer: Siemens Model No.: Ultramat 21p Serial No.: AO4-254 Method: NDIR Range (ppmv) 0-300
<b>Unit No. - 10:</b> Manufacturer: Thermo Environmental (TECO) Model No.: 48H Serial No.: 38391-257 Method: NDIR/GFC Range (ppmv) 0-50, 100, 200, 500, 1000, 2000, 5000, 10000, 20000, 50000	<b>Unit No. - 11:</b> Manufacturer: Thermo Environmental (TECO) Model No.: 48H Serial No.: 35226-249 Method: NDIR/GFC Range (ppmv) 0-50, 100, 200, 500, 1000, 2000, 5000, 10000, 20000, 50000
<b>Unit No - 13:</b> Manufacturer: Thermo Environmental (TECO) Model No.: 48 Serial No.: 48-15970-159 Method: NDIR/GFC Range (ppmv) 0-1, 2, 5, 10, 20, 50, 100, 200, 500, 1000	<b>Unit No - 14:</b> Manufacturer: Thermo Environmental (TECO) Model No.: 48 Serial No.: 48-23925-213 Method: NDIR/GFC Range (ppmv) 0-1, 2, 5, 10, 20, 50, 100, 200, 500, 1000

## BAAQMD Method ST-1B / Ammonia Integrated Sampling

**Principle:** A gas sample is drawn through a solution of 0.1N hydrochloric acid which absorbs the ammonia.

### Pre-Test Procedures:

- ❖ Add 100 ml of 0.1N to each of two impingers; retain 100 ml of the HCl to analyze as a blank
- ❖ Leak-test sampling train by starting the pump, plugging the probe, and adjusting the pump inlet vacuum to 10 inches Hg. The leak rate must not exceed 0.6 liter/min (0.02 CFM). Release plug, stop pump.
- ❖ Record initial "dry test meter" reading and barometric pressure.
- ❖ In the absence of stratification, sample at single point, otherwise select sampling traverse points.



### Sampling Procedure:

- ❖ Each test run shall be of **30 minute** duration or 90% of the batch time, whichever is less
- ❖ Position probe at the sampling point and start the pump
- ❖ Sample at constant rate of **14.3 liter/min (0.5 CFM)**, using rotameter to establish initial rate only
- ❖ Record "dry test meter" volume and temperature and "impinger outlet" temperature at 5-min intervals
- ❖ Maintain impinger temperature at 7°C (45°F) or less by adding ice as necessary
- ❖ At the conclusion of each run, stop pump, remove probe from stack and record final reading
- ❖ Point probe upward and purge sample train with ambient air

**Analysis:** Individually analyze the hydrochloric acid solutions and the blank for total ammonia content.

### Calculations:

Standard sample volume	Ammonia Concentration
$V_o = (17.71)V_m P_b / T_m$ <p><b>Where:</b>  <math>V_o</math> = Corrected sample volume to std conditions;  <math>T_m</math> = Average run meter temperature (°R);  <math>V_m</math> = Uncorrected meter volume (ft<sup>3</sup>);  <math>P_b</math> = Barometric pressure (inches Hg);                      17.71 = Constant correcting to 70°F &amp; 29.92 in.Hg.</p>	$C = 5.02 \times 10^4 W / V_o$ <p><b>Where:</b>  <math>C</math> = Ammonia concentration, ppmv on a dry basis;  <math>W</math> = Total weight of ammonia in the impinger catch, for each run, (g);  <math>5.02 \times 10^4</math> = Constant derived from the MW and correcting to standard conditions.</p>

**QUALITY ASSURANCE**



**CERTIFICATE OF ANALYSIS / EPA PROTOCOL GAS**

**Customer & Order Information**

PRAXAIR PKG OXNARD CA HPS  
455 E WOOLEY RD  
OXNARD CA 93030

Certificate Issuance Date: 08/09/2019  
Praxair Order Number: 71041947  
Part Number: NI NO12ME-AS  
Customer PO Number: 78028536

Fill Date: 07/24/2019  
Lot Number: 70085920504  
Cylinder Style & Outlet: AS CGA 650  
Cylinder Pressure and Volume: 2000 psig 140 ft<sup>3</sup>

**Certified Concentration**

Expiration Date:	08/09/2022	NIST Traceable
Cylinder Number:	DT0011049	Expanded Uncertainty
11.95 ppm	Nitric oxide	± 0.8 %
Balance	Nitrogen	

**ProSpec EZ Cert.**



**For Reference Only:** NOx 11.99 ppm

**Certification Information:** Certification Date: 08/09/2019 Term: 36 Months Expiration Date: 08/09/2022

The cylinder was certified according to the 2012 EPA Traceability Protocol, Document #EPA-600/R-12/531, using Procedure G1.  
Do Not Use this Standard if Pressure is less than 100 PSIG.

**Analytical Data:** (R=Reference Standard, Z=Zero Gas, C=Gas Candidate)

1. Component: Nitric oxide  
Requested Concentration: 12 ppm  
Certified Concentration: 11.95 ppm  
Instrument Used: Thermo Electron 42i-LS S/N 1030645077  
Analytical Method: Chemluminescence  
Last Multipoint Calibration: 07/12/2019

Reference Standard: Type / Cylinder #: GMS / CC457197  
Concentration / Uncertainty: 19.91 ppm ± 0.668%  
Expiration Date: 03/28/2022  
Traceable to: SRM # / Sample # / Cylinder #: APEX1181149 / N/A / APEX1181149  
SRM Concentration / Uncertainty: 20.03 ppm / ± 0.10 ppm  
SRM Expiration Date: 01/27/2020

First Analysis Data:				Date			
Z:	0	R:	19.91	C:	11.94	Conc:	11.94
R:	19.9	Z:	0	C:	11.93	Conc:	11.93
Z:	0	C:	11.94	R:	19.92	Conc:	11.94
UOM: ppm				Mean Test Assay: 11.94 ppm			

Second Analysis Data:				Date			
Z:	0	R:	19.01	C:	11.96	Conc:	11.96
R:	19.9	Z:	0	C:	11.98	Conc:	11.98
Z:	0	C:	11.97	R:	19.92	Conc:	11.97
UOM: ppm				Mean Test Assay: 11.97 ppm			

Analyzed By

Henry Koung

Certified By

Lecanna Flores



DocNumber: 253580



Praxair Distribution, Inc.  
 5700 S. Alameda Street  
 Los Angeles CA 90058  
 Tel: 323-585-2164  
 Fax: 714-542-6689  
 PGVP ID: F22019

**CERTIFICATE OF ANALYSIS / EPA PROTOCOL GAS**

**Customer & Order Information**

PRAXAIR PKG OXNARD CA HPS  
 455 E WOOLEY RD  
 OXNARD CA 93030

Certificate Issuance Date: 06/27/2019  
 Praxair Order Number: 70998288  
 Part Number: NI NO20ME-AS  
 Customer PO Number: 78881976

Fill Date: 06/07/2019  
 Lot Number: 70086915805  
 Cylinder Style & Outlet: AS CGA 660  
 Cylinder Pressure and Volume: 2000 psig 140 ltr

**Certified Concentration**

Expiration Date:	06/27/2022	NIST Traceable
Cylinder Number:	CC724456	Expanded Uncertainty
20.1 ppm	Nitric oxide	± 0.5 %
Balance	Nitrogen	

**ProSpec EZ Cert**



**For Reference Only:** NOx 20.2 ppm

**Certification Information:** Certification Date: 06/27/2019 Term: 36 Months Expiration Date: 06/27/2022

This cylinder was certified according to the 2012 EPA Traceability Protocol, Document #EPA-600/R-12/531, using Procedure G1.  
 Do Not Use this Standard if Pressure is less than 100 PSIG.

**Analytical Data:** (R=Reference Standard, Z=Zero Gas, C=Gas Candidate)

1. Component: Nitric oxide  
 Requested Concentration: 20 ppm  
 Certified Concentration: 20.1 ppm  
 Instrument Used: Thermo Electron 42i-LS SN 1030645077  
 Analytical Method: Chemiluminescence  
 Last Multipoint Calibration: 06/10/2019

Reference Standard: Type / Cylinder #: GMIS / CC2919  
 Concentration / Uncertainty: 18.80 ppm ±0.504%  
 Expiration Date: 03/28/2022  
 Traceable to: SRM # / Sample # / Cylinder #: APEX1161149 / NA / APEX1161149  
 SRM Concentration / Uncertainty: 20.03 ppm / ±0.10 ppm  
 SRM Expiration Date: 01/27/2020

First Analysis Data:				Date
Z: 0	R: 19.9	C: 20	Conc: 20	06/20/2019
R: 19.88	Z: 0	C: 20.1	Conc: 20.1	
Z: 0	C: 20.1	R: 19.91	Conc: 20.1	
UOM: ppm		Mean Test Assay: 20.1		ppm

Second Analysis Data:				Date
Z: 0	R: 19.9	C: 20.1	Conc: 20.1	06/27/2019
R: 19.89	Z: 0	C: 20.2	Conc: 20.2	
Z: 0	C: 20.1	R: 19.89	Conc: 20.1	
UOM: ppm		Mean Test Assay: 20.1		ppm

Analyzed By: Henry Koung

Certified By: Lecanna Flores





**CERTIFICATE OF ANALYSIS / EPA PROTOCOL GAS**

**Customer & Order Information**

PRAXAIR PKG OXNARD CA HPS  
455 E WOOLEY RD  
OXNARD CA 93030

Certificate Modification Date: 11/15/2018  
Praxair Order Number: 70784160  
Part Number: EV NICDOXE90-AS

Fill Date: 11/09/2018  
Lot Number: 70086831305  
Cylinder Style & Outlet: AS CGA 580  
Cylinder Pressure and Volume: 2000 psig 140 ft<sup>3</sup>

**Certified Concentration**

Expiration Date:	11/15/2026	NIST Traceable
Cylinder Number:	CC134279	Expanded Uncertainty
4.01 %	Carbon dioxide	± 0.4 %
11.97 %	Oxygen	± 0.1 %
Balance	Nitrogen	

*ProSpec EZ Cert*



**Certification Information:**

Certification Date: 11/15/2018

Term: 96 Months

Expiration Date: 11/15/2026

This cylinder was certified according to the 2012 EPA Traceability Protocol, Document #EPA-600/R-12/531, using Procedure G1.  
Do Not Use this Standard if Pressure is less than 100 PSIG.

CO2 responses have been corrected for Oxygen IR Broadening effect. O2 responses have been corrected for CO2 interference.

**Analytical Data:**

(R=Reference Standard, Z=Zero Gas, C=Gas Candidate)

1. Component:

Carbon dioxide

Requested Concentration: 4 %  
Certified Concentration: 4.01 %  
Instrument Used: Horiba VIA-510 S/N 20C194WK  
Analytical Method: NDIR  
Last Multipoint Calibration: 10/23/2018

First Analysis Data:		Date	
Z: 0	R: 6.96	C: 4	Conc: 4
R: 6.96	Z: 0	C: 4.01	Conc: 4.01
Z: 0	C: 4.01	R: 6.96	Conc: 4.01
UOM: %		Mean Test Assay: 4.01 %	

Reference Standard: Type / Cylinder #: GMIS / CC243762  
Concentration / Uncertainty: 6.96 % ± 0.208%  
Expiration Date: 08/07/2023

Traceable to: SRM # / Sample # / Cylinder #: SRM 1674b / 7-H-07 / FF10831  
SRM Concentration / Uncertainty: 6.944% / ± 0.013%  
SRM Expiration Date: 06/17/2019

Second Analysis Data:		Date	
Z: 0	R: 0	C: 0	Conc: 0
R: 0	Z: 0	C: 0	Conc: 0
Z: 0	C: 0	R: 0	Conc: 0
UOM: %		Mean Test Assay: %	

2. Component:

Oxygen

Requested Concentration: 12 %  
Certified Concentration: 11.97 %  
Instrument Used: OXYMAT 5E  
Analytical Method: Paramagnetic  
Last Multipoint Calibration: 11/05/2018

First Analysis Data:		Date	
Z: 0	R: 20.86	C: 11.96	Conc: 11.96
R: 20.86	Z: 0	C: 11.97	Conc: 11.97
Z: 0	C: 11.97	R: 20.86	Conc: 11.97
UOM: %		Mean Test Assay: 11.97 %	

Reference Standard: Type / Cylinder #: GMIS / CC75874  
Concentration / Uncertainty: 20.86 % ± 0.111%  
Expiration Date: 11/07/2025

Traceable to: SRM # / Sample # / Cylinder #: SRM 2659a / 71-E-19 / FF22331  
SRM Concentration / Uncertainty: 20.863% / ± 0.021%  
SRM Expiration Date: 08/23/2021

Second Analysis Data:		Date	
Z: 0	R: 0	C: 0	Conc: 0
R: 0	Z: 0	C: 0	Conc: 0
Z: 0	C: 0	R: 0	Conc: 0
UOM: %		Mean Test Assay: %	

Analyzed By

*Daniel Burns*  
Daniel Burns

Certified By

*Jose Vasquez*  
Jose Vasquez



**Praxair**  
 5700 South Alameda Street  
 Los Angeles, CA 90058  
 Tel: (323) 585-2154 Fax: (714) 542-6689  
 PGV#ID: F22015

DocNumber: 000081223

## CERTIFICATE OF ANALYSIS / EPA PROTOCOL GAS

**Customer & Order Information:**

PRAXAIR WHSE OXNARD CA  
 455 E WOOLEY RD  
 OXNARD CA 930300

Praxair Order Number: 31577381  
 Customer P. O. Number: 05587065  
 Customer Reference Number:

Fill Date: 6/19/2015  
 Part Number: EV NIDOXE98-AS  
 Lot Number: 1095169C9  
 Cylinder Size & Outlet: AS CGA 590  
 Cylinder Pressure & Volume: 2000 psig 140 cu ft.

**Certified Concentration:**

Expiration Date:	6/23/2023	NIST Traceable
Cylinder Number:	CC144381	Analytical Uncertainty:
7.93 %	CARBON DIOXIDE	± 0.5 %
20.13 %	OXYGEN	± 0.1 %
Balance: NITROGEN		

**Certification Information:** Certification Date: 6/23/2015 Term: 96 Months Expiration Date: 6/23/2023

This cylinder was certified according to the 2012 EPA Traceability Protocol, Document #EPA-600/R-12/531, using Procedure G1. Do Not Use this Standard if Pressure is less than 100 PSIG.

O2 responses have been corrected for CO2 interference.

**Analytical Data:**

(R=Reference Standard, Z=Zero Gas, C=Gas Candidate)

**1. Component: CARBON DIOXIDE**

Requested Concentration: 0 %  
 Certified Concentration: 7.93 %  
 Instrument Used: Horiba VIA-S10 S/N 2807014  
 Analytical Method: NDIR  
 Last Multipoint Calibration: 6/15/2015

<b>First Analysis Data:</b>		<b>Date:</b> 6/23/2015	
Z: 0	R: 9.87	C: 7.93	Conc: 7.927
R: 9.87	Z: 0	C: 7.93	Conc: 7.927
Z: 0	C: 7.93	R: 9.87	Conc: 7.927
UOM: %	<b>Mean Test Assay:</b>		7.927 %

Reference Standard Type: GMIS  
 Ref. Std. Cylinder #: SA17695  
 Ref. Std. Conc: 9.87 %  
 Ref. Std. Traceable to SRM #: 1674b  
 SRM Sample #: 7-H-07  
 SRM Cylinder #: FF10831

<b>Second Analysis Data:</b>		<b>Date:</b>	
Z: 0	R: 0	C: 0	Conc: 0
R: 0	Z: 0	C: 0	Conc: 0
Z: 0	C: 0	R: 0	Conc: 0
UOM: %	<b>Mean Test Assay:</b>		0 %

**2. Component: OXYGEN**

Requested Concentration: 20 %  
 Certified Concentration: 20.13 %  
 Instrument Used: OXYMAT SE  
 Analytical Method: PARAMAGNETIC  
 Last Multipoint Calibration: 5/29/2015

<b>First Analysis Data:</b>		<b>Date:</b> 6/23/2015	
Z: 0	R: 19.88	C: 20.12	Conc: 20.127
R: 19.88	Z: 0	C: 20.12	Conc: 20.127
Z: 0	C: 20.12	R: 19.88	Conc: 20.127
UOM: %	<b>Mean Test Assay:</b>		20.127 %

Reference Standard Type: GMIS  
 Ref. Std. Cylinder #: SA16022  
 Ref. Std. Conc: 19.90 %  
 Ref. Std. Traceable to SRM #: 2659a  
 SRM Sample #: 71-E-19  
 SRM Cylinder #: FF22331

<b>Second Analysis Data:</b>		<b>Date:</b>	
Z: 0	R: 0	C: 0	Conc: 0
R: 0	Z: 0	C: 0	Conc: 0
Z: 0	C: 0	R: 0	Conc: 0
UOM: %	<b>Mean Test Assay:</b>		0 %

Analyzed by:

Certified by:

Information contained herein has been prepared at your request by qualified experts within Praxair Distribution, Inc. While we believe that the information is accurate within the limits of the analytical methods employed and is complete to the extent of the specific analyses performed, we make no warranty or representation as to the suitability of the use of the information for any purpose. The information is offered with the understanding that any use of the information is at the sole discretion and risk of the user. In no event shall the liability of Praxair Distribution, Inc. arising out of the use of the information contained herein exceed the fee established for providing such information.



Praxair  
 5700 South Alameda Street  
 Los Angeles, CA 90058  
 Tel: (323) 585-2154 Fax: (714) 542-6689  
 PGVPID: F22016

DocNumber: 000094146

**CERTIFICATE OF ANALYSIS / EPA PROTOCOL GAS**

**Customer & Order Information:**

PRAXAIR WHSE OXNARD CA  
 455 E WOOLEY RD  
 OXNARD CA 93030

Praxair Order Number: 34250102  
 Customer P. O. Number: 06017500  
 Customer Reference Number:

Fill Date: 5/5/2016  
 Part Number: NI CO20ME-AS  
 Lot Number: 109612606  
 Cylinder Style & Outlet: AS CGA 350  
 Cylinder Pressure & Volume: 2000 psig 140 cu. ft.

**Certified Concentration:**

Expiration Date:	5/13/2024	NIST Traceable
Cylinder Number:	DT0012296	Analytical Uncertainty:
20.5 ppm	CARBON MONOXIDE	± 0.4 %
Balance	NITROGEN	

**Certification Information:** Certification Date: 5/13/2016 Term: 96 Months Expiration Date: 5/13/2024

This cylinder was certified according to the 2012 EPA Traceability Protocol, Document #EPA-600/R-12/531, using Procedure G1. Do Not Use this Standard if Pressure is less than 100 PSIG.

**Analytical Data:**

(R=Reference Standard, Z=Zero Gas, C=Gas Candidate)

**1. Component: CARBON MONOXIDE**

Requested Concentration: 20 ppm  
 Certified Concentration: 20.5 ppm  
 Instrument Used: Horiba VIA-510 S/N 576976015  
 Analytical Method: NDIR  
 Last Multipoint Calibration: 5/3/2016

Reference Standard Type: GMIS  
 Ref. Std. Cylinder #: CC341229  
 Ref. Std. Conc: 25.04 ppm  
 Ref. Std. Traceable to SRM #: 2635a  
 SRM Sample #: 58-E-34  
 SRM Cylinder #: FF10666

First Analysis Data:				Date:			
Z:	0	R:	25	C:	20.4	Conc:	20.351
R:	25.1	Z:	0	C:	20.5	Conc:	20.451
Z:	0	C:	20.6	R:	25.2	Conc:	20.551
UOM:	ppm		Mean Test Assay:	20.451 ppm			

Second Analysis Data:				Date:			
Z:	0	R:	0	C:	0	Conc:	0
R:	0	Z:	0	C:	0	Conc:	0
Z:	0	C:	0	R:	0	Conc:	0
UOM:	ppm		Mean Test Assay:	0 ppm			

Analyzed by:

Ying Yu

Certified by:

Maria Soberanis

Information contained herein has been prepared at your request by qualified experts within Praxair Distribution, Inc. While we believe that the information is accurate within the limits of the analytical methods employed and is complete to the extent of the specific analyses performed, we make no warranty or representation as to the suitability of the use of the information for any purpose. The information is offered with the understanding that any use of the information is at the sole discretion and risk of the user. In no event shall the liability of Praxair Distribution, Inc., arising out of the use of the information contained herein exceed the fee established for providing such information.



Praxair  
 5700 South Alameda Street  
 Los Angeles, CA 90058  
 Tel: (323) 585-2154 Fax: (714) 542-6689  
 PGVPID: F22017

DocNumber: 000105446

**CERTIFICATE OF ANALYSIS / EPA PROTOCOL GAS**

**Customer & Order Information:**

PRAXAIR PKG OXNARD CA HPS  
 455 E WOOLEY RD  
 OXNARD CA 93030

Praxair Order Number: 70207603  
 Customer P. O. Number:  
 Customer Reference Number:

Fill Date: 2/22/2017  
 Part Number: NI CO4DME-AS  
 Lot Number: 109705308  
 Cylinder Style & Outlet: AS CGA 350  
 Cylinder Pressure & Volume: 2000 psig 140 cu. ft.

**Certified Concentration:**

Expiration Date:	3/7/2025	NIST Traceable
Cylinder Number:	CC272657	Analytical Uncertainty:
40.0 ppm	CARBON MONOXIDE	± 0.5 %
Balance	NITROGEN	

**Certification Information:** Certification Date: 3/7/2017 Term: 96 Months Expiration Date: 3/7/2025  
 This cylinder was certified according to the 2012 EPA Traceability Protocol, Document #EPA-600/R-12/531, using Procedure G1. Do Not Use this Standard if Pressure is less than 100 PSIG.

**Analytical Data:**

(R=Reference Standard, Z=Zero Gas, C=Gas Candidate)

1. Component: CARBON MONOXIDE

Requested Concentration: 40 ppm  
 Certified Concentration: 40.0 ppm  
 Instrument Used: Horiba VIA-510 SIN 576B76015  
 Analytical Method: NDIR  
 Last Multipoint Calibration: 3/1/2017

Reference Standard Type: GMIS  
 Ref. Std. Cylinder #: CC188877  
 Ref. Std. Conc: 50.3 ppm  
 Ref. Std. Traceable to SRM #: 1578c  
 SRM Sample #: 04-1-41  
 SRM Cylinder #: FF18402

First Analysis Data:		Date: 3/7/2017	
Z:	0	R:	50.2
R:	50	Z:	0
Z:	0	C:	39.8
UOM:	ppm	R:	49.9
		Conc:	40.012
		Conc:	40.012
		Conc:	40.012
		Mean Test Assay:	40.012 ppm

Second Analysis Data:		Date:	
Z:	0	R:	0
R:	0	Z:	0
Z:	0	C:	0
UOM:	ppm	R:	0
		Conc:	0
		Conc:	0
		Conc:	0
		Mean Test Assay:	0 ppm

Analyzed by:

Certified by:

Nassim Haddad

Information contained herein has been prepared at your request by qualified experts within Praxair Distribution, Inc. While we believe that the information is accurate within the limits of the analytical methods employed and is complete to the extent of the specific analyses performed, we make no warranty or representation as to the suitability of the use of the information for any purpose. The information is offered with the understanding that any use of the information is at the sole discretion and risk of the user. In no event shall the liability of Praxair Distribution, Inc., arising out of the use of the information contained herein exceed the fee established for providing such information.

**DRY GAS METER CALIBRATION**

Standard Pressure  
Standard Temperature  
Ambient pressure  
Ambient temperature

29.92 in. hg.  
60 F  
29.96 in. hg.  
65 F

Unit Number J  
Date: 1/4/2021  
Leak Check: .004 @ 20"

ΔH in. H2O	TIME min.	WET GAS VOL. cf	DRY GAS VOL. in/out cf	Temperature				*Y	†ΔH@ in. H2O
				W.G. AVG F	D.G. IN F	D.G. OUT F	D.G. AVG. F		
0.75	9.94	5.000	818.537		62.0	61.0	61.8	1.0034	1.6419
			823.518	61.0	63.0	61.0			
0.75	9.94	5.000	823.518		61.0	61.0	61.8	1.0030	1.6419
			828.501	61.0	63.0	62.0			
0.75	9.96	5.000	828.501		62.0	62.0	62.5	1.0126	1.6462
			833.444	61.0	64.0	62.0			
1.50	7.30	5.000	833.512		63.0	62.0	62.5	1.0042	1.7686
			838.487	61.0	63.0	62.0			
1.50	7.32	5.000	838.487		63.0	63.0	63.0	1.0054	1.7766
			843.461	61.0	64.0	62.0			
1.50	7.32	5.000	843.461		63.0	63.0	63.3	1.0051	1.7792
			848.434	61.5	64.0	63.0			
2.25	5.92	5.000	848.513		64.0	63.0	63.5	1.0078	1.7447
			853.466	61.5	64.0	63.0			
2.25	5.92	5.000	853.466		64.0	63.0	64.0	1.0078	1.7464
			858.419	62.0	65.0	64.0			
2.25	5.95	5.000	858.419		65.0	64.0	64.5	1.0083	1.7625
			863.374	62.0	65.0	64.0			
3.00	5.18	5.000	863.374		65.0	64.0	64.8	1.0085	1.7871
			868.382	63.0	66.0	64.0			
3.00	5.18	5.000	868.382		65.0	64.0	64.8	1.0099	1.7871
			873.313	63.0	66.0	64.0			
3.00	5.15	5.000	873.313		66.0	64.0	64.8	1.0091	1.7664
			878.248	63.0	66.0	64.0			
3.75	4.63	5.000	878.248		66.0	64.0	65.0	1.0119	1.7838
			883.190	63.0	66.0	64.0			
3.75	4.62	5.000	883.190		65.0	64.0	64.5	1.0104	1.7846
			888.098	64.0	65.0	64.0			
3.75	4.62	5.000	888.098		65.0	64.0	64.5	1.0100	1.7846
			893.008	64.0	65.0	64.0			
AVERAGE								1.0078	1.7468

Validity checks:

\* Y(max - min) ≤ .02 ?  
† |ΔH@ - ΔH@ avg. | ≤ .20 in. H2O ?

Meter Factor: 1.0078

ΔH@ : 1.7468

Calibration by: FT

Reviewed by: KK

**EQUATIONS USED:**

$$Y = (VWG * PBAR * (TDGavg + 460)) / ((VDG * (PBAR + (\Delta H / 13.6)) * (TWGavg + 460)))$$

$$\Delta H@ = ((0.0319 * \Delta H) / (PBAR * (TDGavg + 460))) * (((TWG + 460) * T) / VWG) ^ 2$$





**Attachment to 6.a. 74.15N1-0157 and P00157PC2-Condition 4 (Nebraska Boiler)**

Condition No.	Response												
	Monthly		Rolling 12-Months			Emissions			Rolling 12 Months				
Month-Year	Natural Gas (scf)	Uptime (hours)	Natural Gas (scf)	Uptime (hours)	Capacity Factor (%)	Nebraska Emissions NOx (tons)	CO (tons)	Cogen Emissions NOx (tons)	CO (tons)	Total Emissions MOx (tons)	CO (tons)	NOx (tons)	CO (tons)
74.15 - 1 & 2	Required source test every 24 months will be conducted during next compliance year (Last test was on 9/22/2020).												
74.15 - 3	Concentration limits do not apply during cold startup for period of three (3) hours.												
74.15 - 4	Records of biennial source test reports will be maintained and submitted to District upon request.												
PC2 - 4	Operate CEMS for NOx and oxygen including zero and span drift checks daily, reporting of breakdown, recording of operational parameters, reporting of violation, provision of CEMS data in an approved format.												
Apr-19	0	0	0	0		0.00	0.000	2.44	3.66	2.44	3.66	44.42	93.28
May-19	0	0	0	0		0.00	0.000	3.82	4.76	3.82	4.76	44.32	87.83
Jun-19	1,013,281	17.32	41,993,023	640	4.581%	0.05	0.426	3.79	5.05	3.84	5.47	43.92	82.91
Jul-19	12,516,147	198.38	59,107,201	890	6.385%	0.33	2.816	2.16	3.30	2.49	6.11	42.98	79.74
Aug-19	0	0	54,309,119	798	5.878%	0.10	0.876	3.38	5.02	3.48	5.90		
Sep-19	0	0	54,309,119	798	5.890%	0.00	0.000	3.89	5.88	3.89	5.88	43.11	66.61
Oct-19	12,356,397	184.65	56,493,986	818	5.924%	0.00	0.000	3.71	5.69	3.71	5.69	43.05	67.17
Nov-19	12,236,751	183.26	55,127,616	783	6.151%	0.28	2.258	2.32	3.65	2.60	5.90	42.58	67.37
Dec-19	0	0	72,147,519	1,023	7.923%	0.21	1.688	3.02	5.64	3.23	7.33	43.02	69.35
Jan-20	0	0	71,042,864	1,008	7.802%	0.07	0.632	3.58	6.05	3.65	6.69	42.72	71.09
Feb-20	1,104,655	15.17	73,750,226	1,050	8.084%	0.23	1.907	3.15	4.01	3.38	5.92	42.12	71.32
Mar-20	0	0	73,750,226	1,050	8.084%	0.00	0.000	3.37	3.83	3.37	3.83	41.76	69.83
Max. Rolling 12 Months	2,707,362	42.09	73,750,226	1,050	8.084%	0.329	2.816	3.95	6.05	3.95	7.33	43.55	71.32
Permit Limit												50.00	97.44
Exceeds Permit Limit?												NO	NO
Excess Emissions (Max. Rolling 12 Months)												-6.45	-26.12
Excess Emissions (Compliance Year)												-7.97	-35.56

Sample Calculation:

Capacity Factor = Annual Fuel Usage \* HHV natural gas / (Equipment rating \* Hours/year)  
 Capacity Factor = (10 MMtcf/yr) \* (1,033 BTU/cf) / ((108 MMBtu/hr) \* (8,760 hr/yr))  
 Capacity Factor = 1.09%

RATA Test on 9/22/2020.  
Nebraska NOx and CO from CEMS.

**New-Indy Oxnard  
Nebraska Boiler Operation  
Compliance Year: 4/1/2020 to 3/31/2021**

<b>Date</b>	<b>Time Online</b>	<b>Operating Hours</b>	<b>Fuel Usage (SCF)</b>
<b>May</b>			
5/4/2020	12:44 PM - 4:59 AM	14.78	1,034,039
5/14/2020	7:26 AM - 12:20 PM, 7:12 PM - 7:00 AM	14.93	936,064
5/15/2020	7:00:00 AM - 18:51 PM	11.85	795,689
<b>June</b>			
6/17/2020	5:37 PM - 7:00 AM	8.3	340,247
6/18/2020	7:00 AM - 7:00 AM	23.02	1,527,273
6/19/2020	7:00 AM - 7:00 AM	24	1,694,141
6/20/2020	7:00 AM - 7:00 AM	24	1,810,080
6/21/2020	7:00 AM - 7:00 AM	24	1,804,827
6/22/2020	7:00 AM - 7:00 AM	24	1,614,900
6/23/2020	7:00 AM - 7:00 AM	24	1,623,879
6/24/2020	7:00 AM - 7:00 AM	24	1,611,216
6/25/2020	7:00 AM - 3:36 PM, 5:25 PM - 7:00 AM	22	1,524,102
6/26/2020	7:00 AM - 6:47 AM, 6:57 AM - 7:00AM	23.83	1,577,716
6/27/2020	7:00 AM - 7:00 AM	24	1,556,195
6/28/2020	7:00 AM - 11:14 PM	16.23	1,164,699
6/29/2020	10:13 AM - 3:59 AM	5.77	278,184
<b>2Q2020</b>	<b>Total</b>	<b>309</b>	<b>18,127,459</b>
<b>July</b>			
7/4/2020	4:15 PM - 7:00 AM	13.8	1,037,106
7/5/2020	7:00 AM - 8:00 PM	13	1,040,361
7/15/2020	4:01 AM -7:00 AM	2.42	143,218
7/15/2020	7:00 AM -7:00 AM	24	1,630,795
7/16/2020	7:00 AM -7:00 AM	24	1,679,286
7/17/2020	7:00 AM -7:00 AM	24	1,844,662
7/18/2020	7:00 AM - 11:54 AM	4.9	342,637
<b>3Q2020</b>	<b>Total</b>	<b>106</b>	<b>7,718,065</b>
<b>Oct</b>			
10/19/2020	1:12 PM - 7:00 AM	17.45	1,148,249
10/20/2020	7:00 AM - 7:00 AM	24	1,671,750
10/21/2020	7:00 AM - 7:00 AM	24	1,716,898
10/22/2020	7:00 AM - 7:00 AM	24	1,708,426
10/23/2020	7:00 AM - 7:00 AM	24	1,758,679
10/24/2020	7:00 AM - 7:00 AM	24	1,805,563
10/25/2020	7:00 AM - 3:42 AM	20.7	1,589,990
10/30/2020	8:18 AM - 7:00 AM	22.7	1,399,035
10/31/2020	7:00 AM - 7:00 AM	24	1,742,672
<b>November</b>			
11/1/2020	7:00 AM - 7:00 AM	24	1,691,660
11/2/2020	7:00 AM - 7:00 AM	24	1,752,246
11/3/2020	7:00 AM - 7:00 AM	24	1,760,755

11/4/2020	7:00 AM - 7:00 AM	24	1,792,929
11/5/2020	7:00 AM - 7:00 AM	24	1,770,931
11/6/2020	7:00 AM - 7:00 AM	24	1,795,404
11/7/2020	7:00 AM - 10:59 AM, 1:16 PM - 1:51 PM	4.57	306,457
<b>December</b>			
12/20/2020	1:34 PM - 7:00 AM	17.2	1,177,508
12/21/2020	7:00 AM - 7:00 AM	24	1,757,073
12/22/2020	7:00 AM - 3:16 PM	8.27	522,811
12/29/2020	9:45 AM - 7:19 PM	9.23	609,451
<b>4Q2020</b>	<b>Total</b>	<b>412.12</b>	<b>29,478,489</b>
<b>January</b>			
1/11/2021	6:38 PM - 7:00 PM, 7:14 PM - 7:19 PM	0.45	16,173
1/12/2021	12:56 PM - 1:03 PM, 2:25 PM - 7:00 AM	16.7	895,741
1/13/2021	7:00 AM - 7:00 AM	24	1,810,298
1/14/2021	7:00 AM - 7:00 AM	24	1,751,611
1/15/2021	7:00 AM - 7:00 AM	24	1,782,639
1/16/2021	7:00 AM - 7:00 AM	24	1,816,438
1/17/2021	7:00 AM - 1:17 PM	6.28	433,461
1/20/2021	10:54 AM - 10:59 AM, 11:03 AM - 7:00 AM	20.03	1,523,416
1/21/2021	7:00 AM - 3:31 PM	8.52	673,410
1/24/2021	AM - 5:45 AM, 5:58 AM - 6:16 AM, 6:34 AM - 7:00 AM	1.13	26,852
1/24/2021	7:00 AM - 9:55 AM	2.92	175,535
1/25/2021	7:50 AM - 7:00 AM	23.17	1,675,476
1/26/2021	7:00 AM - 12:47 PM	5.78	372,009
<b>March</b>			
3/20/2021	9:40 AM - 11:55 AM, 12:07 PM - 7:00 AM	21.13	1,510,295
3/21/2021	7:00 AM - 5:51 PM	10.85	819,318
3/24/2021	7:44 AM - 8:11 AM, 8:28 AM - 4:57 PM	8.93	300,626
3/31/2021	7:51 AM - 8:11 AM, 8:37 AM - 10:28 AM	1.18	77,122
<b>1Q2021</b>	<b>Total</b>	<b>223.07</b>	<b>15,660,422</b>



**COMPLIANCE EMISSIONS TESTING  
NEW INDY OXNARD  
NEBRASKA BOILER - N8E95S1  
VCAPCD PTO #0157-281  
TESTED ON: OCTOBER 22, 2020**

**Prepared for:  
NEW INDY OXNARD  
5936 Perkins Road,  
Oxnard, CA 93033**

**Attn: Robyn Lebrilla**

**Facility Location:  
NEW INDY OXNARD  
5936 Perkins Road,  
Oxnard, CA 93033**

**Submitted to:  
Ventura County Air Pollution Control District  
669 County Square Drive  
Ventura, California 93003**

**Attn: Ed Swede**

**Prepared by:  
AIRx Testing Services, Inc.  
2472 Eastman Avenue #34  
Ventura, CA 93003**

**Job No.: 23022**

**Lab No.: 220-113b**

Two handwritten signatures are present. The top signature is in dark ink and appears to be 'Ken Kennepohl'. The bottom signature is in lighter ink and appears to be 'Tom Porter'. Both signatures are written over a horizontal line.

**Ken Kennepohl; Senior Engineer**

**Tom Porter; Vice President of Testing Services**

**Submitted:  
November 3, 2020**

Table of Contents

<u>Section</u>	<u>Page</u>
Source Test Results	1
Introduction	2
Calculations	4
NO <sub>x</sub> and CO Emissions	5
Bias Adjustments	6
F-Factor Exhaust Gas Flowrate	7
Field Data and Strip Charts	8
CEM Data	9
Strip Charts	11
Facility DAS Printout	13
Boiler Data Sheet	14
Fuel Meter Data	15
Instrument Information	16
Extractive Gas Sample System	17
CEM Instrument List	18
Quality Assurance	21
Calibration Gas Certification Sheets	22

**SUMMARY OF RESULTS**  
**New Indy**  
**Oxnard**  
**Nebraska Boiler**  
**10/22/2020**

**TABLE 1-1. TEST CONDITIONS**

PARAMETER	UNITS	Run #1	Average	Method used
Stack Gas Flowrate	dscfm	12,581	12,581	On-Site Measurements
Fuel Usage	dscfm	1078.1	1078.1	Rated BTU
Heat Input during Test	MMBtu/hr	67.9	<u>Unit Description</u> <b>Nebraska Boiler</b> <b>108 MMBtu/hr.</b>	

**TABLE 1-2. SOURCE TEST RESULTS**

POLLUTANT	UNITS	EMISSIONS	Average	Allowable Limits	District Rule
		Run #1			
Nitrogen Oxide <i>Actual Measured</i>	ppmv	24.8	24.8	-	PTO
	ppmv @ 3% O2	27.1	27.1	40	
	lb/hr	2.23	2.23	-	
	lb/MMBtu	0.033	0.033	-	
Carbon Monoxide <i>10% of Full Scale</i>	ppmv	< 5.0	< 5.0	-	PTO
	ppmv @ 3% O2	< 5.5	< 5.5	400	
	lb/hr	< 0.45	< 0.45	-	
	lb/MMBtu	< 0.0066	< 0.007	-	
Carbon Monoxide <i>Actual Measured</i>	ppmv	0.2	0.2	-	PTO
	ppmv @ 3% O2	0.2	0.2	400	
	lb/hr	0.0090	0.0090	-	
	lb/MMBtu	0.00013	0.00013	-	
Oxygen	%	4.5	4.5	-	-
FGR	Hz	40.22	40.22	-	-

<b>INTRODUCTION</b>
---------------------

## 1.0 INTRODUCTION

fired 108 MMBTU Nebraska Boiler N8E95S1. The unit is located at the New Indy Plant at 5936 Perkins Road, in Oxnard. The boiler is used for the generation of steam for specific facility operations. The tests were required to satisfy requirements detailed in PTO #0157-281 and Rule 74.15.E.2.

**PROCESS DETAILS:** The Nebraska Boiler is rated for 108 MMBtu/hr heat input and exhausts into 44" exhaust stack. The sample ports are three (3) inch males with A and B distances of 84" and 186" respectively (1.9 diameters upstream and 4.0 diameters downstream).

**CONTINUOUS MONITORS:** NOX, CO & O2: Instrumental testing was performed in accordance with CARB Method 100 to determine the concentrations of NOx, CO and O2. A total of 60 minutes of data was obtained while the boiler was operating at a normal (auto) load. A continuous sample of the exhaust gas was taken from the oven stack exhaust through a refrigerated sample gas conditioner and transported under negative pressure through Teflon tubing to a distribution panel. The nitrogen oxides were checked using an API Model 200EH chemiluminescent analyzer. The carbon monoxide was checked with an API Model 300EH NDIR W/GFC analyzer and the oxygen was checked with a Servomex Model 1400 paramagnetic analyzer. Data was collected at one (1) minute intervals and also recorded on a 10" strip chart recorder.


During all tests, the NOx, CO and O2 analyzers were operated on the 50 ppmv (NOx), 50 ppmv (CO) and 10% (O2) scales, respectively.

There were no deviations from CARB Method 100 methodology during the test.


**FLOWRATE:** Stack flow calculations were performed using fuel consumption data obtained from a dedicated fuel meter and the stack oxygen, in accordance with EPA Method 19.

All reporting and calculations have been performed using VCAPCD standard conditions of 68°F and 29.92 inches of Hg. If you have any questions regarding this test program or report, please contact the undersigned at (805) 644-1099.

Respectfully Submitted  
AIRx Testing Services, Inc.

  
Ken Kennepohl  
Senior Engineer

Reviewed by



Tom Porter  
Vice President of Testing Services

CC: Ventura County APCD  
Attn: Ed Swede



<b>CALCULATIONS</b>
---------------------

**CONTINUOUS EMISSIONS MONITORING - CARB METHOD 1-100**

Client : New Indy  
 Site : Oxnard  
 Unit : Nebraska Boiler

Date : 10/22/2020  
 Job# : 23022  
 Lab# : 220-113B

**FIELD DATA**

Standard Temperature:  °F

**Drift Corrected Emissions Data**

<i>Actual</i>	NOx	<input type="text" value="24.81"/>	ppmv
<i>10% FS</i>	CO	<input type="text" value="&lt; 5.0"/>	ppmv
<i>Actual</i>	CO	<input type="text" value="0.17"/>	ppmv
	O2	<input type="text" value="4.52"/>	%

**Boiler and Flow Data**

Rated Heat Input	<input type="text" value="108"/>	MMBtu/hr
Fuel F-Factor	<input type="text" value="8,710"/>	dscf/MM Btu
Q std	<input type="text" value="12581"/>	dscfm

**Equations used;**

NOx or CO @ 3% O2 = [ppmv] \* (17.9/(20.9-O2%))  
 $T_{fact} = (10^{-6} * (29.92 / (21.85 * (460 + T_{std})))) * 60$   
 lb/hr = [ppmv] \* Q \* MW \* (Tfact)  
 lb/MMBtu = [ppmv] \* MW \* F-Factor \* (Tfact / 60) \* (20.9 / (20.9 - O2%))  
 MW NOx = 46, MW CO = 28

**CALCULATED EMISSIONS**

<i>Actual</i>	NOx @ 3% O2, ppmv	<b>27.12</b>
	NOx, lb/hr	<b>2.23</b>
	NOx, lb/MMBtu	<b>0.033</b>
<i>10% FS</i>	CO @ 3% O2, ppmv	<b>&lt; 5.5</b>
	CO, lb/hr	<b>&lt; 0.45</b>
	CO lb/MMBtu	<b>&lt; 0.0066</b>
<i>Actual</i>	CO @ 3% O2, ppmv	<b>0.18</b>
	CO, lb/hr	<b>0.0090</b>
	CO, lb/MMBtu	<b>0.00013</b>

**CALIBRATION ERROR, SYSTEM BIAS, DRIFT CORRECTIONS**

Client : New Indy  
 Site : Oxnard  
 Unit : Nebraska Boiler

Date : 10/22/2020  
 Job# : 23022  
 Lab# : 220-113B

	O2	CO	NOx	Allow.
	Values in ppm except O2 and CO2 in %			Er. (%)
<b>Analyzer Calibration Error</b>				
Instrument Span Range (Range)	10	50	50	
High Calibration Value from Cylinder	8.05	40.4	40.0	
Analyzer Reads (Ca)	8.0	40.5	40.0	
<b>Analyzer Calibration Error (%) = ((Cylinder Value - Ca)/Range)x100</b>	<b>0.2</b>	<b>-0.1</b>	<b>0.0</b>	<b>2</b>
Mid Calibration value from Cylinder	3.98	20.5	20.2	
Analyzer Reads (Ca)	4.0	20.3	20.2	
<b>Analyzer Calibration Error (%) = ((Cylinder Value - Ca)/Range)x100</b>	<b>-0.1</b>	<b>0.5</b>	<b>0.0</b>	<b>2</b>
Zero Calibration Value from Cylinder	0.0	0.0	0.0	
Analyzer Reads (Ca)	0.0	-0.1	0.0	
<b>Analyzer Calibration Error (%) = ((Cylinder Value - Ca)/Range)x100</b>	<b>-0.1</b>	<b>0.2</b>	<b>0.0</b>	<b>2</b>
Calibration Gas (Mid or High) used (Cma)	3.98	20.5	20.2	
<b>Pre-Test Data</b>				
Zero Gas Analyzer Response, (pre-test) (Ca)	0.0	-0.1	0.0	
Zero Gas System Response, (pre-test) (Cib)	0.0	-0.1	0.0	
<b>Sampling System BIAS Check (%) = ((Ca - Cib) / Range) x 100</b>	<b>0.0</b>	<b>-0.1</b>	<b>0.1</b>	<b>5</b>
Calibration Gas Analyzer Response, (pre-test) (Ca)	3.99	20.3	20.2	
Calibration Gas System Response, (pre-test) (Cib)	4.0	20.5	20.2	
<b>Sampling System BIAS Check (%) = ((Ca - Cib) / Range) x 100</b>	<b>0.1</b>	<b>-0.4</b>	<b>0.0</b>	<b>5</b>
<b>Run 1 Data</b>				
Zero Gas System Response, post-run (Cfb1)	0.0	-0.1	0.0	
Calibration Gas System Response, post-run (Cfb1)	4.0	20.4	20.2	
Source Test Average Value (Cavg)	<b>4.51</b>	<b>0.11</b>	<b>24.77</b>	
<b>Zero Drift (%) = ((Cib - Cfb1)/Range)x100</b>	<b>-0.2</b>	<b>0.0</b>	<b>-0.1</b>	<b>3</b>
<b>Calibration Drift (%) = ((Cib - Cfb1)/Range)x100</b>	<b>0.1</b>	<b>0.0</b>	<b>0.1</b>	<b>3</b>
<b>Zero Sampling System BIAS (%) = ((Ca-Cfb1)/Range)x100</b>	<b>-0.2</b>	<b>-0.1</b>	<b>0.0</b>	<b>5</b>
<b>Calibration Sampling System BIAS (%) = ((Ca-Cfb1)/Range)x100</b>	<b>0.2</b>	<b>-0.4</b>	<b>0.1</b>	<b>5</b>
<b>Average Zero Response (Co) = (Cib+Cfb1)/2</b>	<b>0.0</b>	<b>-0.1</b>	<b>0.0</b>	
<b>Average Calibration Response (Cm) = (Cib+Cfb1)/2</b>	<b>4.0</b>	<b>20.4</b>	<b>20.2</b>	
<b>Bias/Drift Corrected Concentration (Cgas) = (Cavg-Co)xCma/(Cm-Co)</b>	<b>4.52</b>	<b>0.17</b>	<b>24.8</b>	

**"F" FACTOR EXHAUST GAS FLOWRATE CALCULATION**

Client : New Indy  
Site : Oxnard  
Unit : Nebraska Boiler

Date : 10/22/2020  
Job# : 23022  
Lab# : 220-113B

"F" Factor, (Natural Gas):  dscf/MMBTU @ 68°F

Standard Temperature, T std:  deg. F

Run #1

Oxygen, %	<input type="text" value="4.52"/>
Fuel Usage, dscfm	<input type="text" value="1078.1"/>
MMBTU/min	<input type="text" value="1.132"/>
Stack Gas Flowrate, dscfm	<input type="text" value="12581"/>

**Equations used:**

Fuel Usage = (Final meter reading [cf] - Initial meter reading [cf]) / elapsed time [min]

If Natural Gas: MMBTU/min = (Fuel Usage [dscfm] \* 1050 Btu/ft<sup>3</sup>) / 10<sup>6</sup>

Stack effluent = "F" Factor \* MMBTU/min \* (20.9 / (20.9 - O<sub>2</sub>))

**FIELD DATA & STRIP CHARTS**

**EMISSIONS TEST - CARB METHOD A.100**

Date: 10/22/2020

Page: 1 of 2

**\*\* Test Information \*\***

Client: New Indy  
 Site: Oxnard  
 County: VCAPCD  
 Unit: Nebraska Boiler  
 Fuel: Natural Gas  
 Stack: Circular

**\*\* Personnel \*\***

AIRx: KK-WH  
 Client: Robyn Lebrilla  
 APCD: Ed Swede  
 Job #: 23022  
 Lab #: 220-113B  
 Flowrate: Fuel Usage

Run Length (min):	60	Inlet ( )	Outlet( )	Fuel ( )	HC's:
No. Points:	1 per stack "t"	S.T.( )	E.I.( )	Data( )	H2S ( ) NH3 ( )
Arrive:	Time 6:30	Baro. 29.85	Temp. 65	Weather. Clear	
Depart:	Time 12:00	Baro. 29.88	Temp. 70	Weather. Clear	

**\*\* Instrument Information \*\***

Instrument	"On"	Unit#	Make/Model
Outlet NOx:	1	3	API 200EH
O2:	1	12	Servomex 1400
CO:	1	3	API 300EM

Recorders: Soltec 6 pen 10 cm/hr

**\*\* Calibration Information \*\***

	Units	Zero	Span	Range	Gas Cyl.#	Gas Flow
NOx: High	ppmv	0.00	40.0	50	DT0028765	1.0
NOx: Mid	ppmv	0.00	20.2	50	CC724456	1.0
NO2	ppmv	0.00	19.0	50	CC3240	1.0
O2: High	%	0.00	8.05	10	CC160530	0.6
O2: Mid	%	0.00	3.98	10	CC60235	0.6
CO: High	ppmv	0.00	40.4	50	CC111187	1.0
CO: Mid	ppmv	0.00	20.5	50	DT0012296	1.0

**\*\* Recorder Information \*\***

	Chanl.	Pen Type	Color
Outlet NOx:	1	Cont.	Black
O2:	2	Cont.	Green
CO:	3	Cont.	Purple

EMISSIONS TEST - CARB Method 1-1u0

Date: 10/22/2020  
 Client: New Indy  
 Site: Oxnard

page: 2 of 2  
 Unit: Nebraska Boiler  
 Run: 1

Start Calibrations @: 12:20  
 Finish Calibrations @: 13:29

Start Run @: 12:25  
 End Run @: 13:25

NO2 Conversion	
Gas Conc.	19.0
Instr. Resp.	18.3
NO2 Convers.	96.3%

**\*\* Measured Emissions Components \*\***

Source: Component: Units:	Outlet NOx ppmv	Outlet CO ppmv	Outlet O2 %
---------------------------------	-----------------------	----------------------	-------------------

**\*\* Raw Emission Data \*\***

NOx, ppm @ 3% C      CO, ppm @ 3% O2

12:25	24.68	0.11	4.63	27.2	0.1
12:26	24.85	0.11	4.46	27.1	0.1
12:27	24.82	0.11	4.45	27.0	0.1
12:28	24.76	0.11	4.44	26.9	0.1
12:29	24.85	0.11	4.60	27.3	0.1
12:30	24.82	0.11	4.77	27.5	0.1
12:31	24.71	0.11	4.29	26.6	0.1
12:32	24.63	0.11	4.49	26.9	0.1
12:33	24.59	0.11	4.49	26.8	0.1
12:34	24.55	0.11	4.49	26.8	0.1
12:35	24.56	0.11	4.31	26.5	0.1
12:36	24.57	0.11	4.47	26.8	0.1
12:37	24.85	0.11	4.60	27.3	0.1
12:38	24.82	0.11	4.77	27.5	0.1
12:39	24.71	0.11	4.29	26.6	0.1
12:40	24.88	0.11	4.25	26.8	0.1
12:41	25.01	0.11	4.68	27.6	0.1
12:42	25.05	0.11	4.42	27.2	0.1
12:43	25.14	0.11	3.98	26.6	0.1
12:44	25.41	0.11	4.53	27.8	0.1
12:45	25.40	0.11	4.59	27.9	0.1
12:46	25.26	0.11	4.55	27.7	0.1
12:47	25.15	0.11	4.48	27.4	0.1
12:48	25.19	0.11	3.95	26.6	0.1
12:49	25.35	0.11	4.41	27.5	0.1
12:50	25.35	0.11	4.56	27.8	0.1
12:51	25.11	0.11	4.55	27.5	0.1
12:52	24.93	0.11	4.52	27.3	0.1
12:53	24.89	0.11	4.24	26.8	0.1
12:54	24.92	0.11	4.41	27.1	0.1
12:55	24.77	0.11	4.82	27.6	0.1
12:56	24.57	0.11	5.44	28.5	0.1
12:57	23.91	0.11	4.49	26.1	0.1
12:58	24.14	0.11	4.77	26.8	0.1
12:59	24.68	0.11	4.63	27.2	0.1
13:00	24.85	0.11	4.46	27.1	0.1
13:01	24.82	0.11	4.45	27.0	0.1
13:02	24.76	0.11	4.44	26.9	0.1
13:03	24.60	0.11	4.44	26.8	0.1
13:04	24.65	0.11	4.49	26.9	0.1
13:05	24.74	0.11	4.95	27.8	0.1
13:06	24.74	0.11	4.49	27.0	0.1
13:07	24.79	0.11	4.27	26.7	0.1
13:08	24.91	0.11	4.77	27.6	0.1
13:09	24.75	0.11	4.57	27.1	0.1
13:10	24.63	0.11	4.49	26.9	0.1
13:11	24.59	0.11	4.49	26.8	0.1
13:12	24.55	0.11	4.49	26.8	0.1
13:13	24.62	0.11	4.66	27.1	0.1
13:14	24.58	0.11	4.79	27.3	0.1
13:15	24.44	0.11	4.31	26.4	0.1
13:16	24.50	0.11	4.49	26.7	0.1
13:17	24.59	0.11	4.76	27.3	0.1
13:18	24.50	0.11	4.32	26.5	0.1
13:19	24.50	0.11	4.31	26.4	0.1
13:20	24.56	0.11	4.31	26.5	0.1
13:21	24.57	0.11	4.47	26.8	0.1
13:22	24.69	0.11	4.83	27.5	0.1
13:23	24.67	0.11	4.30	26.6	0.1
13:24	24.63	0.11	4.28	26.5	0.1
13:25	24.73	0.11	4.81	27.5	0.1

Port Change:

<b>Avg Meas Conc.</b>	<b>24.77</b>	<b>0.11</b>	<b>4.51</b>	<b>27.1</b>	<b>0.1</b>
-----------------------	--------------	-------------	-------------	-------------	------------

NO<sub>x</sub> 1919  
CO 1915

NO<sub>x</sub> 3198

NO<sub>x</sub> in Funtial Bios Col

0480

NO<sub>x</sub> 1915  
CO 1916

Bios

CO 1914  
CELLIST

NO<sub>x</sub> 3198  
CO 1915

NO<sub>x</sub> 3198  
CO 1915

NO<sub>x</sub> 1919

CO 1910

NO<sub>2</sub> → NO Conversion

NO<sub>x</sub> 1919

CO 1915

NO<sub>2</sub> → NO

Funtial Internal University Col

NO<sub>x</sub> 1919  
CO 1915

NO<sub>x</sub> 1919

New Entry

10.2.2019  
10.2.2019

leakback




Start  
Completed  
12:25  
70°  
89.58

Co. 0-50  
05-50  
Next 0-10

Post-Bites cad

Co 205  
02 385  
Next No. 2



*Compliance*

NOx, ppm (3% O2)		Raw NOx, ppm		O2 (%)		Gas Flow (SCFH)	
931A1391_CNOX		931A1391.AT		931A1392.AT		931F1374B.N	
22-Oct-20 12:28:00	28.38	22-Oct-20 12:25:00	26.54	22-Oct-20 12:25:00	4.16	22-Oct-20 12:25:00	69,924
22-Oct-20 12:28:00	27.78	22-Oct-20 12:28:00	25.82	22-Oct-20 12:28:00	4.41	22-Oct-20 12:28:00	70,134
22-Oct-20 12:27:00	27.70	22-Oct-20 12:27:00	25.96	22-Oct-20 12:27:00	4.13	22-Oct-20 12:27:00	71,280
22-Oct-20 12:28:00	28.14	22-Oct-20 12:28:00	26.36	22-Oct-20 12:28:00	4.13	22-Oct-20 12:28:00	71,270
22-Oct-20 12:29:00	28.23	22-Oct-20 12:29:00	26.47	22-Oct-20 12:29:00	4.10	22-Oct-20 12:29:00	71,228
22-Oct-20 12:30:00	28.27	22-Oct-20 12:30:00	28.29	22-Oct-20 12:30:00	4.25	22-Oct-20 12:30:00	69,191
22-Oct-20 12:31:00	28.23	22-Oct-20 12:31:00	28.01	22-Oct-20 12:31:00	4.40	22-Oct-20 12:31:00	70,930
22-Oct-20 12:32:00	28.01	22-Oct-20 12:32:00	26.52	22-Oct-20 12:32:00	3.95	22-Oct-20 12:32:00	71,205
22-Oct-20 12:33:00	28.18	22-Oct-20 12:33:00	25.98	22-Oct-20 12:33:00	4.40	22-Oct-20 12:33:00	68,984
22-Oct-20 12:34:00	28.13	22-Oct-20 12:34:00	25.97	22-Oct-20 12:34:00	4.37	22-Oct-20 12:34:00	70,426
22-Oct-20 12:35:00	28.09	22-Oct-20 12:35:00	26.58	22-Oct-20 12:35:00	3.97	22-Oct-20 12:35:00	72,872
22-Oct-20 12:36:00	28.60	22-Oct-20 12:36:00	26.76	22-Oct-20 12:36:00	4.11	22-Oct-20 12:36:00	70,153
22-Oct-20 12:37:00	28.49	22-Oct-20 12:37:00	26.43	22-Oct-20 12:37:00	4.28	22-Oct-20 12:37:00	70,471
22-Oct-20 12:38:00	28.32	22-Oct-20 12:38:00	26.64	22-Oct-20 12:38:00	4.05	22-Oct-20 12:38:00	72,606
22-Oct-20 12:39:00	28.31	22-Oct-20 12:39:00	27.07	22-Oct-20 12:39:00	3.79	22-Oct-20 12:39:00	71,998
22-Oct-20 12:40:00	28.42	22-Oct-20 12:40:00	26.53	22-Oct-20 12:40:00	4.19	22-Oct-20 12:40:00	70,082
22-Oct-20 12:41:00	28.57	22-Oct-20 12:41:00	26.56	22-Oct-20 12:41:00	4.23	22-Oct-20 12:41:00	70,033
22-Oct-20 12:42:00	28.35	22-Oct-20 12:42:00	26.49	22-Oct-20 12:42:00	4.17	22-Oct-20 12:42:00	70,093
22-Oct-20 12:43:00	28.09	22-Oct-20 12:43:00	26.36	22-Oct-20 12:43:00	4.10	22-Oct-20 12:43:00	71,386
22-Oct-20 12:44:00	28.12	22-Oct-20 12:44:00	26.99	22-Oct-20 12:44:00	3.72	22-Oct-20 12:44:00	72,437
22-Oct-20 12:45:00	28.58	22-Oct-20 12:45:00	26.73	22-Oct-20 12:45:00	4.16	22-Oct-20 12:45:00	70,149
22-Oct-20 12:46:00	28.83	22-Oct-20 12:46:00	26.67	22-Oct-20 12:46:00	4.22	22-Oct-20 12:46:00	70,142
22-Oct-20 12:47:00	28.68	22-Oct-20 12:47:00	26.87	22-Oct-20 12:47:00	4.24	22-Oct-20 12:47:00	70,097
22-Oct-20 12:48:00	28.39	22-Oct-20 12:48:00	26.45	22-Oct-20 12:48:00	4.22	22-Oct-20 12:48:00	70,989
22-Oct-20 12:49:00	28.61	22-Oct-20 12:49:00	26.92	22-Oct-20 12:49:00	4.00	22-Oct-20 12:49:00	71,827
22-Oct-20 12:50:00	28.70	22-Oct-20 12:50:00	26.80	22-Oct-20 12:50:00	4.12	22-Oct-20 12:50:00	69,554
22-Oct-20 12:51:00	28.61	22-Oct-20 12:51:00	26.33	22-Oct-20 12:51:00	4.42	22-Oct-20 12:51:00	69,505
22-Oct-20 12:52:00	28.78	22-Oct-20 12:52:00	26.74	22-Oct-20 12:52:00	4.23	22-Oct-20 12:52:00	71,282
22-Oct-20 12:53:00	28.57	22-Oct-20 12:53:00	26.91	22-Oct-20 12:53:00	4.04	22-Oct-20 12:53:00	70,871
22-Oct-20 12:54:00	28.00	22-Oct-20 12:54:00	26.71	22-Oct-20 12:54:00	4.42	22-Oct-20 12:54:00	69,515
22-Oct-20 12:55:00	28.83	22-Oct-20 12:55:00	26.87	22-Oct-20 12:55:00	4.34	22-Oct-20 12:55:00	70,712
22-Oct-20 12:56:00	28.88	22-Oct-20 12:56:00	26.99	22-Oct-20 12:56:00	4.17	22-Oct-20 12:56:00	70,872
22-Oct-20 12:57:00	28.56	22-Oct-20 12:57:00	26.76	22-Oct-20 12:57:00	4.13	22-Oct-20 12:57:00	70,855
22-Oct-20 12:58:00	28.09	22-Oct-20 12:58:00	26.34	22-Oct-20 12:58:00	4.11	22-Oct-20 12:58:00	70,833
22-Oct-20 12:59:00	28.43	22-Oct-20 12:59:00	26.63	22-Oct-20 12:59:00	4.14	22-Oct-20 12:59:00	70,870
22-Oct-20 13:00:00	28.66	22-Oct-20 13:00:00	26.77	22-Oct-20 13:00:00	4.16	22-Oct-20 13:00:00	69,833
22-Oct-20 13:01:00	28.07	22-Oct-20 13:01:00	26.45	22-Oct-20 13:01:00	4.55	22-Oct-20 13:01:00	68,848
22-Oct-20 13:02:00	28.44	22-Oct-20 13:02:00	26.47	22-Oct-20 13:02:00	4.24	22-Oct-20 13:02:00	71,107
22-Oct-20 13:03:00	28.28	22-Oct-20 13:03:00	26.65	22-Oct-20 13:03:00	4.03	22-Oct-20 13:03:00	71,380
22-Oct-20 13:04:00	28.78	22-Oct-20 13:04:00	26.63	22-Oct-20 13:04:00	4.34	22-Oct-20 13:04:00	68,738
22-Oct-20 13:05:00	28.84	22-Oct-20 13:05:00	26.79	22-Oct-20 13:05:00	4.28	22-Oct-20 13:05:00	70,386
22-Oct-20 13:06:00	28.12	22-Oct-20 13:06:00	26.26	22-Oct-20 13:06:00	4.18	22-Oct-20 13:06:00	70,448
22-Oct-20 13:07:00	28.37	22-Oct-20 13:07:00	26.45	22-Oct-20 13:07:00	4.21	22-Oct-20 13:07:00	70,502
22-Oct-20 13:08:00	27.93	22-Oct-20 13:08:00	26.04	22-Oct-20 13:08:00	4.21	22-Oct-20 13:08:00	70,498
22-Oct-20 13:09:00	28.76	22-Oct-20 13:09:00	26.83	22-Oct-20 13:09:00	4.20	22-Oct-20 13:09:00	69,133
22-Oct-20 13:10:00	28.45	22-Oct-20 13:10:00	26.20	22-Oct-20 13:10:00	4.42	22-Oct-20 13:10:00	68,247
22-Oct-20 13:11:00	27.90	22-Oct-20 13:11:00	26.25	22-Oct-20 13:11:00	4.08	22-Oct-20 13:11:00	71,059
22-Oct-20 13:12:00	28.04	22-Oct-20 13:12:00	26.40	22-Oct-20 13:12:00	4.04	22-Oct-20 13:12:00	69,782
22-Oct-20 13:13:00	28.88	22-Oct-20 13:13:00	26.59	22-Oct-20 13:13:00	4.42	22-Oct-20 13:13:00	68,300
22-Oct-20 13:14:00	28.80	22-Oct-20 13:14:00	26.87	22-Oct-20 13:14:00	4.08	22-Oct-20 13:14:00	71,124
22-Oct-20 13:15:00	28.08	22-Oct-20 13:15:00	26.50	22-Oct-20 13:15:00	4.01	22-Oct-20 13:15:00	71,180
22-Oct-20 13:16:00	27.96	22-Oct-20 13:16:00	26.33	22-Oct-20 13:16:00	4.04	22-Oct-20 13:16:00	71,216
22-Oct-20 13:17:00	28.14	22-Oct-20 13:17:00	26.48	22-Oct-20 13:17:00	4.04	22-Oct-20 13:17:00	70,187
22-Oct-20 13:18:00	28.41	22-Oct-20 13:18:00	26.11	22-Oct-20 13:18:00	4.45	22-Oct-20 13:18:00	68,753
22-Oct-20 13:19:00	28.27	22-Oct-20 13:19:00	26.47	22-Oct-20 13:19:00	4.14	22-Oct-20 13:19:00	71,093
22-Oct-20 13:20:00	28.40	22-Oct-20 13:20:00	26.83	22-Oct-20 13:20:00	3.99	22-Oct-20 13:20:00	71,065
22-Oct-20 13:21:00	28.32	22-Oct-20 13:21:00	26.34	22-Oct-20 13:21:00	4.25	22-Oct-20 13:21:00	68,352
22-Oct-20 13:22:00	28.53	22-Oct-20 13:22:00	26.47	22-Oct-20 13:22:00	4.29	22-Oct-20 13:22:00	69,984
22-Oct-20 13:23:00	28.26	22-Oct-20 13:23:00	26.68	22-Oct-20 13:23:00	4.00	22-Oct-20 13:23:00	71,900
22-Oct-20 13:24:00	27.95	22-Oct-20 13:24:00	26.55	22-Oct-20 13:24:00	3.80	22-Oct-20 13:24:00	70,520
22-Oct-20 13:25:00	28.86	22-Oct-20 13:25:00	26.84	22-Oct-20 13:25:00	4.38	22-Oct-20 13:25:00	69,001
<b>Average Stack NOx, ppm (3% O2)</b>	<b>28.26</b>	<b>Average Stack Raw NOx, ppm</b>	<b>26.43</b>	<b>Average Stack O2 (%)</b>	<b>4.15</b>	<b>Average Gas Flow (SCFH)</b>	<b>70,770</b>



# AIR Testing Inc.

## BOILER DATA SHEET

Client: New Indy

Site: OKWARD

Unit: Nebraska Boiler

Date: 10.22.20

Job #: 23002

Lab #: 220-113

Make: Nebraska

S/N: 20-1983

Model: N8E4551

BTU/hr Rating: 108

MMBTU/hr

Fuel Type: NG

FGR Equipped: Yes

Stack Diameter: 44"


"A" Distance: 7'

"B" Distance: 15.5'

Port Size: 2.5"

M or F: F

Offset: 3"



Boiler Load: Run 1: Normal Run 2: \_\_\_\_\_ Run 3: \_\_\_\_\_ % of Rating

FGR Setting: Run 1: 40.2 H2 Run 2: \_\_\_\_\_ Run 3: \_\_\_\_\_

Notes: \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Fuel Meter Data Sheet**

Client:	<u>New Energy</u>	Date:	<u>10-22-20</u>
Site:	<u>Oxnard</u>	Client #:	<u>22030</u>
Unit:	<u>Nebraska Boiler</u>	Lab #:	<u>220-113</u>

Serial #:	<u>71187</u>		
Manufacturer:	<u>Rosemount</u>		
Type of fuel:	Nat Gas <input checked="" type="checkbox"/>	Diesel ( )	Other ( )
Read out in units of:	ccf ( )	Gal ( )	Btu ( )
	mcf ( )	L ( )	Therm ( )
	m3 ( )	ft3 ( )	SCFH <input checked="" type="checkbox"/>
Temperature corrected:	<input checked="" type="checkbox"/> Yes ( ) No ( )	Cor. Temp:	
Pressure corrected:	<input checked="" type="checkbox"/> Yes ( ) No ( )	Cor. Pres.:	

**Run #1** SCFH

Start Time:	<u>12:25</u>	Stop Time:	<u>13:25</u>
Meter Reading:	<u>65368.2</u>	Meter Reading:	<u>63499.4</u>
Fuel temp.:	_____ °F	Elapsed Time:	<u>60 mins</u>
Fuel pres.:	_____ psi	Fuel Usage:	<u>Avg = 64683.8 SCFH / 60 = 1078.06 SCFM</u>

AVG

**Run #2**

Start Time:	_____	Stop Time:	_____
Meter Reading:	_____	Meter Reading:	_____
Fuel temp.:	_____ °F	Elapsed Time:	_____
Fuel pres.:	_____ psi	Fuel Usage:	_____

**Run #3**

Start Time:	_____	Stop Time:	_____
Meter Reading:	_____	Meter Reading:	_____
Fuel temp.:	_____ °F	Elapsed Time:	_____
Fuel pres.:	_____ psi	Fuel Usage:	_____

FGR 40.22 NZ

<b>INSTRUMENT INFORMATION</b>
-------------------------------




## Method 100 – Sample Train Assembly

### Probe:

1. AIRx Testing Services, Inc. uses a stainless steel sampling probe.
2. The probe is connected to the sample conditioner using a heated (if necessary) Teflon sampling line.

### Sample Conditioner (Condenser System):

1. The sample conditioner consists of a moisture knock-out bottle immersed in an ice bath.
2. All parts of the conditioner exposed to the sample are either glass, stainless steel, or Teflon.
3. The sample conditioner is setup so that the sample gas is not bubbled through the condensate.
4. A temperature gauge is used to determine the temperature of the condenser outlet.



### Manifold Panel & Strip Chart Recorder:

1. The sample gas exiting the condenser is transported to the Instrument manifold panel.
2. The sample manifold directs the sample gas through to the various analyzers.
3. All components of the system that contact the sample are either stainless steel, Teflon, or glass.
4. The data is continuously recorded on a Soltec 10" strip chart recorder with a 1% resolution.

### Calibration System:

1. The system is composed of the Analyzer Calibration and the System BIAS Check.
2. Certified Calibration gases are supplied to the analyzers for calibration purposes.
3. The selection of zero or span calibration gases or sample gas is accomplished by operation of selector knobs located on the main flow control panel.

**Nitrogen Oxides:** – Chemiluminescent Analyzer: Based on the chemiluminescent reaction of NO and ozone to form NO<sub>2</sub> in an excited state. Light emission is monitored through an optical filter by a high sensitivity photomultiplier tube, the output of which is electronically processed so it is linearly proportional to the NO concentration. The output is in units of ppmv.

**Carbon Monoxide:** – Non-Dispersive Infrared (NDIR) Analyzer: Radiation from an infrared source is passed through a narrow band-pass filter and a multiple optical pass sample cell where absorption by the sample gas occurs. The infrared radiation exits the sample cell and falls on a solid state infrared detector. The output is in units of ppmv.

**Oxygen:** – Electrochemical Analyzer: Oxygen in the flue gas sample diffuses through a Teflon membrane and is reduced on the surface of the cathode. A corresponding oxidation occurs at the anode and an electric current proportional to the concentration of oxygen is produced. The output is in units of percent O<sub>2</sub> by volume.

**Carbon Dioxide:** – Non Dispersive Infrared (NDIR) Analyzer: The instrument measures the differential infrared energy absorbed from energy beams passed through a reference cell (containing a gas selected to have minimal absorption of infrared energy in the wavelength absorbed by CO<sub>2</sub>) and a sample cell through which the sample gas flows continuously. The output is in units of percent CO<sub>2</sub> by volume.

## SOURCE EMISSION INSTRUMENTATION LIST

### OXIDES OF NITROGEN

<b>Unit No. - 1:</b> Manufacturer: API Model No.: 200 EH Serial No.: 233 Method: Chemiluminescence Range (ppmv) 0-5000	<b>Unit No. - 6:</b> Manufacturer: API Model No.: 200 A Serial No.: 1013 Method: Chemiluminescence Range (ppmv) 0-50
<b>Unit No. - 2:</b> Manufacturer: API Model No.: 200 EH Serial No.: 234 Method: Chemiluminescence Range (ppmv) 0-5000	<b>Unit No. - 7:</b> Manufacturer: Thermo Environmental (TECO) Model No.: 10AR Serial No.: 25559-221 Method: Chemiluminescence Range (ppmv) 0-2.5, 10, 25, 100, 250, 1000, 2500, 10000
<b>Unit No. - 3:</b> Manufacturer: API Model No.: 200 EH Serial No.: 109 Method: Chemiluminescence Range (ppmv) 0-5000	<b>Unit No. - 8:</b> Manufacturer: Thermo Environmental (TECO) Model No.: 10AR Serial No.: 38586-258 Method: Chemiluminescence Range (ppmv) 0-2.5, 10, 25, 100, 250, 1000, 2500, 10000
<b>Unit No. - 4:</b> Manufacturer: API Model No.: 200 EH Serial No.: 442 Method: Chemiluminescence Range (ppmv) 0-5000	
<b>Unit No. - 5:</b> Manufacturer: API Model No.: 200 EH Serial No.: 441 Method: Chemiluminescence Range (ppmv) 0-5000	



## SOURCE EMISSION INSTRUMENTATION LIST

### CARBON MONOXIDE

<b>Unit No. - 1:</b> Manufacturer: API Model No.: 300 EM Serial No.: 239 Method: NDIR/GFC Range (ppmv) 0-5, 10, 25, 50, 100, 250, 500, 1000, 2500, 5000	<b>Unit No. - 2:</b> Manufacturer: API Model No.: 300 EM Serial No.: 240 Method: NDIR/GFC Range (ppmv) 0-5, 10, 25, 50, 100, 250, 500, 1000, 2500, 5000
<b>Unit No. - 3:</b> Manufacturer: API Model No.: 300 EM Serial No.: 104 Method: NDIR/GFC Range (ppmv) 0-5, 10, 25, 50, 100, 250, 500, 1000, 2500, 5000	<b>Unit No - 5:</b> Manufacturer: Thermo Environmental (TECO) Model No.: 48H Serial No.: 25184-219 Method: NDIR/GFC Range (ppmv) 0-50, 100, 200, 500, 1000, 2000, 5000, 10000, 20000, 50000
<b>Unit No. - 6:</b> Manufacturer: Thermo Environmental (TECO) Model No.: 48H Serial No.: 29031-233 Method: NDIR/GFC Range (ppmv) 0-50, 100, 200, 500, 1000, 2000, 5000, 10000, 20000, 50000	<b>Unit No. - 7:</b> Manufacturer: Siemens Model No.: Ultramat 21p Serial No.: AO4-254 Method: NDIR Range (ppmv) 0-300
<b>Unit No. - 10:</b> Manufacturer: Thermo Environmental (TECO) Model No.: 48H Serial No.: 38391-257 Method: NDIR/GFC Range (ppmv) 0-50, 100, 200, 500, 1000, 2000, 5000, 10000, 20000, 50000	<b>Unit No. - 11:</b> Manufacturer: Thermo Environmental (TECO) Model No.: 48H Serial No.: 35226-249 Method: NDIR/GFC Range (ppmv) 0-50, 100, 200, 500, 1000, 2000, 5000, 10000, 20000, 50000
<b>Unit No - 13:</b> Manufacturer: Thermo Environmental (TECO) Model No.: 48 Serial No.: 48-15970-159 Method: NDIR/GFC Range (ppmv) 0-1, 2, 5, 10, 20, 50, 100, 200, 500, 1000	<b>Unit No - 14:</b> Manufacturer: Thermo Environmental (TECO) Model No.: 48 Serial No.: 48-23925-213 Method: NDIR/GFC Range (ppmv) 0-1, 2, 5, 10, 20, 50, 100, 200, 500, 1000



## SOURCE EMISSION INSTRUMENTATION LIST

### OXYGEN

<b>Unit No. - 5:</b> Manufacturer: California Analytical Instruments Model No.: 110P Serial No.: T02034 Method: Paramagnetic Range (%) 0-5, 10, 25	<b>Unit No. - 13:</b> Manufacturer: Servomex Model No.: 1400 Serial No.: X1420/B707 Method: Paramagnetic Range (%) 0-25
<b>Unit No. - 7:</b> Manufacturer: Teledyne Model No.: 320-AX Serial No.: 108742 Method: Electrochemical Range (%) 0-5, 10, 25	
<b>Unit No. - 9:</b> Manufacturer: Servomex Model No.: 1400 Serial No.: 01420/B701/730 Method: Paramagnetic Range (%) 0-25, 100	
<b>Unit No. - 10:</b> Manufacturer: Servomex Model No.: 1400 Serial No.: 01420/B308 Method: Paramagnetic Range (%) 0-25	
<b>Unit No. - 11:</b> Manufacturer: Teledyne Model No.: 320-A Serial No.: 111211 Method: Electrochemical Range (%) 0-5, 10, 25	
<b>Unit No. - 12:</b> Manufacturer: Servomex Model No.: 1400 Serial No.: 01420/B7103 Method: Paramagnetic Range (%) 0-25, 100	



<p><b>QUALITY ASSURANCE</b></p>
---------------------------------





Making our planet more productive

DocNumber: 253580



Praxair Distribution, Inc.
5700 S. Alameda Street
Los Angeles CA 90058
Tel: 323-585-2164
Fax: 714-542-6689
PGVP ID: F22019

CERTIFICATE OF ANALYSIS / EPA PROTOCOL GAS

Customer & Order Information

PRAXAIR PKG OXNARD CA HPS
455 E WOOLEY RD
OXNARD CA 93030

Certificate Issuance Date: 06/27/2019

Praxair Order Number: 70998288
Part Number: NI NO20ME-AS
Customer PO Number: 78981976

Fill Date: 06/07/2019

Lot Number: 70086915805
Cylinder Style & Outlet: AS CGA 560
Cylinder Pressure and Volume: 2000 psig 140 ltr

Certified Concentration

Table with 3 columns: Parameter, Value, and Note. Includes Expiration Date (06/27/2022), Cylinder Number (CC724456), and Certified Concentration (20.1 ppm Nitric oxide, ± 0.5 %).

ProSpec EZ Cert



For Reference Only: NOx 20.2 ppm

Certification Information: Certification Date: 06/27/2019 Term: 36 Months Expiration Date: 06/27/2022

This cylinder was certified according to the 2012 EPA Traceability Protocol, Document #EPA-600/R-12/531, using Procedure G1.
Do Not Use this Standard if Pressure is less than 100 PSIG.

Analytical Data:

(R=Reference Standard, Z=Zero Gas, C=Gas Candidate)

1. Component: Nitric oxide
Requested Concentration: 20 ppm
Certified Concentration: 20.1 ppm
Instrument Used: Thermo Electron 421-LS S/N 1030645077
Analytical Method: Chemiluminescence
Last Multipoint Calibration: 08/10/2019

Reference Standard: Type / Cylinder #: GMIS / CC2919
Concentration / Uncertainty: 19.90 ppm ±0.504%
Expiration Date: 03/28/2022
Traceable to: SRM # / Sample # / Cylinder #: APEX1161149 / NA / APEX1161149
SRM Concentration / Uncertainty: 20.03 ppm / ±0.10 ppm
SRM Expiration Date: 01/27/2020

Table: First Analysis Data. Columns: Z, R, C, Conc. Values: Z: 0, R: 19.9, C: 20, Conc: 20.1

Table: Second Analysis Data. Columns: Z, R, C, Conc. Values: Z: 0, R: 19.89, C: 20.1, Conc: 20.1

Analyzed By

Henry Koung

Certified By

Leeanna Flores

Information contained herein has been prepared at your request by qualified experts within Praxair Distribution, Inc. While we believe that the information is accurate within the limits of the analytical methods employed and is complete to the extent of the specific analyses performed, we make no warranty or representation as to the suitability of the use of the information for any purpose. The information is offered with the understanding that any use of the information is at the sole discretion and risk of the user. In no event shall the liability of Praxair Distribution, Inc., arising out of the use of the information contained herein exceed the fee established for providing such information.



**CERTIFICATE OF ANALYSIS / EPA PROTOCOL GAS**

**Customer & Order Information**

PRAXAIR PKG OXNARD CA HPS  
455 E WOOLEY RD  
OXNARD CA 93075

Certificate Modification Date: 09/18/2018  
Praxair Order Number: 70717073  
Part Number: NI NO40ME-AS

Fill Date: 09/05/2018  
Lot Number: 70088824602  
Cylinder Style & Outlet: AS CGA 680  
Cylinder Pressure and Volume: 2000 psig 140 ft<sup>3</sup>

**Certified Concentration**

Expiration Date:	09/18/2021	NIST Traceable
Cylinder Number:	DT0028765	Expanded Uncertainty
39.9 ppm	Nitric oxide	± 0.8 %
Balance	Nitrogen	

**ProSpec EZ Cert**



**For Reference Only:**

NOx 40.0 ppm

**Certification Information:**

Certification Date: 09/18/2018

Term: 36 Months

Expiration Date: 09/18/2021

This cylinder was certified according to the 2012 EPA Traceability Protocol, Document #EPA-600/R-12/531, using Procedure G1.  
Do Not Use this Standard if Pressure is less than 100 PSIG.

**Analytical Data:**

(R=Reference Standard, Z=Zero Gas, C=Gas Candidate)

1. Component:

Nitric oxide

Requested Concentration: 40 ppm  
Certified Concentration: 39.9 ppm  
Instrument Used: Thermo Electron 42i-LS S/N 1030645077  
Analytical Method: Chemiluminescence  
Last Multipoint Calibration: 09/10/2018

Reference Standard: Type / Cylinder #: GMIS / CC505947  
Concentration / Uncertainty: 49.44 ppm ±0.714%  
Expiration Date: 08/08/2021  
Traceable to: SRM # / Sample # / Cylinder #: 1683b / 45-V-37 / CAL017888  
SRM Concentration / Uncertainty: 48.79 / ±0.34 ppm  
SRM Expiration Date: 03/25/2019

First Analysis Data:		Date	
Z: 0	R: 49.4	C: 39.8	Conc: 39.8
R: 49.4	Z: 0	C: 39.9	Conc: 39.9
Z: 0	C: 39.9	R: 49.4	Conc: 39.9
UOM: ppm		Mean Test Assay: 39.9 ppm	

Second Analysis Data:		Date	
Z: 0	R: 49.4	C: 39.7	Conc: 39.7
R: 49.4	Z: 0	C: 39.9	Conc: 39.9
Z: 0	C: 39.9	R: 49.4	Conc: 39.9
UOM: ppm		Mean Test Assay: 39.9 ppm	

Analyzed By

Henry Koung

Certified By

Amalia Reed



Praxair  
 5700 South Alameda Street  
 Los Angeles, CA 90058  
 Tel: (323) 585-2154 Fax: (714) 542-6689  
 PGVPID: F22017

DocNumber: 000105588

**CERTIFICATE OF ANALYSIS / EPA PROTOCOL GAS**

**Customer & Order Information:**

PRAXAIR PKG OXNARD CA HPS  
 455 E WOOLEY RD  
 OXNARD CA 93030

Praxair Order Number: 70207599  
 Customer P. O. Number:  
 Customer Reference Number:

Fill Date: 2/27/2017  
 Part Number: NI CD803E-AS  
 Lot Number: 109705811  
 Cylinder Style & Outlet: AS CGA 580  
 Cylinder Pressure & Volume: 2000 psig 140 cu. ft.

**Certified Concentration:**

Expiration Date:	3/8/2025	NIST Traceable
Cylinder Number:	CC60235	Analytical Uncertainty:
7.98 %	CARBON DIOXIDE	± 0.4 %
3.98 %	OXYGEN	± 0.4 %
Balance	NITROGEN	

**Certification Information:** Certification Date: 3/8/2017 Term: 96 Months Expiration Date: 3/8/2025  
 This cylinder was certified according to the 2012 EPA Traceability Protocol, Document #EPA-600/R-12/531, using Procedure G1. Do Not Use this Standard if Pressure is less than 100 PSIG.

CO2 responses have been corrected for Oxygen IR Broadening effect. O2 responses have been corrected for CO2 interference.

**Analytical Data:**

(R=Reference Standard, Z=Zero Gas, C=Gas Candidate)

**1. Component: CARBON DIOXIDE**

Requested Concentration: 8 %  
 Certified Concentration: 7.98 %  
 Instrument Used: Horiba VIA-510 S/N 574763031  
 Analytical Method: NDIR  
 Last Multipoint Calibration: 2/13/2017

Reference Standard Type: GMIS  
 Ref. Std. Cylinder #: DT0008914  
 Ref. Std. Conc: 7.00 %  
 Ref. Std. Traceable to SRM #: 1674b  
 SRM Sample #: 7-H-07  
 SRM Cylinder #: FF10631

<b>First Analysis Data:</b>		<b>Date:</b> 3/8/2017	
Z: 0	R: 6.98	C: 7.97	Conc: 7.985
R: 6.98	Z: 0	C: 7.97	Conc: 7.985
Z: 0	C: 7.97	R: 7	Conc: 7.985
UOM: %	Mean Test Assay:		7.985 %

<b>Second Analysis Date:</b>		<b>Date:</b>	
Z: 0	R: 0	C: 0	Conc: 0
R: 0	Z: 0	C: 0	Conc: 0
Z: 0	C: 0	R: 0	Conc: 0
UOM: %	Mean Test Assay:		0 %

**2. Component: OXYGEN**

Requested Concentration: 4 %  
 Certified Concentration: 3.98 %  
 Instrument Used: OXYMAT 5E  
 Analytical Method: PARAMAGNETIC  
 Last Multipoint Calibration: 3/3/2017

Reference Standard Type: GMIS  
 Ref. Std. Cylinder #: CC244935  
 Ref. Std. Conc: 5.01 %  
 Ref. Std. Traceable to SRM #: 2658a  
 SRM Sample #: 72-D-28  
 SRM Cylinder #: CAL016862

<b>First Analysis Data:</b>		<b>Date:</b> 3/8/2017	
Z: 0	R: 5.01	C: 3.97	Conc: 3.975
R: 5	Z: 0	C: 3.97	Conc: 3.975
Z: 0	C: 3.97	R: 5	Conc: 3.975
UOM: %	Mean Test Assay:		3.975 %

<b>Second Analysis Date:</b>		<b>Date:</b>	
Z: 0	R: 0	C: 0	Conc: 0
R: 0	Z: 0	C: 0	Conc: 0
Z: 0	C: 0	R: 0	Conc: 0
UOM: %	Mean Test Assay:		0 %

Analyzed by:

Ying Yu

Ying Yu

Certified by:

Nassim Haddad

Information contained herein has been prepared at your request by qualified experts within Praxair Distribution, Inc. While we believe that the information is accurate within the limits of the analytical methods employed and is complete to the extent of the specific analyses performed, we make no warranty or representation as to the suitability of the use of the information for any purpose. The information is offered with the understanding that any use of the information is at the sole discretion and risk of the user. In no event shall the liability of Praxair Distribution, Inc., arising out of the use of the information contained herein exceed the fee established for providing such information.



**AIR LIQUIDE**

Air Liquide America  
Specialty Gases LLC



**RATA CLASS**

*Guaranteed +/- 1% Accuracy*

8832 DICE ROAD, SANTA FE SPRINGS, CA 90670-2518

Phone: 800-323-2212

Fax: 562-464-5262

**CERTIFICATE OF ACCURACY: EPA Protocol Gas**

Assay Laboratory - PGVP Vendor ID: A52013

Document #: 51261426-004

Customer

AIR LIQUIDE AMERICA SPECIALTY GASES LLC  
8832 DICE ROAD  
SANTA FE SPRINGS, CA 90670-2518

P.O. No.: AIRX

ALHC-HCF-FRESNO-MEDGAS (LOC 24337)  
3703 BAGLEY AVENUE  
TRANSFER ACCOUNT  
FRESNO CA 93725  
US

**ANALYTICAL INFORMATION Gas Type : CO2,O2,BALN**

This certification was performed according to EPA Traceability Protocol For Assay & Certification of Gaseous Calibration Standards; Procedure G-1.

EPA/600/R-12/531; May 2012. Do not use this standard if pressure is less than 100 psig.

Cylinder Number: CC160530 Certification Date: 02Jul2013 Exp. Date: 03Jul2021  
Cylinder Pressure: 2000 PSIG Batch No: SBO0073281

COMPONENT	CERTIFIED CONCENTRATION (Moles)		ACCURACY ABSOLUTE / RELATIVE			
	Value	%	Absolute	%	Relative	%
CARBON DIOXIDE	20.1	%	.13	%	1	.6 %
OXYGEN	8.05	%	.06	%	1	.7 %
NITROGEN	BALANCE					

**TRACEABILITY**

**REFERENCE STANDARD**

COMPONENT	CONCENTRATION	UNCERTAINTY	CYLINDER	TYPE/SRM SAMPLE	EXP. DATE
CARBON DIOXIDE	17.87 %	0.110 %	K026459	NTRM 1800/	01Mar2019
OXYGEN	10.03 %	0.070 %	K027803	NTRM 2858/	01Feb2016

**ANALYTICAL METHOD**

1st Analysis: 02Jul2013

COMPONENT	INSTRUMENT	ANALYTICAL PRINCIPLE	CALIBRATED	CONCENTRATION
CARBON DIOXIDE	GC-TCD/HP 6890 II/3140A38374	GC	20Jun2013	20.11 %
OXYGEN	GC-TCD/HP 5890 II/3140A38374	GC	22Jun2013	8.051 %

APPROVED BY: \_\_\_\_\_

*[Signature]*  
DC



Praxair  
 5700 South Alameda Street  
 Los Angeles, CA 90058  
 Tel: (323) 585-2154 Fax: (714) 542-6689  
 PGVPID: F22016

DocNumber: 000094146

**CERTIFICATE OF ANALYSIS / EPA PROTOCOL GAS**

**Customer & Order Information:**

PRAXAIR WHSE OXNARD CA  
 455 E WOOLEY RD  
 OXNARD CA 930300

Praxair Order Number: 34250102  
 Customer P. O. Number: 06017500  
 Customer Reference Number:

Fill Date: 5/5/2016  
 Part Number: NI CO20ME-AS  
 Lot Number: 109612606  
 Cylinder Style & Outlet: AS CGA 350  
 Cylinder Pressure & Volume: 2000 psig 140 cu. ft.

**Certified Concentration:**

Expiration Date:	6/13/2024	NIST Traceable
Cylinder Number:	DT0012296	Analytical Uncertainty:
20.5 ppm	CARBON MONOXIDE	± 0.4 %
Balance	NITROGEN	

**Certification Information:** Certification Date: 5/13/2016 Term: 96 Months Expiration Date: 5/13/2024  
 This cylinder was certified according to the 2012 EPA Traceability Protocol, Document #EPA-600/R-12/531, using Procedure G1. Do Not Use this Standard if Pressure is less than 100 PSIG.

**Analytical Data:** (R=Reference Standard, Z=Zero Gas, C=Gas Candidate)

1. Component: CARBON MONOXIDE

Requested Concentration: 20 ppm  
 Certified Concentration: 20.5 ppm  
 Instrument Used: Horiba VIA-510 S/N 576876015  
 Analytical Method: NDIR  
 Last Multipoint Calibration: 5/3/2016

Reference Standard Type: GMIS  
 Ref. Std. Cylinder #: CC341229  
 Ref. Std. Conc: 25.04 ppm  
 Ref. Std. Traceable to SRM #: 2835a  
 SRM Sample #: 68-E-34  
 SRM Cylinder #: FF10868

First Analysis Data:		Date: 5/13/2016	
Z:	0	R:	25
C:	20.4	Conc:	20.351
R:	25.1	Z:	0
C:	20.5	Conc:	20.451
Z:	0	C:	20.6
R:	25.2	Conc:	20.551
UOM:	ppm	Mean Test Assay:	20.451 ppm

Second Analysis Data:		Date:	
Z:	0	R:	0
C:	0	Conc:	0
R:	0	Z:	0
C:	0	Conc:	0
Z:	0	C:	0
R:	0	Conc:	0
UOM:	ppm	Mean Test Assay:	0 ppm

Analyzed by:

Ying Yu

Certified by:

Maria Soberanis

Information contained herein has been prepared at your request by qualified experts within Praxair Distribution, Inc. While we believe that the information is accurate within the limits of the analytical methods employed and is complete to the extent of the specific analyses performed, we make no warranty or representation as to the suitability of the use of the information for any purpose. The information is offered with the understanding that any use of the information is at the sole discretion and risk of the user. In no event shall the liability of Praxair Distribution, Inc., arising out of the use of the information contained herein exceed the fee established for providing such information.





Praxair  
 5700 South Alameda Street  
 Los Angeles, CA 90058  
 Tel: (323) 585-2154 Fax: (714) 542-6689  
 PGPVID: F22016

DocNumber: 000094152

**CERTIFICATE OF ANALYSIS / EPA PROTOCOL GAS**

**Customer & Order Information:**

PRAXAIR WHSE OXNARD CA  
 455 E WOOLEY RD  
 OXNARD CA 93030

Praxair Order Number: 34250113  
 Customer P. O. Number: 06017502  
 Customer Reference Number:

Fill Date: 5/5/2016  
 Part Number: NI CO40ME-AS  
 Lot Number: 109612607  
 Cylinder Style & Outlet: AS CGA 350  
 Cylinder Pressure & Volume: 2000 psig 140 cu. ft.

**Certified Concentration:**

Expiration Date:	5/13/2024	NIST Traceable
Cylinder Number:	CC111187	Analytical Uncertainty:
40.4 ppm	CARBON MONOXIDE	± 0.8 %
Balance	NITROGEN	

**Certification Information:** Certification Date: 5/13/2016 Term: 96 Months Expiration Date: 5/13/2024

This cylinder was certified according to the 2012 EPA Traceability Protocol, Document #EPA-600/R-12/531, using Procedure G1. Do Not Use this Standard if Pressure is less than 100 PSIG.

**Analytical Data:**

(R=Reference Standard, Z=Zero Gas, C=Gas Candidate)

**1. Component: CARBON MONOXIDE**

Requested Concentration: 40 ppm  
 Certified Concentration: 40.4 ppm  
 Instrument Used: Horiba VIA-S10 S/N 576876015  
 Analytical Method: NDIR  
 Last Multiport Calibration: 5/13/2016

Reference Standard Type: GMIS  
 Ref. Std. Cylinder #: CC341229  
 Ref. Std. Conc: 25.04 ppm  
 Ref. Std. Traceable to SRM #: 2635a  
 SRM Sample #: 58-E-34  
 SRM Cylinder #: FF10666

First Analysis Data:		Date:	5/13/2016	
Z:	0	R:	25	C: 40.2 Conc: 40.589
R:	24.7	Z:	0	C: 40 Conc: 40.387
Z:	0	C:	39.9	R: 24.7 Conc: 40.286
UOM:	ppm	Mean Test Assay:	40.421 ppm	

Second Analysis Data:		Date:		
Z:	0	R:	0	C: 0 Conc: 0
R:	0	Z:	0	C: 0 Conc: 0
Z:	0	C:	0	R: 0 Conc: 0
UOM:	ppm	Mean Test Assay:	0 ppm	

Analyzed by:

Maria Soberanis

Certified by:

Ying Yu

Information contained herein has been prepared at your request by qualified experts within Praxair Distribution, Inc. While we believe that the information is accurate within the limits of the analytical methods employed and is complete to the extent of the specific analyses performed, we make no warranty or representation as to the suitability of the use of the information for any purpose. The information is offered with the understanding that any use of the information is at the sole discretion and risk of the user. In no event shall the liability of Praxair Distribution, Inc., arising out of the use of the information contained herein exceed the fee established for providing such information.

**AIRx TESTING SERVICES, INC.**

**RATA EMISSIONS TESTING  
NEW INDY OXNARD  
N8E9551 – NEBRASKA BOILER  
NOx Analyzer - Horiba VA-5000 (New)  
VCAPCD PTO #0157-281  
Test Date: October 21, 2020**

**Prepared for:  
NEW INDY OXNARD  
5936 Perkins Road,  
Oxnard, CA 93033**

**Attn: Robyn Lebrilla**

**Submitted to:  
Ventura County Air Pollution Control District  
669 County Square Drive  
Ventura, California 93003**

**Attn: Ed Swede**

**Prepared by:  
AIRx Testing Services, Inc.  
2472 Eastman Avenue #34  
Ventura, CA 93003**

**Job No.: 23022**

**Lab No.: 220-112**



**Ken Kennepohl; Senior Engineer**

**Tom Porter; Vice President of Testing Services**

**Submitted:  
November 3, 2020**

Table of Contents

<u>Section</u>	<u>Page</u>
RATA Results	1
Introduction	2
Calculations	5
RATA Bias Adjustments	6
CEM Certification	8
Nebraska New Analyzer Certification	10
Field Data and Strip Charts	13
RATA CEM Data	14
Strip Charts	24
Facility CEM Data Sheets	29
Boiler Data Sheet	38
EPA Method 20 Response Time	39
Description of Methods Used	40
EPA Method 7	41
CARB Method 100	42
Instrument Information	44
Extractive Gas Sample System	45
CEM Instrument List	46
Quality Assurance	48
Calibration Gas Certification Sheets	49

# AIR TESTING SERVICES, INC.

New Indy  
Nebraska Boiler New NOx  
10/21/2020

CEMS RATA  
Calculations

Run	AIRx Testing - Reference Method		
	NOx ppmv	O2 Dry %	NOx ppmv @ 3% O2
1	23.37	4.21	25.07
2	23.80	4.28	25.63
3	24.36	4.50	26.58
4	23.96	4.45	26.07
5	23.90	4.57	26.19
6	23.74	4.67	26.19
7	23.33	4.70	25.78
8	23.02	4.73	25.48
9	22.37	4.70	24.71

Run	New Indy	Nebraska	CEMS
	NOx ppmv	O2 Dry %	NOx ppmv @ 3%
1	24.11	4.73	26.69
2	24.11	4.78	26.78
3	24.19	4.82	26.94
4	23.98	4.89	26.81
5	24.27	4.92	27.18
6	24.40	4.94	27.37
7	24.26	4.83	27.03
8	24.33	4.77	27.01
9	24.21	4.82	26.95

Run	Reference Method - CEM, Difference		
	NOx ppmv	O2 Dry %	NOx ppmv @ 3%
1	-0.74	-0.52	-1.62
2	-0.31	-0.50	-1.15
3	0.17	-0.32	-0.36
4	-0.02	-0.44	-0.74
5	-0.37	-0.35	-0.99
6	-0.66	-0.27	-1.18
7	-0.93	-0.13	-1.25
8	-1.31	-0.04	-1.53
9	-1.84	-0.12	-2.24

Arithmetic Mean, d  
Standard Deviation, Sd  
Confidence Coefficient, CC  
Avg Reference Method, RM  
Relative Accuracy, RA

-0.67	-0.30	-1.23
0.63	0.17	0.54
0.49	0.13	0.41
23.54	4.53	25.74
4.91	9.55	6.38

NOTE: Calculations based on "Code of Federal Regulations 40", 1988, Part 60, Appendix B, Specification 2, p. 939.

**INTRODUCTION**

## 1.0 INTRODUCTION

On October 21, 2020 AIRx Testing Services conducted the RATA source test on a natural gas fired 108/MMBTU Nebraska Boiler N8E9551. The unit is located at the New Indy Plant at 5936 Perkins Road, in Oxnard. The boiler is used for the generation of steam for specific facility operations. The testing was required to satisfy requirements detailed in PTO #0157-281.

The objective of the testing was to determine the relative accuracy (RATA) of the facility's boiler CEM equipment. Operational data reported by New Indy personnel included NO<sub>x</sub> readings from their new Horiba VA-5000 analyzer and O<sub>2</sub> readings from their Rosemont analyzer.

**PROCESS DETAILS:** The Nebraska Boiler is rated for 108 MMBtu/hr heat input and exhausts into 44" exhaust stack. The sample was taken off the existing facility CEM system NO<sub>x</sub>/O<sub>2</sub> probe with it being located at A and B distances of 360" and 26" respectively (8.2 diameters upstream and 0.6 diameters downstream).

**CEM DATA:** CEM data was recorded by the Data Acquisition System (DAS) during the test program in the form of one (1) minute averages. The one (1) minute data readings were then printed out and averaged to obtain data for comparison to New Indy's CEM data. The facility CEM data was obtained by plant personnel during the test program and provided to the test team. The CEM data monitored NO<sub>x</sub> and O<sub>2</sub> concentrations and also correcting the NO<sub>x</sub> to 3% oxygen.

**CONTINUOUS MONITORS:** NO<sub>x</sub> and O<sub>2</sub> sampling utilized a stainless steel sampling probe connected with a Teflon sampling line to a sample conditioner. The dry gases were then transported through a Teflon line to the sample gas flow panel for distribution to the individual analyzers. Three (3) 63 minute sampling runs for NO<sub>x</sub> and O<sub>2</sub> were performed in accordance with EPA Method 7e and CARB Method 100. NO<sub>x</sub> measurements were made utilizing an API 200EH chemiluminescent analyzer. Oxygen concentration was measured utilizing a Servomex Model 1400 paramagnetic analyzer. Initial three (3) point calibrations were performed on the analyzers. Subsequent calibrations were performed through the sample system probe tip (Bias checks). All initial bias checks correlated with the initial instrument calibrations to within 2% of analyzer range. Bias calibrations were performed before and after each test run and used to correct the emissions data for any analyzer drift. All monitor calibrations were performed with EPA Protocol 1 calibration gases.

During all tests, the NO<sub>x</sub> and O<sub>2</sub> RM analyzers were operated on the 50 ppmv and 10% scales, respectively.

**TEST DAY DISCUSSION:** Initial certification of the facilities' new Horiba NO<sub>x</sub> analyzer was conducted on 10-20-2020 using EPA method 7e. It was determined that the analyzer met all % error limits required by the method. These results can be found in the "Calculations" section of the report. On 10-21-2020, three (3) 63 minute runs were performed and broken into three (3) 21 minute rata test runs. This is acceptable according to the test methodology and by Ed Swede with the VCAPCD. The three test runs were conducted while the boiler was operating under normal conditions.

**1.0 INTRODUCTION (cont)**

All monitored data was continuously recorded on a 10" analog chart recorder. Chart speed was maintained at 10 cm/hr. Analyzer output was also recorded with a PC based Data Acquisition System (DAS) and imported into spreadsheets for determination of average run concentrations.

All reporting and calculations have been performed using VCAPCD standard conditions of 68°F and 29.92 inches of Hg. If you have any questions regarding this test program or report, please contact the undersigned at (805) 644-1099.

Respectfully Submitted  
AIRx Testing Services, Inc.



Ken Kennepohl  
Senior Engineer

CC: Ventura County APCD  
Attn: Ed Swede

Reviewed by:



Tom Porter  
Vice President of Testing Services

<b>CALCULATIONS</b>
---------------------



## RATA Bias Adjustment

Facility: New Indy  
 Source: Nebraska Boiler New NOx  
 Date: 10/21/20

### Run No. 1

Parameter	Measured Conc. (ppm,%)	Reference Span gas (ppm,%)	Initial Zero (ppm,%)	Final Zero (ppm,%)	Average Zero (ppm,%)	Initial Span (ppm,%)	Final Span (ppm,%)	Average Span (ppm,%)	Adjusted Conc. (ppm,%)
NOx	23.37	20.2	0.0	0.0	0.0	20.3	20.2	20.2	<b>23.37</b>
O2	4.31	3.98	0.0	0.1	0.0	4.0	4.0	4.1	<b>4.21</b>

### Run No. 2

Parameter	Measured Conc. (ppm,%)	Reference Span gas (ppm,%)	Initial Zero (ppm,%)	Final Zero (ppm,%)	Average Zero (ppm,%)	Initial Span (ppm,%)	Final Span (ppm,%)	Average Span (ppm,%)	Adjusted Conc. (ppm,%)
NOx	23.80	20.2	0.0	0.0	0.0	20.3	20.2	20.2	<b>23.80</b>
O2	4.30	3.98	0.0	0.0	0.0	4.0	4.0	4.0	<b>4.28</b>

### Run No. 3

Parameter	Measured Conc. (ppm,%)	Reference Span gas (ppm,%)	Initial Zero (ppm,%)	Final Zero (ppm,%)	Average Zero (ppm,%)	Initial Span (ppm,%)	Final Span (ppm,%)	Average Span (ppm,%)	Adjusted Conc. (ppm,%)
NOx	24.36	20.2	0.0	0.0	0.0	20.3	20.2	20.2	<b>24.36</b>
O2	4.51	3.98	0.0	0.0	0.0	4.0	4.0	4.0	<b>4.50</b>

### Run No. 4

Parameter	Measured Conc. (ppm,%)	Reference Span gas (ppm,%)	Initial Zero (ppm,%)	Final Zero (ppm,%)	Average Zero (ppm,%)	Initial Span (ppm,%)	Final Span (ppm,%)	Average Span (ppm,%)	Adjusted Conc. (ppm,%)
NOx	23.93	20.2	0.0	0.0	0.0	20.2	20.2	20.2	<b>23.96</b>
O2	4.46	3.98	0.0	0.0	0.0	4.0	4.0	4.0	<b>4.45</b>

### Run No. 5

Parameter	Measured Conc. (ppm,%)	Reference Span gas (ppm,%)	Initial Zero (ppm,%)	Final Zero (ppm,%)	Average Zero (ppm,%)	Initial Span (ppm,%)	Final Span (ppm,%)	Average Span (ppm,%)	Adjusted Conc. (ppm,%)
NOx	23.86	20.2	0.0	0.0	0.0	20.2	20.2	20.2	<b>23.90</b>
O2	4.58	3.98	0.0	0.0	0.0	4.0	4.0	4.0	<b>4.57</b>

## RATA Bias Adjustment

Facility: New Indy  
 Source: Nebraska Boiler New N  
 Date: 10/21/20

### Run No. 6

Parameter	Measured Conc. (ppm.%)	Reference Span gas (ppm.%)	Initial Zero (ppm.%)	Final Zero (ppm.%)	Average Zero (ppm.%)	Initial Span (ppm.%)	Final Span (ppm.%)	Average Span (ppm.%)	Adjusted Conc. (ppm.%)
NOx	23.71	20.2	0.0	0.0	0.0	20.2	20.2	20.2	23.74
O2	4.68	3.98	0.0	0.0	0.0	4.0	4.0	4.0	4.67

### Run No. 7

Parameter	Measured Conc. (ppm.%)	Reference Span gas (ppm.%)	Initial Zero (ppm.%)	Final Zero (ppm.%)	Average Zero (ppm.%)	Initial Span (ppm.%)	Final Span (ppm.%)	Average Span (ppm.%)	Adjusted Conc. (ppm.%)
NOx	23.35	20.2	0.0	0.0	0.0	20.2	20.3	20.2	23.33
O2	4.70	3.98	0.0	0.0	0.0	4.0	4.0	4.0	4.70

### Run No. 8

Parameter	Measured Conc. (ppm.%)	Reference Span gas (ppm.%)	Initial Zero (ppm.%)	Final Zero (ppm.%)	Average Zero (ppm.%)	Initial Span (ppm.%)	Final Span (ppm.%)	Average Span (ppm.%)	Adjusted Conc. (ppm.%)
NOx	23.05	20.2	0.0	0.0	0.0	20.2	20.3	20.2	23.02
O2	4.73	3.98	0.0	0.0	0.0	4.0	4.0	4.0	4.73

### Run No. 9

Parameter	Measured Conc. (ppm.%)	Reference Span gas (ppm.%)	Initial Zero (ppm.%)	Final Zero (ppm.%)	Average Zero (ppm.%)	Initial Span (ppm.%)	Final Span (ppm.%)	Average Span (ppm.%)	Adjusted Conc. (ppm.%)
NOx	22.39	20.2	0.0	0.0	0.0	20.2	20.3	20.2	22.37
O2	4.70	3.98	0.0	0.0	0.0	4.0	4.0	4.0	4.70



### C.E.M. CERTIFICATION (Nox)

Facility: New Indy

Date: 10/20/2020

Job No: 220-112

Analyzer: Nox

Source: Nebraska Boiler

Analyzer: Horiba VA-5000

Analyzer Range: 0-100

Cylinder #: DT0028766 Nox  
Conc: 40.0

Cylinder #: CC3240 NO2  
Conc: 19.01

Cylinder #: CC105295 Conc: 81.43

High	% Diff	Cal Gas	Analyzer Reading	Start	Stop
1	-0.074	81.43	81.37	8:50	8:55
2	1.01	81.43	82.25	9:07	9:09
3	0.442	81.43	81.79	9:23	9:25

Low	% Diff	Cal Gas	Analyzer Reading	Start	Stop
1	0.225	40.0	40.09	8:58	9:01
2	-0.375	40.0	39.85	9:11	9:12
3	0.00	40.0	40.00	9:26	9:32

Zero	% Diff	Cal Gas	Analyzer Reading	Start	Stop
1	-0.08	0.0	-0.08	8:43	8:45
2	-0.10	0.0	-0.10	9:04	9:06
3	-0.09	0.0	-0.09	9:15	9:16

#### NO2 to NO Converter Efficiency (%)

NO2	Cal Gas	Analyzer Reading	Efficiency (%)	Time
1	19.01	17.45	91.8	10:22
2	19.01	17.60	92.6	10:33
3	19.01	17.62	92.7	10:44

$\% \text{ Diff} = ((\text{Cylinder} - \text{Analyzer Reading}) / \text{Cylinder}) * 100$

# NEBRASKA NEW ANALYZER CERTIFICATION

10/20/2020 8:30 10/20/2020 9:15 1 m

Raw NOx, ppm		O2 (%)		Gas Flow (SCFH)	
931AI393.AT		931AI392.AT		931FI374B.ft	
20-Oct-20 08:30:00	23.31	20-Oct-20 08:30:00	3.98	20-Oct-20 08:30:00	62,641
20-Oct-20 08:31:00	23.17	20-Oct-20 08:31:00	3.82	20-Oct-20 08:31:00	64,200
20-Oct-20 08:32:00	23.31	20-Oct-20 08:32:00	3.54	20-Oct-20 08:32:00	65,158
20-Oct-20 08:33:00	23.46	20-Oct-20 08:33:00	3.70	20-Oct-20 08:33:00	63,394
20-Oct-20 08:34:00	22.04	20-Oct-20 08:34:00	3.86	20-Oct-20 08:34:00	62,731
20-Oct-20 08:35:00	21.86	20-Oct-20 08:35:00	3.83	20-Oct-20 08:35:00	62,674
20-Oct-20 08:36:00	21.74	20-Oct-20 08:36:00	3.99	20-Oct-20 08:36:00	61,322
20-Oct-20 08:37:00	23.08	20-Oct-20 08:37:00	4.07	20-Oct-20 08:37:00	61,781
20-Oct-20 08:38:00	4.95	20-Oct-20 08:38:00	3.78	20-Oct-20 08:38:00	63,284
20-Oct-20 08:39:00	0.03	20-Oct-20 08:39:00	3.64	20-Oct-20 08:39:00	64,802
20-Oct-20 08:40:00	-0.03	20-Oct-20 08:40:00	3.54	20-Oct-20 08:40:00	64,419
20-Oct-20 08:41:00	-0.06	20-Oct-20 08:41:00	3.76	20-Oct-20 08:41:00	62,339
20-Oct-20 08:42:00	-0.05	20-Oct-20 08:42:00	3.81	20-Oct-20 08:42:00	62,264
20-Oct-20 08:43:00	-0.08	20-Oct-20 08:43:00	3.79	20-Oct-20 08:43:00	64,183
20-Oct-20 08:44:00	-0.04	20-Oct-20 08:44:00	4.15	20-Oct-20 08:44:00	65,035
20-Oct-20 08:45:00	-0.08	20-Oct-20 08:45:00	4.51	20-Oct-20 08:45:00	64,263
20-Oct-20 08:46:00	-0.07	20-Oct-20 08:46:00	4.72	20-Oct-20 08:46:00	62,803
20-Oct-20 08:47:00	16.74	20-Oct-20 08:47:00	4.95	20-Oct-20 08:47:00	62,001
20-Oct-20 08:48:00	79.87	20-Oct-20 08:48:00	4.97	20-Oct-20 08:48:00	82,448
20-Oct-20 08:49:00	81.27	20-Oct-20 08:49:00	4.73	20-Oct-20 08:49:00	64,135
20-Oct-20 08:50:00	81.37	20-Oct-20 08:50:00	4.54	20-Oct-20 08:50:00	63,977
20-Oct-20 08:51:00	81.34	20-Oct-20 08:51:00	4.66	20-Oct-20 08:51:00	62,811
20-Oct-20 08:52:00	59.08	20-Oct-20 08:52:00	4.90	20-Oct-20 08:52:00	61,672
20-Oct-20 08:53:00	23.67	20-Oct-20 08:53:00	4.78	20-Oct-20 08:53:00	63,464
20-Oct-20 08:54:00	32.48	20-Oct-20 08:54:00	4.52	20-Oct-20 08:54:00	65,285
20-Oct-20 08:55:00	40.62	20-Oct-20 08:55:00	4.48	20-Oct-20 08:55:00	64,433
20-Oct-20 08:56:00	39.60	20-Oct-20 08:56:00	4.76	20-Oct-20 08:56:00	62,541
20-Oct-20 08:57:00	39.52	20-Oct-20 08:57:00	4.88	20-Oct-20 08:57:00	62,728
20-Oct-20 08:58:00	40.13	20-Oct-20 08:58:00	4.72	20-Oct-20 08:58:00	64,616
20-Oct-20 08:59:00	40.11	20-Oct-20 08:59:00	4.48	20-Oct-20 08:59:00	64,976
20-Oct-20 09:00:00	40.08	20-Oct-20 09:00:00	4.71	20-Oct-20 09:00:00	63,214
20-Oct-20 09:01:00	40.09	20-Oct-20 09:01:00	4.91	20-Oct-20 09:01:00	62,519
20-Oct-20 09:02:00	21.12	20-Oct-20 09:02:00	4.79	20-Oct-20 09:02:00	64,097
20-Oct-20 09:03:00	0.08	20-Oct-20 09:03:00	4.50	20-Oct-20 09:03:00	65,255
20-Oct-20 09:04:00	-0.10	20-Oct-20 09:04:00	4.52	20-Oct-20 09:04:00	65,202
20-Oct-20 09:05:00	-0.13	20-Oct-20 09:05:00	4.55	20-Oct-20 09:05:00	65,225
20-Oct-20 09:06:00	57.61	20-Oct-20 09:06:00	4.57	20-Oct-20 09:06:00	65,263
20-Oct-20 09:07:00	82.257	20-Oct-20 09:07:00	4.67	20-Oct-20 09:07:00	64,984
20-Oct-20 09:08:00	82.54	20-Oct-20 09:08:00	4.79	20-Oct-20 09:08:00	64,248
20-Oct-20 09:09:00	82.62	20-Oct-20 09:09:00	4.80	20-Oct-20 09:09:00	63,740
20-Oct-20 09:10:00	63.08	20-Oct-20 09:10:00	4.93	20-Oct-20 09:10:00	63,040
20-Oct-20 09:11:00	39.72	20-Oct-20 09:11:00	4.72	20-Oct-20 09:11:00	65,088
20-Oct-20 09:12:00	39.65	20-Oct-20 09:12:00	4.73	20-Oct-20 09:12:00	62,463
20-Oct-20 09:13:00	19.95	20-Oct-20 09:13:00	5.24	20-Oct-20 09:13:00	55,338
20-Oct-20 09:14:00	0.07	20-Oct-20 09:14:00	5.80	20-Oct-20 09:14:00	52,911
<b>Average Stack Raw NOx, ppm</b>		<b>Average Stack O2 (%)</b>		<b>Average Gas Flow (SCFH)</b>	
28.59		4.31		63,394	

① ZERO

② High

① LOW

② ZERO

③ High

④ LOW

# NEBRASKA NEW ANALYZER CERTIFICATION

10/20/2020 9:15 10/20/2020 10:00 1 m

③ 2=LO  
High  
LOW  
RESPONSE TIME NOx  
①  
②  
③

Raw NOx, ppm		O2 (%)		Gas Flow (SCFH)	
931AI393.AT		931AI392.AT		931FI374B.ft	
20-Oct-20 09:15:00	-0.09	20-Oct-20 09:15:00	6.01	20-Oct-20 09:15:00	50,336
20-Oct-20 09:16:00	-0.15	20-Oct-20 09:16:00	6.30	20-Oct-20 09:16:00	48,620
20-Oct-20 09:17:00	-0.16	20-Oct-20 09:17:00	6.33	20-Oct-20 09:17:00	49,037
20-Oct-20 09:18:00	-0.15	20-Oct-20 09:18:00	5.71	20-Oct-20 09:18:00	52,449
20-Oct-20 09:19:00	50.28	20-Oct-20 09:19:00	5.11	20-Oct-20 09:19:00	53,573
20-Oct-20 09:20:00	82.89	20-Oct-20 09:20:00	4.85	20-Oct-20 09:20:00	55,897
20-Oct-20 09:21:00	83.15	20-Oct-20 09:21:00	4.45	20-Oct-20 09:21:00	56,398
20-Oct-20 09:22:00	81.86	20-Oct-20 09:22:00	4.78	20-Oct-20 09:22:00	53,791
20-Oct-20 09:23:00	81.79	20-Oct-20 09:23:00	5.62	20-Oct-20 09:23:00	50,532
20-Oct-20 09:24:00	81.78	20-Oct-20 09:24:00	5.88	20-Oct-20 09:24:00	51,582
20-Oct-20 09:25:00	80.46	20-Oct-20 09:25:00	5.20	20-Oct-20 09:25:00	54,819
20-Oct-20 09:26:00	40.22	20-Oct-20 09:26:00	4.81	20-Oct-20 09:26:00	55,260
20-Oct-20 09:27:00	39.42	20-Oct-20 09:27:00	4.83	20-Oct-20 09:27:00	55,221
20-Oct-20 09:28:00	39.84	20-Oct-20 09:28:00	4.98	20-Oct-20 09:28:00	53,818
20-Oct-20 09:29:00	40.14	20-Oct-20 09:29:00	5.35	20-Oct-20 09:29:00	51,608
20-Oct-20 09:30:00	40.00	20-Oct-20 09:30:00	5.99	20-Oct-20 09:30:00	49,074
20-Oct-20 09:31:00	40.00	20-Oct-20 09:31:00	6.22	20-Oct-20 09:31:00	48,488
20-Oct-20 09:32:00	39.84	20-Oct-20 09:32:00	6.26	20-Oct-20 09:32:00	48,034
20-Oct-20 09:33:00	22.37	20-Oct-20 09:33:00	6.24	20-Oct-20 09:33:00	48,682
20-Oct-20 09:34:00	0.12	20-Oct-20 09:34:00	5.41	20-Oct-20 09:34:00	53,954
20-Oct-20 09:35:00	18.11	20-Oct-20 09:35:00	4.33	20-Oct-20 09:35:00	56,846
20-Oct-20 09:36:00	81.22	20-Oct-20 09:36:00	4.16	20-Oct-20 09:36:00	56,866
20-Oct-20 09:37:00	81.72	20-Oct-20 09:37:00	4.27	20-Oct-20 09:37:00	56,537
20-Oct-20 09:38:00	48.55	20-Oct-20 09:38:00	4.44	20-Oct-20 09:38:00	55,668
20-Oct-20 09:39:00	0.35	20-Oct-20 09:39:00	4.83	20-Oct-20 09:39:00	55,089
20-Oct-20 09:40:00	0.33	20-Oct-20 09:40:00	4.76	20-Oct-20 09:40:00	54,360
20-Oct-20 09:41:00	19.69	20-Oct-20 09:41:00	4.72	20-Oct-20 09:41:00	55,287
20-Oct-20 09:42:00	79.86	20-Oct-20 09:42:00	4.38	20-Oct-20 09:42:00	58,414
20-Oct-20 09:43:00	81.41	20-Oct-20 09:43:00	4.24	20-Oct-20 09:43:00	58,785
20-Oct-20 09:44:00	80.19	20-Oct-20 09:44:00	4.26	20-Oct-20 09:44:00	58,704
20-Oct-20 09:45:00	10.66	20-Oct-20 09:45:00	4.49	20-Oct-20 09:45:00	56,337
20-Oct-20 09:46:00	0.03	20-Oct-20 09:46:00	5.09	20-Oct-20 09:46:00	53,465
20-Oct-20 09:47:00	-0.12	20-Oct-20 09:47:00	5.39	20-Oct-20 09:47:00	52,148
20-Oct-20 09:48:00	28.56	20-Oct-20 09:48:00	5.72	20-Oct-20 09:48:00	50,918
20-Oct-20 09:49:00	80.62	20-Oct-20 09:49:00	5.91	20-Oct-20 09:49:00	49,933
20-Oct-20 09:50:00	81.24	20-Oct-20 09:50:00	6.14	20-Oct-20 09:50:00	49,032
20-Oct-20 09:51:00	52.34	20-Oct-20 09:51:00	6.29	20-Oct-20 09:51:00	48,208
20-Oct-20 09:52:00	0.37	20-Oct-20 09:52:00	6.08	20-Oct-20 09:52:00	50,789
20-Oct-20 09:53:00	-0.10	20-Oct-20 09:53:00	4.86	20-Oct-20 09:53:00	56,116
20-Oct-20 09:54:00	-0.16	20-Oct-20 09:54:00	4.18	20-Oct-20 09:54:00	58,618
20-Oct-20 09:55:00	2.62	20-Oct-20 09:55:00	4.15	20-Oct-20 09:55:00	58,759
20-Oct-20 09:56:00	13.58	20-Oct-20 09:56:00	4.16	20-Oct-20 09:56:00	58,740
20-Oct-20 09:57:00	12.34	20-Oct-20 09:57:00	4.30	20-Oct-20 09:57:00	58,763
20-Oct-20 09:58:00	0.01	20-Oct-20 09:58:00	4.48	20-Oct-20 09:58:00	59,123
20-Oct-20 09:59:00	-0.27	20-Oct-20 09:59:00	4.40	20-Oct-20 09:59:00	60,869
<b>Average Stack Raw NOx, ppm</b>	<b>40.77</b>	<b>Average Stack O2 (%)</b>	<b>5.14</b>	<b>Average Gas Flow (SCFH)</b>	<b>53,626</b>

**NEBRASKA NEW ANALYZER CERTIFICATION**

10/20/2020 10:00 10/20/2020 10:45 1 m

Raw NOx, ppm		O2 (%)		Gas Flow (SCFH)	
931AI393.AT		931AI392.AT		931FI374B.ft	
20-Oct-20 10:00:00	-0.23	20-Oct-20 10:00:00	4.35	20-Oct-20 10:00:00	61,005
20-Oct-20 10:01:00	-0.22	20-Oct-20 10:01:00	4.38	20-Oct-20 10:01:00	61,011
20-Oct-20 10:02:00	-0.05	20-Oct-20 10:02:00	4.36	20-Oct-20 10:02:00	60,869
20-Oct-20 10:03:00	0.58	20-Oct-20 10:03:00	4.34	20-Oct-20 10:03:00	59,817
20-Oct-20 10:04:00	10.07	20-Oct-20 10:04:00	4.74	20-Oct-20 10:04:00	55,865
20-Oct-20 10:05:00	14.47	20-Oct-20 10:05:00	5.00	20-Oct-20 10:05:00	57,430
20-Oct-20 10:06:00	14.62	20-Oct-20 10:06:00	4.78	20-Oct-20 10:06:00	59,453
20-Oct-20 10:07:00	15.35	20-Oct-20 10:07:00	4.99	20-Oct-20 10:07:00	60,826
20-Oct-20 10:08:00	16.07	20-Oct-20 10:08:00	5.02	20-Oct-20 10:08:00	62,520
20-Oct-20 10:09:00	16.16	20-Oct-20 10:09:00	4.83	20-Oct-20 10:09:00	62,458
20-Oct-20 10:10:00	16.00	20-Oct-20 10:10:00	4.88	20-Oct-20 10:10:00	61,391
20-Oct-20 10:11:00	16.17	20-Oct-20 10:11:00	4.86	20-Oct-20 10:11:00	60,190
20-Oct-20 10:12:00	16.38	20-Oct-20 10:12:00	4.73	20-Oct-20 10:12:00	59,397
20-Oct-20 10:13:00	16.47	20-Oct-20 10:13:00	5.05	20-Oct-20 10:13:00	56,692
20-Oct-20 10:14:00	16.51	20-Oct-20 10:14:00	5.08	20-Oct-20 10:14:00	58,277
20-Oct-20 10:15:00	16.59	20-Oct-20 10:15:00	5.01	20-Oct-20 10:15:00	59,715
20-Oct-20 10:16:00	12.94	20-Oct-20 10:16:00	4.89	20-Oct-20 10:16:00	61,477
20-Oct-20 10:17:00	6.87	20-Oct-20 10:17:00	4.81	20-Oct-20 10:17:00	61,740
20-Oct-20 10:18:00	16.94	20-Oct-20 10:18:00	4.75	20-Oct-20 10:18:00	61,665
20-Oct-20 10:19:00	17.25	20-Oct-20 10:19:00	4.76	20-Oct-20 10:19:00	60,762
20-Oct-20 10:20:00	17.40	20-Oct-20 10:20:00	4.78	20-Oct-20 10:20:00	58,957
20-Oct-20 10:21:00	17.49	20-Oct-20 10:21:00	5.03	20-Oct-20 10:21:00	56,327
20-Oct-20 10:22:00	17.45	20-Oct-20 10:22:00	5.46	20-Oct-20 10:22:00	54,423
20-Oct-20 10:23:00	17.40	20-Oct-20 10:23:00	5.47	20-Oct-20 10:23:00	54,791
20-Oct-20 10:24:00	14.18	20-Oct-20 10:24:00	5.21	20-Oct-20 10:24:00	55,958
20-Oct-20 10:25:00	1.51	20-Oct-20 10:25:00	5.09	20-Oct-20 10:25:00	56,045
20-Oct-20 10:26:00	0.69	20-Oct-20 10:26:00	5.09	20-Oct-20 10:26:00	56,070
20-Oct-20 10:27:00	6.43	20-Oct-20 10:27:00	5.10	20-Oct-20 10:27:00	56,030
20-Oct-20 10:28:00	16.89	20-Oct-20 10:28:00	5.08	20-Oct-20 10:28:00	56,086
20-Oct-20 10:29:00	17.18	20-Oct-20 10:29:00	5.05	20-Oct-20 10:29:00	56,062
20-Oct-20 10:30:00	17.29	20-Oct-20 10:30:00	5.04	20-Oct-20 10:30:00	56,110
20-Oct-20 10:31:00	17.36	20-Oct-20 10:31:00	4.92	20-Oct-20 10:31:00	57,389
20-Oct-20 10:32:00	17.39	20-Oct-20 10:32:00	4.80	20-Oct-20 10:32:00	58,563
20-Oct-20 10:33:00	17.60	20-Oct-20 10:33:00	4.62	20-Oct-20 10:33:00	59,493
20-Oct-20 10:34:00	13.75	20-Oct-20 10:34:00	4.68	20-Oct-20 10:34:00	60,280
20-Oct-20 10:35:00	1.18	20-Oct-20 10:35:00	4.56	20-Oct-20 10:35:00	60,899
20-Oct-20 10:36:00	0.61	20-Oct-20 10:36:00	4.56	20-Oct-20 10:36:00	60,921
20-Oct-20 10:37:00	7.30	20-Oct-20 10:37:00	4.61	20-Oct-20 10:37:00	61,040
20-Oct-20 10:38:00	16.98	20-Oct-20 10:38:00	4.58	20-Oct-20 10:38:00	61,000
20-Oct-20 10:39:00	17.30	20-Oct-20 10:39:00	4.48	20-Oct-20 10:39:00	60,722
20-Oct-20 10:40:00	17.45	20-Oct-20 10:40:00	4.53	20-Oct-20 10:40:00	60,107
20-Oct-20 10:41:00	17.53	20-Oct-20 10:41:00	4.47	20-Oct-20 10:41:00	59,259
20-Oct-20 10:42:00	17.53	20-Oct-20 10:42:00	4.75	20-Oct-20 10:42:00	57,286
20-Oct-20 10:43:00	17.60	20-Oct-20 10:43:00	4.81	20-Oct-20 10:43:00	58,668
20-Oct-20 10:44:00	17.62	20-Oct-20 10:44:00	4.71	20-Oct-20 10:44:00	60,607
<b>Average Stack Raw NOx, ppm</b>	<b>12.65</b>	<b>Average Stack O2 (%)</b>	<b>4.89</b>	<b>Average Gas Flow (SCFH)</b>	<b>58,636</b>

NOx raw data checks

①

②

③

**FIELD DATA & STRIP CHARTS**

**EMISSIONS TEST - SCAQMD Method 100.1**

Date: 10/21/20

Job #: 220-112

Client #: 23022

**\*\* Test Information \*\***

Client: New Indy  
County: Ventura  
Site(s): Oxnard, CA  
Unit(s): Nebraska Boiler New NOx

**\*\* Personnel \*\***

AIRx: KK/WH  
Client: RL  
APCD: ES

Run Length:	32	Inlet ( )	Outlet(X)	Fuel ( )	HC's( )	
No. Points:	Single	S.T.(X)	E.I.( )	Data( )	H2S ( )	NH3 ( )
	Time	Baro.	Temp.		Weather.	
Arrive:	6:30 AM	29.81	66		Clear	
Depart:	12:30 PM	29.84	70		Clear	

**\*\* Instrument Information \*\***

Instrument	"On"	Unit#	Make/Model
Outlet NOx:	1	3	API 200EH
O2:	1	12	Servomex 1400

Recorders: Soltec 6 pin 10 cm/hr

**\*\* Calibration Information \*\***

	Units	Zero	Span	Range	Gas Cyl.#	Gas Flow
NOx:	ppmv	0.0	20.2	50	CC724456	1.0
NOx:	ppmv	0.0	40.0	50	DT0028766	1.0
NO2:	ppmv	0.0	19.0	50	CC3240	1.0
O2:	%	0.0	3.98	10	CC60235	0.6
O2:	%	0.0	8.05	10	CC160530	0.6

**\*\* Recorder Information \*\***

		Chanl.	Pen Type	Color
Outlet NOx:	ppmv	1	Cont.	Black
O2:	%	2	Cont.	Green



CLIENT: New Indy      JOB# 220-112  
 PLANT: Oxnard, CA      RUN# RATA 1  
 DATE: 10/21/2020      UNIT ID: Nebraska Boiler New NOx  
 ENGINEER: KK/WH      RUN START: 8:15 AM

NOx, ppm	O2, %	NOx, ppm @ 3% O2	TIME
22.57	2.98	22.54	8:15
22.73	3.92	23.96	8:16
23.35	4.07	24.83	8:17
23.70	4.20	25.40	8:18
23.75	4.33	25.66	8:19
23.72	3.51	24.42	8:20
23.69	5.21	27.03	8:21
23.19	5.35	26.69	8:22
22.44	5.26	25.68	8:23
22.14	4.25	23.80	8:24
22.58	4.66	24.89	8:25
23.19	4.63	25.51	8:26
23.61	4.66	26.02	8:27
23.74	4.04	25.20	8:28
23.67	4.09	25.20	8:29
23.58	4.21	25.29	8:30
23.69	4.84	26.40	8:31
23.76	4.45	25.85	8:32
23.61	3.95	24.93	8:33
23.57	3.90	24.82	8:34
23.93	3.98	25.32	8:35
23.98	4.43	26.06	8:36
<b>23.37</b>	<b>4.31</b>	<b>25.25</b>	<b>Averages</b>

CLIENT: New Indy      JOB# 220-112  
 PLANT: Oxnard, CA      RUN# RATA 2  
 DATE: 10/21/2020      UNIT ID: Nebraska Boiler New NOx  
 ENGINEER: KK/WH      RUN START: 8:37 AM

NOx, ppm	O2, %	NOx, ppm @ 3% O2	TIME
23.80	4.32	25.69	8:37
23.74	4.42	25.79	8:38
23.90	4.25	25.69	8:39
24.11	4.23	25.89	8:40
24.05	4.26	25.87	8:41
23.96	4.17	25.64	8:42
23.83	4.15	25.47	8:43
23.74	4.17	25.40	8:44
23.81	4.24	25.58	8:45
23.73	4.53	25.95	8:46
23.60	4.25	25.37	8:47
23.78	4.27	25.60	8:48
24.02	4.26	25.84	8:49
23.96	4.38	25.96	8:50
23.64	4.29	25.48	8:51
23.64	4.45	25.72	8:52
23.81	4.65	26.23	8:53
23.61	4.21	25.32	8:54
23.43	4.26	25.20	8:55
23.70	4.22	25.43	8:56
23.70	4.32	25.59	8:57
24.00	4.20	25.72	8:58
<b>23.80</b>	<b>4.30</b>	<b>25.66</b>	<b>Averages</b>

CLIENT:	New Indy	JOB#	220-112
PLANT:	Oxnard, CA	RUN#	RATA 3
DATE:	10/21/2020	UNIT ID:	Nebraska Boiler New NOx
ENGINEER:	KK/WH	RUN START:	8:59 AM

NOx, ppm	O2, %	NOx, ppm (@ 3% O2	Time
24.16	4.42	26.24	8:59
24.02	4.32	25.93	9:00
23.96	4.44	26.06	9:01
24.34	4.52	26.60	9:02
24.47	4.63	26.92	9:03
24.47	4.63	26.92	9:04
24.06	4.53	26.31	9:05
24.19	4.45	26.32	9:06
24.50	4.69	27.05	9:07
24.50	4.58	26.87	9:08
24.18	4.25	26.00	9:09
24.23	4.26	26.06	9:10
24.60	4.21	26.38	9:11
24.56	4.86	27.41	9:12
24.33	4.68	26.85	9:13
24.36	4.46	26.52	9:14
24.64	4.34	26.63	9:15
24.56	4.61	26.99	9:16
24.39	4.40	26.46	9:17
24.31	4.34	26.28	9:18
24.52	4.79	27.24	9:19
24.48	4.91	27.40	9:20
<b>24.36</b>	<b>4.51</b>	<b>26.61</b>	<b>Averages</b>

CLIENT:	New Indy	JOB#	220-112
PLANT:	Oxnard, CA	RUN#	RATA 4
DATE:	10/21/2020	UNIT ID:	Nebraska Boiler New NOx
ENGINEER:	KK/WH	RUN START:	9:32 AM

NOx, ppm	O2, %	NOx, ppm @ 3% O2	TIME
24.15	4.49	26.34	9:32
24.16	4.61	26.55	9:33
23.95	4.60	26.30	9:34
23.87	4.51	26.07	9:35
23.90	4.44	25.99	9:36
23.94	4.42	26.01	9:37
23.94	4.42	26.01	9:38
23.89	4.42	25.95	9:39
23.88	4.44	25.97	9:40
23.83	4.70	26.33	9:41
23.73	4.54	25.96	9:42
23.94	4.53	26.18	9:43
24.07	4.60	26.43	9:44
23.85	4.33	25.76	9:45
23.81	3.94	25.13	9:46
24.13	4.50	26.34	9:47
24.06	4.72	26.62	9:48
23.84	4.34	25.77	9:49
23.90	4.46	26.02	9:50
23.91	4.70	26.42	9:51
23.80	4.20	25.51	9:52
23.86	4.11	25.44	9:53
<b>23.93</b>	<b>4.46</b>	<b>26.05</b>	<b>Averages</b>

CLIENT:	New Indy	JOB#	220-112
PLANT:	Oxnard, CA	RUN#	RATA 5
DATE:	10/21/2020	UNIT ID:	Nebraska Boiler New NOx
ENGINEER:	KK/WH	RUN START:	9:54 AM

NOx, ppm	O2, %	NOx, ppm @ 3% O2	TIME
24.03	4.14	25.66	9:54
24.03	4.63	26.44	9:55
23.87	4.80	26.54	9:56
23.66	4.67	26.09	9:57
23.71	4.61	26.05	9:58
23.84	4.22	25.58	9:59
23.95	4.61	26.32	10:00
24.06	4.69	26.57	10:01
24.06	4.68	26.55	10:02
24.00	4.69	26.50	10:03
23.96	4.67	26.43	10:04
23.95	4.65	26.38	10:05
23.92	4.64	26.33	10:06
23.90	4.40	25.93	10:07
23.93	4.22	25.68	10:08
23.90	4.71	26.42	10:09
23.85	4.74	26.42	10:10
23.62	4.75	26.18	10:11
23.53	4.54	25.74	10:12
23.63	4.33	25.53	10:13
23.79	4.51	25.98	10:14
23.81	4.76	26.41	10:15
<b>23.86</b>	<b>4.58</b>	<b>26.17</b>	<b>Averages</b>

CLIENT:	New Indy	JOB#	220-112
PLANT:	Oxnard, CA	RUN#	RATA 6
DATE:	10/21/2020	UNIT ID:	Nebraska Boiler New NOx
ENGINEER:	KK/WH	RUN START:	10:17 AM

NOx, ppm	O2, %	NOx, ppm @ 3% O2	TIME
23.69	4.40	25.70	10:17
23.85	4.56	26.13	10:18
23.89	4.25	25.68	10:19
24.12	3.91	25.41	10:20
24.36	4.88	27.22	10:21
24.06	5.42	27.82	10:22
23.33	4.77	25.89	10:24
23.23	4.69	25.65	10:25
23.51	5.12	26.67	10:26
23.46	4.66	25.86	10:27
23.49	4.42	25.51	10:28
23.76	4.76	26.35	10:29
23.82	4.81	26.50	10:30
23.78	4.25	25.57	10:31
23.86	4.62	26.23	10:32
23.85	5.35	27.45	10:33
23.55	4.65	25.94	10:34
23.41	4.58	25.68	10:35
23.50	4.60	25.81	10:36
23.62	4.59	25.92	10:37
23.71	4.61	26.05	10:38
23.81	5.02	26.84	10:39
<b>23.71</b>	<b>4.68</b>	<b>26.18</b>	<b>Averages</b>

CLIENT:	New Indy	JOB#	220-112
PLANT:	Oxnard, CA	RUN#	RATA 7
DATE:	10/21/2020	UNIT ID:	Nebraska Boiler New NOx
ENGINEER:	KK/WH	RUN START:	10:50 AM

NOx, ppm	O2, %	NOx, ppm @ 3% O2	TIME
23.67	4.80	26.32	10:50
23.59	5.11	26.74	10:51
23.45	4.57	25.70	10:52
23.48	4.58	25.75	10:53
23.49	4.60	25.80	10:54
23.43	4.61	25.75	10:55
23.37	4.66	25.76	10:56
23.34	4.70	25.79	10:57
23.28	4.68	25.69	10:58
23.21	4.71	25.66	10:59
23.28	5.14	26.44	11:00
23.24	4.82	25.87	11:01
23.23	4.28	25.02	11:02
23.43	4.51	25.59	11:03
23.51	4.90	26.30	11:04
23.29	4.89	26.04	11:05
23.08	4.64	25.41	11:06
23.10	4.60	25.37	11:07
23.23	4.60	25.51	11:08
23.31	4.60	25.60	11:09
23.43	4.94	26.28	11:10
23.35	4.50	25.49	11:11
<b>23.35</b>	<b>4.70</b>	<b>25.81</b>	<b>Averages</b>

CLIENT: New-Indy      JOB# 220-112  
 PLANT: Oxnard, CA      RUN# RATA 8  
 DATE: 10/21/2020      UNIT ID: Nebraska Boiler New NOx  
 ENGINEER: KK/WH      RUN START: 11:12 AM


NOx, ppm	O2, %	NOx, ppm @ 3% O2	TIME
23.36	5.14	26.53	11:12
23.11	4.98	25.98	11:13
22.85	4.46	24.88	11:14
23.02	4.25	24.75	11:15
23.28	4.55	25.49	11:16
23.32	4.91	26.11	11:17
23.23	4.91	26.00	11:18
23.10	4.72	25.56	11:19
23.00	4.64	25.32	11:20
22.95	4.65	25.28	11:21
23.02	4.65	25.36	11:22
23.07	4.72	25.52	11:23
23.06	5.01	25.98	11:24
22.89	4.48	24.95	11:25
22.93	4.35	24.80	11:26
23.14	5.10	26.22	11:27
22.93	4.92	25.69	11:28
22.65	4.31	24.44	11:29
22.82	4.31	24.62	11:30
23.17	4.98	26.05	11:31
22.97	5.09	26.01	11:32
23.17	4.98	26.05	11:33
<b>23.05</b>	<b>4.73</b>	<b>25.53</b>	<b>Averages</b>




CLIENT:	New Indy	JOB#	220-112
PLANT:	Oxnard, CA	RUN#	RATA 9
DATE:	10/21/2020	UNIT ID:	Nebraska Boiler New NOx
ENGINEER:	KK/WH	RUN START:	11:34 AM

NOx, ppm	O2, %	NOx, ppm @ 3% O2	TIME
22.97	5.09	26.01	11:34
22.57	4.37	24.44	11:35
22.67	4.50	24.74	11:36
22.74	5.17	25.88	11:37
22.48	4.60	24.69	11:38
22.38	4.27	24.09	11:39
22.62	4.55	24.76	11:40
22.69	4.98	25.51	11:41
22.51	5.01	25.36	11:42
22.26	4.74	24.66	11:43
22.21	4.64	24.45	11:44
22.20	4.63	24.42	11:45
22.15	4.64	24.38	11:46
22.22	4.63	24.45	11:47
22.29	4.61	24.49	11:48
22.29	4.59	24.46	11:49
22.33	4.82	24.86	11:50
22.26	5.07	25.17	11:51
22.13	4.57	24.26	11:52
22.21	4.38	24.07	11:53
22.33	4.90	24.98	11:54
22.16	4.60	24.34	11:55
<b>22.39</b>	<b>4.70</b>	<b>24.75</b>	<b>Averages</b>












10V/div

100ns/div

Post Processed



Start  
9:31  
9:57  
10:16  
10:31

69  
28.62




Start  
7:59  
10:50

29.32

6.8

Post Bias Cell

2:00  
14:45  
8:02



OK

Next

New-Indy Oxnard, LLC  
 Nebraska Boiler CEMS RATA TEST - November 21, 2020

10/21/2020 8:15 10/21/2020 8:37 1 m

K-1

NOx, ppm (3% O2)		Raw NOx, ppm		O2 (%)		Gas Flow (SCFH)	
931AI391_CNOX		931AI393.AT		931AI392.AT		931FI374B.ft	
21-Oct-20 08:15:00	26.60	21-Oct-20 08:15:00	24.70	21-Oct-20 08:15:00	4.28	21-Oct-20 08:15:00	83,564
21-Oct-20 08:16:00	27.11	21-Oct-20 08:16:00	24.90	21-Oct-20 08:16:00	4.46	21-Oct-20 08:16:00	77,223
21-Oct-20 08:17:00	27.79	21-Oct-20 08:17:00	24.12	21-Oct-20 08:17:00	5.36	21-Oct-20 08:17:00	65,197
21-Oct-20 08:18:00	27.12	21-Oct-20 08:18:00	23.21	21-Oct-20 08:18:00	5.58	21-Oct-20 08:18:00	61,133
21-Oct-20 08:19:00	26.84	21-Oct-20 08:19:00	23.16	21-Oct-20 08:19:00	5.44	21-Oct-20 08:19:00	63,215
21-Oct-20 08:20:00	25.55	21-Oct-20 08:20:00	23.24	21-Oct-20 08:20:00	4.62	21-Oct-20 08:20:00	70,946
21-Oct-20 08:21:00	25.51	21-Oct-20 08:21:00	23.68	21-Oct-20 08:21:00	4.36	21-Oct-20 08:21:00	73,917
21-Oct-20 08:22:00	26.26	21-Oct-20 08:22:00	24.24	21-Oct-20 08:22:00	4.39	21-Oct-20 08:22:00	74,145
21-Oct-20 08:23:00	26.40	21-Oct-20 08:23:00	24.36	21-Oct-20 08:23:00	4.42	21-Oct-20 08:23:00	73,123
21-Oct-20 08:24:00	26.88	21-Oct-20 08:24:00	24.26	21-Oct-20 08:24:00	4.74	21-Oct-20 08:24:00	70,639
21-Oct-20 08:25:00	26.56	21-Oct-20 08:25:00	24.02	21-Oct-20 08:25:00	4.71	21-Oct-20 08:25:00	72,729
21-Oct-20 08:26:00	26.29	21-Oct-20 08:26:00	24.27	21-Oct-20 08:26:00	4.39	21-Oct-20 08:26:00	75,115
21-Oct-20 08:27:00	27.01	21-Oct-20 08:27:00	24.60	21-Oct-20 08:27:00	4.60	21-Oct-20 08:27:00	70,527
21-Oct-20 08:28:00	26.96	21-Oct-20 08:28:00	23.95	21-Oct-20 08:28:00	5.00	21-Oct-20 08:28:00	71,267
21-Oct-20 08:29:00	26.61	21-Oct-20 08:29:00	24.17	21-Oct-20 08:29:00	4.64	21-Oct-20 08:29:00	72,728
21-Oct-20 08:30:00	26.62	21-Oct-20 08:30:00	24.24	21-Oct-20 08:30:00	4.60	21-Oct-20 08:30:00	72,779
21-Oct-20 08:31:00	26.92	21-Oct-20 08:31:00	24.41	21-Oct-20 08:31:00	4.67	21-Oct-20 08:31:00	71,126
21-Oct-20 08:32:00	27.13	21-Oct-20 08:32:00	24.09	21-Oct-20 08:32:00	5.00	21-Oct-20 08:32:00	69,784
21-Oct-20 08:33:00	26.73	21-Oct-20 08:33:00	23.93	21-Oct-20 08:33:00	4.88	21-Oct-20 08:33:00	71,768
21-Oct-20 08:34:00	26.42	21-Oct-20 08:34:00	24.17	21-Oct-20 08:34:00	4.53	21-Oct-20 08:34:00	73,917
21-Oct-20 08:35:00	26.72	21-Oct-20 08:35:00	24.41	21-Oct-20 08:35:00	4.55	21-Oct-20 08:35:00	72,089
21-Oct-20 08:36:00	27.07	21-Oct-20 08:36:00	24.28	21-Oct-20 08:36:00	4.84	21-Oct-20 08:36:00	70,340
<b>Average Stack NOx, ppm (3% O2)</b>	<b>26.69</b>	<b>Average Stack Raw NOx, ppm</b>	<b>24.11</b>	<b>Average Stack O2 (%)</b>	<b>4.73</b>	<b>Average Gas Flow (SCFH)</b>	<b>71,694</b>

New-Indy Oxnard, LLC  
 Nebraska Boiler CEMS RATA TEST - November 21, 2020

10/21/2020 8:37 10/21/2020 8:59 1 m

K-2

NOx, ppm (3% O2)		Raw NOx, ppm		O2 (%)		Gas Flow (SCFH)	
931AI391_CNOX		931AI393.AT		931AI392.AT		931FI374B.R	
21-Oct-20 08:37:00	26.87	21-Oct-20 08:37:00	24.08	21-Oct-20 08:37:00	4.86	21-Oct-20 08:37:00	71,348
21-Oct-20 08:38:00	26.57	21-Oct-20 08:38:00	24.12	21-Oct-20 08:38:00	4.65	21-Oct-20 08:38:00	72,488
21-Oct-20 08:39:00	26.73	21-Oct-20 08:39:00	24.22	21-Oct-20 08:39:00	4.68	21-Oct-20 08:39:00	72,472
21-Oct-20 08:40:00	26.84	21-Oct-20 08:40:00	24.29	21-Oct-20 08:40:00	4.69	21-Oct-20 08:40:00	72,420
21-Oct-20 08:41:00	27.15	21-Oct-20 08:41:00	24.38	21-Oct-20 08:41:00	4.83	21-Oct-20 08:41:00	70,226
21-Oct-20 08:42:00	27.11	21-Oct-20 08:42:00	23.99	21-Oct-20 08:42:00	5.06	21-Oct-20 08:42:00	69,889
21-Oct-20 08:43:00	26.55	21-Oct-20 08:43:00	23.94	21-Oct-20 08:43:00	4.76	21-Oct-20 08:43:00	72,787
21-Oct-20 08:44:00	26.51	21-Oct-20 08:44:00	24.17	21-Oct-20 08:44:00	4.58	21-Oct-20 08:44:00	73,025
21-Oct-20 08:45:00	26.72	21-Oct-20 08:45:00	24.26	21-Oct-20 08:45:00	4.64	21-Oct-20 08:45:00	71,512
21-Oct-20 08:46:00	26.90	21-Oct-20 08:46:00	23.94	21-Oct-20 08:46:00	4.97	21-Oct-20 08:46:00	70,499
21-Oct-20 08:47:00	26.38	21-Oct-20 08:47:00	23.93	21-Oct-20 08:47:00	4.68	21-Oct-20 08:47:00	74,015
21-Oct-20 08:48:00	26.56	21-Oct-20 08:48:00	24.37	21-Oct-20 08:48:00	4.48	21-Oct-20 08:48:00	72,766
21-Oct-20 08:49:00	27.39	21-Oct-20 08:49:00	24.14	21-Oct-20 08:49:00	5.12	21-Oct-20 08:49:00	68,594
21-Oct-20 08:50:00	26.60	21-Oct-20 08:50:00	23.80	21-Oct-20 08:50:00	4.89	21-Oct-20 08:50:00	71,406
21-Oct-20 08:51:00	26.68	21-Oct-20 08:51:00	23.99	21-Oct-20 08:51:00	4.80	21-Oct-20 08:51:00	71,451
21-Oct-20 08:52:00	26.55	21-Oct-20 08:52:00	23.91	21-Oct-20 08:52:00	4.78	21-Oct-20 08:52:00	72,493
21-Oct-20 08:53:00	26.37	21-Oct-20 08:53:00	24.16	21-Oct-20 08:53:00	4.50	21-Oct-20 08:53:00	73,587
21-Oct-20 08:54:00	27.20	21-Oct-20 08:54:00	24.26	21-Oct-20 08:54:00	4.91	21-Oct-20 08:54:00	69,450
21-Oct-20 08:55:00	26.71	21-Oct-20 08:55:00	23.76	21-Oct-20 08:55:00	4.98	21-Oct-20 08:55:00	71,853
21-Oct-20 08:56:00	26.54	21-Oct-20 08:56:00	24.06	21-Oct-20 08:56:00	4.67	21-Oct-20 08:56:00	73,802
21-Oct-20 08:57:00	26.69	21-Oct-20 08:57:00	24.30	21-Oct-20 08:57:00	4.60	21-Oct-20 08:57:00	72,835
21-Oct-20 08:58:00	27.41	21-Oct-20 08:58:00	24.31	21-Oct-20 08:58:00	5.02	21-Oct-20 08:58:00	68,023
<b>Average Stack NOx, ppm (3% O2)</b>	<b>26.78</b>	<b>Average Stack Raw NOx, ppm</b>	<b>24.11</b>	<b>Average Stack O2 (%)</b>	<b>4.78</b>	<b>Average Gas Flow (SCFH)</b>	<b>71,668</b>



New-Indy Oxnard, LLC  
 Nebraska Boiler CEMS RATA TEST - November 21, 2020

10/21/2020 8:59 10/21/2020 9:21 1 m

R-3

NOx, ppm (3% O2)		Raw NOx, ppm		O2 (%)		Gas Flow (SCFH)	
931AI391_CNOX		931AI393.AT		931AI392.AT		931FI374B.ft	
21-Oct-20 08:59:00	26.90	21-Oct-20 08:59:00	23.70	21-Oct-20 08:59:00	5.14	21-Oct-20 08:59:00	70,522
21-Oct-20 09:00:00	26.56	21-Oct-20 09:00:00	23.94	21-Oct-20 09:00:00	4.77	21-Oct-20 09:00:00	72,649
21-Oct-20 09:01:00	26.56	21-Oct-20 09:01:00	24.14	21-Oct-20 09:01:00	4.63	21-Oct-20 09:01:00	73,570
21-Oct-20 09:02:00	26.70	21-Oct-20 09:02:00	24.30	21-Oct-20 09:02:00	4.61	21-Oct-20 09:02:00	72,830
21-Oct-20 09:03:00	27.36	21-Oct-20 09:03:00	24.31	21-Oct-20 09:03:00	4.99	21-Oct-20 09:03:00	68,778
21-Oct-20 09:04:00	26.85	21-Oct-20 09:04:00	23.75	21-Oct-20 09:04:00	5.06	21-Oct-20 09:04:00	71,722
21-Oct-20 09:05:00	26.64	21-Oct-20 09:05:00	24.26	21-Oct-20 09:05:00	4.60	21-Oct-20 09:05:00	73,393
21-Oct-20 09:06:00	27.27	21-Oct-20 09:06:00	24.66	21-Oct-20 09:06:00	4.71	21-Oct-20 09:06:00	70,961
21-Oct-20 09:07:00	27.61	21-Oct-20 09:07:00	24.23	21-Oct-20 09:07:00	5.18	21-Oct-20 09:07:00	67,946
21-Oct-20 09:08:00	26.63	21-Oct-20 09:08:00	23.88	21-Oct-20 09:08:00	4.85	21-Oct-20 09:08:00	72,304
21-Oct-20 09:09:00	26.63	21-Oct-20 09:09:00	24.36	21-Oct-20 09:09:00	4.53	21-Oct-20 09:09:00	73,230
21-Oct-20 09:10:00	26.94	21-Oct-20 09:10:00	24.45	21-Oct-20 09:10:00	4.65	21-Oct-20 09:10:00	71,274
21-Oct-20 09:11:00	27.21	21-Oct-20 09:11:00	24.20	21-Oct-20 09:11:00	4.97	21-Oct-20 09:11:00	70,089
21-Oct-20 09:12:00	26.95	21-Oct-20 09:12:00	24.04	21-Oct-20 09:12:00	4.94	21-Oct-20 09:12:00	71,074
21-Oct-20 09:13:00	26.61	21-Oct-20 09:13:00	24.14	21-Oct-20 09:13:00	4.67	21-Oct-20 09:13:00	73,543
21-Oct-20 09:14:00	26.84	21-Oct-20 09:14:00	24.49	21-Oct-20 09:14:00	4.57	21-Oct-20 09:14:00	72,128
21-Oct-20 09:15:00	27.56	21-Oct-20 09:15:00	24.28	21-Oct-20 09:15:00	5.13	21-Oct-20 09:15:00	67,773
21-Oct-20 09:16:00	26.87	21-Oct-20 09:16:00	23.82	21-Oct-20 09:16:00	5.04	21-Oct-20 09:16:00	70,853
21-Oct-20 09:17:00	26.81	21-Oct-20 09:17:00	24.15	21-Oct-20 09:17:00	4.77	21-Oct-20 09:17:00	72,254
21-Oct-20 09:18:00	26.74	21-Oct-20 09:18:00	24.31	21-Oct-20 09:18:00	4.63	21-Oct-20 09:18:00	73,479
21-Oct-20 09:19:00	26.92	21-Oct-20 09:19:00	24.45	21-Oct-20 09:19:00	4.65	21-Oct-20 09:19:00	72,192
21-Oct-20 09:20:00	27.44	21-Oct-20 09:20:00	24.37	21-Oct-20 09:20:00	5.00	21-Oct-20 09:20:00	70,185
<b>Average Stack NOx, ppm (3% O2)</b>	<b>26.94</b>	<b>Average Stack Raw NOx, ppm</b>	<b>24.19</b>	<b>Average Stack O2 (%)</b>	<b>4.82</b>	<b>Average Gas Flow (SCFH)</b>	<b>71,489</b>

New-Indy Oxnard, LLC  
 Nebraska Boiler CEMS RATA TEST - November 21, 2020

10/21/2020 9:31 10/21/2020 9:53 1 m

*Rata 4*

NOx, ppm (3% O2)		Raw NOx, ppm		O2 (%)		Gas Flow (SCFH)	
931AI391_CNOX		931AI393.AT		931AI392.AT		931FI374B.ft	
21-Oct-20 09:31:00	26.56	21-Oct-20 09:31:00	23.88	21-Oct-20 09:31:00	4.80	21-Oct-20 09:31:00	71,529
21-Oct-20 09:32:00	26.64	21-Oct-20 09:32:00	23.98	21-Oct-20 09:32:00	4.79	21-Oct-20 09:32:00	71,506
21-Oct-20 09:33:00	26.60	21-Oct-20 09:33:00	23.97	21-Oct-20 09:33:00	4.78	21-Oct-20 09:33:00	71,566
21-Oct-20 09:34:00	26.59	21-Oct-20 09:34:00	23.94	21-Oct-20 09:34:00	4.79	21-Oct-20 09:34:00	71,617
21-Oct-20 09:35:00	26.75	21-Oct-20 09:35:00	24.09	21-Oct-20 09:35:00	4.79	21-Oct-20 09:35:00	71,449
21-Oct-20 09:36:00	27.23	21-Oct-20 09:36:00	24.15	21-Oct-20 09:36:00	5.02	21-Oct-20 09:36:00	69,063
21-Oct-20 09:37:00	26.88	21-Oct-20 09:37:00	23.75	21-Oct-20 09:37:00	5.08	21-Oct-20 09:37:00	70,389
21-Oct-20 09:38:00	26.46	21-Oct-20 09:38:00	23.94	21-Oct-20 09:38:00	4.70	21-Oct-20 09:38:00	72,478
21-Oct-20 09:39:00	26.90	21-Oct-20 09:39:00	24.13	21-Oct-20 09:39:00	4.84	21-Oct-20 09:39:00	70,053
21-Oct-20 09:40:00	26.85	21-Oct-20 09:40:00	23.80	21-Oct-20 09:40:00	5.03	21-Oct-20 09:40:00	70,307
21-Oct-20 09:41:00	26.58	21-Oct-20 09:41:00	23.85	21-Oct-20 09:41:00	4.83	21-Oct-20 09:41:00	72,453
21-Oct-20 09:42:00	26.44	21-Oct-20 09:42:00	24.12	21-Oct-20 09:42:00	4.57	21-Oct-20 09:42:00	72,459
21-Oct-20 09:43:00	27.20	21-Oct-20 09:43:00	24.09	21-Oct-20 09:43:00	5.05	21-Oct-20 09:43:00	68,131
21-Oct-20 09:44:00	26.61	21-Oct-20 09:44:00	23.52	21-Oct-20 09:44:00	5.08	21-Oct-20 09:44:00	70,281
21-Oct-20 09:45:00	26.60	21-Oct-20 09:45:00	23.92	21-Oct-20 09:45:00	4.80	21-Oct-20 09:45:00	71,029
21-Oct-20 09:46:00	26.94	21-Oct-20 09:46:00	23.99	21-Oct-20 09:46:00	4.96	21-Oct-20 09:46:00	68,540
21-Oct-20 09:47:00	26.80	21-Oct-20 09:47:00	23.63	21-Oct-20 09:47:00	5.12	21-Oct-20 09:47:00	70,113
21-Oct-20 09:48:00	26.79	21-Oct-20 09:48:00	24.19	21-Oct-20 09:48:00	4.75	21-Oct-20 09:48:00	71,716
21-Oct-20 09:49:00	26.77	21-Oct-20 09:49:00	24.19	21-Oct-20 09:49:00	4.72	21-Oct-20 09:49:00	71,804
21-Oct-20 09:50:00	26.99	21-Oct-20 09:50:00	24.35	21-Oct-20 09:50:00	4.75	21-Oct-20 09:50:00	70,786
21-Oct-20 09:51:00	27.44	21-Oct-20 09:51:00	24.13	21-Oct-20 09:51:00	5.14	21-Oct-20 09:51:00	68,893
21-Oct-20 09:52:00	27.25	21-Oct-20 09:52:00	23.96	21-Oct-20 09:52:00	5.16	21-Oct-20 09:52:00	69,728
<b>Average Stack NOx, ppm (3% O2)</b>	<b>26.81</b>	<b>Average Stack Raw NOx, ppm</b>	<b>23.98</b>	<b>Average Stack O2 (%)</b>	<b>4.89</b>	<b>Average Gas Flow (SCFH)</b>	<b>70,722</b>

*Rata 5*

**New-Indy Oxnard, LLC  
Nebraska Boiler CEMS RATA TEST - November 21, 2020**

10/21/2020 9:53 10/21/2020 10:15 1 m

NOx, ppm (3% O2)		Raw NOx, ppm		O2 (%)		Gas Flow (SCFH)	
931AI391_CNOX		931AI393.AT		931AI392.AT		931FI374B.ft	
21-Oct-20 09:53:00	27.13	21-Oct-20 09:53:00	24.09	21-Oct-20 09:53:00	5.01	21-Oct-20 09:53:00	70,441
21-Oct-20 09:54:00	27.08	21-Oct-20 09:54:00	24.08	21-Oct-20 09:54:00	4.98	21-Oct-20 09:54:00	71,472
21-Oct-20 09:55:00	26.84	21-Oct-20 09:55:00	24.37	21-Oct-20 09:55:00	4.65	21-Oct-20 09:55:00	71,883
21-Oct-20 09:56:00	27.31	21-Oct-20 09:56:00	24.38	21-Oct-20 09:56:00	4.93	21-Oct-20 09:56:00	70,133
21-Oct-20 09:57:00	27.45	21-Oct-20 09:57:00	24.38	21-Oct-20 09:57:00	5.01	21-Oct-20 09:57:00	70,099
21-Oct-20 09:58:00	27.55	21-Oct-20 09:58:00	24.42	21-Oct-20 09:58:00	5.04	21-Oct-20 09:58:00	70,012
21-Oct-20 09:59:00	27.51	21-Oct-20 09:59:00	24.39	21-Oct-20 09:59:00	5.03	21-Oct-20 09:59:00	70,073
21-Oct-20 10:00:00	27.46	21-Oct-20 10:00:00	24.40	21-Oct-20 10:00:00	4.99	21-Oct-20 10:00:00	70,076
21-Oct-20 10:01:00	27.30	21-Oct-20 10:01:00	24.33	21-Oct-20 10:01:00	4.95	21-Oct-20 10:01:00	70,067
21-Oct-20 10:02:00	27.23	21-Oct-20 10:02:00	24.26	21-Oct-20 10:02:00	4.95	21-Oct-20 10:02:00	70,085
21-Oct-20 10:03:00	26.88	21-Oct-20 10:03:00	24.21	21-Oct-20 10:03:00	4.79	21-Oct-20 10:03:00	72,131
21-Oct-20 10:04:00	27.01	21-Oct-20 10:04:00	24.55	21-Oct-20 10:04:00	4.63	21-Oct-20 10:04:00	71,312
21-Oct-20 10:05:00	27.34	21-Oct-20 10:05:00	24.28	21-Oct-20 10:05:00	5.00	21-Oct-20 10:05:00	69,731
21-Oct-20 10:06:00	27.25	21-Oct-20 10:06:00	24.18	21-Oct-20 10:06:00	5.02	21-Oct-20 10:06:00	69,715
21-Oct-20 10:07:00	27.23	21-Oct-20 10:07:00	24.15	21-Oct-20 10:07:00	5.02	21-Oct-20 10:07:00	69,756
21-Oct-20 10:08:00	27.05	21-Oct-20 10:08:00	24.09	21-Oct-20 10:08:00	4.95	21-Oct-20 10:08:00	71,065
21-Oct-20 10:09:00	27.01	21-Oct-20 10:09:00	24.37	21-Oct-20 10:09:00	4.75	21-Oct-20 10:09:00	71,465
21-Oct-20 10:10:00	27.06	21-Oct-20 10:10:00	24.30	21-Oct-20 10:10:00	4.81	21-Oct-20 10:10:00	70,026
21-Oct-20 10:11:00	27.19	21-Oct-20 10:11:00	23.99	21-Oct-20 10:11:00	5.11	21-Oct-20 10:11:00	69,915
21-Oct-20 10:12:00	26.96	21-Oct-20 10:12:00	24.05	21-Oct-20 10:12:00	4.94	21-Oct-20 10:12:00	70,660
21-Oct-20 10:13:00	26.84	21-Oct-20 10:13:00	24.05	21-Oct-20 10:13:00	4.86	21-Oct-20 10:13:00	72,363
21-Oct-20 10:14:00	27.18	21-Oct-20 10:14:00	24.52	21-Oct-20 10:14:00	4.75	21-Oct-20 10:14:00	70,097
<b>Average Stack NOx, ppm (3% O2)</b>	<b>27.18</b>	<b>Average Stack Raw NOx, ppm</b>	<b>24.27</b>	<b>Average Stack O2 (%)</b>	<b>4.92</b>	<b>Average Gas Flow (SCFH)</b>	<b>70,572</b>

Rata 6

New-Indy Oxnard, LLC  
 Nebraska Boiler CEMS RATA TEST - November 21, 2020

10/21/2020 10:15 10/21/2020 10:37 1 m

NOx, ppm (3% O2)		Raw NOx, ppm		O2 (%)		Gas Flow (SCFH)	
931AI391_CNOX		931AI393.AT		931AI392.AT		931FI374B.ft	
21-Oct-20 10:15:00	26.94	21-Oct-20 10:15:00	24.16	21-Oct-20 10:15:00	4.83	21-Oct-20 10:15:00	76,095
21-Oct-20 10:16:00	27.00	21-Oct-20 10:16:00	24.94	21-Oct-20 10:16:00	4.36	21-Oct-20 10:16:00	78,293
21-Oct-20 10:17:00	27.85	21-Oct-20 10:17:00	24.87	21-Oct-20 10:17:00	4.91	21-Oct-20 10:17:00	70,327
21-Oct-20 10:18:00	27.82	21-Oct-20 10:18:00	24.01	21-Oct-20 10:18:00	5.45	21-Oct-20 10:18:00	68,200
21-Oct-20 10:19:00	26.95	21-Oct-20 10:19:00	23.87	21-Oct-20 10:19:00	5.04	21-Oct-20 10:19:00	71,328
21-Oct-20 10:20:00	27.02	21-Oct-20 10:20:00	24.15	21-Oct-20 10:20:00	4.91	21-Oct-20 10:20:00	71,182
21-Oct-20 10:21:00	27.46	21-Oct-20 10:21:00	24.08	21-Oct-20 10:21:00	5.19	21-Oct-20 10:21:00	69,277
21-Oct-20 10:22:00	26.97	21-Oct-20 10:22:00	23.93	21-Oct-20 10:22:00	5.02	21-Oct-20 10:22:00	71,920
21-Oct-20 10:23:00	26.95	21-Oct-20 10:23:00	24.30	21-Oct-20 10:23:00	4.77	21-Oct-20 10:23:00	72,739
21-Oct-20 10:24:00	27.26	21-Oct-20 10:24:00	24.42	21-Oct-20 10:24:00	4.87	21-Oct-20 10:24:00	70,836
21-Oct-20 10:25:00	27.46	21-Oct-20 10:25:00	24.21	21-Oct-20 10:25:00	5.13	21-Oct-20 10:25:00	70,942
21-Oct-20 10:26:00	26.79	21-Oct-20 10:26:00	24.29	21-Oct-20 10:26:00	4.69	21-Oct-20 10:26:00	74,482
21-Oct-20 10:27:00	27.28	21-Oct-20 10:27:00	24.72	21-Oct-20 10:27:00	4.68	21-Oct-20 10:27:00	72,018
21-Oct-20 10:28:00	28.12	21-Oct-20 10:28:00	24.49	21-Oct-20 10:28:00	5.31	21-Oct-20 10:28:00	67,691
21-Oct-20 10:29:00	27.29	21-Oct-20 10:29:00	24.26	21-Oct-20 10:29:00	5.02	21-Oct-20 10:29:00	71,329
21-Oct-20 10:30:00	27.21	21-Oct-20 10:30:00	24.48	21-Oct-20 10:30:00	4.80	21-Oct-20 10:30:00	71,757
21-Oct-20 10:31:00	27.28	21-Oct-20 10:31:00	24.49	21-Oct-20 10:31:00	4.83	21-Oct-20 10:31:00	71,860
21-Oct-20 10:32:00	27.47	21-Oct-20 10:32:00	24.60	21-Oct-20 10:32:00	4.87	21-Oct-20 10:32:00	71,914
21-Oct-20 10:33:00	27.35	21-Oct-20 10:33:00	24.59	21-Oct-20 10:33:00	4.81	21-Oct-20 10:33:00	71,744
21-Oct-20 10:34:00	27.73	21-Oct-20 10:34:00	24.68	21-Oct-20 10:34:00	4.98	21-Oct-20 10:34:00	69,537
21-Oct-20 10:35:00	27.94	21-Oct-20 10:35:00	24.63	21-Oct-20 10:35:00	5.12	21-Oct-20 10:35:00	69,156
21-Oct-20 10:36:00	27.94	21-Oct-20 10:36:00	24.63	21-Oct-20 10:36:00	5.12	21-Oct-20 10:36:00	69,156
<b>Average Stack NOx, ppm (3% O2)</b>	<b>27.37</b>	<b>Average Stack Raw NOx, ppm</b>	<b>24.40</b>	<b>Average Stack O2 (%)</b>	<b>4.94</b>	<b>Average Gas Flow (SCFH)</b>	<b>71,445</b>

New-Indy Oxnard, LLC  
 Nebraska Boiler CEMS RATA TEST - November 21, 2020

*Rata 7*

10/21/2020 10:50 10/21/2020 11:12 1 m

NOx, ppm (3% O2)		Raw NOx, ppm		O2 (%)		Gas Flow (SCFH)	
931AI391_CNOX		931AI393.AT		931AI392.AT		931FI374B.R	
21-Oct-20 10:50:00	26.92	21-Oct-20 10:50:00	24.21	21-Oct-20 10:50:00	4.80	21-Oct-20 10:50:00	71,287
21-Oct-20 10:51:00	26.87	21-Oct-20 10:51:00	24.16	21-Oct-20 10:51:00	4.80	21-Oct-20 10:51:00	71,334
21-Oct-20 10:52:00	26.93	21-Oct-20 10:52:00	24.16	21-Oct-20 10:52:00	4.84	21-Oct-20 10:52:00	71,325
21-Oct-20 10:53:00	26.85	21-Oct-20 10:53:00	24.18	21-Oct-20 10:53:00	4.78	21-Oct-20 10:53:00	71,364
21-Oct-20 10:54:00	26.84	21-Oct-20 10:54:00	24.09	21-Oct-20 10:54:00	4.83	21-Oct-20 10:54:00	71,315
21-Oct-20 10:55:00	26.92	21-Oct-20 10:55:00	24.10	21-Oct-20 10:55:00	4.87	21-Oct-20 10:55:00	71,374
21-Oct-20 10:56:00	26.95	21-Oct-20 10:56:00	24.14	21-Oct-20 10:56:00	4.88	21-Oct-20 10:56:00	71,381
21-Oct-20 10:57:00	27.17	21-Oct-20 10:57:00	24.30	21-Oct-20 10:57:00	4.89	21-Oct-20 10:57:00	70,045
21-Oct-20 10:58:00	27.45	21-Oct-20 10:58:00	24.10	21-Oct-20 10:58:00	5.18	21-Oct-20 10:58:00	68,906
21-Oct-20 10:59:00	26.77	21-Oct-20 10:59:00	24.00	21-Oct-20 10:59:00	4.85	21-Oct-20 10:59:00	71,754
21-Oct-20 11:00:00	26.59	21-Oct-20 11:00:00	24.29	21-Oct-20 11:00:00	4.56	21-Oct-20 11:00:00	72,812
21-Oct-20 11:01:00	27.10	21-Oct-20 11:01:00	24.44	21-Oct-20 11:01:00	4.76	21-Oct-20 11:01:00	70,429
21-Oct-20 11:02:00	27.22	21-Oct-20 11:02:00	24.18	21-Oct-20 11:02:00	5.00	21-Oct-20 11:02:00	69,735
21-Oct-20 11:03:00	27.12	21-Oct-20 11:03:00	24.08	21-Oct-20 11:03:00	4.99	21-Oct-20 11:03:00	70,238
21-Oct-20 11:04:00	26.76	21-Oct-20 11:04:00	24.08	21-Oct-20 11:04:00	4.80	21-Oct-20 11:04:00	71,221
21-Oct-20 11:05:00	26.84	21-Oct-20 11:05:00	24.22	21-Oct-20 11:05:00	4.76	21-Oct-20 11:05:00	71,190
21-Oct-20 11:06:00	27.16	21-Oct-20 11:06:00	24.54	21-Oct-20 11:06:00	4.74	21-Oct-20 11:06:00	71,170
21-Oct-20 11:07:00	27.22	21-Oct-20 11:07:00	24.60	21-Oct-20 11:07:00	4.73	21-Oct-20 11:07:00	70,813
21-Oct-20 11:08:00	27.45	21-Oct-20 11:08:00	24.42	21-Oct-20 11:08:00	4.97	21-Oct-20 11:08:00	69,469
21-Oct-20 11:09:00	26.76	21-Oct-20 11:09:00	24.28	21-Oct-20 11:09:00	4.67	21-Oct-20 11:09:00	72,869
21-Oct-20 11:10:00	26.97	21-Oct-20 11:10:00	24.88	21-Oct-20 11:10:00	4.52	21-Oct-20 11:10:00	71,688
21-Oct-20 11:11:00	27.73	21-Oct-20 11:11:00	24.50	21-Oct-20 11:11:00	5.08	21-Oct-20 11:11:00	67,230
<b>Average Stack NOx, ppm (3% O2)</b>	<b>27.03</b>	<b>Average Stack Raw NOx, ppm</b>	<b>24.26</b>	<b>Average Stack O2 (%)</b>	<b>4.83</b>	<b>Average Gas Flow (SCFH)</b>	<b>70,861</b>

*Rata 8*

New-Indy Oxnard, LLC  
Nebraska Boiler CEMS RATA TEST - November 21, 2020

10/21/2020 11:12 10/21/2020 11:34 1 m

NOx, ppm (3% O2)		Raw NOx, ppm		O2 (%)		Gas Flow (SCFH)	
931AI391_CNOX		931AI393.AT		931AI392.AT		931FI374B.ft	
21-Oct-20 11:12:00	27.15	21-Oct-20 11:12:00	24.16	21-Oct-20 11:12:00	4.98	21-Oct-20 11:12:00	70,143
21-Oct-20 11:13:00	26.82	21-Oct-20 11:13:00	24.38	21-Oct-20 11:13:00	4.64	21-Oct-20 11:13:00	72,493
21-Oct-20 11:14:00	26.70	21-Oct-20 11:14:00	24.53	21-Oct-20 11:14:00	4.46	21-Oct-20 11:14:00	72,812
21-Oct-20 11:15:00	27.23	21-Oct-20 11:15:00	24.70	21-Oct-20 11:15:00	4.67	21-Oct-20 11:15:00	70,744
21-Oct-20 11:16:00	27.46	21-Oct-20 11:16:00	24.45	21-Oct-20 11:16:00	4.96	21-Oct-20 11:16:00	69,886
21-Oct-20 11:17:00	27.34	21-Oct-20 11:17:00	24.35	21-Oct-20 11:17:00	4.95	21-Oct-20 11:17:00	69,932
21-Oct-20 11:18:00	27.06	21-Oct-20 11:18:00	24.26	21-Oct-20 11:18:00	4.85	21-Oct-20 11:18:00	71,084
21-Oct-20 11:19:00	26.96	21-Oct-20 11:19:00	24.30	21-Oct-20 11:19:00	4.76	21-Oct-20 11:19:00	71,141
21-Oct-20 11:20:00	27.05	21-Oct-20 11:20:00	24.33	21-Oct-20 11:20:00	4.81	21-Oct-20 11:20:00	71,131
21-Oct-20 11:21:00	27.23	21-Oct-20 11:21:00	24.46	21-Oct-20 11:21:00	4.82	21-Oct-20 11:21:00	71,090
21-Oct-20 11:22:00	27.29	21-Oct-20 11:22:00	24.51	21-Oct-20 11:22:00	4.82	21-Oct-20 11:22:00	70,533
21-Oct-20 11:23:00	27.54	21-Oct-20 11:23:00	24.39	21-Oct-20 11:23:00	5.04	21-Oct-20 11:23:00	69,653
21-Oct-20 11:24:00	26.89	21-Oct-20 11:24:00	24.32	21-Oct-20 11:24:00	4.71	21-Oct-20 11:24:00	72,587
21-Oct-20 11:25:00	26.80	21-Oct-20 11:25:00	24.58	21-Oct-20 11:25:00	4.48	21-Oct-20 11:25:00	71,968
21-Oct-20 11:26:00	27.34	21-Oct-20 11:26:00	24.41	21-Oct-20 11:26:00	4.92	21-Oct-20 11:26:00	67,573
21-Oct-20 11:27:00	26.89	21-Oct-20 11:27:00	23.84	21-Oct-20 11:27:00	5.03	21-Oct-20 11:27:00	69,973
21-Oct-20 11:28:00	26.45	21-Oct-20 11:28:00	24.17	21-Oct-20 11:28:00	4.55	21-Oct-20 11:28:00	73,276
21-Oct-20 11:29:00	26.61	21-Oct-20 11:29:00	24.45	21-Oct-20 11:29:00	4.47	21-Oct-20 11:29:00	72,791
21-Oct-20 11:30:00	27.19	21-Oct-20 11:30:00	24.45	21-Oct-20 11:30:00	4.79	21-Oct-20 11:30:00	68,496
21-Oct-20 11:31:00	27.14	21-Oct-20 11:31:00	23.87	21-Oct-20 11:31:00	5.16	21-Oct-20 11:31:00	69,141
21-Oct-20 11:32:00	26.42	21-Oct-20 11:32:00	24.03	21-Oct-20 11:32:00	4.61	21-Oct-20 11:32:00	72,801
21-Oct-20 11:33:00	26.65	21-Oct-20 11:33:00	24.41	21-Oct-20 11:33:00	4.50	21-Oct-20 11:33:00	71,250
<b>Average Stack NOx, ppm (3% O2)</b>	<b>27.01</b>	<b>Average Stack Raw NOx, ppm</b>	<b>24.33</b>	<b>Average Stack O2 (%)</b>	<b>4.77</b>	<b>Average Gas Flow (SCFH)</b>	<b>70,932</b>

*Kata*

New-Indy Oxnard, LLC  
 Nebraska Boiler CEMS RATA TEST - November 21, 2020

10/21/2020 11:34 10/21/2020 11:56 1 m

NOx, ppm (3% O2)		Raw NOx, ppm		O2 (%)		Gas Flow (SCFH)	
931AI391_CNOX		931AI393.AT		931AI392.AT		931FI374B.ft	
21-Oct-20 11:34:00	27.28	21-Oct-20 11:34:00	24.07	21-Oct-20 11:34:00	5.10	21-Oct-20 11:34:00	68,276
21-Oct-20 11:35:00	26.69	21-Oct-20 11:35:00	23.86	21-Oct-20 11:35:00	4.89	21-Oct-20 11:35:00	71,683
21-Oct-20 11:36:00	26.45	21-Oct-20 11:36:00	24.18	21-Oct-20 11:36:00	4.54	21-Oct-20 11:36:00	73,108
21-Oct-20 11:37:00	26.78	21-Oct-20 11:37:00	24.36	21-Oct-20 11:37:00	4.82	21-Oct-20 11:37:00	71,479
21-Oct-20 11:38:00	27.31	21-Oct-20 11:38:00	24.22	21-Oct-20 11:38:00	5.02	21-Oct-20 11:38:00	69,811
21-Oct-20 11:39:00	27.22	21-Oct-20 11:39:00	24.07	21-Oct-20 11:39:00	5.09	21-Oct-20 11:39:00	69,769
21-Oct-20 11:40:00	26.92	21-Oct-20 11:40:00	23.99	21-Oct-20 11:40:00	4.94	21-Oct-20 11:40:00	71,077
21-Oct-20 11:41:00	26.77	21-Oct-20 11:41:00	24.09	21-Oct-20 11:41:00	4.79	21-Oct-20 11:41:00	71,284
21-Oct-20 11:42:00	26.77	21-Oct-20 11:42:00	24.14	21-Oct-20 11:42:00	4.77	21-Oct-20 11:42:00	71,353
21-Oct-20 11:43:00	26.84	21-Oct-20 11:43:00	24.21	21-Oct-20 11:43:00	4.78	21-Oct-20 11:43:00	71,356
21-Oct-20 11:44:00	27.25	21-Oct-20 11:44:00	24.51	21-Oct-20 11:44:00	4.80	21-Oct-20 11:44:00	71,360
21-Oct-20 11:45:00	26.98	21-Oct-20 11:45:00	24.30	21-Oct-20 11:45:00	4.77	21-Oct-20 11:45:00	71,381
21-Oct-20 11:46:00	26.81	21-Oct-20 11:46:00	24.27	21-Oct-20 11:46:00	4.75	21-Oct-20 11:46:00	71,398
21-Oct-20 11:47:00	27.09	21-Oct-20 11:47:00	24.37	21-Oct-20 11:47:00	4.80	21-Oct-20 11:47:00	70,172
21-Oct-20 11:48:00	27.37	21-Oct-20 11:48:00	24.12	21-Oct-20 11:48:00	5.12	21-Oct-20 11:48:00	69,120
21-Oct-20 11:49:00	26.85	21-Oct-20 11:49:00	24.03	21-Oct-20 11:49:00	4.88	21-Oct-20 11:49:00	71,598
21-Oct-20 11:50:00	26.67	21-Oct-20 11:50:00	24.39	21-Oct-20 11:50:00	4.53	21-Oct-20 11:50:00	72,349
21-Oct-20 11:51:00	27.12	21-Oct-20 11:51:00	24.38	21-Oct-20 11:51:00	4.81	21-Oct-20 11:51:00	69,804
21-Oct-20 11:52:00	26.90	21-Oct-20 11:52:00	24.05	21-Oct-20 11:52:00	4.90	21-Oct-20 11:52:00	71,567
21-Oct-20 11:53:00	26.65	21-Oct-20 11:53:00	24.32	21-Oct-20 11:53:00	4.57	21-Oct-20 11:53:00	73,132
21-Oct-20 11:54:00	26.86	21-Oct-20 11:54:00	24.51	21-Oct-20 11:54:00	4.56	21-Oct-20 11:54:00	71,416
<b>Average Stack NOx, ppm (3% O2)</b>	<b>26.95</b>	<b>Average Stack Raw NOx, ppm</b>	<b>24.21</b>	<b>Average Stack O2 (%)</b>	<b>4.82</b>	<b>Average Gas Flow (SCFH)</b>	<b>71,018</b>



# AIR Testing Inc.

## BOILER DATA SHEET

Client: New Indy  
 Site: OXNARD  
 Unit: NEBRASKA BOILER

Date: 10-21-20  
 Job #: 220-012  
 Lab #: 23002

Make: NEBRASKA  
 S/N: 2D-1983  
 Model: N8E9551  
 BTU/hr Rating: 10.8 MMBTU/hr  
 Fuel Type: NG  
 FGR Equipped: YES

Stack Diameter: 44"  
 "A" Distance: 30'  
 "B" Distance: 26"  
 Port Size: 3/8"  
 M or F: HOLE  
 Offset: —



Boiler Load: Run 1: Normal Run 2: \_\_\_\_\_ Run 3: ▶ % of Rating  
 FGR Setting: Run 1: 40' Run 2: 40 Run 3: 40

Notes: \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_



**EPA METHOD 20 MEASUREMENT  
SYSTEM PERFORMANCE TIME**

**RESPONSE TIME**

DATE OF TEST: 10-21-20

PROBE LOCATION: Stack

ANALYZER TYPE & MODEL #

NOx: API 200EH

CO: —

O2: SERVOMEX 1400

Span Gas Concentration/Analyzer Full Scale Setting:

NOx; Gas, ppmv: 40.0

Full Scale setting, ppmv 50

CO; Gas, ppmv: —

Full Scale setting, ppmv —

O2; Gas, %: 8.05

Full Scale setting, % 25

UPSCALE TIME - Analyzer + Sampling System:

Trial #	Nox	CO	O2
1	<u>56</u>	<u>—</u>	<u>46</u> seconds
2	<u>55</u>	<u>—</u>	<u>45</u> seconds
3	<u>55</u>	<u>—</u>	<u>46</u> seconds
Average Response	<u>56</u>	<u>—</u>	<u>46</u> seconds

DOWNSCALE - Analyzer + Sampling System:

Trial #	Nox	CO	O2
1	<u>53</u>	<u>—</u>	<u>43</u> seconds
2	<u>53</u>	<u>—</u>	<u>43</u> seconds
3	<u>53</u>	<u>—</u>	<u>43</u> seconds
Average Response	<u>53</u>	<u>—</u>	<u>43</u> seconds

Slower Average 56 — 46 seconds

**EPA METHOD 20 MEASUREMENT  
SYSTEM PERFORMANCE TIME**

**RESPONSE TIME**

DATE OF TEST: 10-20-20

PROBE LOCATION: Stack

ANALYZER TYPE & MODEL #

NOx: Honiba VA 5000

CO: —

O2: —

Span Gas Concentration/Analyzer Full Scale Setting:

NOx; Gas, ppmv: 81.43

Full Scale setting, ppmv 100

CO; Gas, ppmv: —

Full Scale setting, ppmv —

O2; Gas, %: —

Full Scale setting, % —

UPSCALE TIME - Analyzer + Sampling System:

Trial #	Nox	CO	O2
1	<u>60</u>	<u>—</u>	<u>—</u> seconds
2	<u>60</u>	<u>—</u>	<u>—</u> seconds
3	<u>60</u>	<u>—</u>	<u>—</u> seconds
Average Response	<u>60</u>	<u>—</u>	<u>—</u> seconds

DOWNSCALE - Analyzer + Sampling System:

Trial #	Nox	CO	O2
1	<u>55</u>	<u>—</u>	<u>—</u> seconds
2	<u>55</u>	<u>—</u>	<u>—</u> seconds
3	<u>55</u>	<u>—</u>	<u>—</u> seconds
Average Response	<u>55</u>	<u>—</u>	<u>—</u> seconds

Slower  
Average

60

—

— seconds

<b>DESCRIPTION OF METHOD(S) USED</b>
--------------------------------------

## EPA Method 7 - Nitrogen Oxides (NOx)

### Method 7: NOx from stationary sources

**Applicability:** Applicable range:  $2 \text{ mg/m}^3 < [\text{NOx}]_{\text{as NO}_2} < 400 \text{ mg/m}^3$  ( $\text{m}^3$  at dry standard conditions)

**Principle:** A grab sample is collected in a evacuated flask containing a dilute sulfuric acid-hydrogen peroxide absorbing solution, and the NOx (except NO) are measured colorimetrically using phenoldisulfonic acid.

### Method 7A: NOx from stationary sources

**Applicability:** Applicable range:  $125 \text{ mg/m}^3 (65 \text{ ppm}) < [\text{NOx}]_{\text{as NO}_2} < 1250 \text{ mg/m}^3 (655 \text{ ppm})$

**Principle:** A grab sample is collected in a evacuated flask containing a dilute sulfuric acid-hydrogen peroxide absorbing solution, and the NOx (except NO) are oxidized to nitrates and measured by ion chromatography.

### Method 7B: NOx from nitric acid plants

**Applicability:** Applicable range:  $57 \text{ mg/m}^3 (30 \text{ ppm}) < [\text{NOx}]_{\text{as NO}_2} < 1550 \text{ mg/m}^3 (786 \text{ ppm})$

**Principle:** A grab sample is collected in a evacuated flask containing a dilute sulfuric acid-hydrogen peroxide absorbing solution, and the NOx (except NO) are measured by ultraviolet absorption.

### Method 7C and 7D: NOx from fossil fuel fired generators, electric utility plants or nitric acid plants

**Applicability:** Applicable range:  $13 \text{ mg/m}^3 (7 \text{ ppm}) < [\text{NOx}]_{\text{as NO}_2} < 1782 \text{ mg/m}^3 (932 \text{ ppm})$

**Principle:** A grab sample is collected from the stack in alkaline-potassium permanganate solution.  $\text{NOx} = \text{NO} + \text{NO}_2$

**7C:** NOx are oxidized to  $\text{NO}_2$  and  $\text{NO}_3$ . Cadmium is used to reduce  $\text{NO}_3$  to  $\text{NO}_2$ , which is analyzed colorimetrically.

**7D:** NOx are oxidized to  $\text{NO}_3$  and analyzed by ion chromatography.

### Method 7E: NOx from stationary source - Using Instrumental Analyzer

**Analytical range:** Determined by instrumental design. A portion of it is selected by choosing the span of the monitoring system. The span shall be selected such that the pollutant gas concentration equivalent to the emission standard is not less than 30% of the span. If at any time during the test, the pollutant concentration exceeds the span, the test is invalid.

**Sensitivity:** The Minimum Detection Limit (MDL) depends on the analytical range, the span, and the signal-to-noise ratio of the measurement system. The MDL should be less than 2% of the span.

**Principle:** A gas sample is continuously extracted from a stack and conveyed to a *chemiluminescent* analyzer for determination of NOx concentration.

**Measurement system:** Sample interface, Gas analyzer, Data recorder.

**NO<sub>2</sub> to NO converter:** Not necessary if the NO<sub>2</sub> portion of the exhaust gas is less than 5% of the total NOx concentration.

**NOx analyzer:** Based on the chemiluminescent reaction of NO and ozone to form NO<sub>2</sub> in an excited state. Light emission is monitored through an optical filter by a high sensitivity photomultiplier tube, the output of which is electrically processed so it is linearly proportional to the NO concentration.

**NOx calibration gases (NO in N<sub>2</sub>):** Three calibration gases shall be used. Ambient air may be used for the zero gas.

## CARB Method 100

### Generalized Procedures For Continuous Gaseous Emission Stack Sampling

**Principle:** A sample of an exhaust gas stream is continuously extracted, conditioned, and analyzed by instruments to determine average emission concentrations. By combining this information with stack gas flowrate and moisture, mass emission rates can be determined.

**Applicability:** Determination of emissions of oxides of nitrogen, carbon monoxide, carbon dioxide, sulfur dioxide, total hydrocarbons, and oxygen from stationary sources flowing gas streams in ducts, stacks, and flues. This method is an alternative to EPA reference methods 3A, 6C, 7E, 10, and 25A and B.

**Range:** The range is selected so that the sample gas concentration for each run is between 10 and 95% of the range for each pollutant of interest.

**Measurement system:** Sample interface, Gas analyzer, Data acquisition

**Probe Nozzle:** Quartz, borosilicate, stainless steel, porcelain or aluminum oxide

**Probe:** Shall have an inside diameter of 6 mm or larger and shall be of one of the materials mentioned above

**Sample line:** Teflon or other material that does not alter the sample gas

**Sample conditioner:** capable of reducing sample gas temperature to 60°F or 20°F lower than ambient temp.

**Probe filter:** internal or external, to prevent accumulation of particulate in the measurement system

**Probe Calibration System:** allows calibration of the instruments by introducing calibration gases

**Sample System Heaters:** if needed, to prevent condensation of water or hydrocarbons

**Gas analyzers:** shall be housed in a temperature-controlled, vibration-free environment:

- CO & CO<sub>2</sub>: Nondispersive infrared analyzer
- O<sub>2</sub> : Paramagnetic or electrochemical (fuel cell) analyzer
- Total HC : Flame ionization detector (FID) or nondispersive infrared analyzer (NDIR)
- NO<sub>x</sub> : Chemiluminescent analyzer, minimum of 90% efficiency for NO<sub>2</sub> conversion
- SO<sub>2</sub> : Infrared or ultraviolet absorption or fluorescence analyzer

**Data recorder:** Provides a permanent record of gas analyzer data. Consist of a strip chart recorder and an electronic data logger capable of integration at ten second interval. Resolution or readability should be 0.5% of range.

**Range:** Upper limit of the gas concentration measurement range displayed on the data recorder

**Calibration gas (CalGas):** A gas of known concentration in an inert diluent gas

- High-Range: 80 to 100% of the range
- Mid-Range: 40 to 60% of the range
- Zero gas: impurity concentration < 0.25% of the range → Nitrogen

**Analyzer calibration error:** The difference between the known concentration of the CalGas and the gas concentration exhibited by the gas analyzer when the CalGas is introduced directly to the analyzer.

**Performance Spec:** less than 2% of the range for the zero, mid-range, and high-range CalGas.

**Sampling system BIAS:** The difference between the gas concentrations exhibited by the measurement system when CalGas is introduced at the sampling probe tip filter and when the same CalGas is introduced directly to the analyzer.

**Performance Spec:** less than 5% of the range for the zero, mid-range, and high-range CalGas.

**Zero Drift:** The difference in the measurement system responses at a zero concentration level during the initial calibration, and final calibration check after a test.

No adjustment to the measurement system is allowed at that point.

**Performance Spec:** less than 3% of the range over the period of each run.

**Calibration Drift (or Span Drift):** The difference in the measurement system responses at a mid-range

concentration level during the initial calibration, and final calibration check after a test.

No adjustment to the measurement system is allowed at that point.

**Performance Spec:** less than 3% of the range over the period of each run.

**Response time:** the time required for the system to display 95% of a step change in gas concentration on the data recorder.

**Interference response:** The output response of the measurement system to a component in the sample gas, other than the gas component being measured.

**Performance Spec:** less than 5% of the range.

#### **Measurement System Performance Test Procedures:**

- **Cleaning/Assembly of sample train:** Flush probe, lines and sample conditioner with Deionized Water, then acetone. Dry with filtered dry air. Assemble as shown in accompanying figure.
- **Pretest Leak Check:** Leak check the vacuum side of the assembly at the maximum pump vacuum.
- **Calibrate analyzers and data recorders:** Allow analyzers to warm up. Adjust system components to achieve manufacturer's recommended sampling rates. Introduce CalGases directly to the instruments and make all necessary adjustments to calibrate the analyzer and data recorder.
- **Analyzer calibration error check:** at the beginning of test run
  1. Introduce zero, mid-range, high range CalGas
  2. Make no adjustments to the system except those necessary to achieve the correct flow rate.
  3. If invalid calibration is exhibited (> 2% of the range), take corrective action and repeat check.
- **Sampling system BIAS check: Mandatory**
  1. Backflush gas through the probe as necessary to prevent particulate buildup
  2. Introduce zero, and either mid-range or high-range (whichever is closest to effluent concentration)
  3. Make no adjustments to the system except those necessary to achieve the correct flow rate.
  4. If invalid calibration is exhibited (> 5% of the range), take corrective action and repeat check.
  5. If adjustment to the analyzer is required, first repeat the analyzer calibration error check, then repeat the sampling system BIAS check.
- **Stratification check:** Select traverse points according to ARB Methods 1 and 2. Multipoint gas sampling (traversing) must be performed unless data is available to demonstrate the mean pollutant concentration is *less than 10% different* from any single point.

#### **Sample Collection:**

1. Insert probe in stack
2. If traversing required, leave probe at each position for at least the system response time + 1 minute.
3. Three sample runs is recommended. See District Rules and permit conditions for special requirements.
4. When test duration exceeds one hour, conduct system BIAS checks every 2 hours. Adjust settings as necessary, mark strip charts and record changes in log books.
5. Strip chart must include the following information: pollutant, source, range, calibration cylinder ID number, chart speeds, date, time, operator.
6. **At end of run, perform sampling system BIAS Check.** Make no adjustments except to flow rates.

**INSTRUMENT INFORMATION**


## Method 100 – Sample Train Assembly

### Probe:

1. AIRx Testing Services, Inc. uses a stainless steel sampling probe.
2. The probe is connected to the sample conditioner using a heated (if necessary) Teflon sampling line.

### Sample Conditioner (Condenser System):

1. The sample conditioner consists of a moisture knock-out bottle immersed in an ice bath.
2. All parts of the conditioner exposed to the sample are either glass, stainless steel, or Teflon.
3. The sample conditioner is setup so that the sample gas is not bubbled through the condensate.
4. A temperature gauge is used to determine the temperature of the condenser outlet.



### Manifold Panel & Strip Chart Recorder:

1. The sample gas exiting the condenser is transported to the instrument manifold panel.
2. The sample manifold directs the sample gas through to the various analyzers.
3. All components of the system that contact the sample are either stainless steel, Teflon, or glass.
4. The data is continuously recorded on a Soltec 10" Strip chart recorder with a 1% resolution.

### Calibration System:

1. The system is composed of the Analyzer Calibration and the System BIAS Check.
2. Certified Calibration gases are supplied to the analyzers for calibration purposes.
3. The selection of zero or span calibration gases or sample gas is accomplished by operation of selector knobs located on the main flow control panel.

Nitrogen Oxides: – Chemiluminescent Analyzer: Based on the chemiluminescent reaction of NO and ozone to form NO<sub>2</sub> in an excited state. Light emission is monitored through an optical filter by a high sensitivity photomultiplier tube, the output of which is electronically processed so it is linearly proportional to the NO concentration. The output is in units of ppmv.

Carbon Monoxide: – Non-Dispersive Infrared (NDIR) Analyzer: Radiation from an infrared source is passed through a narrow band-pass filter and a multiple optical pass sample cell where absorption by the sample gas occurs. The infrared radiation exits the sample cell and falls on a solid state infrared detector. The output is in units of ppmv.

Oxygen: – Electrochemical Analyzer: Oxygen in the flue gas sample diffuses through a Teflon membrane and is reduced on the surface of the cathode. A corresponding oxidation occurs at the anode and an electric current proportional to the concentration of oxygen is produced. The output is in units of percent O<sub>2</sub> by volume.

Carbon Dioxide: – Non-Dispersive Infrared (NDIR) Analyzer: The instrument measures the differential in infrared energy absorbed from energy beams passed through a reference cell (containing a gas selected to have minimal absorption of infrared energy in the wavelength absorbed by CO<sub>2</sub>) and a sample cell through which the sample gas flows continuously. The output is in units of percent CO<sub>2</sub> by volume.



**SOURCE EMISSION INSTRUMENTATION LIST**

**OXYGEN**

<p><b><u>Unit No. - 5:</u></b>                  Manufacturer: California Analytical Instruments                  Model No.: 110P                  Serial No.: T02034                  Method: Paramagnetic                  Range (%): 0-5, 10, 25</p>	<p><b><u>Unit No. - 13:</u></b>                  Manufacturer: Servomex                  Model No.: 1400                  Serial No.: X1420/B707                  Method: Paramagnetic                  Range (%): 0-25</p>
<p><b><u>Unit No. - 7:</u></b>                  Manufacturer: Teledyne                  Model No.: 320-AX                  Serial No.: 108742                  Method: Electrochemical                  Range (%): 0-5, 10, 25</p>	
<p><b><u>Unit No. - 9:</u></b>                  Manufacturer: Servomex                  Model No.: 1400                  Serial No.: 01420/B701/730                  Method: Paramagnetic                  Range (%): 0-25, 100</p>	
<p><b><u>Unit No. - 10:</u></b>                  Manufacturer: Servomex                  Model No.: 1400                  Serial No.: 01420/B308                  Method: Paramagnetic                  Range (%): 0-25</p>	
<p><b><u>Unit No. - 11:</u></b>                  Manufacturer: Teledyne                  Model No.: 320-A                  Serial No.: 111211                  Method: Electrochemical                  Range (%): 0-5, 10, 25</p>	
<p><b><u>Unit No. - 12:</u></b>                  Manufacturer: Servomex                  Model No.: 1400                  Serial No.: 01420/B7103                  Method: Paramagnetic                  Range (%): 0-25, 100</p>	



**SOURCE-EMISSION-INSTRUMENTATION LIST**

**OXIDES OF NITROGEN**

<p><b>Unit No. - 1:</b>                  Manufacturer: API                  Model No.: 200 EH                  Serial No.: 233                  Method: Chemiluminescence                  Range (ppmv) 0-5000</p>	<p><b>Unit No. - 6:</b>                  Manufacturer: API                  Model No.: 200 A                  Serial No.: 1013                  Method: Chemiluminescence                  Range (ppmv) 0-50</p>
<p><b>Unit No. - 2:</b>                  Manufacturer: API                  Model No.: 200 EH                  Serial No.: 234                  Method: Chemiluminescence                  Range (ppmv) 0-5000</p>	<p><b>Unit No. - 7:</b>                  Manufacturer: Thermo Environmental (TECO)                  Model No.: 10AR                  Serial No.: 25559-221                  Method: Chemiluminescence                  Range (ppmv) 0-2.5, 10, 25, 100, 250, 1000, 2500, 10000</p>
<p><b>Unit No. - 3:</b>                  Manufacturer: API                  Model No.: 200 EH                  Serial No.: 109                  Method: Chemiluminescence                  Range (ppmv) 0-5000</p>	<p><b>Unit No. - 8:</b>                  Manufacturer: Thermo Environmental (TECO)                  Model No.: 10AR                  Serial No.: 38586-258                  Method: Chemiluminescence                  Range (ppmv) 0-2.5, 10, 25, 100, 250, 1000, 2500, 10000</p>
<p><b>Unit No. - 4:</b>                  Manufacturer: API                  Model No.: 200 EH                  Serial No.: 442                  Method: Chemiluminescence                  Range (ppmv) 0-5000</p>	
<p><b>Unit No. - 5:</b>                  Manufacturer: API                  Model No.: 200 EH                  Serial No.: 441                  Method: Chemiluminescence                  Range (ppmv) 0-5000</p>	



**QUALITY ASSURANCE**



Praxair  
 5700 South Alameda Street  
 Los Angeles, CA 90058  
 Tel: (323) 585-2154 Fax: (714) 542-6689  
 PGVPID: F22017

DocNumber: 000105588

## CERTIFICATE OF ANALYSIS / EPA PROTOCOL GAS

**Customer & Order Information:**

PRAXAIR PKG OXNARD CA HPS  
 455 E WOOLEY RD  
 OXNARD CA 93030

Praxair Order Number: 70207599  
 Customer P. O. Number:  
 Customer Reference Number:

Fill Date: 2/27/2017  
 Part Number: NI CD803E-AS  
 Lot Number: 109705811  
 Cylinder Style & Outlet: AS CGA 580  
 Cylinder Pressure & Volume: 2000 psig 140 cu. ft.

**Certified Concentration:**

Expiration Date:	3/8/2025	NIST Traceable
Cylinder Number:	CC60235	Analytical Uncertainty:
7.98 %	CARBON DIOXIDE	± 0.4 %
3.98 %	OXYGEN	± 0.4 %
Balance	NITROGEN	

**Certification Information:** Certification Date: 3/8/2017 Term: 96 Months Expiration Date: 3/8/2025  
 This cylinder was certified according to the 2012 EPA Traceability Protocol, Document #EPA-600/R-12/531, using Procedure G1. Do Not Use this Standard if Pressure is less than 100 PSIG.

CO2 responses have been corrected for Oxygen IR Broadening effect. O2 responses have been corrected for CO2 interference.

**Analytical Data:**

(R=Reference Standard, Z=Zero Gas, C=Gas Candidate)

**1. Component: CARBON DIOXIDE**

Requested Concentration: 8 %  
 Certified Concentration: 7.98 %  
 Instrument Used: Horiba VIA-510 S/N 574763031  
 Analytical Method: NDIR  
 Last Multipoint Calibration: 2/13/2017

Reference Standard Type: GMS  
 Ref. Std. Cylinder #: DT0008914  
 Ref. Std. Conc: 7.00 %  
 Ref. Std. Traceable to SRM #: 1674b  
 SRM Sample #: 7-H-07  
 SRM Cylinder #: FF10631

<b>First Analysis Data:</b>		<b>Date:</b> 3/8/2017	
Z: 0	R: 6.98	C: 7.97	Conc: 7.985
R: 6.98	Z: 0	C: 7.97	Conc: 7.985
Z: 0	C: 7.97	R: 7	Conc: 7.985
UOM: %	Mean Test Assay:		7.985 %

<b>Second Analysis Data:</b>		<b>Date:</b>	
Z: 0	R: 0	C: 0	Conc: 0
R: 0	Z: 0	C: 0	Conc: 0
Z: 0	C: 0	R: 0	Conc: 0
UOM: %	Mean Test Assay:		0 %

**2. Component: OXYGEN**

Requested Concentration: 4 %  
 Certified Concentration: 3.98 %  
 Instrument Used: OXYMAT 5E  
 Analytical Method: PARAMAGNETIC  
 Last Multipoint Calibration: 3/3/2017

Reference Standard Type: GMS  
 Ref. Std. Cylinder #: CC244935  
 Ref. Std. Conc: 5.01%  
 Ref. Std. Traceable to SRM #: 2658a  
 SRM Sample #: 72-D-28  
 SRM Cylinder #: CAL016862

<b>First Analysis Data:</b>		<b>Date:</b> 3/8/2017	
Z: 0	R: 5.01	C: 3.97	Conc: 3.975
R: 5	Z: 0	C: 3.97	Conc: 3.975
Z: 0	C: 3.97	R: 5	Conc: 3.975
UOM: %	Mean Test Assay:		3.975 %

<b>Second Analysis Data:</b>		<b>Date:</b>	
Z: 0	R: 0	C: 0	Conc: 0
R: 0	Z: 0	C: 0	Conc: 0
Z: 0	C: 0	R: 0	Conc: 0
UOM: %	Mean Test Assay:		0 %

Analyzed by:

Ying Yu

  
 Ying Yu

Certified by:

  
 Nassim Haddad

Information contained herein has been prepared at your request by qualified experts within Praxair Distribution, Inc. While we believe that the information is accurate within the limits of the analytical methods employed and is complete to the extent of the specific analyses performed, we make no warranty or representation as to the suitability of the use of the information for any purpose. The information is offered with the understanding that any use of the information is at the sole discretion and risk of the user. In no event shall the liability of Praxair Distribution, Inc., arising out of the use of the information contained herein exceed the fee established for providing such information.



**AIR LIQUIDE**

Air Liquide America  
Specialty Gases LLC



Intertek

**RATA CLASS**

*Guaranteed +/- 1% Accuracy*

8832 DICE ROAD, SANTA FE SPRINGS, CA 90670-2516

Phone: 800-323-2212

Fax: 562-464-5262

**CERTIFICATE OF ACCURACY: EPA Protocol Gas**

Assay Laboratory - PGVP Vendor ID: A52013

Document.# : 51261426-004

Customer

AIR LIQUIDE AMERICA SPECIALTY GASES LLC P.O. No.: AIRX  
8832 DICE ROAD  
SANTA FE SPRINGS, CA 90670-2516

ALHC-HCF-FRESNO-MEDGAS (LOC 24337)  
3703 BAGLEY AVENUE  
TRANSFER ACCOUNT  
FRESNO CA 93725  
US.

**ANALYTICAL INFORMATION Gas Type : CO2,O2,BALN**

This certification was performed according to EPA Traceability Protocol For Assay & Certification of Gaseous Calibration Standards; Procedure G-1.

EPA/600/R-12/531; May 2012. Do not use this standard if pressure is less than 100 psig.

Cylinder Number: CC160530  
Cylinder Pressure: 2000 PSIG

Certification Date: 02Jul2013

Exp. Date: 03Jul2021  
Batch No: SBO0073261

COMPONENT	CERTIFIED CONCENTRATION (Moles)		ACCURACY ABSOLUTE / RELATIVE			
	Value	%	Value	%	Value	%
CARBON DIOXIDE	20.1	%	.13	%	/	.6 %
OXYGEN	8.05	%	.06	%	/	.7 %
NITROGEN	BALANCE					

**TRACEABILITY**

REFERENCE STANDARD

COMPONENT	CONCENTRATION	UNCERTAINTY	CYLINDER	TYPE/SRM SAMPLE	EXP. DATE
CARBON DIOXIDE	17.87 %	0.110 %	K026459	NTRM 1800/	01Mar2019
OXYGEN	10.03 %	0.070 %	K027803	NTRM 2658/	01Feb2016

ANALYTICAL METHOD

1st Analysis: 02Jul2013

COMPONENT	INSTRUMENT	ANALYTICAL PRINCIPLE	CALIBRATED	CONCENTRATION
CARBON DIOXIDE	GC-TCO/HP 589D II/J140A38374	GC	28Jun2013	20.11 %
OXYGEN	GC-TCO/HP 589D II/J140A38374	GC	22Jun2013	8.051 %

APPROVED BY:

*DC*  
DC

DocNumber: 253580



Praxair Distribution, Inc.  
 5700 S. Alameda Street  
 Los Angeles CA 90058  
 Tel: 323-585-2164  
 Fax: 714-542-6689  
 PGVP ID: F22019

**CERTIFICATE OF ANALYSIS / EPA PROTOCOL GAS**

**Customer & Order Information**

PRAXAIR PKG OXNARD CA HPS  
 455 E WOOLEY RD  
 OXNARD CA 93030

Certificate Issuance Date: 06/27/2019

Praxair Order Number: 70898288

Part Number: NI NO20ME-AS

Customer PO Number: 78981976

Fill Date: 06/07/2019

Lot Number: 70086915805

Cylinder Style & Outlet: AS

CGA 680

Cylinder Pressure and Volume: 2000 psig 140 ltr

**Certified Concentration**

Expiration Date:	06/27/2022	NIST Traceable
Cylinder Number:	CC724466	Expanded Uncertainty
20.1 ppm	Nitric oxide	± 0.5 %
Balance	Nitrogen	

**ProSpec EZ Cert**



**For Reference Only:** NOx 20.2 ppm

**Certification Information:** Certification Date: 06/27/2019 Term: 36 Months Expiration Date: 06/27/2022

This cylinder was certified according to the 2012 EPA Traceability Protocol, Document #EPA-600/R-12/531, using Procedure G1.  
 Do Not Use this Standard if Pressure is less than 100 PSIG.

**Analytical Data:**

(R=Reference Standard, Z=Zero Gas, C=Gas Candidate)

1. Component:

Nitric oxide

Reference Standard:

Type / Cylinder #: GMIS / CC2819

Requested Concentration: 20 ppm

Concentration / Uncertainty: 19.90 ppm ±0.504%

Certified Concentration: 20.1 ppm

Expiration Date: 03/28/2022

Instrument Used: Thermo Electron 42i-LS S/N 1030645077

Traceable to: SRM # / Sample # / Cylinder #: APEX1161149 / NA / APEX1161149

Analytical Method: Chemiluminescence

SRM Concentration / Uncertainty: 20.03 ppm / ±0.10 ppm

Last Multipoint Calibration: 06/10/2019

SRM Expiration Date: 01/27/2020

First Analysis Data:				Date
Z: 0	R: 19.9	C: 20	Conc: 20	06/20/2019
R: 19.88	Z: 0	C: 20.1	Conc: 20.1	
Z: 0	C: 20.1	R: 19.91	Conc: 20.1	
UOM: ppm	Mean Test Assay: 20.1 ppm			

Second Analysis Data:				Date
Z: 0	R: 19.0	C: 20.1	Conc: 20.1	06/27/2019
R: 19.89	Z: 0	C: 20.2	Conc: 20.2	
Z: 0	C: 20.1	R: 19.89	Conc: 20.1	
UOM: ppm	Mean Test Assay: 20.1 ppm			

Analyzed By

Henry Koung

Certified By

Leeanna Flores



**Customer & Order Information:**

PRAXAIR PKG OXNARD CA HPS  
455 E WOOLEY RD,  
OXNARD, CA 93030-7224  
Praxair Order Number: 71378203  
Customer PO Number: 79369772

Certificate Issuance Date: 7/10/2020  
Certification Date: 7/10/2020  
Lot Number: 70086018201  
Part Number: NI NX20MP-AS  
DocNumber: 232651  
Expiration Date: 7/10/2021

**CERTIFICATE OF ANALYSIS**  
*Primary Standard*

Component	Requested Concentration (Molar)	Certified Concentration (Molar)	Analytical Reference	Analytical Uncertainty
Nitrogen dioxide (as NOx)	20 ppm	19.01 ppm	1	± 1%
Nitrogen	Balance	Balance		

Cylinder Style: AS  
Cylinder Pressure @ 70 F: 2000 psig  
Cylinder Volume: 142 ft<sup>3</sup>  
Valve Outlet Connection: CGA 660  
Cylinder Number(s): CC3240

Fill Date: 6/30/2020  
Analysis Date: 7/10/2020  
Filling Method: Gravimetric

Comments: This mixture contains 2% Oxygen.

Analyst: Leeanna Flores

Approved-Signer: Henry Koung

**Key to Analytical Techniques:**

Reference	Analytical Instrument - Analytical Principle
1	MKS MG2031 - FTIR

The gas calibration cylinder standard prepared by Praxair Distribution, Inc. is considered a certified standard. It is prepared by gravimetric, volumetric, or partial pressure techniques. The calibration standard provided is certified against Praxair Distribution, Inc. Reference Materials which are traceable to the International System of Units (SI) through either weights traceable to the National Institute of Standards and Technology (NIST) or Measurement Canada, or through NIST Standard Reference Materials or equivalent where available.

Note: All expressions for concentration (e.g., % or ppm) are for gas phase, by volume (e.g., ppmv) unless otherwise noted. Analytical uncertainty is expressed as a Relative % unless otherwise noted.

**IMPORTANT**

The information contained herein has been prepared at your request by personnel within Praxair Distribution, Inc. While we believe the information is accurate within the limits of the analytical methods employed and is complete to the extent of the specific analyses performed, we make no warranty or representation as to the suitability of the use of the information for any particular purpose. The information is offered with the understanding that any use of the information is at the sole discretion and risk of the user. In no event shall liability of Praxair Distribution, Inc. arising out of the use of the information contained herein exceed the fee established for providing such information.

DocNumber: 224276



Praxair Distribution, Inc.  
 5700 S. Alameda Street  
 Los Angeles CA 90058  
 Tel: 323-585-2154  
 Fax: 714-542-6689  
 PGVP ID: F22018

**CERTIFICATE OF ANALYSIS / EPA PROTOCOL GAS**

**Customer & Order Information**

PRAXAIR PKG OXNARD CA HPS  
 455 E WOOLEY RD  
 OXNARD CA 93030

Certificate Modification Date: 09/18/2018  
 Praxair Order Number: 70717073  
 Part Number: NI NO40ME-AS

Fill Date: 08/05/2018  
 Lot Number: 70088824602  
 Cylinder Style & Outlet: AS CGA 680  
 Cylinder Pressure and Volume: 2000 psig 140 ltr

**Certified Concentration**

Expiration Date:	09/18/2021	NIST Traceable
Cylinder Number:	DT0028765	Expanded Uncertainty
39.9 ppm	Nitric oxide	± 0.8 %
Balance	Nitrogen	

**ProSpec EZ Cert**



**For Reference Only:** NOx 40.0 ppm

**Certification Information:** Certification Date: 09/18/2018 Term: 36 Months Expiration Date: 09/18/2021

This cylinder was certified according to the 2012 EPA Traceability Protocol, Document #EPA-600/R-12/531, using Procedure G1.  
 Do Not Use this Standard if Pressure is less than 100 PSIG.

**Analytical Data:**

(R=Reference Standard, Z=Zero Gas, C=Gas Candidate)

1. Component: Nitric oxide  
 Requested Concentration: 40 ppm  
 Certified Concentration: 39.9 ppm  
 Instrument Used: Thermo Electron 42i-LS S/N 1030845077  
 Analytical Method: Chemiluminescence  
 Last Multipoint Calibration: 09/10/2018

Reference Standard: Type / Cylinder #: GMIS / CC505947  
 Concentration / Uncertainty: 49.44 ppm ±0.714%  
 Expiration Date: 08/08/2021  
 Traceable to: SRM # / Sample # / Cylinder #: 1683b / 45-V-37 / CAL017868  
 SRM Concentration / Uncertainty: 48.79 / ±0.34 ppm  
 SRM Expiration Date: 03/25/2019

First Analysis Date:		Date	
Z: 0	R: 49.4	C: 30.0	Conc: 30.8
R: 49.4	Z: 0	C: 39.9	Conc: 39.9
Z: 0	C: 39.9	R: 49.4	Conc: 39.9
UOM: ppm		Mean Test Assay: 39.9 ppm	

Second Analysis Date:		Date	
Z: 0	R: 49.4	C: 36.7	Conc: 25.7
R: 49.4	Z: 0	C: 39.9	Conc: 39.9
Z: 0	C: 39.9	R: 49.4	Conc: 39.9
UOM: ppm		Mean Test Assay: 39.9 ppm	

Analyzed By

Henry Koung

Certified By

Amanda Beal



Nebraska Boiler  
Changed: 09-29-20

**CERTIFICATE OF ANALYSIS**  
**Grade of Product: EPA Protocol**

Part Number: E02NI99E15A01M5 Reference Number: 48-401725601-1  
Cylinder Number: CC105295 Cylinder Volume: 144.3 CF  
Laboratory: 124 - Los Angeles (SAP) - CA Cylinder Pressure: 2015 PSIG  
PGVP Number: B32020 Valve Outlet: 660  
Gas Code: NO,NOX,BALN Certification Date: Feb 13, 2020

Expiration Date: Feb 13, 2028

Certification performed in accordance with "EPA Traceability Protocol for Assay and Certification of Gaseous Calibration Standards (May 2012)" document EPA 800/R-12/531, using the assay procedures listed. Analytical Methodology does not require correction for analytical interference. This cylinder has a total analytical uncertainty as stated below with a confidence level of 95%. There are no significant impurities which affect the use of this calibration mixture. All concentrations are on a mole/mole basis unless otherwise noted.

Do Not Use This Cylinder below 100 psig, i.e. 0.7 megapascals.

**ANALYTICAL RESULTS**

Component	Requested Concentration	Actual Concentration	Protocol Method	Total Relative Uncertainty	Assay Dates
NOX	82.00 PPM	81.43 PPM	G1	+/- 1.0% NIST Traceable	02/06/2020, 02/13/2020
NITRIC OXIDE	82.00 PPM	81.23 PPM	G1	+/- 1.1% NIST Traceable	02/06/2020, 02/13/2020
NITROGEN	Balance				

**CALIBRATION STANDARDS**

Type	Lot ID	Cylinder No	Concentration	Uncertainty	Expiration Date
NTRM	13010422	KAL004028	97.6 PPM NITRIC OXIDE/NITROGEN	+/- 0.8%	Jul 23, 2025
PRM	12386	D885025	9.91 PPM NITROGEN DIOXIDE/AIR	+/- 2.0%	Feb 20, 2020
GMIS	401203436104	CC513876	4.653 PPM NITROGEN DIOXIDE/NITROGEN	+/- 2.1%	May 02, 2022

The SRM, PRM or RGM-noted above is only in reference to the GMIS used in the assay and not part of the analysis.

**ANALYTICAL EQUIPMENT**

Instrument/Make/Model	Analytical Principle	Last Multipoint Calibration
Nicolet 6700 AHR0801551 NO	FTIR	Jan 21, 2020
Nicolet 6700 AHR0801551 NO2	FTIR	Jan 21, 2020

Triad Data Available Upon Request



*[Signature]*  
Approved for Release

**AIR TESTING SERVICES, INC.**

**RATA EMISSIONS TESTING  
NEW INDY OXNARD  
N8E9551 – NEBRASKA BOILER  
NOx Analyzer - Horiba VA-50 (Old)  
VCAPCD PTO #0157-281  
Test Date: October 22, 2020**

**Prepared for:  
NEW INDY OXNARD  
5936 Perkins Road,  
Oxnard, CA 93033**

**Attn: Robyn Lebrilla**

**Submitted to:  
Ventura County Air Pollution Control District  
669 County Square Drive  
Ventura, California 93003**

**Attn: Ed Swede**

**Prepared by:  
AIRx Testing Services, Inc.  
2472 Eastman Avenue #34  
Ventura, CA 93003**

**Job No.: 23022**

**Lab No.: 220-113**

  
Tom Porter  
Tom Porter

**Wesley Hart; Source Test Engineer**

**Tom Porter; Vice President of Testing Services**

**Submitted:  
November 3, 2020**

Table of Contents

<i>Section</i>	<i>Page</i>
RATA Results	1
Introduction	2
Calculations	5
RATA Bias Adjustments	6
Field Data and Strip Charts	8
RATA CEM Data	9
Strip Charts	19
Facility CEM Data Sheets	23
Boiler Data Sheet	32
R.M. Response Times	33
Descriptions of Methods Used	34
EPA Method 7E	35
CARB Method 100	36
Instrument Information	38
Extractive Gas Sample System	39
CEM Instrument List	40
Quality Assurance	42
Calibration Gas Certification Sheets	43

# AIR TESTING SERVICES, INC.

New Indy  
Nebraska Boiler  
10/22/2020

CEMS RATA  
Calculations

Run	AIRx Testing - Reference Method		
	NOx ppmv	O2 Dry %	NOx ppmv @ 3% O2
1	25.17	4.39	27.29
2	24.56	4.46	26.75
3	24.47	4.43	26.59
4	24.62	4.45	26.79
5	25.18	4.48	27.45
6	24.86	4.46	27.07
7	26.03	4.54	28.48
8	26.04	4.47	28.38
9	25.70	4.50	28.04

Run	New Indy	Nebraska	CEMS
	NOx ppmv	O2 Dry %	NOx ppmv @ 3%
1	28.06	4.86	31.31
2	27.99	4.89	31.30
3	27.90	4.87	31.16
4	27.69	4.85	30.89
5	27.39	4.87	30.60
6	27.20	4.80	30.24
7	26.89	4.91	30.10
8	26.66	4.82	29.66
9	26.50	4.86	29.58

Run	Reference Method - CEM, Difference		
	NOx ppmv	O2 Dry %	NOx ppmv @ 3%
1	-2.89	-0.47	-4.02
2	-3.43	-0.43	-4.55
3	-3.43	-0.44	-4.57
4	-3.07	-0.40	-4.10
5	-2.21	-0.39	-3.15
6	-2.34	-0.34	-3.17
7	-0.86	-0.37	-1.62
8	-0.62	-0.35	-1.28
9	-0.80	-0.36	-1.54

Arithmetic Mean, d  
Standard Deviation, Sd  
Confidence Coefficient, CC  
Avg Reference Method, RM  
Relative Accuracy, RA

-2.18	-0.39	-3.11
1.15	0.04	1.32
0.88	0.03	1.02
25.18	4.47	27.43
12.17	9.58	15.06

NOTE: Calculations based on "Code of Federal Regulations 40", 1988, Part 60, Appendix B, Specification 2, p. 939.

**INTRODUCTION**

## 1.0 INTRODUCTION

On October 22, 2020 AIRx Testing Services conducted RATA source testing on a natural gas fired 108/MMBTU Nebraska Boiler N8E9551. The unit is located at the New Indy Plant at 5936 Perkins Road, in Oxnard. The boiler is used for the generation of steam for specific facility operations. The tests were required to satisfy requirements detailed in PTO #0157-281.

The objective of the testing was to determine the relative accuracy (RATA) of the facility's boiler CEM equipment (Old NOx Analyzer). Operational data reported by New Indy personnel included NOx readings from their Horiba analyzer and O2 readings from their Rosemont analyzer.

**PROCESS DETAILS:** The Nebraska Boiler is rated for 108 MMBtu/hr heat input and exhausts into 44" exhaust stack. The sample was taken off the existing facility CEM system NOx/O2 probe with it being located at A and B distances of 360" and 26" respectively (8.2 diameters upstream and 0.6 diameters downstream).

**CEM DATA:** CEM data was recorded by the Data Acquisition System (DAS) during the test program in the form of one (1) minute averages. The one (1) minute data readings were then printed out and averaged to obtain data for comparison to New Indy's CEM data. The facility CEM data was obtained by plant personnel during the test program and provided to the test team. The CEM data monitored NOx and O2 concentrations and also correcting the NOx to 3% oxygen.

**CONTINUOUS MONITORS:** NOx and O2 sampling utilized a stainless steel sampling probe connected with a Teflon sampling line to a sample conditioner. The dry gases were then transported through a Teflon line to the sample gas flow panel for distribution to the individual analyzers. Three (3) 63 minute sampling runs for NOx and O2 were performed in accordance with EPA Method 7e and CARB Method 100. NOx measurements were made utilizing an API 200EH chemiluminescent analyzer. Oxygen concentration was measured utilizing a Servomex Model 1400 paramagnetic analyzer. Initial three (3) point calibrations were performed on the analyzers. Subsequent calibrations were performed through the sample system probe tip (Bias checks). All initial bias checks correlated with the initial instrument calibrations to within 2% of analyzer range. Bias calibrations were performed before and after each test run and used to correct the emissions data for any analyzer drift. All monitor calibrations were performed with EPA Protocol 1 calibration gases.

During all tests, the NOx and O2 RM analyzers were operated on the 50 ppmv and 10% scales, respectively.

**TEST DAY DISCUSSION:** On 10-22-2020, RATA testing was performed on the facilities' old NOx Analyzer (Horiba VA-50). Three (3) 63 minute runs were performed and broken into three 21 minute RATA test runs. This is acceptable according to the test methodology and by Ed Swede with the VCAPCD. The three test runs were conducted while the boiler was operating under normal conditions.

**1.0 INTRODUCTION (cont)**

All monitored data was continuously recorded on a 10" analog chart recorder. Chart speed was maintained at 10 cm/hr. Analyzer output was also recorded with a PC based Data Acquisition System (DAS) and imported into spreadsheets for determination of average run concentrations.

All reporting and calculations have been performed using VCAPCD standard conditions of 68°F and 29.92 inches of Hg. If you have any questions regarding this test program or report, please contact the undersigned at (805) 644-1099.

Respectfully Submitted  
AIRx Testing Services, Inc.



Wesley Hart  
Source Test Engineer

Reviewed by:



Tom Porter  
Vice President of Testing Services

CC: Ventura County APCD  
Attn: Ed Swede

<b>CALCULATIONS</b>
---------------------



## RATA Bias Adjustment

Facility: New Indy  
 Source: Nebraska Boiler  
 Date: 10/22/20

### Run No. 1

Parameter	Measured Conc. (ppm,%)	Reference Span gas (ppm,%)	Initial Zero (ppm,%)	Final Zero (ppm,%)	Average Zero (ppm,%)	Initial Span (ppm,%)	Final Span (ppm,%)	Average Span (ppm,%)	Adjusted Conc. (ppm,%)
NOx	25.14	20.2	0.0	0.0	0.0	20.2	20.2	20.2	25.17
O2	4.50	3.98	0.0	0.0	0.0	4.0	4.0	4.1	4.39

### Run No. 2

Parameter	Measured Conc. (ppm,%)	Reference Span gas (ppm,%)	Initial Zero (ppm,%)	Final Zero (ppm,%)	Average Zero (ppm,%)	Initial Span (ppm,%)	Final Span (ppm,%)	Average Span (ppm,%)	Adjusted Conc. (ppm,%)
NOx	24.53	20.2	0.0	0.0	0.0	20.2	20.2	20.2	24.56
O2	4.50	3.98	0.0	0.0	0.0	4.0	4.0	4.0	4.46

### Run No. 3

Parameter	Measured Conc. (ppm,%)	Reference Span gas (ppm,%)	Initial Zero (ppm,%)	Final Zero (ppm,%)	Average Zero (ppm,%)	Initial Span (ppm,%)	Final Span (ppm,%)	Average Span (ppm,%)	Adjusted Conc. (ppm,%)
NOx	24.44	20.2	0.0	0.0	0.0	20.2	20.2	20.2	24.47
O2	4.47	3.98	0.0	0.0	0.0	4.0	4.0	4.0	4.43

### Run No. 4

Parameter	Measured Conc. (ppm,%)	Reference Span gas (ppm,%)	Initial Zero (ppm,%)	Final Zero (ppm,%)	Average Zero (ppm,%)	Initial Span (ppm,%)	Final Span (ppm,%)	Average Span (ppm,%)	Adjusted Conc. (ppm,%)
NOx	24.53	20.2	0.0	0.0	0.0	20.2	20.1	20.1	24.62
O2	4.47	3.98	0.0	0.0	0.0	4.0	4.0	4.0	4.45

### Run No. 5

Parameter	Measured Conc. (ppm,%)	Reference Span gas (ppm,%)	Initial Zero (ppm,%)	Final Zero (ppm,%)	Average Zero (ppm,%)	Initial Span (ppm,%)	Final Span (ppm,%)	Average Span (ppm,%)	Adjusted Conc. (ppm,%)
NOx	25.09	20.2	0.0	0.0	0.0	20.2	20.1	20.1	25.18
O2	4.50	3.98	0.0	0.0	0.0	4.0	4.0	4.0	4.48

## RATA Bias Adjustment

Facility: New Indy  
 Source: Nebraska Boiler  
 Date: 10/22/20

### Run No. 6

Parameter	Measured Conc. (ppm,%)	Reference Span gas (ppm.%)	Initial Zero (ppm.%)	Final Zero (ppm.%)	Average Zero (ppm.%)	Initial Span (ppm.%)	Final Span (ppm.%)	Average Span (ppm.%)	Adjusted Conc. (ppm.%)
NOx	24.76	20.2	0.0	0.0	0.0	20.2	20.1	20.1	<b>24.86</b>
O2	4.48	3.98	0.0	0.0	0.0	4.0	4.0	4.0	<b>4.46</b>

### Run No. 7

Parameter	Measured Conc. (ppm,%)	Reference Span gas (ppm.%)	Initial Zero (ppm.%)	Final Zero (ppm.%)	Average Zero (ppm.%)	Initial Span (ppm.%)	Final Span (ppm.%)	Average Span (ppm.%)	Adjusted Conc. (ppm.%)
NOx	25.90	20.2	0.0	0.0	0.0	20.1	20.1	20.1	<b>26.03</b>
O2	4.55	3.98	0.0	0.0	0.0	4.0	4.0	4.0	<b>4.54</b>

### Run No. 8

Parameter	Measured Conc. (ppm,%)	Reference Span gas (ppm.%)	Initial Zero (ppm.%)	Final Zero (ppm.%)	Average Zero (ppm.%)	Initial Span (ppm.%)	Final Span (ppm.%)	Average Span (ppm.%)	Adjusted Conc. (ppm.%)
NOx	25.91	20.2	0.0	0.0	0.0	20.1	20.1	20.1	<b>26.04</b>
O2	4.48	3.98	0.0	0.0	0.0	4.0	4.0	4.0	<b>4.47</b>

### Run No. 9

Parameter	Measured Conc. (ppm,%)	Reference Span gas (ppm.%)	Initial Zero (ppm.%)	Final Zero (ppm.%)	Average Zero (ppm.%)	Initial Span (ppm.%)	Final Span (ppm.%)	Average Span (ppm.%)	Adjusted Conc. (ppm.%)
NOx	25.57	20.2	0.0	0.0	0.0	20.1	20.1	20.1	<b>25.70</b>
O2	4.50	3.98	0.0	0.0	0.0	4.0	4.0	4.0	<b>4.50</b>