



February 5, 2024

Mr. Keith Macias
Manager, Compliance Division
Ventura County Air Quality Management District
4567 Telephone Road, 2nd Floor
Ventura, CA 93003

**Subject: Submission of Part 70 Permit Annual Compliance Certification
McGrath Peaker Generating Station, Permit No. 07891**


Dear Mr. Macias,

Southern California Edison Company (SCE) is submitting the Part 70 Permit Annual Compliance Certification for McGrath Peaker Generating Station, Permit #07891, for the period between January 1, 2023 and December 31, 2023. During the subject period there were no equipment breakdowns, permit deviations, emergency conditions, or activities involving the demolition of asbestos-containing material at the facility. All monitoring required by the permit was conducted.

Also, please find enclosed the Annual Compliance Certification Signature Cover Form, Permit Attachment Forms, Source Test Summary Form, and supporting documents.

If you have any questions regarding these reports, please contact Ali Aleshaiker at (909) 353-9609 or by email at Ali.Aleshaiker@sce.com.

Sincerely,

DocuSigned by:

8CDD88E198D4CA...

Dan Golden
Principal Manager
Generation, Western Operations

Enclosures

cc: Ms. Roshni Brahmhatt
Enforcement & Compliance Enforcement Division
EPA Region 9
75 Hawthorne Street
San Francisco, CA 94105

Certificate Of Completion

Envelope Id: EBDE25F18C8841B9AB919B64DF8E1137
 Subject: Complete with DocuSign: McGrath 2023 Cover Letter.docx
 Custom Envelope Field:
 Source Envelope:
 Document Pages: 1
 Certificate Pages: 1
 AutoNav: Enabled
 EnvelopeId Stamping: Enabled
 Time Zone: (UTC-08:00) Pacific Time (US & Canada)

Status: Completed

Envelope Originator:
 Ali Aleshaiker
 P.O. Box 700
 Rosemead, CA 91770
 ALI.ALESHAIKER@SCE.COM
 IP Address: 163.116.248.39

Record Tracking

Status: Original
 2/5/2024 12:58:01 PM


Holder: Ali Aleshaiker
 ALI.ALESHAIKER@SCE.COM

Location: DocuSign

Signer Events

Dan Golden
 dan.golden@sce.com
 Principal Manager, Generation
 Southern California Edison
 Security Level: Email, Account Authentication
 (None)

Signature



Signature Adoption: Pre-selected Style
 Using IP Address: 163.116.248.48

Timestamp

Sent: 2/5/2024 12:58:25 PM
 Viewed: 2/5/2024 12:59:10 PM
 Signed: 2/5/2024 12:59:14 PM

Electronic Record and Signature Disclosure:
 Not Offered via DocuSign

In Person Signer Events	Signature	Timestamp
Editor Delivery Events	Status	Timestamp
Agent Delivery Events	Status	Timestamp
Intermediary Delivery Events	Status	Timestamp
Certified Delivery Events	Status	Timestamp
Carbon Copy Events	Status	Timestamp
Witness Events	Signature	Timestamp
Notary Events	Signature	Timestamp
Envelope Summary Events	Status	Timestamps
Envelope Sent	Hashed/Encrypted	2/5/2024 12:58:25 PM
Certified Delivered	Security Checked	2/5/2024 12:59:10 PM
Signing Complete	Security Checked	2/5/2024 12:59:14 PM
Completed	Security Checked	2/5/2024 12:59:14 PM
Payment Events	Status	Timestamps



Ventura County
Air Pollution
Control District

**ANNUAL COMPLIANCE CERTIFICATION
SIGNATURE COVER FORM**

TV Permit # 07891

A copy of each Annual Compliance Certification shall be submitted to EPA, Region 9, at the following address:


Ms. Roshni Brahmhatt
Enforcement & Compliance Enforcement Division
EPA Region 9
75 Hawthorne Street
San Francisco, CA 94105

Confidentiality

All information in a Part 70 permit compliance certification is public information. The Part 70 permit is also public information.

Certification by Responsible Official

I certify that, based on information and belief formed after reasonable inquiry, the statements and information in this compliance certification are true, accurate, and complete.

<p>Signature and Title of Responsible Official:</p>  <p>Title: Principal Manager, Generation - Western Operations</p>	<p>Date: <u>1-29-24</u></p>
--	-----------------------------

<p>Time Period Covered by Compliance Certification</p> <p><u>01</u> / <u>01</u> / <u>2023</u> (MM/DD/YY) to <u>12</u> / <u>31</u> / <u>2023</u> (MM/DD/YY)</p>
--



Ventura County
Air Pollution
Control District

ANNUAL COMPLIANCE CERTIFICATION PERMIT ATTACHMENT FORM

Period Covered by Compliance Certification: 01 / 01 / 23 (MM/DD/YY) to 12 / 31 / 23 (MM/DD/YY)

<p>A. Attachment # or Permit Condition #: 7891-T1-161 Conditions # 1-5</p>	<p>D. Frequency of monitoring: Annual Source Test and Continuous Emissions Monitoring</p>
<p>B. Description: Gas Turbine Emissions Limits: - NOx emissions shall not exceed 2.5 ppmvd @ 15% O2 - NOx emissions shall not exceed 25 ppm @ 15% O2 (4-hr rolling avg) - ROC emissions shall not exceed 2.0 ppmvd @ 15% O2 - CO emissions shall not exceed 6.0 ppmvd @ 15% O2 - NH3 emissions shall not exceed 5.0 ppmvd @ 15% O2</p>	<p>E. Source test reference method, if applicable. Attach Source Test Summary Form, if applicable</p> <p>Source Test Summary Form attached.</p>
<p>C. Method of monitoring: Continuous Emissions Monitoring. Annual compliance source tests performed on 8/22/23.</p>	<p>F. Currently in Compliance? (Y or N): <u>Y</u></p> <p>G. Compliance Status? (C or I): <u>C</u></p> <p>H. *Excursions, exceedances, or other non-compliance? (Y or N): <u>N</u> *If yes, attach Deviation Summary Form</p>

<p>A. Attachment # or Permit Condition #: 7891-T1-161 Conditions # 6-12</p>	<p>D. Frequency of monitoring: Continuous monitoring</p>
<p>B. Description: Continuous Emissions Monitor - Permittee shall: -Install, Operate, Maintain, and Calibrate CEMS pursuant to Rule 74.23; Rule 103, NSPS KKKK and 40 CFR 75. -Promptly report emission violations as indicated by the CEMS -Maintain permanent CEMS records. -Maintain records of all maintenance activities</p>	<p>E. Source test reference method, if applicable. Attach Source Test Summary Form, if applicable</p> <p>N/A</p>
<p>C. Method of monitoring: Attachment #1: CEMS emissions and natural gas usage records. Attachment #2: CEMS maintenance record. Attachment #3: CEMS calibrations record. Attachment #4: SCR and CO catalyst temperature and pressure devices calibrations records. Attachment #5: Fuel and Ammonia flow-meters calibrations records.</p>	<p>F. Currently in Compliance? (Y or N): <u>Y</u></p> <p>G. Compliance Status? (C or I): <u>C</u></p> <p>H. *Excursions, exceedances, or other non-compliance? (Y or N): <u>N</u> *If yes, attach Deviation Summary Form</p>

<p>A. Attachment # or Permit Condition #: 7891-T1-161 Conditions # 13,14</p>	<p>D. Frequency of monitoring: Continuous monitoring</p>
<p>B. Description: Permittee shall submit operating records pursuant to Rule 74.23.E of: -Actual fuel consumption or operating hour records for the past 12 months; -Annual source test and control system operating parameters Permittee shall submit excess emissions and monitoring report every 6 months pursuant to 40 CFR 60, Subpart KKKK, including 4-hour rolling NOx average</p>	<p>E. Source test reference method, if applicable. Attach Source Test Summary Form, if applicable</p> <p>Source Test Summary Form attached</p>
<p>C. Method of monitoring: An in-line fuel flow meter is used to monitor actual fuel consumption. Attachment #1 includes rolling twelve months total gas consumption. Annual compliance source tests were performed on 8/22/23. Test reports have been submitted to the District. Excess emissions and monitoring systems reports have been submitted to the District</p>	<p>F. Currently in Compliance? (Y or N): <u>Y</u></p> <p>G. Compliance Status? (C or I): <u>C</u></p> <p>H. *Excursions, exceedances, or other non-compliance? (Y or N): <u>N</u> *If yes, attach Deviation Summary Form</p>



ANNUAL COMPLIANCE CERTIFICATION PERMIT ATTACHMENT FORM

Period Covered by Compliance Certification: 01 / 01 / 23 (MM/DD/YY) to 12 / 31 / 23 (MM/DD/YY)

<p>A. Attachment # or Permit Condition #: 7891-T2 Conditions # 1-2</p>	<p>D. Frequency of monitoring: None for PUC quality gas</p>
<p>B. Description: 40 CFR 60 KKKK: SO2 emissions shall not exceed 0.9 lbs/MW-hr or total sulfur in fuel shall not exceed 0.06 lbs/MMBTU heat input</p>	<p>E. Source test reference method, if applicable. Attach Source Test Summary Form, if applicable N/A</p>
<p>C. Method of monitoring: Only PUC quality gas supplied by Southern California Gas Co. is combusted in the turbine, therefore the facility is in compliance with Rule 64 and 40 CFR 60 Subpart KKKK pursuant to 7891-T2 condition #2.</p>	<p>F. Currently in Compliance? (Y or N): <u>Y</u> G. Compliance Status? (C or I): <u>C</u> H. *Excursions, exceedances, or other non-compliance? (Y or N): <u>N</u> <small>*If yes, attach Deviation Summary Form</small></p>

<p>A. Attachment # or Permit Condition #: 7891-T2 Conditions # 3,4</p>	<p>D. Frequency of monitoring: None for PUC quality gas</p>
<p>B. Description: Pursuant to Rule 54.B.1, no person shall discharge sulfur compounds calculated as SO2 in excess of (a) 300 ppmv at point of discharge; (b) 0.25 ppmv, 1-hr average and/or 0.04 ppmv, 24-hr average at ground or sea level.</p>	<p>E. Source test reference method, if applicable. Attach Source Test Summary Form, if applicable N/A</p>
<p>C. Method of monitoring: Combustion of PUC quality natural gas ensures compliance with this requirement. No additional monitoring was required pursuant to 07891-T2 condition #3. No source testing pursuant to condition #4 is required.</p>	<p>F. Currently in Compliance? (Y or N): <u>Y</u> G. Compliance Status? (C or I): <u>C</u> H. *Excursions, exceedances, or other non-compliance? (Y or N): <u>N</u> <small>*If yes, attach Deviation Summary Form</small></p>

<p>A. Attachment # or Permit Condition #: 07891-Engine-161 Condition #1</p>	<p>D. Frequency of monitoring: Continuous</p>
<p>B. Description: The permitted emissions for the 924 BHP Waukesha engine are based on a NOx emission level of 1.25 grams/BHP-hr. Permittee shall maintain manufacturer's emission documentation that the engine meets this emission level.</p>	<p>E. Source test reference method, if applicable. Attach Source Test Summary Form, if applicable N/A</p>
<p>C. Method of monitoring: Attachment #6 is the generator and engine specification. Emissions Performance vs load curve on page 7 of the attachment shows the engine complies with NOx emissions level of 1.25 grams/BHP-hr.</p>	<p>F. Currently in Compliance? (Y or N): <u>Y</u> G. Compliance Status? (C or I): <u>C</u> H. *Excursions, exceedances, or other non-compliance? (Y or N): <u>N</u> <small>*If yes, attach Deviation Summary Form</small></p>



ANNUAL COMPLIANCE CERTIFICATION PERMIT ATTACHMENT FORM

Period Covered by Compliance Certification: 01 / 01 / 23 (MM/DD/YY) to 12 / 31 / 23 (MM/DD/YY)

<p>A. Attachment # or Permit Condition #: 07891-Engine-161 Conditions 2 & 7</p>	<p>D. Frequency of monitoring:</p> <p style="text-align: center; font-weight: bold;">Monthly</p>
<p>B. Description:</p> <p>-Engine shall be used only when electrical power fails, except for testing and maintenance; -Engine recordkeeping requirement</p>	<p>E. Source test reference method, if applicable. Attach Source Test Summary Form, if applicable</p> <p style="text-align: center; font-weight: bold;">N/A</p>
<p>C. Method of monitoring:</p> <p>Attachment #7, engine operating log is used to document all engine operations including emergency use hours and testing/maintenance use hours. The log shows the engine in question operated 4.5 hours in 2023.</p>	<p>F. Currently in Compliance? (Y or N): <u>Y</u></p> <p>G. Compliance Status? (C or I): <u>C</u></p> <p>H. *Excursions, exceedances, or other non-compliance? (Y or N): <u>N</u> *If yes, attach Deviation Summary Form</p>

<p>A. Attachment # or Permit Condition #: 07891-Engine-161 Conditions # 3, 4, 5, 6 & 7</p>	<p>D. Frequency of monitoring:</p> <p style="text-align: center; font-weight: bold;">Monthly</p>
<p>B. Description:</p> <p>-Engine shall not operate more than 200 hrs/yr; -Engine shall be equipped with a non-resettable time meter; -Annual compliance certification shall include engine mfr., model number, operator ID and location. -Total engine hours shall be reported annually.</p>	<p>E. Source test reference method, if applicable. Attach Source Test Summary Form, if applicable</p> <p style="text-align: center; font-weight: bold;">N/A</p>
<p>C. Method of monitoring:</p> <p>Engine operating log is used to document all engine operations including emergency use hours and testing/maintenance use hours. (Attachment #7, engine operating log indicates manufacturer, model number, operator ID and location of use). The engine operated 4.5 hours in 2023.</p>	<p>F. Currently in Compliance? (Y or N): <u>Y</u></p> <p>G. Compliance Status? (C or I): <u>C</u></p> <p>H. *Excursions, exceedances, or other non-compliance? (Y or N): <u>N</u> *If yes, attach Deviation Summary Form</p>

<p>A. Attachment # or Permit Condition #: PO07891PC-111 Condition 1</p>	<p>D. Frequency of monitoring:</p> <p style="text-align: center; font-weight: bold;">Continuous monitoring</p>
<p>B. Description:</p> <p>Annual natural gas limit for turbine operation shall not exceed 1,667 MMSCF/yr.</p>	<p>E. Source test reference method, if applicable. Attach Source Test Summary Form, if applicable</p> <p style="text-align: center; font-weight: bold;">N/A</p>
<p>C. Method of monitoring:</p> <p>An in-line fuel flow meter calibrated annually is used to monitor natural gas combusted in the turbine. Attachment #1 includes the 12-month natural gas consumption for turbine operation, which indicated 116.90 mmscf of natural gas was combusted in the gas turbine in 2023.</p>	<p>F. Currently in Compliance? (Y or N): <u>Y</u></p> <p>G. Compliance Status? (C or I): <u>C</u></p> <p>H. *Excursions, exceedances, or other non-compliance? (Y or N): <u>N</u> *If yes, attach Deviation Summary Form</p>



ANNUAL COMPLIANCE CERTIFICATION PERMIT ATTACHMENT FORM

Period Covered by Compliance Certification: 01 / 01 / 23 (MM/DD/YY) to 12 / 31 / 23 (MM/DD/YY)

<p>A. Attachment # or Permit Condition #: PO07891PC1-111 Condition 2</p>	<p>D. Frequency of monitoring: Continuous Monitoring</p>
<p>B. Description: Turbine annual NOx emissions shall not exceed 4.81 tons/yr. Facility annual NOx emissions shall not exceed 4.99 tons/yr. A rolling 12-month record shall be maintained.</p>	<p>E. Source test reference method, if applicable. Attach Source Test Summary Form, if applicable N/A</p>
<p>C. Method of monitoring: Continuous Emissions Monitoring, Attachment #1 includes the rolling 12-month NOx emissions for 2023, which indicates 1,101 lbs of NOx were emitted from the gas turbine (0.55 tons). Attachment #7 indicates 11.46 lbs of NOx was emitted from the black-start generator engine (0.01 tons). These totals are below the permit limits.</p>	<p>F. Currently in Compliance? (Y or N): <u>Y</u> G. Compliance Status? (C or I): <u>C</u> H. *Excursions, exceedances, or other non-compliance? (Y or N): <u>N</u> <small>*If yes, attach Deviation Summary Form</small></p>

<p>A. Attachment # or Permit Condition #: PO07891PC1-111 Condition 3</p>	<p>D. Frequency of monitoring: Monthly</p>
<p>B. Description: The 924 BHP Waukesha natural gas engine shall not be used for more than 200 hours per year.</p>	<p>E. Source test reference method, if applicable. Attach Source Test Summary Form, if applicable N/A</p>
<p>C. Method of monitoring: A totalizing hour meter documents engine operated hours. All engine operation is documented in an operation log. Attachment #7, engine operating log showed the engine operated 4.5 hours in 2023.</p>	<p>F. Currently in Compliance? (Y or N): <u>Y</u> G. Compliance Status? (C or I): <u>C</u> H. *Excursions, exceedances, or other non-compliance? (Y or N): <u>N</u> <small>*If yes, attach Deviation Summary Form</small></p>

<p>A. Attachment # or Permit Condition #: PO07891PC1-111 Condition 4</p>	<p>D. Frequency of monitoring: N/A</p>
<p>B. Description: The LM-6000 SPRINT gas turbine and the 924 BHP Waukesha engine shall be fired on PUC regulated natural gas.</p>	<p>E. Source test reference method, if applicable. Attach Source Test Summary Form, if applicable N/A</p>
<p>C. Method of monitoring: Southern California Gas Company supplies only PUC quality natural gas to McGrath Generating Station.</p>	<p>F. Currently in Compliance? (Y or N): <u>Y</u> G. Compliance Status? (C or I): <u>C</u> H. *Excursions, exceedances, or other non-compliance? (Y or N): <u>N</u> <small>*If yes, attach Deviation Summary Form</small></p>



ANNUAL COMPLIANCE CERTIFICATION PERMIT ATTACHMENT FORM

Period Covered by Compliance Certification: 01 / 01 / 23 (MM/DD/YY) to 12 / 31 / 23 (MM/DD/YY)

<p>A. Attachment # or Permit Condition #: PO07891PC1-111 Condition 5</p>	<p>D. Frequency of monitoring:</p> <p style="text-align: center; font-size: 1.2em;">N/A</p>
<p>B. Description:</p> <p>The 10,500 gallon ammonia storage tank shall be equipped with a pressure vacuum relief valve set at 50 psig and shall be vented to the vessel from which it is being filled during all filling operations.</p>	<p>E. Source test reference method, if applicable. Attach Source Test Summary Form, if applicable</p> <p style="text-align: center; font-size: 1.2em;">N/A</p>
<p>C. Method of monitoring:</p> <p>-Visual observation during filling to verify the tank is vented to the filling vessel. -Attachment #8 is the pressure vacuum relief valve calibration sheet.</p>	<p>F. Currently in Compliance? (Y or N): <u>Y</u></p> <p>G. Compliance Status? (C or I): <u>C</u></p> <p>H. *Excursions, exceedances, or other non-compliance? (Y or N): <u>N</u></p> <p><small>*If yes, attach Deviation Summary Form</small></p>

<p>A. Attachment # or Permit Condition #: PO07891-111 Condition 6</p>	<p>D. Frequency of monitoring:</p> <p style="text-align: center; font-size: 1.2em;">N/A</p>
<p>B. Description:</p> <p>Exempted solvents, coatings, adhesives, and lubricants.</p>	<p>E. Source test reference method, if applicable. Attach Source Test Summary Form, if applicable</p> <p style="text-align: center; font-size: 1.2em;">N/A</p>
<p>C. Method of monitoring:</p> <p>A list of all solvents and coatings used at the facility is maintained. Attachment #9 is the coating and solvent usage record for 2023.</p>	<p>F. Currently in Compliance? (Y or N): <u>Y</u></p> <p>G. Compliance Status? (C or I): <u>C</u></p> <p>H. *Excursions, exceedances, or other non-compliance? (Y or N): <u>N</u></p> <p><small>*If yes, attach Deviation Summary Form</small></p>

<p>A. Attachment # or Permit Condition #: VCAPCD Rule 50</p>	<p>D. Frequency of monitoring:</p> <p>Routine surveillance, annual certification</p>
<p>B. Description:</p> <p>Rule 50 - Opacity: No visible emissions for a period or periods greater than 3 minutes in any one hour which are as dark or darker in shade as No. 1 on the Ringelmann Chart, or equivalent to 20% opacity or greater.</p>	<p>E. Source test reference method, if applicable. Attach Source Test Summary Form, if applicable</p> <p style="text-align: center; font-size: 1.2em;">EPA Method 9</p>
<p>C. Method of monitoring:</p> <p>Annual certification indicates the emissions units at the facility comply with the applicable sections of Rule 50. Attachment #10 is a copy of the 2023 opacity survey for the gas turbine and Black-start Generator performed on 8/22/23.</p>	<p>F. Currently in Compliance? (Y or N): <u>Y</u></p> <p>G. Compliance Status? (C or I): <u>C</u></p> <p>H. *Excursions, exceedances, or other non-compliance? (Y or N): <u>N</u></p> <p><small>*If yes, attach Deviation Summary Form</small></p>



ANNUAL COMPLIANCE CERTIFICATION PERMIT ATTACHMENT FORM

Period Covered by Compliance Certification: 01 / 01 / 23 (MM/DD/YY) to 12 / 31 / 23 (MM/DD/YY)

<p>A. Attachment # or Permit Condition #: VCAPCD Rules 54.B.1 & 54.B.2</p>	<p>D. Frequency of monitoring: Upon request</p>
<p>B. Description: -Stationary IC engine & gas turbine operators shall not discharge sulfur compounds in excess of 300 ppm by vol (SO₂) at 15% O₂. -Sulfur concentration at ground level or at any point at or beyond property line shall not exceed 0.25 ppmv 1-hr average or 0.04 ppmv 24-hr average.</p>	<p>E. Source test reference method, if applicable. Attach Source Test Summary Form, if applicable EPA Method 6, 6A, 6C, 8, 15, 16A, 16B, or SCAQMD method 307-91</p>
<p>C. Method of monitoring: -Only PUC quality gas is combusted at the facility in compliance with Rule 64 and by extension Rule 54.B.1; -Sulfur concentration at point of discharge and ground level concentration of SO₂ shall be monitored upon District's request.</p>	<p>F. Currently in Compliance? (Y or N): <u>Y</u> G. Compliance Status? (C or I): <u>C</u> H. *Excursions, exceedances, or other non-compliance? (Y or N): <u>N</u> <i>*If yes, attach Deviation Summary Form</i></p>

<p>A. Attachment # or Permit Condition #: VCAPCD Rule 55 - Fugitive Dust</p>	<p>D. Frequency of monitoring: Routine Surveillance</p>
<p>B. Description: Fugitive dust emissions resulting from any operation, disturbed surface area or man made conditions shall not be visible beyond the midpoint of an adjacent street. Opacity shall be less than 20 percent and track out shall be less than 25 feet.</p>	<p>E. Source test reference method, if applicable. Attach Source Test Summary Form, if applicable N/A</p>
<p>C. Method of monitoring: On site operations monitoring: All applicable sources of dust at the facility are operating in compliance with Rule 55. Facility did not import gravel in 2023 for landscaping purposes.</p>	<p>F. Currently in Compliance? (Y or N): <u>Y</u> G. Compliance Status? (C or I): <u>C</u> H. *Excursions, exceedances, or other non-compliance? (Y or N): <u>N</u> <i>*If yes, attach Deviation Summary Form</i></p>

<p>A. Attachment # or Permit Condition #: VCAPCD Rule 57.1</p>	<p>D. Frequency of monitoring: Upon request</p>
<p>B. Description: Particulate Matter emission from fuel burning equipment shall not exceed 0.12 lbs. per million BTU.</p>	<p>E. Source test reference method, if applicable. Attach Source Test Summary Form, if applicable CARB Method 5</p>
<p>C. Method of monitoring: -Per District Analysis dated Dec. 3rd, 1997, Gas Turbine emission factor was determined to be 0.0419 lb./MMBTU, less than 0.12 lbs./MMBTU limit specified in Rule 57.1 -PM source test will be conducted upon request by the District.</p>	<p>F. Currently in Compliance? (Y or N): <u>Y</u> G. Compliance Status? (C or I): <u>C</u> H. *Excursions, exceedances, or other non-compliance? (Y or N): <u>N</u> <i>*If yes, attach Deviation Summary Form</i></p>



**ANNUAL COMPLIANCE CERTIFICATION
PERMIT ATTACHMENT FORM**

Period Covered by Compliance Certification: 01 / 01 / 23 (MM/DD/YY) to 12 / 31 / 23 (MM/DD/YY)

A. Attachment # or Permit Condition #: VCAPCD Rule 64.B.1	D. Frequency of monitoring:
B. Description: Sulfur content of fuels: No person shall burn gaseous fuel containing sulfur compounds in excess of 50 grains per 100 cubic feet of fuel (788 ppmv).	None for PUC quality gas
C. Method of monitoring: Southern California Gas Company supplies only PUC quality natural gas to McGrath Peaker. No additional monitoring required.	E. Source test reference method, if applicable. Attach Source Test Summary Form, if applicable N/A F. Currently in Compliance? (Y or N): <u>Y</u> G. Compliance Status? (C or I): <u>C</u> H. *Excursions, exceedances, or other non-compliance? (Y or N): <u>N</u> *If yes, attach Deviation Summary Form

A. Attachment # or Permit Condition #: VCAPCD Rule 74.6	D. Frequency of monitoring:
B. Description: Surface Cleaning and Degreasing: comply with VOC content limits and maintain records of usage. The exemption for aerosol usage is less than 160 fl oz/day.	N/A
C. Method of monitoring: An annual log is used to document surface cleaning and degreasing activities. Attachment #9 is the coating, solvent, adhesive, and aerosol usage log. All aerosol use was < 160 fl oz/day in 2023. Isopropyl alcohol was used for fiber optic cleaning and exempt.	E. Source test reference method, if applicable. Attach Source Test Summary Form, if applicable N/A F. Currently in Compliance? (Y or N): <u>Y</u> G. Compliance Status? (C or I): <u>C</u> H. *Excursions, exceedances, or other non-compliance? (Y or N): <u>N</u> *If yes, attach Deviation Summary Form

A. Attachment # or Permit Condition #: VCAPCD Rule 74.11.1	D. Frequency of monitoring:
B. Description: Large Water Heaters and Small Boilers: comply with NOx emission limits for subject equipment. Maintain list of equipment.	N/A
C. Method of monitoring: There are no large water heaters or small boilers installed at the facility.	E. Source test reference method, if applicable. Attach Source Test Summary Form, if applicable N/A F. Currently in Compliance? (Y or N): <u>Y</u> G. Compliance Status? (C or I): <u>C</u> H. *Excursions, exceedances, or other non-compliance? (Y or N): <u>N</u> *If yes, attach Deviation Summary Form



ANNUAL COMPLIANCE CERTIFICATION PERMIT ATTACHMENT FORM

Period Covered by Compliance Certification: 01 / 01 / 23 (MM/DD/YY) to 12 / 31 / 23 (MM/DD/YY)

<p>A. Attachment # or Permit Condition #: VCAPCD Rule 74.22</p>	<p>D. Frequency of monitoring:</p> <p style="font-size: 1.2em; text-align: center;">N/A</p>
<p>B. Description: Natural Gas Fired Fan-Type Central Furnaces: comply with NOx emission limits for subject equipment. Maintain list of equipment.</p>	<p>E. Source test reference method, if applicable. Attach Source Test Summary Form, if applicable</p> <p style="font-size: 1.2em; text-align: center;">N/A</p>
<p>C. Method of monitoring: There are no natural gas fired fan-type central furnaces installed at the facility.</p>	<p>F. Currently in Compliance? (Y or N): <u>Y</u></p> <p>G. Compliance Status? (C or I): <u>C</u></p> <p>H. *Excursions, exceedances, or other non-compliance? (Y or N): <u>N</u></p> <p><small>*If yes, attach Deviation Summary Form</small></p>

<p>A. Attachment # or Permit Condition #: VCAPCD Rule 74.1</p>	<p>D. Frequency of monitoring:</p> <p style="font-size: 1.2em; text-align: center;">N/A</p>
<p>B. Description: Abrasive Blasting: Comply with visible emissions standards and methods for abrasive blasting operations. Maintain records.</p>	<p>E. Source test reference method, if applicable. Attach Source Test Summary Form, if applicable</p> <p style="font-size: 1.2em; text-align: center;">N/A</p>
<p>C. Method of monitoring: No abrasive blasting operation was performed at McGrath Peaker in 2023.</p>	<p>F. Currently in Compliance? (Y or N): <u>Y</u></p> <p>G. Compliance Status? (C or I): <u>C</u></p> <p>H. *Excursions, exceedances, or other non-compliance? (Y or N): <u>N</u></p> <p><small>*If yes, attach Deviation Summary Form</small></p>

<p>A. Attachment # or Permit Condition #: VCAPCD Rule 74.2</p>	<p>D. Frequency of monitoring:</p> <p style="font-size: 1.2em; text-align: center;">N/A</p>
<p>B. Description: Architectural Coatings: Comply with VOC content limits and maintain records for architectural coating use.</p>	<p>E. Source test reference method, if applicable. Attach Source Test Summary Form, if applicable</p> <p style="font-size: 1.2em; text-align: center;">N/A</p>
<p>C. Method of monitoring: Coatings used at the facility in 2023 were in compliance with the VOC content limits in Rule 74.2 or exempt (small container). Attachment #9 is the coating and solvent usage log for 2023.</p>	<p>F. Currently in Compliance? (Y or N): <u>Y</u></p> <p>G. Compliance Status? (C or I): <u>C</u></p> <p>H. *Excursions, exceedances, or other non-compliance? (Y or N): <u>N</u></p> <p><small>*If yes, attach Deviation Summary Form</small></p>



ANNUAL COMPLIANCE CERTIFICATION PERMIT ATTACHMENT FORM

Period Covered by Compliance Certification: 01 / 01 / 23 (MM/DD/YY) to 12 / 31 / 23 (MM/DD/YY)

<p>A. Attachment # or Permit Condition #: VCAPCD Rule 74.4.D</p>	<p>D. Frequency of monitoring:</p> <p style="font-size: 1.2em; text-align: center;">N/A</p>
<p>B. Description:</p> <p>Cutback Asphalt: Comply with organic compound limit (0.5%) for road oils applied for street paving or maintenance.</p>	<p>E. Source test reference method, if applicable. Attach Source Test Summary Form, if applicable</p> <p style="font-size: 1.2em; text-align: center;">N/A</p>
<p>C. Method of monitoring:</p> <p>Cut back asphalt activities were not performed at the facility in 2023.</p>	<p>F. Currently in Compliance? (Y or N): <u>Y</u></p> <p>G. Compliance Status? (C or I): <u>C</u></p> <p>H. *Excursions, exceedances, or other non-compliance? (Y or N): <u>N</u></p> <p><small>*If yes, attach Deviation Summary Form</small></p>

<p>A. Attachment # or Permit Condition #: 40 CFR Part 61, Subpart M</p>	<p>D. Frequency of monitoring:</p> <p style="font-size: 1.2em; text-align: center;">N/A</p>
<p>B. Description:</p> <p>National Emission Standards for Asbestos: Comply with applicable requirements for demolition/renovation activities.</p>	<p>E. Source test reference method, if applicable. Attach Source Test Summary Form, if applicable</p> <p style="font-size: 1.2em; text-align: center;">N/A</p>
<p>C. Method of monitoring:</p> <p>Asbestos demolition/renovation activities were not performed at the facility in 2023.</p>	<p>F. Currently in Compliance? (Y or N): <u>Y</u></p> <p>G. Compliance Status? (C or I): <u>C</u></p> <p>H. *Excursions, exceedances, or other non-compliance? (Y or N): <u>N</u></p> <p><small>*If yes, attach Deviation Summary Form</small></p>

<p>A. Attachment # or Permit Condition #:</p>	<p>D. Frequency of monitoring:</p>
<p>B. Description:</p>	<p>E. Source test reference method, if applicable. Attach Source Test Summary Form, if applicable</p>
<p>C. Method of monitoring:</p>	<p>F. Currently in Compliance? (Y or N): _____</p> <p>G. Compliance Status? (C or I): _____</p> <p>H. *Excursions, exceedances, or other non-compliance? (Y or N): _____</p> <p><small>*If yes, attach Deviation Summary Form</small></p>



Ventura County
Air Pollution
Control District

ANNUAL COMPLIANCE CERTIFICATION SOURCE TEST SUMMARY FORM

Period Covered by Compliance Certification: 01 / 01 / 23 (MM/DD/YY) to 12 / 31 / 23 (MM/DD/YY)

A. Emission Unit Description: GE LM-6000 PC SPRINT Gas Turbine			B. Pollutant: NOx
C. Measured Emission Rate: 2.04 ppm @ 15% O2 3.05 lbs/hr	D. Limited Emission Rate: 2.5 ppm @ 15% O2 50 lbs/hr	E. Specific Source Test or Monitoring Record Citation: 07891-T1-161, Condition 1a	F. Test Date: 08/22/23

A. Emission Unit Description: GE LM-6000 PC SPRINT Gas Turbine			B. Pollutant: ROC
C. Measured Emission Rate: <0.96 ppm @ 15% O2 <0.50 lbs/hr	D. Limited Emission Rate: 2.0 ppm @ 15% O2 1.38 lbs/hr	E. Specific Source Test or Monitoring Record Citation: 07891-T1-161, Condition 1b	F. Test Date: 08/22/23

A. Emission Unit Description: GE LM-6000 PC SPRINT Gas Turbine			B. Pollutant: CO
C. Measured Emission Rate: 1.14 ppm @ 15% O2 1.04 lbs/hr	D. Limited Emission Rate: 6.0 ppm @ 15% O2 15.80 lbs/hr	E. Specific Source Test or Monitoring Record Citation: 07891-T1-161, Condition 1c	F. Test Date: 08/22/23

A. Emission Unit Description: GE LM-6000 PC SPRINT Gas Turbine			B. Pollutant: NH3
C. Measured Emission Rate: 0.81 ppm @15% O2 0.45 lbs/hr	D. Limited Emission Rate: 5.0 ppm @ 15% O2 3.44 lbs/hr	E. Specific Source Test or Monitoring Record Citation: 07891-T1-161, Condition 1d	F. Test Date: 08/22/23

A. Emission Unit Description:			B. Pollutant:
C. Measured Emission Rate:	D. Limited Emission Rate:	E. Specific Source Test or Monitoring Record Citation:	F. Test Date:

Attachment 1

Emissions and Natural Gas Fuel Records

SCE McGrath Peaker
Oxnard CA
McGrath- Multi Purpose Report
2023

Month	Unit On-Time	Number of Starts	Water Injection klb	NOx lbs	CO lbs	PM lbs	SOx lbs	ROC lbs	Gas Flow mmscf	Gross Megawatt Hours	Net Megawatt Hours
Jan 2023	5.61	4	88.1	45	10	24	1	1	2.27	240.4	235.5
Feb 2023	6.72	4	116.3	47	12	29	1	3	2.76	295.9	290.0
Mar 2023	20.99	4	369.5	93	17	95	5	8	9.12	984.8	945.2
Apr 2023	11.51	5	196.7	71	18	51	3	4	4.93	518.4	509.0
May 2023	22.89	7	340.0	76	29	87	4	9	8.44	839.7	808.7
Jun 2023	10.48	3	170.8	42	15	45	3	3	4.28	455.4	447.3
Jul 2023	42.15	12	623.2	138	69	169	10	17	16.26	1708.3	1668.6
Aug 2023	48.47	11	740.5	150	82	196	12	20	18.79	1976.4	1931.3
Sep 2023	33.41	11	512.4	120	65	135	11	11	13.07	1378.5	1346.9
Oct 2023	32.12	10	497.4	106	55	134	8	15	12.80	1362.5	1330.4
Nov 2023	24.42	9	381.2	103	38	101	8	9	9.79	1053.4	1030.2
Dec 2023	34.42	8	619.7	110	51	150	7	14	14.39	1560.0	1527.8
Total	293.19	88	4655.8	1101	461	1216	73	114	116.90	12373.7	12070.9

SCE McGrath Peaker

Oxnard CA

McGrath- 12-Month Rolling Mass Emissions Report

December 2023 12-Month Rolling

12-Month Rolling Emission Limits

NOx tons - 4.81
PM tons - 8.64

CO tons - 11.79
NH3 Slip tons - 5.96

SOx tons - 0.5
ROC tons - 2.24

Month	NOx lbs	SOx lbs	CO lbs	PM lbs	NH3 Slip lbs (Bias Adjusted)	ROC lbs
Jan 2023	45	1	10	24	10	1
Feb 2023	47	1	12	29	15	3
Mar 2023	93	5	17	95	28	8
Apr 2023	71	3	18	51	23	4
May 2023	76	4	29	87	49	9
Jun 2023	42	3	15	45	7	3
Jul 2023	138	10	69	169	42	17
Aug 2023	150	12	82	196	45	20
Sep 2023	120	11	65	135	37	11
Oct 2023	106	8	55	134	39	15
Nov 2023	103	8	38	101	21	9
Dec 2023	110	7	51	150	30	14
12-Mo Roll	0.55 tons	0.0 tons	0.23 tons	0.61 tons	0.2 tons	0.06 tons

SOx Emission Factor - 0.6 lb/mmscf
PM Emission Factor - 10.37 lb/mmscf
ROC Emission Factor - 0.96 lb/mmscf

SCE McGrath Peaker

Oxnard CA

McGrath- 12-Month Rolling Fuel Usage & Start Ups Report

December 2023 12-Month Rolling

Month	Monthly Gas Flow mmscf	12-Month Rolling Gas Flow mmscf	Monthly # of Startups	12-Month Rolling # of Startups	Monthly Unit On-Time	12-Month Rolling Unit On-Time
Jan 2023	2.27	291.65	4	205	5.61	694.33
Feb 2023	2.76	276.66	4	193	6.72	659.13
Mar 2023	9.12	256.57	4	171	20.99	611.68
Apr 2023	4.93	212.30	5	144	11.51	508.16
May 2023	8.44	216.67	7	144	22.89	517.21
Jun 2023	4.28	188.77	3	123	10.48	451.44
Jul 2023	16.26	168.41	12	111	42.15	406.90
Aug 2023	18.79	137.86	11	96	48.47	339.61
Sep 2023	13.07	136.49	11	97	33.41	338.33
Oct 2023	12.80	127.64	10	94	32.12	315.59
Nov 2023	9.79	119.54	9	90	24.42	298.20
Dec 2023	14.39	116.90	8	88	34.42	293.19

Attachment 2
CEMS Maintenance Records

McGrath Peaker Quarterly Checks

Quality Control/Quality Assurance Plan
Checklist for CEMS Shelter Inspection

Quarterly QA/QC Inspections				
	1st Quarter	2nd Quarter	3rd Quarter	4th Quarter
Date:	3/29/2023	6/28/2023	9/28/2023	12/18/2023
Technician's Name:	White	White	White	White
Sample System Checks				
Perform Weekly and Monthly Inspections	√	√	√	√
Inspect Sample Pressure, if declining, perform one or more of the following:	√	√	√	√
Perform Probe maintenance	N/A	N/A	N/A	N/A
Clean and Inspect Probe Filter Element and Chamber. Replace Filter if needed	N/A	N/A	N/A	N/A
Verify if Probe Box Heater is operating (Amp Clamp)	√	√	√	√
Replace NH3 NOx Dessicant Media	N/A	N/A	N/A	N/A
If Sample Flow is low then leak check Sample Pump	N/A	N/A	N/A	N/A
Perform CEMS sample system leak check	N/A	N/A	N/A	N/A
Perform general housekeeping duties. Dust/clean all equipment surfaces.	√	√	√	√
Analyzer Checks				
Visually check for obvious defects such as loose connectors, loose fittings, cracked or clogged teflon lines, and excessive dust or dirt accumulation. Dirt accumulation can cause overheating or component failure and may provide a conducting path for electricity	√	√	√	√
Clean inside of each instrument by vacuuming	√	√	√	√
Clean all Analyzer cooling fans/filters	√	√	√	√
NOx Analyzer				
If there is excessive noise, drifting zero or span values, low response or a combination of all, the Reaction Cell may need to be cleaned.	N/A	N/A	N/A	N/A
NO Norm Offset	-20 to 150 mV	0.1	0.3	0.4
NO Stability	≤ 0.001 PPM at Zero Air for 10 minutes	√	√	√
NO2 Stability	≤ 0.001 PPM at Zero Air for 10 minutes	√	√	√
NOx Norm Offset	-20 to 150 mV	0.1	0.4	0.4
NOx Stability	≤ 0.001 PPM at Zero Air for 10 minutes	√	√	√
CO/O2 Analyzer				
Examine and clean pneumatic lines as needed.	√	√	√	√
O2 STB	< 1% during Zero Air after 10 minutes	√	√	√
Caution: Observe all safety warnings from manufacturers manual				
NOx/NH3 Analyzer				
If there is excessive noise, drifting zero or span values, low response or a combination of all, the Reaction Cell may need to be cleaned.	N/A	N/A	N/A	N/A
NO Norm Offset	-20 to 150 mV	0.0	0.2	0.2
NO Stability	≤ 0.001 PPM at Zero Air for 10 minutes	√	√	√
NO2 Stability	≤ 0.001 PPM at Zero Air for 10 minutes	√	√	√
NOx Norm Offset	-20 to 150 mV	0.0	0.4	0.5
NOx Stability	≤ 0.001 PPM at Zero Air for 10 minutes	√	√	√
REMARKS:				

Mark as either **Acceptable "√"**, **Corrective action Required "X"**, or **Actual Readings, where required**
Note: Report all deficiencies to the Control Operator immediately. Log corrective actions in the CEMS shelter log book.

Location:	JAN	FEB	MAR	APR	MAY	JUN	
Date:	1/30/2023	2/28/2023	3/29/2023	4/29/2023	5/1/2023	6/1/2023	
Technician's Name:	White	White	White	White	Mitchell	White	
Sample System Checks							
Perform Weekly Inspections	✓	✓	✓	✓	✓	✓	
Check NH3 NOx Dessicant Media. Replace as necessary	N/A	N/A	N/A	N/A	N/A	N/A	
Plan for the upcoming Linearity/CGA.	✓	✓	✓	✓	✓	✓	
Check gas bottles pressures >500 psig. Verify expiration dates.	✓	✓	✓	✓	✓	✓	
Order new bottles as needed keeping in mind the lead time may be several weeks.	✓	✓	✓	✓	✓	✓	
Check incoming instrument air filter. Open purge valve for at least 1 minute, then close.	✓	✓	✓	✓	✓	✓	
DAHS Checks							
If enabled, check/change backup media (removable hard drive, network location, etc.).	✓	✓	✓	✓	✓	✓	
Verify that automatic backups to the iFIX workstation have occurred for the month. Log on to the iFIX HMI D:\WindowsImageBackup\pkdahs1br and check for date and time to make sure the backups are current.	N/A	N/A	N/A	N/A	N/A	N/A	
Verify there is more than 10GB of free disk space available in the iFIX workstation.	N/A	N/A	N/A	N/A	N/A	N/A	
Diagnostic Checks: (Dashboard) Mark as either Acceptable/Checked "V", Corrective action Required "X", or Actual Readings, where required							
NOx Analyzer							
Auto Zero	-20 to 150 mV	-1.8	-1.7	-1.8	-1.8	-1.7	-1.6
Box Temp	32 ± 5 °C	30.4	29.8	30.0	30.2	31.9	29.9
Conv Temp	700 ± 15 °C	699.3	698.5	699.5	698.3	699.1	699.9
HVPS	400 to 900 V, nominal 500 V ± 50	447	447	447	447	447	447
NO Slope	1.00 ± 0.3 PPM/mV	1.149	1.200	1.210	1.210	1.210	1.161
NO Slope (Range 2)	1.00 ± 0.3 PPM/mV	1.104	1.139	1.138	1.138	1.138	1.064
NOx Slope	1.00 ± 0.3 PPM/mV	1.134	1.158	1.171	1.171	1.171	1.092
NOx Slope (Range 2)	1.00 ± 0.3 PPM/mV	1.110	1.133	1.135	1.135	1.135	1.067
O3 Flow (Ozone)	250 ± 25 cc/minute	245	246	244	246	243	255
PMT Signal	-20 to 150 mV at Zero Air	0.8	-1.6	-1.6	-1.6	-0.4	-1.5
	0 to 5,000 mV at Span Gas Concentration	2756.9	2714.8	2617.0	2577.1	2610.6	3062.9
PMT Temp	7.0 ± 2 °C	6.9	6.9	6.9	6.9	6.9	6.9
Rx Cell Pressure	less than 10 in. HgA or, Barometric pressure if the pump is off	5.5	5.6	5.7	5.9	5.0	5.0
Rx Cell Temp	50 ± 1 °C	50.0	50.0	50.0	50.0	50.0	50.0
Sample Flow	250 ± 25 cc/minute	255	256	255	255	254	254
Sample Pressure	Barometric Pressure ± 1 in. Hg-A	30.0	30.1	30.0	30.1	29.9	29.9
Required Check:	Check or Replace Particulate Filter	✓	✓	✓	✓	✓	✓
CO/O2 Analyzer							
Bench Temp	50 ± 5 °C	48.0	48.0	48.0	48.0	48.0	48.0
Box Temp	30 ± 5 °C	31.4	30.8	31.0	30.5	32.7	31.1
CO Offset 1	0 ± 0.5 during Zero Cal	-0.008	-0.008	-0.008	-0.008	-0.008	-0.030
CO Offset 2	0 ± 0.5 during Zero Cal	-0.008	-0.008	-0.008	-0.008	-0.008	-0.030
CO Slope 1	1.00 ± 0.3 during Span Cal	1.024	1.024	1.029	1.064	1.064	1.137
CO Slope 2	1.00 ± 0.3 during Span Cal	0.979	0.981	0.981	0.981	0.981	1.077
Meas Detector	4500 mV ± 300 mV during Zero Cal	2606.6	2466.5	4138.7	3902.4	3885.4	4263.2
MR Ratio	1.15 - 1.200 during Zero Cal	1.203	1.203	1.203	1.203	1.203	1.181
O2 Cell Temperature	50 ± 5 °C	50.0	50.0	50.0	50.0	50.0	50.0
O2 Offset	< 1%	0.594	0.594	0.594	0.594	0.594	0.337
O2 Slope	1.000 ± 0.3	1.040	1.040	1.040	1.040	1.040	1.071
PHT Drive	< 4800 mV during Sample	3855.3	3834.8	3885.1	3941.1	3961.6	1868.4
Ref Detector	< 4800 mV during Sample	2183.2	2068.4	3445.6	3264.9	3237.3	3682.7
Sample Flow	800 ± 50 cc/minute	855	853	852	858	836	790
Sample Pressure	29.0 ± 1 in. Hg-A during Sample	29.0	29.1	29.0	29.1	28.9	29.8
Sample Temp	50 ± 5 °C	48.0	48.1	48.0	47.8	47.7	48.2
Wheel Temp	70 ± 5 °C	67.9	68.0	68.3	68.3	67.7	68.1
Required Check:	Check or Replace Particulate Filter	✓	✓	✓	✓	✓	✓
NOx/NH3 Analyzer							
Auto Zero	-20 to 150 mV	-0.7	-0.7	-0.8	-0.8	-0.6	-0.5
Box Temp	32 ± 5 °C	30.2	30.1	30.4	30.0	32.1	29.3
Conv Temp	700 ± 15 °C	699.2	698.6	699.8	699.7	697.5	699.1
HVPS	400 to 900 V, nominal 500 V ± 50	429	429	429	429	429	429
NO Slope	1.00 ± 0.3 PPM/mV	1.349	1.432	1.493	1.652	1.652	1.202
NO Slope (Range 2)	1.00 ± 0.3 PPM/mV	1.343	1.405	1.495	1.596	1.596	1.179
NOx Slope	1.00 ± 0.3 PPM/mV	1.381	1.469	1.533	1.682	1.682	1.232
NOx Slope (Range 2)	1.00 ± 0.3 PPM/mV	1.352	1.422	1.511	1.612	1.612	1.190
O3 Flow (Ozone)	250 ± 25 cc/minute	252	253	252	253	251	255
PMT Signal	-20 to 150 mV at Zero Air	0.1	-0.6	-0.7	-0.6	0.3	-0.5
	0 to 5,000 mV at Span Gas Concentration	2254.9	2131.6	1957.2	1810.2	1846.6	2622.0
PMT Temp	7.0 ± 2 °C	6.6	6.6	6.6	6.6	6.6	6.6
Rx Cell Pressure	less than 10 in. Hg-A or, Barometric pressure if the pump is off	5.6	5.7	5.8	5.9	5.7	5.3
Rx Cell Temp	50 ± 1 °C	50.0	50.0	50.0	50.0	50.0	50.0
Sample Flow	250 ± 25 cc/minute	263	264	263	264	262	263
Sample Pressure	Barometric Pressure ± 1 in. Hg-A	30.0	30.1	30.0	30.1	29.9	30.0
Required Check:	Check or Replace Particulate Filter	✓	✓	✓	✓	✓	✓
REMARKS:	Note: All deficiencies must be reported to the Control Operator immediately. Corrective Action performed shall be logged in the CEMS shelter log book.						

Location:	JUL	AUG	SEP	OCT	NOV	DEC	
Date:	7/3/2023	8/1/2023	9/5/2023	10/30/2023	11/29/2023	12/18/2023	
Technician's Name:	White	White	White	White	White	White	
Sample System Checks							
Check NOx analyzer desiccant media. Replace as necessary	N/A	N/A	N/A	N/A	N/A	N/A	
Plan for the upcoming Linearity/CGA. Check gas bottles pressures >500 psig. Verify expiration dates. Order new bottles as needed keeping in mind the lead time may be several weeks.	✓	✓	✓	✓	✓	✓	
Check incoming instrument air filter	✓	✓	✓	✓	✓	✓	
DAHS Checks							
If enabled, check/change backup media (removable hard drive, network location, etc.).	✓	✓	✓	✓	✓	✓	
Verify that automatic backups to the iFIX workstation have occurred for the month. Log on to the iFIX HMI D:\WindowsImageBackup\pkdahs1br and check for date and time to make sure the backups are current.	N/A	N/A	N/A	N/A	N/A	N/A	
Verify there is more than 10GB of free disk space available in the iFIX workstation.	N/A	N/A	N/A	N/A	N/A	N/A	
Diagnostic Checks: (Dashboard) Mark as either Acceptable/Checked "V", Corrective action Required "X", or Actual Readings, where required							
NOx Analyzer							
Auto Zero	-20 to 150 mV	-1.6	-1.6	-1.5	-1.5	-1.4	-1.5
Box Temp	32 ± 5 °C	29.3	29.8	27.6	28.2	27.8	27.2
Conv Temp	700 ± 15 °C	697.3	699.5	700.8	700.1	699.7	701.1
HVPS	400 to 900 V, nominal 500 V ± 50	447	447	447	447	447	447
NO Slope	1.00 ± 0.3 PPM/mV	1.161	1.161	1.098	1.098	1.098	1.098
NO Slope (Range 2)	1.00 ± 0.3 PPM/mV	1.060	1.060	1.047	1.047	1.039	1.049
NOx Slope	1.00 ± 0.3 PPM/mV	1.092	1.092	1.068	1.068	1.068	1.068
NOx Slope (Range 2)	1.00 ± 0.3 PPM/mV	1.061	1.061	1.042	1.042	1.028	1.035
O3 Flow (Ozone)	250 ± 25 cc/minute	255	255	254	257	256	257
PMT Signal	-20 to 150 mV at Zero Air	-1.5	-1.3	-1.4	-1.5	-1.3	-0.4
	0 to 5,000 mV at Span Gas Concentration	3043.0	3015.0	3287.3	3245.4	3273.0	3205.7
PMT Temp	7.0 ± 2 °C	6.9	6.9	6.9	6.9	6.9	6.9
Rx Cell Pressure	less than 10 in. Hg-A or, Barometric pressure if the pump is off	5.1	4.8	4.7	4.7	4.7	4.7
Rx Cell Temp	50 ± 1 °C	50.0	50.0	50.0	50.0	50.0	50.0
Sample Flow	250 ± 25 cc/minute	254	255	254	255	254	255
Sample Pressure	Barometric Pressure ± 1 in. Hg-A	29.9	30.0	29.9	30.0	29.9	30.0
Required Check:	Check or Replace Particulate Filter	✓	✓	✓	✓	✓	✓
CO/O2 Analyzer							
Bench Temp	50 ± 5 °C	48.0	48.0	48.0	48.0	48.0	48.0
Box Temp	30 ± 5 °C	30.3	30.8	30.2	30.8	30.2	30.1
CO Offset 1	0 ± 0.5 during Zero Cal	-0.030	-0.030	-0.030	-0.030	-0.030	-0.030
CO Offset 2	0 ± 0.5 during Zero Cal	-0.030	-0.030	-0.030	-0.029	-0.029	-0.029
CO Slope 1	1.00 ± 0.3 during Span Cal	1.117	1.141	1.138	1.127	1.160	1.134
CO Slope 2	1.00 ± 0.3 during Span Cal	1.077	1.077	1.077	1.075	1.070	1.070
Meas Detector	4500 mV ± 300 mV during Zero Cal	3762.0	2918.5	2962.6	2714.0	2322.5	4492.4
MR Ratio	1.15 - 1.200 during Zero Cal	1.181	1.181	1.168	1.181	1.180	1.180
O2 Cell Temperature	50 ± 5 °C	50.0	50.0	50.0	50.0	50.0	50.0
O2 Offset	< 1%	0.337	0.337	0.337	0.447	0.447	0.447
O2 Slope	1.000 ± 0.3	1.071	1.071	1.071	1.073	1.073	1.073
PHT Drive	< 4800 mV during Sample	1870.9	1867.6	1845.4	1847.1	1856.2	1856.2
Ref Detector	< 4800 mV during Sample	3198.4	2404.2	2566.3	2316.4	1985.1	3824.1
Sample Flow	800 ± 50 cc/minute	790	798	792	807	791	796
Sample Pressure	29.0 ± 1 in. Hg-A during Sample	29.8	29.9	29.7	29.9	29.7	29.8
Sample Temp	50 ± 5 °C	48.4	48.3	48.5	48.4	48.3	48.4
Wheel Temp	70 ± 5 °C	68.0	68.1	68.2	67.9	68.0	68.1
Required Check:	Check or Replace Particulate Filter	✓	✓	✓	✓	✓	✓
NOx/NH3 Analyzer							
Auto Zero	-20 to 150 mV	-0.5	-0.5	-0.5	-0.4	-0.4	-0.5
Box Temp	32 ± 5 °C	29.1	29.2	28.6	28.9	28.7	29.0
Conv Temp	700 ± 15 °C	700.3	700.4	699.6	698.8	700.5	699.6
HVPS	400 to 900 V, nominal 500 V ± 50	429	429	429	429	429	429
NO Slope	1.00 ± 0.3 PPM/mV	1.282	1.323	1.365	1.425	1.425	1.425
NO Slope (Range 2)	1.00 ± 0.3 PPM/mV	1.271	1.295	1.325	1.375	1.408	1.440
NOx Slope	1.00 ± 0.3 PPM/mV	1.304	1.358	1.403	1.469	1.467	1.467
NOx Slope (Range 2)	1.00 ± 0.3 PPM/mV	1.287	1.310	1.339	1.385	1.414	1.446
O3 Flow (Ozone)	250 ± 25 cc/minute	255	255	254	256	255	255
PMT Signal	-20 to 150 mV at Zero Air	-0.4	-0.5	-0.2	-0.4	-0.4	-0.1
	0 to 5,000 mV at Span Gas Concentration	2430.6	2384.2	2351.2	228.2	2238.1	2212.6
PMT Temp	7.0 ± 2 °C	6.6	6.6	6.6	6.6	6.6	6.6
Rx Cell Pressure	less than 10 in. Hg-A or, Barometric pressure if the pump is off	5.3	5.3	5.2	5.3	5.4	5.4
Rx Cell Temp	50 ± 1 °C	50.0	50.0	50.0	50.0	50.0	50.0
Sample Flow	250 ± 25 cc/minute	263	264	263	264	263	264
Sample Pressure	Barometric Pressure ± 1 in. Hg-A	30.0	30.1	30.0	30.1	30.0	30.1
Required Check:	Check or Replace Particulate Filter	✓	✓	✓	✓	✓	✓
REMARKS:							

Note: All deficiencies must be reported to the Control Operator immediately. Corrective Action performed shall be logged in the CEMS shelter log book.

Location: McGrath	Tag ID	Limits	Date
Technician's Name: Jason White			1/5/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares			
O2/CO High Span, NOX Zero	SV1	>150 PSI	1660
NOX Low Span/CO Low Span	SV2	>150 PSI	1630
NOX High Span, O2/CO Zero	SV3	>150 PSI	1620
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Sample Line Temperature	TC1	250°F	250
Sample NH3 Temperature	TC2	760°C	760
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	5.9
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	5.6
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	5.4
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	7.1
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	73.0
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X
Moisture Sensor B/Filter	MS-2		X
Operational Status of Sample Pumps (2)		Check if ok	X
Operational Status of Condensate Drain Pump		Check if ok	X
Check LED Status of Sample Cooler		Check if ok	X
NH3 Scrubber Drain	HV-4	Drain weekly	X
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
System Flow	FM-1	3-5 LPM	3.2
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.75
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.55
CO Analyzer	FM-5	1.2 - 1.7 LPM	1.95
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A
System Flow	FM-7	3-5 LPM	4.2
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.55
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.35
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Check DAHS for normal operation. Is system logging data?		Check if ok	X
Verify no Alarms in DAHS		Check if ok	X
Check printer status		Check if ok	X
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X
No yellow "WARNING" status indicated?		Check if ok	X
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	N/A
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.			

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath	Tag ID	Limits	Date
Technician's Name: Jason White			1/12/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares			
O2/CO High Span, NOX Zero	SV1	>150 PSI	1520
NOX Low Span/CO Low Span	SV2	>150 PSI	1500
NOX High Span, O2/CO Zero	SV3	>150 PSI	1450
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Sample Line Temperature	TC1	250°F	250
Sample NH3 Temperature	TC2	760°C	760
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	5.5
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	5.8
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	5.4
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	7.1
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	72.0
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X
Moisture Sensor B/Filter	MS-2		X
Operational Status of Sample Pumps (2)		Check if ok	X
Operational Status of Condensate Drain Pump		Check if ok	X
Check LED Status of Sample Cooler		Check if ok	X
NH3 Scrubber Drain	HV-4	Drain weekly	X
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
System Flow	FM-1	3-5 LPM	3.2
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.75
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.55
CO Analyzer	FM-5	1.2 - 1.7 LPM	1.95
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A
System Flow	FM-7	3-5 LPM	4.2
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.55
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.35
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Check DAHS for normal operation. Is system logging data?		Check if ok	X
Verify no Alarms in DAHS		Check if ok	X
Check printer status		Check if ok	X
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X
No yellow "WARNING" status indicated?		Check if ok	X
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	X
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	X
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.			

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath		Tag ID	Limits	Date
Technician's Name: Jason White				1/19/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares				
O2/CO High Span, NOX Zero	SV1	>150 PSI	1170	
NOX Low Span/CO Low Span	SV2	>150 PSI	1190	
NOX High Span, O2/CO Zero	SV3	>150 PSI	1080	
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Sample Line Temperature	TC1	250°F	250	
Sample NH3 Temperature	TC2	760°C	760	
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	6.8	
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	5.7	
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	5.4	
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	7.1	
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X	
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X	
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	72.5	
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X	
Moisture Sensor B/Filter	MS-2		X	
Operational Status of Sample Pumps (2)		Check if ok	X	
Operational Status of Condensate Drain Pump		Check if ok	X	
Check LED Status of Sample Cooler		Check if ok	X	
NH3 Scrubber Drain	HV-4	Drain weekly	X	
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
System Flow	FM-1	3-5 LPM	3.2	
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.75	
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.60	
CO Analyzer	FM-5	1.2 - 1.7 LPM	1.95	
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A	
System Flow	FM-7	3-5 LPM	4.0	
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.55	
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.35	
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Check DAHS for normal operation. Is system logging data?		Check if ok	X	
Verify no Alarms in DAHS		Check if ok	X	
Check printer status		Check if ok	X	
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X	
No yellow "WARNING" status indicated?		Check if ok	X	
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	N/A	
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A	
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.				

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath		Tag ID	Limits	Date
Technician's Name: Jason White				1/26/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares				
O2/CO High Span, NOX Zero	SV1	>150 PSI	990	
NOX Low Span/CO Low Span	SV2	>150 PSI	990	
NOX High Span, O2/CO Zero	SV3	>150 PSI	840	
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Sample Line Temperature	TC1	250°F	250	
Sample NH3 Temperature	TC2	760°C	760	
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	5.5	
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	5.0	
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	5.4	
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	7.2	
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X	
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X	
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	70.5	
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X	
Moisture Sensor B/Filter	MS-2		X	
Operational Status of Sample Pumps (2)		Check if ok	X	
Operational Status of Condensate Drain Pump		Check if ok	X	
Check LED Status of Sample Cooler		Check if ok	X	
NH3 Scrubber Drain	HV-4	Drain weekly	X	
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
System Flow	FM-1	3-5 LPM	3.2	
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.75	
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.60	
CO Analyzer	FM-5	1.2 - 1.7 LPM	1.95	
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A	
System Flow	FM-7	3-5 LPM	4.2	
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.55	
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.35	
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Check DAHS for normal operation. Is system logging data?		Check if ok	X	
Verify no Alarms in DAHS		Check if ok	X	
Check printer status		Check if ok	X	
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X	
No yellow "WARNING" status indicated?		Check if ok	X	
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	N/A	
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A	
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.				

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath		Tag ID	Limits	Date
Technician's Name: Jason White				2/2/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares				
O2/CO High Span, NOX Zero	SV1	>150 PSI	1920	
NOX Low Span/CO Low Span	SV2	>150 PSI	1910	
NOX High Span, O2/CO Zero	SV3	>150 PSI	2020	
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Sample Line Temperature	TC1	250°F	250	
Sample NH3 Temperature	TC2	760°C	760	
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	5.4	
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	6.0	
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	5.0	
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	7.2	
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X	
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X	
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	72.0	
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X	
Moisture Sensor B/Filter	MS-2		X	
Operational Status of Sample Pumps (2)		Check if ok	X	
Operational Status of Condensate Drain Pump		Check if ok	X	
Check LED Status of Sample Cooler		Check if ok	X	
NH3 Scrubber Drain	HV-4	Drain weekly	X	
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
System Flow	FM-1	3-5 LPM	3.3	
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.80	
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.65	
CO Analyzer	FM-5	1.2 - 1.7 LPM	1.95	
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A	
System Flow	FM-7	3-5 LPM	4.3	
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.55	
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.35	
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Check DAHS for normal operation. Is system logging data?		Check if ok	X	
Verify no Alarms in DAHS		Check if ok	X	
Check printer status		Check if ok	X	
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X	
No yellow "WARNING" status indicated?		Check if ok	X	
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	X	
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	X	
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.				

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath	Tag ID	Limits	Date
Technician's Name: Jason White			2/9/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares			
O2/CO High Span, NOX Zero	SV1	>150 PSI	1680
NOX Low Span/CO Low Span	SV2	>150 PSI	1650
NOX High Span, O2/CO Zero	SV3	>150 PSI	1750
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Sample Line Temperature	TC1	250°F	250
Sample NH3 Temperature	TC2	760°C	760
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	6.5
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	6.0
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	6.3
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	7.1
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	72.5
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X
Moisture Sensor B/Filter	MS-2		X
Operational Status of Sample Pumps (2)		Check if ok	X
Operational Status of Condensate Drain Pump		Check if ok	X
Check LED Status of Sample Cooler		Check if ok	X
NH3 Scrubber Drain	HV-4	Drain weekly	X
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
System Flow	FM-1	3-5 LPM	3.3
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.80
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.65
CO Analyzer	FM-5	1.2 - 1.7 LPM	1.95
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A
System Flow	FM-7	3-5 LPM	4.2
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.55
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.35
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Check DAHS for normal operation. Is system logging data?		Check if ok	X
Verify no Alarms in DAHS		Check if ok	X
Check printer status		Check if ok	X
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X
No yellow "WARNING" status indicated?		Check if ok	X
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	N/A
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.			

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath	Tag ID	Limits	Date
Technician's Name: Jason White			2/16/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares			
O2/CO High Span, NOX Zero	SV1	>150 PSI	1500
NOX Low Span/CO Low Span	SV2	>150 PSI	1440
NOX High Span, O2/CO Zero	SV3	>150 PSI	1500
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Sample Line Temperature	TC1	250°F	250
Sample NH3 Temperature	TC2	760°C	760
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	5.5
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	5.9
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	5.1
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	7.2
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	72.5
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X
Moisture Sensor B/Filter	MS-2		X
Operational Status of Sample Pumps (2)		Check if ok	X
Operational Status of Condensate Drain Pump		Check if ok	X
Check LED Status of Sample Cooler		Check if ok	X
NH3 Scrubber Drain	HV-4	Drain weekly	X
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
System Flow	FM-1	3-5 LPM	3.2
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.80
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.65
CO Analyzer	FM-5	1.2 - 1.7 LPM	2.00
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A
System Flow	FM-7	3-5 LPM	4.1
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.55
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.35
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Check DAHS for normal operation. Is system logging data?		Check if ok	X
Verify no Alarms in DAHS		Check if ok	X
Check printer status		Check if ok	X
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X
No yellow "WARNING" status indicated?		Check if ok	X
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	N/A
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.			

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath	Tag ID	Limits	Date
Technician's Name: Jason White			2/26/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares			
O2/CO High Span, NOX Zero	SV1	>150 PSI	1100
NOX Low Span/CO Low Span	SV2	>150 PSI	1130
NOX High Span, O2/CO Zero	SV3	>150 PSI	1130
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Sample Line Temperature	TC1	250°F	250
Sample NH3 Temperature	TC2	760°C	760
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	5.6
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	5.8
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	5.1
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	7.1
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	73.0
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X
Moisture Sensor B/Filter	MS-2		X
Operational Status of Sample Pumps (2)		Check if ok	X
Operational Status of Condensate Drain Pump		Check if ok	X
Check LED Status of Sample Cooler		Check if ok	X
NH3 Scrubber Drain	HV-4	Drain weekly	X
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
System Flow	FM-1	3-5 LPM	3.3
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.80
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.75
CO Analyzer	FM-5	1.2 - 1.7 LPM	2.00
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A
System Flow	FM-7	3-5 LPM	4.1
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.55
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.45
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Check DAHS for normal operation. Is system logging data?		Check if ok	X
Verify no Alarms in DAHS		Check if ok	X
Check printer status		Check if ok	X
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X
No yellow "WARNING" status indicated?		Check if ok	X
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	N/A
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.			

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath	Tag ID	Limits	Date
Technician's Name: Jason White			3/2/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares			
O2/CO High Span, NOX Zero	SV1	>150 PSI	930
NOX Low Span/CO Low Span	SV2	>150 PSI	960
NOX High Span, O2/CO Zero	SV3	>150 PSI	940
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Sample Line Temperature	TC1	250°F	250
Sample NH3 Temperature	TC2	760°C	760
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	5.4
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	6.0
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	5.0
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	7.0
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	70.5
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X
Moisture Sensor B/Filter	MS-2		X
Operational Status of Sample Pumps (2)		Check if ok	X
Operational Status of Condensate Drain Pump		Check if ok	X
Check LED Status of Sample Cooler		Check if ok	X
NH3 Scrubber Drain	HV-4	Drain weekly	X
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
System Flow	FM-1	3-5 LPM	3.2
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.80
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.65
CO Analyzer	FM-5	1.2 - 1.7 LPM	2.00
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A
System Flow	FM-7	3-5 LPM	4.2
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.55
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.35
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Check DAHS for normal operation. Is system logging data?		Check if ok	X
Verify no Alarms in DAHS		Check if ok	X
Check printer status		Check if ok	X
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X
No yellow "WARNING" status indicated?		Check if ok	X
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	N/A
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.			

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath	Tag ID	Limits	Date
Technician's Name: Jason White			3/6/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares			
O2/CO High Span, NOX Zero	SV1	>150 PSI	830
NOX Low Span/CO Low Span	SV2	>150 PSI	810
NOX High Span, O2/CO Zero	SV3	>150 PSI	790
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Sample Line Temperature	TC1	250°F	250
Sample NH3 Temperature	TC2	760°C	760
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	5.4
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	5.8
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	5.4
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	7.0
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	71.0
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X
Moisture Sensor B/Filter	MS-2		X
Operational Status of Sample Pumps (2)		Check if ok	X
Operational Status of Condensate Drain Pump		Check if ok	X
Check LED Status of Sample Cooler		Check if ok	X
NH3 Scrubber Drain	HV-4	Drain weekly	X
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
System Flow	FM-1	3-5 LPM	3.2
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.80
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.65
CO Analyzer	FM-5	1.2 - 1.7 LPM	1.95
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A
System Flow	FM-7	3-5 LPM	4.3
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.55
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.35
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Check DAHS for normal operation. Is system logging data?		Check if ok	X
Verify no Alarms in DAHS		Check if ok	X
Check printer status		Check if ok	X
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X
No yellow "WARNING" status indicated?		Check if ok	X
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	N/A
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.			

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath	Tag ID	Limits	Date
Technician's Name: Jason White			3/16/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares			
O2/CO High Span, NOX Zero	SV1	>150 PSI	530
NOX Low Span/CO Low Span	SV2	>150 PSI	540
NOX High Span, O2/CO Zero	SV3	>150 PSI	480
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Sample Line Temperature	TC1	250°F	250
Sample NH3 Temperature	TC2	760°C	760
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	5.6
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	5.8
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	5.3
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	7.1
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	72.5
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X
Moisture Sensor B/Filter	MS-2		X
Operational Status of Sample Pumps (2)		Check if ok	X
Operational Status of Condensate Drain Pump		Check if ok	X
Check LED Status of Sample Cooler		Check if ok	X
NH3 Scrubber Drain	HV-4	Drain weekly	X
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
System Flow	FM-1	3-5 LPM	3.2
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.80
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.65
CO Analyzer	FM-5	1.2 - 1.7 LPM	2.00
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A
System Flow	FM-7	3-5 LPM	4.1
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.55
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.35
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Check DAHS for normal operation. Is system logging data?		Check if ok	X
Verify no Alarms in DAHS		Check if ok	X
Check printer status		Check if ok	X
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X
No yellow "WARNING" status indicated?		Check if ok	X
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	N/A
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.			

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath	Tag ID	Limits	Date
Technician's Name: Jason White			3/22/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares			
O2/CO High Span, NOX Zero	SV1	>150 PSI	1960
NOX Low Span/CO Low Span	SV2	>150 PSI	1900
NOX High Span, O2/CO Zero	SV3	>150 PSI	2110
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Sample Line Temperature	TC1	250°F	250
Sample NH3 Temperature	TC2	760°C	760
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	4.9
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	5.6
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	5.0
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	7.0
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	71.0
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X
Moisture Sensor B/Filter	MS-2		X
Operational Status of Sample Pumps (2)		Check if ok	X
Operational Status of Condensate Drain Pump		Check if ok	X
Check LED Status of Sample Cooler		Check if ok	X
NH3 Scrubber Drain	HV-4	Drain weekly	X
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
System Flow	FM-1	3-5 LPM	3.2
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.75
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.60
CO Analyzer	FM-5	1.2 - 1.7 LPM	2.00
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A
System Flow	FM-7	3-5 LPM	4.2
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.55
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.35
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Check DAHS for normal operation. Is system logging data?		Check if ok	X
Verify no Alarms in DAHS		Check if ok	X
Check printer status		Check if ok	X
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X
No yellow "WARNING" status indicated?		Check if ok	X
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	X
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	X
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.			

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath	Tag ID	Limits	Date
Technician's Name: Jason White			3/29/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares			
O2/CO High Span, NOX Zero	SV1	>150 PSI	1560
NOX Low Span/CO Low Span	SV2	>150 PSI	1620
NOX High Span, O2/CO Zero	SV3	>150 PSI	1650
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Sample Line Temperature	TC1	250°F	250
Sample NH3 Temperature	TC2	760°C	760
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	5.4
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	5.5
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	5.3
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	6.9
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	73.0
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X
Moisture Sensor B/Filter	MS-2		X
Operational Status of Sample Pumps (2)		Check if ok	X
Operational Status of Condensate Drain Pump		Check if ok	X
Check LED Status of Sample Cooler		Check if ok	X
NH3 Scrubber Drain	HV-4	Drain weekly	X
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
System Flow	FM-1	3-5 LPM	3.2
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.70
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.55
CO Analyzer	FM-5	1.2 - 1.7 LPM	2.00
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A
System Flow	FM-7	3-5 LPM	4.2
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.55
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.35
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Check DAHS for normal operation. Is system logging data?		Check if ok	X
Verify no Alarms in DAHS		Check if ok	X
Check printer status		Check if ok	X
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X
No yellow "WARNING" status indicated?		Check if ok	X
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	N/A
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.			

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath		Tag ID	Limits	Date
Technician's Name: Jason White				4/6/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares				
O2/CO High Span, NOX Zero	SV1	>150 PSI	1360	
NOX Low Span/CO Low Span	SV2	>150 PSI	1380	
NOX High Span, O2/CO Zero	SV3	>150 PSI	1330	
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Sample Line Temperature	TC1	250°F	250	
Sample NH3 Temperature	TC2	760°C	760	
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	5.3	
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	5.7	
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	5.2	
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	7.0	
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X	
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X	
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	73.0	
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X	
Moisture Sensor B/Filter	MS-2		X	
Operational Status of Sample Pumps (2)		Check if ok	X	
Operational Status of Condensate Drain Pump		Check if ok	X	
Check LED Status of Sample Cooler		Check if ok	X	
NH3 Scrubber Drain	HV-4	Drain weekly	X	
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
System Flow	FM-1	3-5 LPM	3.2	
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.70	
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.55	
CO Analyzer	FM-5	1.2 - 1.7 LPM	2.00	
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A	
System Flow	FM-7	3-5 LPM	4.2	
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.55	
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.35	
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Check DAHS for normal operation. Is system logging data?		Check if ok	X	
Verify no Alarms in DAHS		Check if ok	X	
Check printer status		Check if ok	X	
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X	
No yellow "WARNING" status indicated?		Check if ok	X	
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	X	
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A	
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.				

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath	Tag ID	Limits	Date
Technician's Name: Jason White			4/13/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares			
O2/CO High Span, NOX Zero	SV1	>150 PSI	1180
NOX Low Span/CO Low Span	SV2	>150 PSI	1160
NOX High Span, O2/CO Zero	SV3	>150 PSI	1090
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Sample Line Temperature	TC1	250°F	250
Sample NH3 Temperature	TC2	760°C	760
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	5.0
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	5.6
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	5.0
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	6.9
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	71.5
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X
Moisture Sensor B/Filter	MS-2		X
Operational Status of Sample Pumps (2)		Check if ok	X
Operational Status of Condensate Drain Pump		Check if ok	X
Check LED Status of Sample Cooler		Check if ok	X
NH3 Scrubber Drain	HV-4	Drain weekly	X
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
System Flow	FM-1	3-5 LPM	3.2
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.70
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.55
CO Analyzer	FM-5	1.2 - 1.7 LPM	1.95
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A
System Flow	FM-7	3-5 LPM	3.9
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.60
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.40
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Check DAHS for normal operation. Is system logging data?		Check if ok	X
Verify no Alarms in DAHS		Check if ok	X
Check printer status		Check if ok	X
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X
No yellow "WARNING" status indicated?		Check if ok	X
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	N/A
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.			

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath	Tag ID	Limits	Date
Technician's Name: Jason White			4/17/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares			
O2/CO High Span, NOX Zero	SV1	>150 PSI	950
NOX Low Span/CO Low Span	SV2	>150 PSI	1000
NOX High Span, O2/CO Zero	SV3	>150 PSI	900
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Sample Line Temperature	TC1	250°F	250
Sample NH3 Temperature	TC2	760°C	760
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	5.4
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	5.7
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	5.3
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	6.9
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	73.0
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X
Moisture Sensor B/Filter	MS-2		X
Operational Status of Sample Pumps (2)		Check if ok	X
Operational Status of Condensate Drain Pump		Check if ok	X
Check LED Status of Sample Cooler		Check if ok	X
NH3 Scrubber Drain	HV-4	Drain weekly	X
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
System Flow	FM-1	3-5 LPM	3.2
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.70
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.55
CO Analyzer	FM-5	1.2 - 1.7 LPM	2.00
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A
System Flow	FM-7	3-5 LPM	4.0
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.55
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.35
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Check DAHS for normal operation. Is system logging data?		Check if ok	X
Verify no Alarms in DAHS		Check if ok	X
Check printer status		Check if ok	X
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X
No yellow "WARNING" status indicated?		Check if ok	X
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	X
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.			

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath	Tag ID	Limits	Date
Technician's Name: Jason White			4/29/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares			
O2/CO High Span, NOX Zero	SV1	>150 PSI	1960
NOX Low Span/CO Low Span	SV2	>150 PSI	1880
NOX High Span, O2/CO Zero	SV3	>150 PSI	2080
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Sample Line Temperature	TC1	250°F	250
Sample NH3 Temperature	TC2	760°C	760
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	4.8
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	5.7
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	4.4
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	6.7
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	71.5
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X
Moisture Sensor B/Filter	MS-2		X
Operational Status of Sample Pumps (2)		Check if ok	X
Operational Status of Condensate Drain Pump		Check if ok	X
Check LED Status of Sample Cooler		Check if ok	X
NH3 Scrubber Drain	HV-4	Drain weekly	X
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
System Flow	FM-1	3-5 LPM	3.2
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.70
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.55
CO Analyzer	FM-5	1.2 - 1.7 LPM	1.95
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A
System Flow	FM-7	3-5 LPM	4.0
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.80
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.45
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Check DAHS for normal operation. Is system logging data?		Check if ok	X
Verify no Alarms in DAHS		Check if ok	X
Check printer status		Check if ok	X
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X
No yellow "WARNING" status indicated?		Check if ok	X
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	X
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	X
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.			

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

McGrath

Peaker Weekly Checks

Quality Control/Quality Assurance Plan
Checklist for CEMS Shelter Inspection

Location:	Tag ID	Limits	Date
Technician's Name: CM			5-1-23
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares			
O2/CO High Span, NOX Zero	SV1	>150 PSI	1900
NOX Low Span/CO Low Span	SV2	>150 PSI	1800
NOX High Span, O2/CO Zero	SV3	>150 PSI	2000
Stack Sample Line - Mark as either Acceptable "✓", Corrective action Required "X", or Actual Readings, where required			
Sample Line Temperature	TC1	250°F	250
Sample NH3 Temperature	TC2	760°C	760
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	5.6
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	5.5
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	5.1
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	6.5
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	✓
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	✓
Visual Checks - Mark as either Acceptable "✓", Corrective action Required "X", or Actual Readings, where required			
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	73
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	✓
Moisture Sensor B/Filter	MS-2		✓
Operational Status of Sample Pumps (2)		Check if ok	✓
Operational Status of Condensate Drain Pump		Check if ok	✓
Check LED Status of Sample Cooler		Check if ok	✓ 38.4
NH3 Scrubber Drain	HV-4	Drain weekly	✓
Sample Flow Meter Readings - Mark as either Acceptable "✓", Corrective action Required "X", or Actual Readings, where required			
System Flow	FM-1	3-5 LPM	3.1
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.6
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.5
CO Analyzer	FM-5	1.2 - 1.7 LPM	2
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	12
System Flow	FM-7	3-5 LPM	4.1
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.7
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.5
DAHS Checks - Mark as either Acceptable "✓", Corrective action Required "X", or Actual Readings, where required			
Check DAHS for normal operation. Is system logging data?		Check if ok	✓
Verify no Alarms in DAHS		Check if ok	✓
Check printer status		Check if ok	✓
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	✓
No yellow "WARNING" status indicated?		Check if ok	✓
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	✓
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	✓
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.			
Perform Monthly Inspection. Ran off hands Calibration All passed.			
Maint made 1404 return to service @ 1440.			

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath	Tag ID	Limits	Date
Technician's Name: Jason White			5/13/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares			
O2/CO High Span, NOX Zero	SV1	>150 PSI	1110
NOX Low Span/CO Low Span	SV2	>150 PSI	1000
NOX High Span, O2/CO Zero	SV3	>150 PSI	910
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Sample Line Temperature	TC1	250°F	250
Sample NH3 Temperature	TC2	760°C	760
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	5.5
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	6.4
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	5.2
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	6.3
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	70.5
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X
Moisture Sensor B/Filter	MS-2		X
Operational Status of Sample Pumps (2)		Check if ok	X
Operational Status of Condensate Drain Pump		Check if ok	X
Check LED Status of Sample Cooler		Check if ok	X
NH3 Scrubber Drain	HV-4	Drain weekly	X
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
System Flow	FM-1	3-5 LPM	3.5
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.85
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.65
CO Analyzer	FM-5	1.2 - 1.7 LPM	2.00
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A
System Flow	FM-7	3-5 LPM	4.7
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.55
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.35
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Check DAHS for normal operation. Is system logging data?		Check if ok	X
Verify no Alarms in DAHS		Check if ok	X
Check printer status		Check if ok	X
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X
No yellow "WARNING" status indicated?		Check if ok	X
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	X
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.			

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath	Tag ID	Limits	Date
Technician's Name: Jason White			5/17/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares			
O2/CO High Span, NOX Zero	SV1	>150 PSI	1000
NOX Low Span/CO Low Span	SV2	>150 PSI	850
NOX High Span, O2/CO Zero	SV3	>150 PSI	760
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Sample Line Temperature	TC1	250°F	250
Sample NH3 Temperature	TC2	760°C	760
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	5.4
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	6.5
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	5.0
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	6.5
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	70.5
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X
Moisture Sensor B/Filter	MS-2		X
Operational Status of Sample Pumps (2)		Check if ok	X
Operational Status of Condensate Drain Pump		Check if ok	X
Check LED Status of Sample Cooler		Check if ok	X
NH3 Scrubber Drain	HV-4	Drain weekly	X
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
System Flow	FM-1	3-5 LPM	3.6
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.80
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.70
CO Analyzer	FM-5	1.2 - 1.7 LPM	2.05
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A
System Flow	FM-7	3-5 LPM	4.6
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.60
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.40
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Check DAHS for normal operation. Is system logging data?		Check if ok	X
Verify no Alarms in DAHS		Check if ok	X
Check printer status		Check if ok	X
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X
No yellow "WARNING" status indicated?		Check if ok	X
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	N/A
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.			

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath	Tag ID	Limits	Date
Technician's Name: Jason White			5/26/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares			
O2/CO High Span, NOX Zero	SV1	>150 PSI	1420
NOX Low Span/CO Low Span	SV2	>150 PSI	1480
NOX High Span, O2/CO Zero	SV3	>150 PSI	1490
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Sample Line Temperature	TC1	250°F	250
Sample NH3 Temperature	TC2	760°C	760
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	4.5
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	6.3
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	4.0
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	6.7
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	69.5
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X
Moisture Sensor B/Filter	MS-2		X
Operational Status of Sample Pumps (2)		Check if ok	X
Operational Status of Condensate Drain Pump		Check if ok	X
Check LED Status of Sample Cooler		Check if ok	X
NH3 Scrubber Drain	HV-4	Drain weekly	X
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
System Flow	FM-1	3-5 LPM	3.6
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.70
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.45
CO Analyzer	FM-5	1.2 - 1.7 LPM	1.85
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A
System Flow	FM-7	3-5 LPM	4.5
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.65
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.40
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Check DAHS for normal operation. Is system logging data?		Check if ok	X
Verify no Alarms in DAHS		Check if ok	X
Check printer status		Check if ok	X
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X
No yellow "WARNING" status indicated?		Check if ok	X
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	X
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	X
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.			

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath	Tag ID	Limits	Date
Technician's Name: Jason White			6/1/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares			
O2/CO High Span, NOX Zero	SV1	>150 PSI	1000
NOX Low Span/CO Low Span	SV2	>150 PSI	1070
NOX High Span, O2/CO Zero	SV3	>150 PSI	1140
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Sample Line Temperature	TC1	250°F	250
Sample NH3 Temperature	TC2	760°C	760
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	4.5
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	6.4
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	3.8
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	6.6
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	69.0
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X
Moisture Sensor B/Filter	MS-2		X
Operational Status of Sample Pumps (2)		Check if ok	X
Operational Status of Condensate Drain Pump		Check if ok	X
Check LED Status of Sample Cooler		Check if ok	X
NH3 Scrubber Drain	HV-4	Drain weekly	X
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
System Flow	FM-1	3-5 LPM	3.7
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.60
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.35
CO Analyzer	FM-5	1.2 - 1.7 LPM	2.10
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A
System Flow	FM-7	3-5 LPM	4.3
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.65
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.40
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Check DAHS for normal operation. Is system logging data?		Check if ok	X
Verify no Alarms in DAHS		Check if ok	X
Check printer status		Check if ok	X
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X
No yellow "WARNING" status indicated?		Check if ok	X
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	N/A
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.			

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath		Tag ID	Limits	Date
Technician's Name: Jason White				6/8/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares				
O2/CO High Span, NOX Zero	SV1	>150 PSI	710	
NOX Low Span/CO Low Span	SV2	>150 PSI	740	
NOX High Span, O2/CO Zero	SV3	>150 PSI	830	
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Sample Line Temperature	TC1	250°F	250	
Sample NH3 Temperature	TC2	760°C	760	
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	4.9	
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	6.3	
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	4.0	
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	6.6	
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X	
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X	
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	70.5	
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X	
Moisture Sensor B/Filter	MS-2		X	
Operational Status of Sample Pumps (2)		Check if ok	X	
Operational Status of Condensate Drain Pump		Check if ok	X	
Check LED Status of Sample Cooler		Check if ok	X	
NH3 Scrubber Drain	HV-4	Drain weekly	X	
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
System Flow	FM-1	3-5 LPM	3.7	
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.60	
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.35	
CO Analyzer	FM-5	1.2 - 1.7 LPM	2.05	
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A	
System Flow	FM-7	3-5 LPM	4.3	
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.65	
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.40	
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Check DAHS for normal operation. Is system logging data?		Check if ok	X	
Verify no Alarms in DAHS		Check if ok	X	
Check printer status		Check if ok	X	
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X	
No yellow "WARNING" status indicated?		Check if ok	X	
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	X	
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A	
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.				

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath	Tag ID	Limits	Date
Technician's Name: Jason White			6/15/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares			
O2/CO High Span, NOX Zero	SV1	>150 PSI	400
NOX Low Span/CO Low Span	SV2	>150 PSI	500
NOX High Span, O2/CO Zero	SV3	>150 PSI	590
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Sample Line Temperature	TC1	250°F	250
Sample NH3 Temperature	TC2	760°C	760
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	5.0
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	6.4
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	3.3
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	6.5
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	71.0
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X
Moisture Sensor B/Filter	MS-2		X
Operational Status of Sample Pumps (2)		Check if ok	X
Operational Status of Condensate Drain Pump		Check if ok	X
Check LED Status of Sample Cooler		Check if ok	X
NH3 Scrubber Drain	HV-4	Drain weekly	X
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
System Flow	FM-1	3-5 LPM	3.7
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.60
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.15
CO Analyzer	FM-5	1.2 - 1.7 LPM	2.05
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A
System Flow	FM-7	3-5 LPM	4.0
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.65
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.40
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Check DAHS for normal operation. Is system logging data?		Check if ok	X
Verify no Alarms in DAHS		Check if ok	X
Check printer status		Check if ok	X
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X
No yellow "WARNING" status indicated?		Check if ok	X
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	N/A
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.			

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath		Tag ID	Limits	Date
Technician's Name: Jason White				6/21/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares				
O2/CO High Span, NOX Zero	SV1	>150 PSI	1810	
NOX Low Span/CO Low Span	SV2	>150 PSI	1880	
NOX High Span, O2/CO Zero	SV3	>150 PSI	1830	
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Sample Line Temperature	TC1	250°F	250	
Sample NH3 Temperature	TC2	760°C	760	
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	5.0	
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	6.2	
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	4.5	
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	6.5	
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X	
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X	
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	71.0	
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X	
Moisture Sensor B/Filter	MS-2		X	
Operational Status of Sample Pumps (2)		Check if ok	X	
Operational Status of Condensate Drain Pump		Check if ok	X	
Check LED Status of Sample Cooler		Check if ok	X	
NH3 Scrubber Drain	HV-4	Drain weekly	X	
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
System Flow	FM-1	3-5 LPM	3.6	
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.60	
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.30	
CO Analyzer	FM-5	1.2 - 1.7 LPM	2.05	
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A	
System Flow	FM-7	3-5 LPM	4.0	
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.65	
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.40	
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Check DAHS for normal operation. Is system logging data?		Check if ok	X	
Verify no Alarms in DAHS		Check if ok	X	
Check printer status		Check if ok	X	
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X	
No yellow "WARNING" status indicated?		Check if ok	X	
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	N/A	
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A	
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.				

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath		Tag ID	Limits	Date
Technician's Name: Jason White				6/28/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares				
O2/CO High Span, NOX Zero	SV1	>150 PSI	1500	
NOX Low Span/CO Low Span	SV2	>150 PSI	1600	
NOX High Span, O2/CO Zero	SV3	>150 PSI	1560	
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Sample Line Temperature	TC1	250°F	250	
Sample NH3 Temperature	TC2	760°C	760	
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	5.0	
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	6.2	
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	4.5	
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	6.5	
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X	
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X	
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	71.0	
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X	
Moisture Sensor B/Filter	MS-2		X	
Operational Status of Sample Pumps (2)		Check if ok	X	
Operational Status of Condensate Drain Pump		Check if ok	X	
Check LED Status of Sample Cooler		Check if ok	X	
NH3 Scrubber Drain	HV-4	Drain weekly	X	
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
System Flow	FM-1	3-5 LPM	3.7	
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.60	
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.35	
CO Analyzer	FM-5	1.2 - 1.7 LPM	2.05	
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A	
System Flow	FM-7	3-5 LPM	4.3	
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.65	
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.40	
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Check DAHS for normal operation. Is system logging data?		Check if ok	X	
Verify no Alarms in DAHS		Check if ok	X	
Check printer status		Check if ok	X	
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X	
No yellow "WARNING" status indicated?		Check if ok	X	
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	N/A	
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A	
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.				

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath		Tag ID	Limits	Date
Technician's Name: Jason White				7/6/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares				
O2/CO High Span, NOX Zero	SV1	>150 PSI	1100	
NOX Low Span/CO Low Span	SV2	>150 PSI	1340	
NOX High Span, O2/CO Zero	SV3	>150 PSI	1270	
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Sample Line Temperature	TC1	250°F	250	
Sample NH3 Temperature	TC2	760°C	760	
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	5.3	
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	6.3	
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	4.6	
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	6.5	
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X	
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X	
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	71.0	
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X	
Moisture Sensor B/Filter	MS-2		X	
Operational Status of Sample Pumps (2)		Check if ok	X	
Operational Status of Condensate Drain Pump		Check if ok	X	
Check LED Status of Sample Cooler		Check if ok	X	
NH3 Scrubber Drain	HV-4	Drain weekly	X	
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
System Flow	FM-1	3-5 LPM	3.8	
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.60	
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.40	
CO Analyzer	FM-5	1.2 - 1.7 LPM	2.05	
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A	
System Flow	FM-7	3-5 LPM	4.2	
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.65	
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.40	
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Check DAHS for normal operation. Is system logging data?		Check if ok	X	
Verify no Alarms in DAHS		Check if ok	X	
Check printer status		Check if ok	X	
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X	
No yellow "WARNING" status indicated?		Check if ok	X	
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	X	
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A	
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.				

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath	Tag ID	Limits	Date
Technician's Name: Jason White			7/13/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares			
O2/CO High Span, NOX Zero	SV1	>150 PSI	870
NOX Low Span/CO Low Span	SV2	>150 PSI	1080
NOX High Span, O2/CO Zero	SV3	>150 PSI	980
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Sample Line Temperature	TC1	250°F	250
Sample NH3 Temperature	TC2	760°C	760
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	4.5
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	6.4
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	3.7
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	6.7
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	68.5
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X
Moisture Sensor B/Filter	MS-2		X
Operational Status of Sample Pumps (2)		Check if ok	X
Operational Status of Condensate Drain Pump		Check if ok	X
Check LED Status of Sample Cooler		Check if ok	X
NH3 Scrubber Drain	HV-4	Drain weekly	X
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
System Flow	FM-1	3-5 LPM	3.8
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.60
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.40
CO Analyzer	FM-5	1.2 - 1.7 LPM	2.05
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A
System Flow	FM-7	3-5 LPM	4.3
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.65
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.40
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Check DAHS for normal operation. Is system logging data?		Check if ok	X
Verify no Alarms in DAHS		Check if ok	X
Check printer status		Check if ok	X
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X
No yellow "WARNING" status indicated?		Check if ok	X
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	N/A
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.			

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath	Tag ID	Limits	Date
Technician's Name: Jason White			7/18/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares			
O2/CO High Span, NOX Zero	SV1	>150 PSI	680
NOX Low Span/CO Low Span	SV2	>150 PSI	880
NOX High Span, O2/CO Zero	SV3	>150 PSI	750
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Sample Line Temperature	TC1	250°F	250
Sample NH3 Temperature	TC2	760°C	760
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	4.9
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	6.4
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	4.1
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	6.5
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	69.0
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X
Moisture Sensor B/Filter	MS-2		X
Operational Status of Sample Pumps (2)		Check if ok	X
Operational Status of Condensate Drain Pump		Check if ok	X
Check LED Status of Sample Cooler		Check if ok	X
NH3 Scrubber Drain	HV-4	Drain weekly	X
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
System Flow	FM-1	3-5 LPM	3.8
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.60
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.40
CO Analyzer	FM-5	1.2 - 1.7 LPM	2.05
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A
System Flow	FM-7	3-5 LPM	3.8
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.65
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.40
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Check DAHS for normal operation. Is system logging data?		Check if ok	X
Verify no Alarms in DAHS		Check if ok	X
Check printer status		Check if ok	X
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X
No yellow "WARNING" status indicated?		Check if ok	X
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	X
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.			

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath	Tag ID	Limits	Date
Technician's Name: Jason White			7/27/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares			
O2/CO High Span, NOX Zero	SV1	>150 PSI	1950
NOX Low Span/CO Low Span	SV2	>150 PSI	1980
NOX High Span, O2/CO Zero	SV3	>150 PSI	1900
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Sample Line Temperature	TC1	250°F	250
Sample NH3 Temperature	TC2	760°C	760
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	5.0
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	6.5
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	4.5
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	6.5
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	69.0
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X
Moisture Sensor B/Filter	MS-2		X
Operational Status of Sample Pumps (2)		Check if ok	X
Operational Status of Condensate Drain Pump		Check if ok	X
Check LED Status of Sample Cooler		Check if ok	X
NH3 Scrubber Drain	HV-4	Drain weekly	X
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
System Flow	FM-1	3-5 LPM	3.8
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.60
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.45
CO Analyzer	FM-5	1.2 - 1.7 LPM	2.05
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A
System Flow	FM-7	3-5 LPM	3.9
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.65
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.40
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Check DAHS for normal operation. Is system logging data?		Check if ok	X
Verify no Alarms in DAHS		Check if ok	X
Check printer status		Check if ok	X
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X
No yellow "WARNING" status indicated?		Check if ok	X
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	N/A
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.			

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath	Tag ID	Limits	Date
Technician's Name: Jason White			8/1/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares			
O2/CO High Span, NOX Zero	SV1	>150 PSI	1660
NOX Low Span/CO Low Span	SV2	>150 PSI	1700
NOX High Span, O2/CO Zero	SV3	>150 PSI	1610
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Sample Line Temperature	TC1	250°F	250
Sample NH3 Temperature	TC2	760°C	760
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	5.2
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	6.4
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	4.7
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	6.5
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	69.5
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X
Moisture Sensor B/Filter	MS-2		X
Operational Status of Sample Pumps (2)		Check if ok	X
Operational Status of Condensate Drain Pump		Check if ok	X
Check LED Status of Sample Cooler		Check if ok	X
NH3 Scrubber Drain	HV-4	Drain weekly	X
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
System Flow	FM-1	3-5 LPM	3.8
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.60
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.45
CO Analyzer	FM-5	1.2 - 1.7 LPM	2.05
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A
System Flow	FM-7	3-5 LPM	4.0
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.65
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.40
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Check DAHS for normal operation. Is system logging data?		Check if ok	X
Verify no Alarms in DAHS		Check if ok	X
Check printer status		Check if ok	X
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X
No yellow "WARNING" status indicated?		Check if ok	X
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	N/A
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.			

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath	Tag ID	Limits	Date
Technician's Name: Jason White			8/10/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares			
O2/CO High Span, NOX Zero	SV1	>150 PSI	1330
NOX Low Span/CO Low Span	SV2	>150 PSI	1370
NOX High Span, O2/CO Zero	SV3	>150 PSI	1300
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Sample Line Temperature	TC1	250°F	250
Sample NH3 Temperature	TC2	760°C	760
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	5.4
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	6.4
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	4.7
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	6.5
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	69.5
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X
Moisture Sensor B/Filter	MS-2		X
Operational Status of Sample Pumps (2)		Check if ok	X
Operational Status of Condensate Drain Pump		Check if ok	X
Check LED Status of Sample Cooler		Check if ok	X
NH3 Scrubber Drain	HV-4	Drain weekly	X
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
System Flow	FM-1	3-5 LPM	3.8
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.70
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.40
CO Analyzer	FM-5	1.2 - 1.7 LPM	2.10
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A
System Flow	FM-7	3-5 LPM	4.0
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.65
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.40
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Check DAHS for normal operation. Is system logging data?		Check if ok	X
Verify no Alarms in DAHS		Check if ok	X
Check printer status		Check if ok	X
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X
No yellow "WARNING" status indicated?		Check if ok	X
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	X
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	X
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.			

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath	Tag ID	Limits	Date
Technician's Name: Jason White			8/17/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares			
O2/CO High Span, NOX Zero	SV1	>150 PSI	990
NOX Low Span/CO Low Span	SV2	>150 PSI	1000
NOX High Span, O2/CO Zero	SV3	>150 PSI	930
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Sample Line Temperature	TC1	250°F	250
Sample NH3 Temperature	TC2	760°C	760
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	5.0
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	6.5
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	4.3
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	6.5
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	68.5
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X
Moisture Sensor B/Filter	MS-2		X
Operational Status of Sample Pumps (2)		Check if ok	X
Operational Status of Condensate Drain Pump		Check if ok	X
Check LED Status of Sample Cooler		Check if ok	X
NH3 Scrubber Drain	HV-4	Drain weekly	X
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
System Flow	FM-1	3-5 LPM	3.9
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.60
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.30
CO Analyzer	FM-5	1.2 - 1.7 LPM	2.05
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A
System Flow	FM-7	3-5 LPM	4.0
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.65
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.40
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Check DAHS for normal operation. Is system logging data?		Check if ok	X
Verify no Alarms in DAHS		Check if ok	X
Check printer status		Check if ok	X
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X
No yellow "WARNING" status indicated?		Check if ok	X
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	N/A
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.			

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath	Tag ID	Limits	Date
Technician's Name: Jason White			8/23/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares			
O2/CO High Span, NOX Zero	SV1	>150 PSI	770
NOX Low Span/CO Low Span	SV2	>150 PSI	800
NOX High Span, O2/CO Zero	SV3	>150 PSI	700
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Sample Line Temperature	TC1	250°F	250
Sample NH3 Temperature	TC2	760°C	760
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	5.1
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	7.5
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	4.6
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	7.5
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	68.0
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X
Moisture Sensor B/Filter	MS-2		X
Operational Status of Sample Pumps (2)		Check if ok	X
Operational Status of Condensate Drain Pump		Check if ok	X
Check LED Status of Sample Cooler		Check if ok	X
NH3 Scrubber Drain	HV-4	Drain weekly	X
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
System Flow	FM-1	3-5 LPM	3.8
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.55
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.40
CO Analyzer	FM-5	1.2 - 1.7 LPM	2.05
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A
System Flow	FM-7	3-5 LPM	4.0
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.65
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.40
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Check DAHS for normal operation. Is system logging data?		Check if ok	X
Verify no Alarms in DAHS		Check if ok	X
Check printer status		Check if ok	X
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X
No yellow "WARNING" status indicated?		Check if ok	X
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	N/A
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.			

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath	Tag ID	Limits	Date
Technician's Name: Jason White			8/31/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares			
O2/CO High Span, NOX Zero	SV1	>150 PSI	1920
NOX Low Span/CO Low Span	SV2	>150 PSI	1970
NOX High Span, O2/CO Zero	SV3	>150 PSI	1980
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Sample Line Temperature	TC1	250°F	250
Sample NH3 Temperature	TC2	760°C	760
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	5.8
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	6.3
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	5.0
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	6.4
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	70.0
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X
Moisture Sensor B/Filter	MS-2		X
Operational Status of Sample Pumps (2)		Check if ok	X
Operational Status of Condensate Drain Pump		Check if ok	X
Check LED Status of Sample Cooler		Check if ok	X
NH3 Scrubber Drain	HV-4	Drain weekly	X
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
System Flow	FM-1	3-5 LPM	3.7
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.60
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.40
CO Analyzer	FM-5	1.2 - 1.7 LPM	2.00
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A
System Flow	FM-7	3-5 LPM	3.9
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.65
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.40
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Check DAHS for normal operation. Is system logging data?		Check if ok	X
Verify no Alarms in DAHS		Check if ok	X
Check printer status		Check if ok	X
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X
No yellow "WARNING" status indicated?		Check if ok	X
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	N/A
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.			

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath		Tag ID	Limits	Date
Technician's Name: Jason White				9/7/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares				
O2/CO High Span, NOX Zero	SV1	>150 PSI	1680	
NOX Low Span/CO Low Span	SV2	>150 PSI	1740	
NOX High Span, O2/CO Zero	SV3	>150 PSI	1730	
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Sample Line Temperature	TC1	250°F	250	
Sample NH3 Temperature	TC2	760°C	760	
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	5.7	
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	6.4	
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	5.2	
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	6.5	
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X	
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X	
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	69.0	
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X	
Moisture Sensor B/Filter	MS-2		X	
Operational Status of Sample Pumps (2)		Check if ok	X	
Operational Status of Condensate Drain Pump		Check if ok	X	
Check LED Status of Sample Cooler		Check if ok	X	
NH3 Scrubber Drain	HV-4	Drain weekly	X	
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
System Flow	FM-1	3-5 LPM	3.8	
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.60	
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.40	
CO Analyzer	FM-5	1.2 - 1.7 LPM	2.00	
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A	
System Flow	FM-7	3-5 LPM	4.0	
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.65	
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.40	
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Check DAHS for normal operation. Is system logging data?		Check if ok	X	
Verify no Alarms in DAHS		Check if ok	X	
Check printer status		Check if ok	X	
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X	
No yellow "WARNING" status indicated?		Check if ok	X	
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	N/A	
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A	
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.				

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath	Tag ID	Limits	Date
Technician's Name: Jason White			9/12/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares			
O2/CO High Span, NOX Zero	SV1	>150 PSI	1420
NOX Low Span/CO Low Span	SV2	>150 PSI	1480
NOX High Span, O2/CO Zero	SV3	>150 PSI	1460
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Sample Line Temperature	TC1	250°F	250
Sample NH3 Temperature	TC2	760°C	760
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	5.5
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	6.4
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	5.0
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	6.4
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	68.5
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X
Moisture Sensor B/Filter	MS-2		X
Operational Status of Sample Pumps (2)		Check if ok	X
Operational Status of Condensate Drain Pump		Check if ok	X
Check LED Status of Sample Cooler		Check if ok	X
NH3 Scrubber Drain	HV-4	Drain weekly	X
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
System Flow	FM-1	3-5 LPM	3.8
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.55
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.40
CO Analyzer	FM-5	1.2 - 1.7 LPM	1.95
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A
System Flow	FM-7	3-5 LPM	4.0
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.65
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.35
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Check DAHS for normal operation. Is system logging data?		Check if ok	X
Verify no Alarms in DAHS		Check if ok	X
Check printer status		Check if ok	X
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X
No yellow "WARNING" status indicated?		Check if ok	X
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	N/A
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.			

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath		Tag ID	Limits	Date
Technician's Name: Jason White				9/19/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares				
O2/CO High Span, NOX Zero	SV1	>150 PSI	1140	
NOX Low Span/CO Low Span	SV2	>150 PSI	1220	
NOX High Span, O2/CO Zero	SV3	>150 PSI	1200	
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Sample Line Temperature	TC1	250°F	250	
Sample NH3 Temperature	TC2	760°C	760	
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	5.6	
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	6.4	
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	5.0	
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	6.3	
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X	
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X	
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	69.5	
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X	
Moisture Sensor B/Filter	MS-2		X	
Operational Status of Sample Pumps (2)		Check if ok	X	
Operational Status of Condensate Drain Pump		Check if ok	X	
Check LED Status of Sample Cooler		Check if ok	X	
NH3 Scrubber Drain	HV-4	Drain weekly	X	
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
System Flow	FM-1	3-5 LPM	3.8	
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.55	
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.40	
CO Analyzer	FM-5	1.2 - 1.7 LPM	1.95	
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A	
System Flow	FM-7	3-5 LPM	4.0	
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.65	
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.40	
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Check DAHS for normal operation. Is system logging data?		Check if ok	X	
Verify no Alarms in DAHS		Check if ok	X	
Check printer status		Check if ok	X	
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X	
No yellow "WARNING" status indicated?		Check if ok	X	
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	N/A	
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A	
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.				

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath		Tag ID	Limits	Date
Technician's Name: Jason White				9/28/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares				
O2/CO High Span, NOX Zero	SV1	>150 PSI	730	
NOX Low Span/CO Low Span	SV2	>150 PSI	810	
NOX High Span, O2/CO Zero	SV3	>150 PSI	710	
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Sample Line Temperature	TC1	250°F	250	
Sample NH3 Temperature	TC2	760°C	760	
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	5.9	
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	6.3	
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	5.5	
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	6.4	
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X	
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X	
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	69.5	
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X	
Moisture Sensor B/Filter	MS-2		X	
Operational Status of Sample Pumps (2)		Check if ok	X	
Operational Status of Condensate Drain Pump		Check if ok	X	
Check LED Status of Sample Cooler		Check if ok	X	
NH3 Scrubber Drain	HV-4	Drain weekly	X	
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
System Flow	FM-1	3-5 LPM	3.8	
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.55	
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.20	
CO Analyzer	FM-5	1.2 - 1.7 LPM	1.95	
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A	
System Flow	FM-7	3-5 LPM	4.0	
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.65	
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.40	
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Check DAHS for normal operation. Is system logging data?		Check if ok	X	
Verify no Alarms in DAHS		Check if ok	X	
Check printer status		Check if ok	X	
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X	
No yellow "WARNING" status indicated?		Check if ok	X	
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	X	
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	X	
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.				

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath		Tag ID	Limits	Date
Technician's Name: Jason White				10/8/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares				
O2/CO High Span, NOX Zero	SV1	>150 PSI	1610	
NOX Low Span/CO Low Span	SV2	>150 PSI	1520	
NOX High Span, O2/CO Zero	SV3	>150 PSI	1650	
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Sample Line Temperature	TC1	250°F	250	
Sample NH3 Temperature	TC2	760°C	760	
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	6.0	
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	6.3	
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	5.6	
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	6.4	
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X	
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X	
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	70.0	
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X	
Moisture Sensor B/Filter	MS-2		X	
Operational Status of Sample Pumps (2)		Check if ok	X	
Operational Status of Condensate Drain Pump		Check if ok	X	
Check LED Status of Sample Cooler		Check if ok	X	
NH3 Scrubber Drain	HV-4	Drain weekly	X	
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
System Flow	FM-1	3-5 LPM	3.8	
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.55	
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.40	
CO Analyzer	FM-5	1.2 - 1.7 LPM	1.95	
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A	
System Flow	FM-7	3-5 LPM	3.6	
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.65	
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.35	
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Check DAHS for normal operation. Is system logging data?		Check if ok	X	
Verify no Alarms in DAHS		Check if ok	X	
Check printer status		Check if ok	X	
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X	
No yellow "WARNING" status indicated?		Check if ok	X	
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	N/A	
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A	
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.				

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath	Tag ID	Limits	Date
Technician's Name: Jason White			10/15/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares			
O2/CO High Span, NOX Zero	SV1	>150 PSI	1490
NOX Low Span/CO Low Span	SV2	>150 PSI	1330
NOX High Span, O2/CO Zero	SV3	>150 PSI	1460
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Sample Line Temperature	TC1	250°F	250
Sample NH3 Temperature	TC2	760°C	760
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	5.9
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	6.2
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	5.6
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	6.4
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	69.5
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X
Moisture Sensor B/Filter	MS-2		X
Operational Status of Sample Pumps (2)		Check if ok	X
Operational Status of Condensate Drain Pump		Check if ok	X
Check LED Status of Sample Cooler		Check if ok	X
NH3 Scrubber Drain	HV-4	Drain weekly	X
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
System Flow	FM-1	3-5 LPM	3.8
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.60
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.20
CO Analyzer	FM-5	1.2 - 1.7 LPM	1.95
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A
System Flow	FM-7	3-5 LPM	4.0
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.65
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.35
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Check DAHS for normal operation. Is system logging data?		Check if ok	X
Verify no Alarms in DAHS		Check if ok	X
Check printer status		Check if ok	X
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X
No yellow "WARNING" status indicated?		Check if ok	X
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	N/A
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.			

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath		Tag ID	Limits	Date
Technician's Name: Jason White				10/21/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares				
O2/CO High Span, NOX Zero	SV1	>150 PSI		1260
NOX Low Span/CO Low Span	SV2	>150 PSI		1060
NOX High Span, O2/CO Zero	SV3	>150 PSI		1270
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Sample Line Temperature	TC1	250°F		250
Sample NH3 Temperature	TC2	760°C		760
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg		5.7
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI		6.3
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg		5.4
Sample Line Pressure	PI-5	3-10 TGT 8 PSI		6.4
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)		X
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)		X
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F		67.5
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter		X
Moisture Sensor B/Filter	MS-2			X
Operational Status of Sample Pumps (2)		Check if ok		X
Operational Status of Condensate Drain Pump		Check if ok		X
Check LED Status of Sample Cooler		Check if ok		X
NH3 Scrubber Drain	HV-4	Drain weekly		X
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
System Flow	FM-1	3-5 LPM		3.9
NOx Analyzer	FM-3	1.2 - 1.7 LPM		1.55
NOx Bypass	FM-4	1.2 - 1.7 LPM		1.15
CO Analyzer	FM-5	1.2 - 1.7 LPM		1.95
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7		N/A
System Flow	FM-7	3-5 LPM		4.0
NOx/NH3 Analyzer	FM-8	1.5 LPM		1.65
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM		1.40
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Check DAHS for normal operation. Is system logging data?		Check if ok		X
Verify no Alarms in DAHS		Check if ok		X
Check printer status		Check if ok		X
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok		X
No yellow "WARNING" status indicated?		Check if ok		X
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A		N/A
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A		N/A
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.				

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath		Tag ID	Limits	Date
Technician's Name: Jason White				10/26/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares				
O2/CO High Span, NOX Zero	SV1	>150 PSI	1150	
NOX Low Span/CO Low Span	SV2	>150 PSI	960	
NOX High Span, O2/CO Zero	SV3	>150 PSI	1190	
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Sample Line Temperature	TC1	250°F	250	
Sample NH3 Temperature	TC2	760°C	760	
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	6.0	
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	6.3	
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	5.9	
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	6.4	
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X	
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X	
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	68.5	
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X	
Moisture Sensor B/Filter	MS-2		X	
Operational Status of Sample Pumps (2)		Check if ok	X	
Operational Status of Condensate Drain Pump		Check if ok	X	
Check LED Status of Sample Cooler		Check if ok	X	
NH3 Scrubber Drain	HV-4	Drain weekly	X	
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
System Flow	FM-1	3-5 LPM	3.9	
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.55	
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.15	
CO Analyzer	FM-5	1.2 - 1.7 LPM	1.95	
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A	
System Flow	FM-7	3-5 LPM	4.0	
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.65	
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.40	
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Check DAHS for normal operation. Is system logging data?		Check if ok	X	
Verify no Alarms in DAHS		Check if ok	X	
Check printer status		Check if ok	X	
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X	
No yellow "WARNING" status indicated?		Check if ok	X	
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	N/A	
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A	
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.				

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath		Tag ID	Limits	Date
Technician's Name: Jason White				11/2/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares				
O2/CO High Span, NOX Zero	SV1	>150 PSI	920	
NOX Low Span/CO Low Span	SV2	>150 PSI	660	
NOX High Span, O2/CO Zero	SV3	>150 PSI	940	
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Sample Line Temperature	TC1	250°F	250	
Sample NH3 Temperature	TC2	760°C	760	
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	5.9	
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	6.4	
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	5.5	
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	6.5	
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X	
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X	
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	68.5	
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X	
Moisture Sensor B/Filter	MS-2		X	
Operational Status of Sample Pumps (2)		Check if ok	X	
Operational Status of Condensate Drain Pump		Check if ok	X	
Check LED Status of Sample Cooler		Check if ok	X	
NH3 Scrubber Drain	HV-4	Drain weekly	X	
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
System Flow	FM-1	3-5 LPM	3.9	
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.55	
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.40	
CO Analyzer	FM-5	1.2 - 1.7 LPM	1.95	
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A	
System Flow	FM-7	3-5 LPM	4.5	
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.65	
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.40	
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Check DAHS for normal operation. Is system logging data?		Check if ok	X	
Verify no Alarms in DAHS		Check if ok	X	
Check printer status		Check if ok	X	
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X	
No yellow "WARNING" status indicated?		Check if ok	X	
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	N/A	
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A	
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.				

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath		Tag ID	Limits	Date
Technician's Name: Jason White				11/9/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares				
O2/CO High Span, NOX Zero	SV1	>150 PSI	1950	
NOX Low Span/CO Low Span	SV2	>150 PSI	1940	
NOX High Span, O2/CO Zero	SV3	>150 PSI	2000	
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Sample Line Temperature	TC1	250°F	250	
Sample NH3 Temperature	TC2	760°C	760	
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	5.9	
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	6.4	
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	5.8	
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	6.5	
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X	
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X	
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	69.5	
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X	
Moisture Sensor B/Filter	MS-2		X	
Operational Status of Sample Pumps (2)		Check if ok	X	
Operational Status of Condensate Drain Pump		Check if ok	X	
Check LED Status of Sample Cooler		Check if ok	X	
NH3 Scrubber Drain	HV-4	Drain weekly	X	
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
System Flow	FM-1	3-5 LPM	3.9	
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.55	
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.40	
CO Analyzer	FM-5	1.2 - 1.7 LPM	1.95	
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A	
System Flow	FM-7	3-5 LPM	4.4	
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.60	
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.35	
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Check DAHS for normal operation. Is system logging data?		Check if ok	X	
Verify no Alarms in DAHS		Check if ok	X	
Check printer status		Check if ok	X	
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X	
No yellow "WARNING" status indicated?		Check if ok	X	
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	X	
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	X	
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.				

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath		Tag ID	Limits	Date
Technician's Name: Jason White				11/16/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares				
O2/CO High Span, NOX Zero	SV1	>150 PSI	1640	
NOX Low Span/CO Low Span	SV2	>150 PSI	1690	
NOX High Span, O2/CO Zero	SV3	>150 PSI	1710	
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Sample Line Temperature	TC1	250°F	250	
Sample NH3 Temperature	TC2	760°C	760	
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	6.0	
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	6.4	
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	6.0	
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	6.4	
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X	
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X	
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	68.5	
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X	
Moisture Sensor B/Filter	MS-2		X	
Operational Status of Sample Pumps (2)		Check if ok	X	
Operational Status of Condensate Drain Pump		Check if ok	X	
Check LED Status of Sample Cooler		Check if ok	X	
NH3 Scrubber Drain	HV-4	Drain weekly	X	
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
System Flow	FM-1	3-5 LPM	3.9	
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.55	
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.15	
CO Analyzer	FM-5	1.2 - 1.7 LPM	1.95	
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A	
System Flow	FM-7	3-5 LPM	4.3	
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.60	
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.35	
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Check DAHS for normal operation. Is system logging data?		Check if ok	X	
Verify no Alarms in DAHS		Check if ok	X	
Check printer status		Check if ok	X	
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X	
No yellow "WARNING" status indicated?		Check if ok	X	
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	N/A	
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A	
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.				

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath	Tag ID	Limits	Date
Technician's Name: Jason White			11/21/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares			
O2/CO High Span, NOX Zero	SV1	>150 PSI	1500
NOX Low Span/CO Low Span	SV2	>150 PSI	1560
NOX High Span, O2/CO Zero	SV3	>150 PSI	1600
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Sample Line Temperature	TC1	250°F	250
Sample NH3 Temperature	TC2	760°C	760
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	6.0
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	6.5
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	6.0
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	6.5
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	70.0
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X
Moisture Sensor B/Filter	MS-2		X
Operational Status of Sample Pumps (2)		Check if ok	X
Operational Status of Condensate Drain Pump		Check if ok	X
Check LED Status of Sample Cooler		Check if ok	X
NH3 Scrubber Drain	HV-4	Drain weekly	X
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
System Flow	FM-1	3-5 LPM	3.9
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.60
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.40
CO Analyzer	FM-5	1.2 - 1.7 LPM	1.95
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A
System Flow	FM-7	3-5 LPM	4.4
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.60
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.40
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Check DAHS for normal operation. Is system logging data?		Check if ok	X
Verify no Alarms in DAHS		Check if ok	X
Check printer status		Check if ok	X
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X
No yellow "WARNING" status indicated?		Check if ok	X
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	N/A
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.			

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath		Tag ID	Limits	Date
Technician's Name: Jason White				11/30/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares				
O2/CO High Span, NOX Zero	SV1	>150 PSI	1230	
NOX Low Span/CO Low Span	SV2	>150 PSI	1320	
NOX High Span, O2/CO Zero	SV3	>150 PSI	1300	
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Sample Line Temperature	TC1	250°F	250	
Sample NH3 Temperature	TC2	760°C	760	
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	5.9	
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	6.3	
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	5.7	
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	6.4	
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X	
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X	
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	68.0	
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X	
Moisture Sensor B/Filter	MS-2		X	
Operational Status of Sample Pumps (2)		Check if ok	X	
Operational Status of Condensate Drain Pump		Check if ok	X	
Check LED Status of Sample Cooler		Check if ok	X	
NH3 Scrubber Drain	HV-4	Drain weekly	X	
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
System Flow	FM-1	3-5 LPM	3.9	
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.55	
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.40	
CO Analyzer	FM-5	1.2 - 1.7 LPM	1.95	
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A	
System Flow	FM-7	3-5 LPM	4.2	
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.60	
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.35	
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Check DAHS for normal operation. Is system logging data?		Check if ok	X	
Verify no Alarms in DAHS		Check if ok	X	
Check printer status		Check if ok	X	
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X	
No yellow "WARNING" status indicated?		Check if ok	X	
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	N/A	
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A	
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.				

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath		Tag ID	Limits	Date
Technician's Name: Jason White				12/5/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares				
O2/CO High Span, NOX Zero	SV1	>150 PSI	1180	
NOX Low Span/CO Low Span	SV2	>150 PSI	1210	
NOX High Span, O2/CO Zero	SV3	>150 PSI	1170	
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Sample Line Temperature	TC1	250°F	250	
Sample NH3 Temperature	TC2	760°C	760	
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	6.0	
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	6.4	
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	6.0	
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	6.3	
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X	
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X	
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	69.5	
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X	
Moisture Sensor B/Filter	MS-2		X	
Operational Status of Sample Pumps (2)		Check if ok	X	
Operational Status of Condensate Drain Pump		Check if ok	X	
Check LED Status of Sample Cooler		Check if ok	X	
NH3 Scrubber Drain	HV-4	Drain weekly	X	
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
System Flow	FM-1	3-5 LPM	3.9	
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.55	
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.40	
CO Analyzer	FM-5	1.2 - 1.7 LPM	1.95	
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A	
System Flow	FM-7	3-5 LPM	4.0	
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.60	
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.35	
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Check DAHS for normal operation. Is system logging data?		Check if ok	X	
Verify no Alarms in DAHS		Check if ok	X	
Check printer status		Check if ok	X	
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X	
No yellow "WARNING" status indicated?		Check if ok	X	
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	N/A	
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A	
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.				

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath		Tag ID	Limits	Date
Technician's Name: Jason White				12/14/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares				
O2/CO High Span, NOX Zero	SV1	>150 PSI	780	
NOX Low Span/CO Low Span	SV2	>150 PSI	890	
NOX High Span, O2/CO Zero	SV3	>150 PSI	720	
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Sample Line Temperature	TC1	250°F	250	
Sample NH3 Temperature	TC2	760°C	760	
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	5.9	
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	6.5	
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	5.8	
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	6.4	
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X	
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X	
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	68.0	
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X	
Moisture Sensor B/Filter	MS-2		X	
Operational Status of Sample Pumps (2)		Check if ok	X	
Operational Status of Condensate Drain Pump		Check if ok	X	
Check LED Status of Sample Cooler		Check if ok	X	
NH3 Scrubber Drain	HV-4	Drain weekly	X	
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
System Flow	FM-1	3-5 LPM	3.9	
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.55	
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.40	
CO Analyzer	FM-5	1.2 - 1.7 LPM	1.95	
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A	
System Flow	FM-7	3-5 LPM	4.1	
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.60	
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.35	
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required				
Check DAHS for normal operation. Is system logging data?		Check if ok	X	
Verify no Alarms in DAHS		Check if ok	X	
Check printer status		Check if ok	X	
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X	
No yellow "WARNING" status indicated?		Check if ok	X	
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	X	
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	X	
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.				

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath	Tag ID	Limits	Date
Technician's Name: Jason White			12/19/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares			
O2/CO High Span, NOX Zero	SV1	>150 PSI	1900
NOX Low Span/CO Low Span	SV2	>150 PSI	1910
NOX High Span, O2/CO Zero	SV3	>150 PSI	1970
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Sample Line Temperature	TC1	250°F	250
Sample NH3 Temperature	TC2	760°C	760
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	6.5
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	6.2
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	6.5
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	6.3
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	69.5
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X
Moisture Sensor B/Filter	MS-2		X
Operational Status of Sample Pumps (2)		Check if ok	X
Operational Status of Condensate Drain Pump		Check if ok	X
Check LED Status of Sample Cooler		Check if ok	X
NH3 Scrubber Drain	HV-4	Drain weekly	X
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
System Flow	FM-1	3-5 LPM	3.9
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.55
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.35
CO Analyzer	FM-5	1.2 - 1.7 LPM	1.95
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A
System Flow	FM-7	3-5 LPM	4.4
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.65
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.40
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Check DAHS for normal operation. Is system logging data?		Check if ok	X
Verify no Alarms in DAHS		Check if ok	X
Check printer status		Check if ok	X
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X
No yellow "WARNING" status indicated?		Check if ok	X
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	X
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	X
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.			

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Location: McGrath	Tag ID	Limits	Date
Technician's Name: Jason White			12/28/2023
Calibration Gas Pressures - Enter readings and submit an order as needed, no less than 2 spares, no more than 5 spares			
O2/CO High Span, NOX Zero	SV1	>150 PSI	1710
NOX Low Span/CO Low Span	SV2	>150 PSI	1710
NOX High Span, O2/CO Zero	SV3	>150 PSI	1730
Stack Sample Line - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Sample Line Temperature	TC1	250°F	250
Sample NH3 Temperature	TC2	760°C	760
Sample Line Pressure/Vacuum	PI-1	<10 TGT 6.0 "Hg	6.5
Sample Line Pressure/Vacuum	PI-2	3-10 TGT 8 PSI	6.2
Sample Line Pressure/Vacuum	PI-4	<10 TGT 7.5 "Hg	6.7
Sample Line Pressure	PI-5	3-10 TGT 8 PSI	6.4
Verify Functionality of Sample Pump A Flow Switch	FS-1	5-7 LPM (set point)	X
Verify Functionality of Sample Pump B Flow Switch	FS-2	5-7 LPM (set point)	X
Visual Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Room Enclosure Temperature	Check HVAC control	72 ± 5 °F	70.5
Moisture Sensor A/Filter	MS-1	Clean and dry, if filter shows buildup and flow levels are dropping, replace filter	X
Moisture Sensor B/Filter	MS-2		X
Operational Status of Sample Pumps (2)		Check if ok	X
Operational Status of Condensate Drain Pump		Check if ok	X
Check LED Status of Sample Cooler		Check if ok	X
NH3 Scrubber Drain	HV-4	Drain weekly	X
Sample Flow Meter Readings - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
System Flow	FM-1	3-5 LPM	3.8
NOx Analyzer	FM-3	1.2 - 1.7 LPM	1.55
NOx Bypass	FM-4	1.2 - 1.7 LPM	1.15
CO Analyzer	FM-5	1.2 - 1.7 LPM	1.95
Cal Gas Flow (only during Calibration)	FM-6	> than FM-1 plus FM-7	N/A
System Flow	FM-7	3-5 LPM	4.1
NOx/NH3 Analyzer	FM-8	1.5 LPM	1.65
NOx/NH3 Analyzer Bypass	FM-9	1.2 - 1.5 LPM	1.40
DAHS Checks - Mark as either Acceptable "√", Corrective action Required "X", or Actual Readings, where required			
Check DAHS for normal operation. Is system logging data?		Check if ok	X
Verify no Alarms in DAHS		Check if ok	X
Check printer status		Check if ok	X
Check Analyzer calibration drift. Did all calibrations pass?		Check if ok	X
No yellow "WARNING" status indicated?		Check if ok	X
If necessary, perform manual calibration for each Analyzer, including NO first, before calibrating NOx on the NOx and NOx/NH3 Analyzers		Check if complete, or N/A	N/A
Perform a full "hands-off" calibration, as needed		Check if ok, or N/A	N/A
REMARKS: After changing a bottle, or any maintenance, such as changing a filter, etc., always perform a manual and "hands off" cal.			

Note: Report all discrepancies to the Control Operator immediately. Log corrective actions performed in the CEMS shelter log book.

Attachment 3
CEMS Calibrations Records

McGrath Calibration Checks

SCE McGrath Peaker

Cal Checks (Unit online) for 1/1/2023 thru 12/31/2023

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	EPA Vendor ID	Allowable Drift	Instrument Span	Results
			Cylinder ID						EPA Gas Type	Codes	
1/2/2023 4:35 PM	NOx ppm	High	Zero	0.0 ppm	11/21/2030	0.0 ppm	0 ppm	B72022	±10 ppm	200 ppm	Unit online; Passed
			CC739507						CO, O2, BALN		
1/2/2023 4:35 PM	NOx ppm	High	Span	179.4 ppm	10/28/2030	177.8 ppm	-1.6 ppm	B32022	±10 ppm	200 ppm	Unit online; Passed
			AAL072879						NO, NOX, BALN		
1/2/2023 4:35 PM	NOx ppm	Low	Zero	0.00 ppm	11/21/2030	0.04 ppm	0.04 ppm	B72022	±0.5 ppm	10 ppm	Unit online; Passed
			CC739507						CO, O2, BALN		
1/2/2023 4:35 PM	NOx ppm	Low	Span	8.98 ppm	10/27/2025	8.85 ppm	-0.13 ppm	B32022	±0.5 ppm	10 ppm	Unit online; Passed
			CC437354						CO, NO, NOX, BALN		
1/2/2023 4:35 PM	75-NOx ppm	High	Zero	0.0 ppm	11/21/2030	0.0 ppm	0 ppm	B72022	±10 ppm	200 ppm	Unit online; Passed
			CC739507						CO, O2, BALN		
1/2/2023 4:35 PM	75-NOx ppm	High	Span	179.4 ppm	10/28/2030	177.8 ppm	-1.6 ppm	B32022	±10 ppm	200 ppm	Unit online; Passed
			AAL072879						NO, NOX, BALN		
1/2/2023 4:35 PM	75-NOx ppm	Low	Zero	0.00 ppm	11/21/2030	0.04 ppm	0.04 ppm	B72022	±5 ppm	10 ppm	Unit online; Passed
			CC739507						CO, O2, BALN		
1/2/2023 4:35 PM	75-NOx ppm	Low	Span	8.98 ppm	10/27/2025	8.85 ppm	-0.13 ppm	B32022	±5 ppm	10 ppm	Unit online; Passed
			CC437354						CO, NO, NOX, BALN		
1/2/2023 4:35 PM	CO ppm	High	Zero	0.0 ppm	10/28/2030	0.0 ppm	0 ppm	B32022	±20 ppm	200 ppm	Unit online; Passed
			AAL072879						NO, NOX, BALN		
1/2/2023 4:35 PM	CO ppm	High	Span	182.7 ppm	11/21/2030	180.8 ppm	-1.9 ppm	B72022	±20 ppm	200 ppm	Unit online; Passed
			CC739507						CO, O2, BALN		
1/2/2023 4:35 PM	CO ppm	Low	Zero	0.00 ppm	10/28/2030	0.00 ppm	0 ppm	B32022	±1 ppm	10 ppm	Unit online; Passed
			AAL072879						NO, NOX, BALN		
1/2/2023 4:35 PM	CO ppm	Low	Span	8.98 ppm	10/27/2025	8.83 ppm	-0.15 ppm	B32022	±1 ppm	10 ppm	Unit online; Passed
			CC437354						CO, NO, NOX, BALN		
1/2/2023 4:35 PM	NH3/NOx ppm	High	Zero	0.0 ppm	11/21/2030	0.0 ppm	0 ppm	B72022	±10 ppm	200 ppm	Unit online; Passed
			CC739507						CO, O2, BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	EPA Vendor ID	Allowable Drift	Instrument Span	Results
1/2/2023 4:35 PM	NH3/NOx ppm	High	Span	179.4 ppm	10/28/2030	178.2 ppm	-1.2 ppm	B32022	±10 ppm NO, NOX, BALN	200 ppm	Unit online; Passed
1/2/2023 4:35 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	11/21/2030	0.11 ppm	0.11 ppm	B72022	±0.5 ppm CO, O2, BALN	10 ppm	Unit online; Passed
1/2/2023 4:35 PM	NH3/NOx ppm	Low	Span	8.98 ppm	10/27/2025	9.03 ppm	0.05 ppm	B32022	±0.5 ppm CO, NO, NOX, BALN	10 ppm	Unit online; Passed
1/2/2023 4:35 PM	O2 %	Single	Zero	0.00%	10/28/2030	0.00%	0%	B32022	±1% NO, NOX, BALN	25%	Unit online; Passed
1/2/2023 4:35 PM	O2 %	Single	Span	22.51%	11/21/2030	22.29%	-0.22%	B72022	±1% CO, O2, BALN	25%	Unit online; Passed
1/2/2023 4:35 PM	75-O2 %	Single	Zero	0.00%	10/28/2030	0.00%	0%	B32022	±1% NO, NOX, BALN	25%	Unit online; Passed
1/2/2023 4:35 PM	75-O2 %	Single	Span	22.51%	11/21/2030	22.29%	-0.22%	B72022	±1% CO, O2, BALN	25%	Unit online; Passed
1/17/2023 6:05 PM	NOx ppm	High	Zero	0.0 ppm	11/21/2030	0.0 ppm	0 ppm	B72022	±10 ppm CO, O2, BALN	200 ppm	Unit online; Passed
1/17/2023 6:05 PM	NOx ppm	High	Span	179.4 ppm	10/28/2030	178.0 ppm	-1.4 ppm	B32022	±10 ppm NO, NOX, BALN	200 ppm	Unit online; Passed
1/17/2023 6:05 PM	NOx ppm	Low	Zero	0.00 ppm	11/21/2030	0.07 ppm	0.07 ppm	B72022	±0.5 ppm CO, O2, BALN	10 ppm	Unit online; Passed
1/17/2023 6:05 PM	NOx ppm	Low	Span	8.98 ppm	10/27/2025	8.86 ppm	-0.12 ppm	B32022	±0.5 ppm CO, NO, NOX, BALN	10 ppm	Unit online; Passed
1/17/2023 6:05 PM	75-NOx ppm	High	Zero	0.0 ppm	11/21/2030	0.0 ppm	0 ppm	B72022	±10 ppm CO, O2, BALN	200 ppm	Unit online; Passed
1/17/2023 6:05 PM	75-NOx ppm	High	Span	179.4 ppm	10/28/2030	178.0 ppm	-1.4 ppm	B32022	±10 ppm NO, NOX, BALN	200 ppm	Unit online; Passed
1/17/2023 6:05 PM	75-NOx ppm	Low	Zero	0.00 ppm	11/21/2030	0.07 ppm	0.07 ppm	B72022	±5 ppm CO, O2, BALN	10 ppm	Unit online; Passed
1/17/2023 6:05 PM	75-NOx ppm	Low	Span	8.98 ppm	10/27/2025	8.86 ppm	-0.12 ppm	B32022	±5 ppm CO, NO, NOX, BALN	10 ppm	Unit online; Passed

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	EPA Vendor ID	Allowable Drift	Instrument Span	Results
1/17/2023 6:05 PM	CO ppm	High	Zero	0.0 ppm	10/28/2030	0.0 ppm	0 ppm	B32022	±20 ppm NO, NOX, BALN	200 ppm	Unit online; Passed
1/17/2023 6:05 PM	CO ppm	High	Span	182.7 ppm	11/21/2030	181.6 ppm	-1.1 ppm	B72022	±20 ppm CO, O2, BALN	200 ppm	Unit online; Passed
1/17/2023 6:05 PM	CO ppm	Low	Zero	0.00 ppm	10/28/2030	0.00 ppm	0 ppm	B32022	±1 ppm NO, NOX, BALN	10 ppm	Unit online; Passed
1/17/2023 6:05 PM	CO ppm	Low	Span	8.98 ppm	10/27/2025	9.06 ppm	0.08 ppm	B32022	±1 ppm CO, NO, NOX, BALN	10 ppm	Unit online; Passed
1/17/2023 6:05 PM	NH3/NOx ppm	High	Zero	0.0 ppm	11/21/2030	0.0 ppm	0 ppm	B72022	±10 ppm CO, O2, BALN	200 ppm	Unit online; Passed
1/17/2023 6:05 PM	NH3/NOx ppm	High	Span	179.4 ppm	10/28/2030	178.7 ppm	-0.7 ppm	B32022	±10 ppm NO, NOX, BALN	200 ppm	Unit online; Passed
1/17/2023 6:05 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	11/21/2030	0.17 ppm	0.17 ppm	B72022	±0.5 ppm CO, O2, BALN	10 ppm	Unit online; Passed
1/17/2023 6:05 PM	NH3/NOx ppm	Low	Span	8.98 ppm	10/27/2025	9.07 ppm	0.09 ppm	B32022	±0.5 ppm CO, NO, NOX, BALN	10 ppm	Unit online; Passed
1/17/2023 6:05 PM	O2 %	Single	Zero	0.00%	10/28/2030	0.00%	0%	B32022	±1% NO, NOX, BALN	25%	Unit online; Passed
1/17/2023 6:05 PM	O2 %	Single	Span	22.51%	11/21/2030	22.29%	-0.22%	B72022	±1% CO, O2, BALN	25%	Unit online; Passed
1/17/2023 6:05 PM	75-O2 %	Single	Zero	0.00%	10/28/2030	0.00%	0%	B32022	±1% NO, NOX, BALN	25%	Unit online; Passed
1/17/2023 6:05 PM	75-O2 %	Single	Span	22.51%	11/21/2030	22.29%	-0.22%	B72022	±1% CO, O2, BALN	25%	Unit online; Passed
2/3/2023 8:05 AM	NOx ppm	High	Zero	0.0 ppm	12/13/2030	0.0 ppm	0 ppm	B72022	±10 ppm CO, O2, BALN	200 ppm	Unit online; Passed
2/3/2023 8:05 AM	NOx ppm	High	Span	182.1 ppm	11/28/2030	181.8 ppm	-0.3 ppm	B32022	±10 ppm NO, NOX, BALN	200 ppm	Unit online; Passed
2/3/2023 8:05 AM	NOx ppm	Low	Zero	0.00 ppm	12/13/2030	0.07 ppm	0.07 ppm	B72022	±0.5 ppm CO, O2, BALN	10 ppm	Unit online; Passed

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
		Cylinder ID		Value	Value	EPA Vendor ID	EPA Vendor ID	EPA Gas Type Codes		
2/3/2023 8:05 AM	NOx ppm	Low	Span	8.89 ppm	8.90 ppm	12/1/2025	0.01 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC470005			B32022		CO, NO, NOX, BALN		
2/3/2023 8:05 AM	75-NOx ppm	High	Zero	0.0 ppm	0.0 ppm	12/13/2030	0 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC415330			B72022		CO, O2, BALN		
2/3/2023 8:05 AM	75-NOx ppm	High	Span	182.1 ppm	181.8 ppm	11/28/2030	-0.3 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC269367			B32022		NO, NOX, BALN		
2/3/2023 8:05 AM	75-NOx ppm	Low	Zero	0.00 ppm	0.07 ppm	12/13/2030	0.07 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC415330			B72022		CO, O2, BALN		
2/3/2023 8:05 AM	75-NOx ppm	Low	Span	8.89 ppm	8.90 ppm	12/1/2025	0.01 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC470005			B32022		CO, NO, NOX, BALN		
2/3/2023 8:05 AM	CO ppm	High	Zero	0.0 ppm	0.0 ppm	11/28/2030	0 ppm	±20 ppm	200 ppm	Unit online; Passed
			CC269367			B32022		NO, NOX, BALN		
2/3/2023 8:05 AM	CO ppm	High	Span	182.0 ppm	180.6 ppm	12/13/2030	-1.4 ppm	±20 ppm	200 ppm	Unit online; Passed
			CC415330			B72022		CO, O2, BALN		
2/3/2023 8:05 AM	CO ppm	Low	Zero	0.00 ppm	0.00 ppm	11/28/2030	0 ppm	±1 ppm	10 ppm	Unit online; Passed
			CC269367			B32022		NO, NOX, BALN		
2/3/2023 8:05 AM	CO ppm	Low	Span	8.95 ppm	9.03 ppm	12/1/2025	0.08 ppm	±1 ppm	10 ppm	Unit online; Passed
			CC470005			B32022		CO, NO, NOX, BALN		
2/3/2023 8:05 AM	NH3/NOx ppm	High	Zero	0.0 ppm	0.0 ppm	12/13/2030	0 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC415330			B72022		CO, O2, BALN		
2/3/2023 8:05 AM	NH3/NOx ppm	High	Span	182.1 ppm	181.4 ppm	11/28/2030	-0.7 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC269367			B32022		NO, NOX, BALN		
2/3/2023 8:05 AM	NH3/NOx ppm	Low	Zero	0.00 ppm	0.16 ppm	12/13/2030	0.16 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC415330			B72022		CO, O2, BALN		
2/3/2023 8:05 AM	NH3/NOx ppm	Low	Span	8.89 ppm	9.03 ppm	12/1/2025	0.14 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC470005			B32022		CO, NO, NOX, BALN		
2/3/2023 8:05 AM	O2 %	Single	Zero	0.00%	0.00%	11/28/2030	0%	±1%	25%	Unit online; Passed
			CC269367			B32022		NO, NOX, BALN		
2/3/2023 8:05 AM	O2 %	Single	Span	22.53%	22.29%	12/13/2030	-0.24%	±1%	25%	Unit online; Passed
			CC415330			B72022		CO, O2, BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
			Cylinder ID	Value	Value		EPA Vendor ID	EPA Gas Type	Codes	
2/3/2023 8:05 AM	75-O2 %	Single	Zero	0.00%	0.00%	11/28/2030	0%	±1%	25%	Unit online; Passed
			CC269367				B32022	NO, NOX, BALN		
2/3/2023 8:05 AM	75-O2 %	Single	Span	22.53%	22.29%	12/13/2030	-0.24%	±1%	25%	Unit online; Passed
			CC415330				B72022	CO, O2, BALN		
2/14/2023 5:21 PM	NOx ppm	High	Zero	0.0 ppm	0.0 ppm	12/13/2030	0 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC415330				B72022	CO, O2, BALN		
2/14/2023 5:21 PM	NOx ppm	High	Span	182.1 ppm	183.1 ppm	11/28/2030	1 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC269367				B32022	NO, NOX, BALN		
2/14/2023 5:21 PM	NOx ppm	Low	Zero	0.00 ppm	0.07 ppm	12/13/2030	0.07 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC415330				B72022	CO, O2, BALN		
2/14/2023 5:21 PM	NOx ppm	Low	Span	8.89 ppm	8.92 ppm	12/11/2025	0.03 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC470005				B32022	CO, NO, NOX, BALN		
2/14/2023 5:21 PM	75-NOx ppm	High	Zero	0.0 ppm	0.0 ppm	12/13/2030	0 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC415330				B72022	CO, O2, BALN		
2/14/2023 5:21 PM	75-NOx ppm	High	Span	182.1 ppm	183.1 ppm	11/28/2030	1 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC269367				B32022	NO, NOX, BALN		
2/14/2023 5:21 PM	75-NOx ppm	Low	Zero	0.00 ppm	0.07 ppm	12/13/2030	0.07 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC415330				B72022	CO, O2, BALN		
2/14/2023 5:21 PM	75-NOx ppm	Low	Span	8.89 ppm	8.92 ppm	12/11/2025	0.03 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC470005				B32022	CO, NO, NOX, BALN		
2/14/2023 5:21 PM	CO ppm	High	Zero	0.0 ppm	0.0 ppm	11/28/2030	0 ppm	±20 ppm	200 ppm	Unit online; Passed
			CC269367				B32022	NO, NOX, BALN		
2/14/2023 5:21 PM	CO ppm	High	Span	182.0 ppm	181.8 ppm	12/13/2030	-0.2 ppm	±20 ppm	200 ppm	Unit online; Passed
			CC415330				B72022	CO, O2, BALN		
2/14/2023 5:21 PM	CO ppm	Low	Zero	0.00 ppm	0.00 ppm	11/28/2030	0 ppm	±1 ppm	10 ppm	Unit online; Passed
			CC269367				B32022	NO, NOX, BALN		
2/14/2023 5:21 PM	CO ppm	Low	Span	8.95 ppm	8.84 ppm	12/11/2025	-0.11 ppm	±1 ppm	10 ppm	Unit online; Passed
			CC470005				B32022	CO, NO, NOX, BALN		
2/14/2023 5:21 PM	NH3/NOx ppm	High	Zero	0.0 ppm	0.0 ppm	12/13/2030	0 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC415330				B72022	CO, O2, BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
		Cylinder ID		Value	Value		EPA Vendor ID	EPA Gas Type	Codes	
2/14/2023 5:21 PM	NH3/NOx ppm	High	Span	182.1 ppm	179.9 ppm	11/28/2030	-2.2 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC269367			B32022		NO, NOX, BALN		
2/14/2023 5:21 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	0.18 ppm	12/13/2030	0.18 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC415330			B72022		CO, O2, BALN		
2/14/2023 5:21 PM	NH3/NOx ppm	Low	Span	8.89 ppm	8.94 ppm	12/11/2025	0.05 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC470005			B32022		CO, NO, NOX, BALN		
2/14/2023 5:21 PM	O2 %	Single	Zero	0.00%	0.00%	11/28/2030	0%	±1%	25%	Unit online; Passed
			CC269367			B32022		NO, NOX, BALN		
2/14/2023 5:21 PM	O2 %	Single	Span	22.53%	22.30%	12/13/2030	-0.23%	±1%	25%	Unit online; Passed
			CC415330			B72022		CO, O2, BALN		
2/14/2023 5:21 PM	75-O2 %	Single	Zero	0.00%	0.00%	11/28/2030	0%	±1%	25%	Unit online; Passed
			CC269367			B32022		NO, NOX, BALN		
2/14/2023 5:21 PM	75-O2 %	Single	Span	22.53%	22.30%	12/13/2030	-0.23%	±1%	25%	Unit online; Passed
			CC415330			B72022		CO, O2, BALN		
2/15/2023 6:05 AM	NOx ppm	High	Zero	0.0 ppm	0.0 ppm	12/13/2030	0 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC415330			B72022		CO, O2, BALN		
2/15/2023 6:05 AM	NOx ppm	High	Span	182.1 ppm	182.3 ppm	11/28/2030	0.2 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC269367			B32022		NO, NOX, BALN		
2/15/2023 6:05 AM	NOx ppm	Low	Zero	0.00 ppm	0.03 ppm	12/13/2030	0.03 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC415330			B72022		CO, O2, BALN		
2/15/2023 6:05 AM	NOx ppm	Low	Span	8.89 ppm	8.83 ppm	12/11/2025	-0.06 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC470005			B32022		CO, NO, NOX, BALN		
2/15/2023 6:05 AM	75-NOx ppm	High	Zero	0.0 ppm	0.0 ppm	12/13/2030	0 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC415330			B72022		CO, O2, BALN		
2/15/2023 6:05 AM	75-NOx ppm	High	Span	182.1 ppm	182.3 ppm	11/28/2030	0.2 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC269367			B32022		NO, NOX, BALN		
2/15/2023 6:05 AM	75-NOx ppm	Low	Zero	0.00 ppm	0.03 ppm	12/13/2030	0.03 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC415330			B72022		CO, O2, BALN		
2/15/2023 6:05 AM	75-NOx ppm	Low	Span	8.89 ppm	8.83 ppm	12/11/2025	-0.06 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC470005			B32022		CO, NO, NOX, BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
			Cylinder ID			EPA Vendor ID	EPA Gas Type Codes			
2/15/2023 6:05 AM	CO ppm	High	Zero	0.0 ppm	0.0 ppm	11/28/2030	0 ppm	±20 ppm	200 ppm	Unit online; Passed
			CC269367			B32022		NO, NOX, BALN		
2/15/2023 6:05 AM	CO ppm	High	Span	182.0 ppm	181.5 ppm	12/13/2030	-0.5 ppm	±20 ppm	200 ppm	Unit online; Passed
			CC415330			B72022		CO, O2, BALN		
2/15/2023 6:05 AM	CO ppm	Low	Zero	0.00 ppm	0.00 ppm	11/28/2030	0 ppm	±1 ppm	10 ppm	Unit online; Passed
			CC269367			B32022		NO, NOX, BALN		
2/15/2023 6:05 AM	CO ppm	Low	Span	8.95 ppm	9.12 ppm	12/1/2025	0.17 ppm	±1 ppm	10 ppm	Unit online; Passed
			CC470005			B32022		CO, NO, NOX, BALN		
2/15/2023 6:05 AM	NH3/NOx ppm	High	Zero	0.0 ppm	0.0 ppm	12/13/2030	0 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC415330			B72022		CO, O2, BALN		
2/15/2023 6:05 AM	NH3/NOx ppm	High	Span	182.1 ppm	179.3 ppm	11/28/2030	-2.8 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC269367			B32022		NO, NOX, BALN		
2/15/2023 6:05 AM	NH3/NOx ppm	Low	Zero	0.00 ppm	0.11 ppm	12/13/2030	0.11 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC415330			B72022		CO, O2, BALN		
2/15/2023 6:05 AM	NH3/NOx ppm	Low	Span	8.89 ppm	8.88 ppm	12/1/2025	-0.01 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC470005			B32022		CO, NO, NOX, BALN		
2/15/2023 6:05 AM	O2 %	Single	Zero	0.00%	0.00%	11/28/2030	0%	±1%	25%	Unit online; Passed
			CC269367			B32022		NO, NOX, BALN		
2/15/2023 6:05 AM	O2 %	Single	Span	22.53%	22.29%	12/13/2030	-0.24%	±1%	25%	Unit online; Passed
			CC415330			B72022		CO, O2, BALN		
2/15/2023 6:05 AM	75-O2 %	Single	Zero	0.00%	0.00%	11/28/2030	0%	±1%	25%	Unit online; Passed
			CC269367			B32022		NO, NOX, BALN		
2/15/2023 6:05 AM	75-O2 %	Single	Span	22.53%	22.29%	12/13/2030	-0.24%	±1%	25%	Unit online; Passed
			CC415330			B72022		CO, O2, BALN		
2/25/2023 1:35 PM	NOx ppm	High	Zero	0.0 ppm	0.0 ppm	12/13/2030	0 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC415330			B72022		CO, O2, BALN		
2/25/2023 1:35 PM	NOx ppm	High	Span	182.1 ppm	181.7 ppm	11/28/2030	-0.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC269367			B32022		NO, NOX, BALN		
2/25/2023 1:35 PM	NOx ppm	Low	Zero	0.00 ppm	0.06 ppm	12/13/2030	0.06 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC415330			B72022		CO, O2, BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	EPA Vendor ID	Allowable Drift	Instrument Span	Results
2/25/2023 1:35 PM	NOx ppm	Low	Span	8.89 ppm	12/1/2025	8.80 ppm	-0.09 ppm	B32022	±0.5 ppm	10 ppm	Unit online; Passed
			CC470005						CO, NO, NOX, BALN		
2/25/2023 1:35 PM	75-NOx ppm	High	Zero	0.0 ppm	12/13/2030	0.0 ppm	0 ppm	B72022	±10 ppm	200 ppm	Unit online; Passed
			CC415330						CO, O2, BALN		
2/25/2023 1:35 PM	75-NOx ppm	High	Span	182.1 ppm	11/28/2030	181.7 ppm	-0.4 ppm	B32022	±10 ppm	200 ppm	Unit online; Passed
			CC269367						NO, NOX, BALN		
2/25/2023 1:35 PM	75-NOx ppm	Low	Zero	0.0 ppm	12/13/2030	0.06 ppm	0.06 ppm	B72022	±5 ppm	10 ppm	Unit online; Passed
			CC415330						CO, O2, BALN		
2/25/2023 1:35 PM	75-NOx ppm	Low	Span	8.89 ppm	12/1/2025	8.80 ppm	-0.09 ppm	B32022	±5 ppm	10 ppm	Unit online; Passed
			CC470005						CO, NO, NOX, BALN		
2/25/2023 1:35 PM	CO ppm	High	Zero	0.0 ppm	11/28/2030	0.0 ppm	0 ppm	B32022	±20 ppm	200 ppm	Unit online; Passed
			CC269367						NO, NOX, BALN		
2/25/2023 1:35 PM	CO ppm	High	Span	182.0 ppm	12/13/2030	182.5 ppm	0.5 ppm	B72022	±20 ppm	200 ppm	Unit online; Passed
			CC415330						CO, O2, BALN		
2/25/2023 1:35 PM	CO ppm	Low	Zero	0.00 ppm	11/28/2030	0.00 ppm	0 ppm	B32022	±1 ppm	10 ppm	Unit online; Passed
			CC269367						NO, NOX, BALN		
2/25/2023 1:35 PM	CO ppm	Low	Span	8.95 ppm	12/1/2025	9.04 ppm	0.09 ppm	B32022	±1 ppm	10 ppm	Unit online; Passed
			CC470005						CO, NO, NOX, BALN		
2/25/2023 1:35 PM	NH3/NOx ppm	High	Zero	0.0 ppm	12/13/2030	0.0 ppm	0 ppm	B72022	±10 ppm	200 ppm	Unit online; Passed
			CC415330						CO, O2, BALN		
2/25/2023 1:35 PM	NH3/NOx ppm	High	Span	182.1 ppm	11/28/2030	181.0 ppm	-1.1 ppm	B32022	±10 ppm	200 ppm	Unit online; Passed
			CC269367						NO, NOX, BALN		
2/25/2023 1:35 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	12/13/2030	0.16 ppm	0.16 ppm	B72022	±0.5 ppm	10 ppm	Unit online; Passed
			CC415330						CO, O2, BALN		
2/25/2023 1:35 PM	NH3/NOx ppm	Low	Span	8.89 ppm	12/1/2025	8.74 ppm	-0.15 ppm	B32022	±0.5 ppm	10 ppm	Unit online; Passed
			CC470005						CO, NO, NOX, BALN		
2/25/2023 1:35 PM	O2 %	Single	Zero	0.00%	11/28/2030	0.00%	0%	B32022	±1%	25%	Unit online; Passed
			CC269367						NO, NOX, BALN		
2/25/2023 1:35 PM	O2 %	Single	Span	22.53%	12/13/2030	22.31%	-0.22%	B72022	±1%	25%	Unit online; Passed
			CC415330						CO, O2, BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	EPA Vendor ID	Allowable Drift	Instrument Span	Results
2/25/2023 1:35 PM	75-O2 %	Single	Zero	0.00%	11/28/2030	0.00%	0%	B32022	±1% NO, NOX, BALN	25%	Unit online; Passed
2/25/2023 1:35 PM	75-O2 %	Single	Span	22.53%	12/13/2030	22.31%	-0.22%	B72022	±1% CO, O2, BALN	25%	Unit online; Passed
3/2/2023 6:05 AM	NOx ppm	High	Zero	0.0 ppm	12/13/2030	0.0 ppm	0 ppm	B72022	±10 ppm CO, O2, BALN	200 ppm	Unit online; Passed
3/2/2023 6:05 AM	NOx ppm	High	Span	182.1 ppm	11/28/2030	181.9 ppm	-0.2 ppm	B32022	±10 ppm NO, NOX, BALN	200 ppm	Unit online; Passed
3/2/2023 6:05 AM	NOx ppm	Low	Zero	0.00 ppm	12/13/2030	0.03 ppm	0.03 ppm	B72022	±0.5 ppm CO, O2, BALN	10 ppm	Unit online; Passed
3/2/2023 6:05 AM	NOx ppm	Low	Span	8.89 ppm	12/11/2025	8.67 ppm	-0.22 ppm	B32022	±0.5 ppm CO, NO, NOX, BALN	10 ppm	Unit online; Passed
3/2/2023 6:05 AM	75-NOx ppm	High	Zero	0.0 ppm	12/13/2030	0.0 ppm	0 ppm	B72022	±10 ppm CO, O2, BALN	200 ppm	Unit online; Passed
3/2/2023 6:05 AM	75-NOx ppm	High	Span	182.1 ppm	11/28/2030	181.9 ppm	-0.2 ppm	B32022	±10 ppm NO, NOX, BALN	200 ppm	Unit online; Passed
3/2/2023 6:05 AM	75-NOx ppm	Low	Zero	0.00 ppm	12/13/2030	0.03 ppm	0.03 ppm	B72022	±5 ppm CO, O2, BALN	10 ppm	Unit online; Passed
3/2/2023 6:05 AM	75-NOx ppm	Low	Span	8.89 ppm	12/11/2025	8.67 ppm	-0.22 ppm	B32022	±5 ppm CO, NO, NOX, BALN	10 ppm	Unit online; Passed
3/2/2023 6:05 AM	CO ppm	High	Zero	0.0 ppm	11/28/2030	0.0 ppm	0 ppm	B32022	±20 ppm NO, NOX, BALN	200 ppm	Unit online; Passed
3/2/2023 6:05 AM	CO ppm	High	Span	182.0 ppm	12/13/2030	181.1 ppm	-0.9 ppm	B72022	±20 ppm CO, O2, BALN	200 ppm	Unit online; Passed
3/2/2023 6:05 AM	CO ppm	Low	Zero	0.00 ppm	11/28/2030	0.00 ppm	0 ppm	B32022	±1 ppm NO, NOX, BALN	10 ppm	Unit online; Passed
3/2/2023 6:05 AM	CO ppm	Low	Span	8.95 ppm	12/11/2025	8.72 ppm	-0.23 ppm	B32022	±1 ppm CO, NO, NOX, BALN	10 ppm	Unit online; Passed
3/2/2023 6:05 AM	NH3/NOx ppm	High	Zero	0.0 ppm	12/13/2030	0.0 ppm	0 ppm	B72022	±10 ppm CO, O2, BALN	200 ppm	Unit online; Passed

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	EPA Vendor ID	Allowable Drift	Instrument Span	Results
3/2/2023 6:05 AM	NH3/NOx ppm	High	Span	182.1 ppm	11/28/2030	181.8 ppm	-0.3 ppm	B32022	±10 ppm NO, NOX, BALN	200 ppm	Unit online; Passed
3/2/2023 6:05 AM	NH3/NOx ppm	Low	Zero	0.00 ppm	12/13/2030	0.13 ppm	0.13 ppm	B72022	±0.5 ppm CO, O2, BALN	10 ppm	Unit online; Passed
3/2/2023 6:05 AM	NH3/NOx ppm	Low	Span	8.89 ppm	12/11/2025	9.00 ppm	0.11 ppm	B32022	±0.5 ppm CO, NO, NOX, BALN	10 ppm	Unit online; Passed
3/2/2023 6:05 AM	O2 %	Single	Zero	0.00%	11/28/2030	0.00%	0%	B32022	±1% NO, NOX, BALN	25%	Unit online; Passed
3/2/2023 6:05 AM	O2 %	Single	Span	22.53%	12/13/2030	22.30%	-0.23%	B72022	±1% CO, O2, BALN	25%	Unit online; Passed
3/2/2023 6:05 AM	75-O2 %	Single	Zero	0.00%	11/28/2030	0.00%	0%	B32022	±1% NO, NOX, BALN	25%	Unit online; Passed
3/2/2023 6:05 AM	75-O2 %	Single	Span	22.53%	12/13/2030	22.30%	-0.23%	B72022	±1% CO, O2, BALN	25%	Unit online; Passed
3/4/2023 7:05 PM	NOx ppm	High	Zero	0.0 ppm	12/13/2030	0.0 ppm	0 ppm	B72022	±10 ppm CO, O2, BALN	200 ppm	Unit online; Passed
3/4/2023 7:05 PM	NOx ppm	High	Span	182.1 ppm	11/28/2030	181.6 ppm	-0.5 ppm	B32022	±10 ppm NO, NOX, BALN	200 ppm	Unit online; Passed
3/4/2023 7:05 PM	NOx ppm	Low	Zero	0.00 ppm	12/13/2030	0.04 ppm	0.04 ppm	B72022	±0.5 ppm CO, O2, BALN	10 ppm	Unit online; Passed
3/4/2023 7:05 PM	NOx ppm	Low	Span	8.89 ppm	12/11/2025	8.70 ppm	-0.19 ppm	B32022	±0.5 ppm CO, NO, NOX, BALN	10 ppm	Unit online; Passed
3/4/2023 7:05 PM	75-NOx ppm	High	Zero	0.0 ppm	12/13/2030	0.0 ppm	0 ppm	B72022	±10 ppm CO, O2, BALN	200 ppm	Unit online; Passed
3/4/2023 7:05 PM	75-NOx ppm	High	Span	182.1 ppm	11/28/2030	181.6 ppm	-0.5 ppm	B32022	±10 ppm NO, NOX, BALN	200 ppm	Unit online; Passed
3/4/2023 7:05 PM	75-NOx ppm	Low	Zero	0.00 ppm	12/13/2030	0.04 ppm	0.04 ppm	B72022	±5 ppm CO, O2, BALN	10 ppm	Unit online; Passed
3/4/2023 7:05 PM	75-NOx ppm	Low	Span	8.89 ppm	12/11/2025	8.70 ppm	-0.19 ppm	B32022	±5 ppm CO, NO, NOX, BALN	10 ppm	Unit online; Passed

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
			Cylinder ID				EPA Vendor ID	EPA Gas Type Codes		
3/4/2023 7:05 PM	CO ppm	High	Zero	0.0 ppm	0.0 ppm	11/28/2030	0 ppm	±20 ppm	200 ppm	Unit online; Passed
			CC269367				B32022	NO, NOX, BALN		
3/4/2023 7:05 PM	CO ppm	High	Span	182.0 ppm	180.3 ppm	12/13/2030	-1.7 ppm	±20 ppm	200 ppm	Unit online; Passed
			CC415330				B72022	CO, O2, BALN		
3/4/2023 7:05 PM	CO ppm	Low	Zero	0.00 ppm	0.00 ppm	11/28/2030	0 ppm	±1 ppm	10 ppm	Unit online; Passed
			CC269367				B32022	NO, NOX, BALN		
3/4/2023 7:05 PM	CO ppm	Low	Span	8.95 ppm	9.07 ppm	12/1/2025	0.12 ppm	±1 ppm	10 ppm	Unit online; Passed
			CC470005				B32022	CO, NO, NOX, BALN		
3/4/2023 7:05 PM	NH3/NOx ppm	High	Zero	0.0 ppm	0.0 ppm	12/13/2030	0 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC415330				B72022	CO, O2, BALN		
3/4/2023 7:05 PM	NH3/NOx ppm	High	Span	182.1 ppm	180.3 ppm	11/28/2030	-1.8 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC269367				B32022	NO, NOX, BALN		
3/4/2023 7:05 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	0.14 ppm	12/13/2030	0.14 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC415330				B72022	CO, O2, BALN		
3/4/2023 7:05 PM	NH3/NOx ppm	Low	Span	8.89 ppm	8.95 ppm	12/1/2025	0.06 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC470005				B32022	CO, NO, NOX, BALN		
3/4/2023 7:05 PM	O2 %	Single	Zero	0.00%	0.00%	11/28/2030	0%	±1%	25%	Unit online; Passed
			CC269367				B32022	NO, NOX, BALN		
3/4/2023 7:05 PM	O2 %	Single	Span	22.53%	22.31%	12/13/2030	-0.22%	±1%	25%	Unit online; Passed
			CC415330				B72022	CO, O2, BALN		
3/4/2023 7:05 PM	75-O2 %	Single	Zero	0.00%	0.00%	11/28/2030	0%	±1%	25%	Unit online; Passed
			CC269367				B32022	NO, NOX, BALN		
3/4/2023 7:05 PM	75-O2 %	Single	Span	22.53%	22.31%	12/13/2030	-0.22%	±1%	25%	Unit online; Passed
			CC415330				B72022	CO, O2, BALN		
3/14/2023 7:35 AM	NOx ppm	High	Zero	0.0 ppm	0.0 ppm	12/13/2030	0 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC415330				B72022	CO, O2, BALN		
3/14/2023 7:35 AM	NOx ppm	High	Span	182.1 ppm	180.9 ppm	11/28/2030	-1.2 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC269367				B32022	NO, NOX, BALN		
3/14/2023 7:35 AM	NOx ppm	Low	Zero	0.00 ppm	0.05 ppm	12/13/2030	0.05 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC415330				B72022	CO, O2, BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
			Cylinder ID	Value	Value	EPA Vendor ID	EPA Gas Type Codes			
3/14/2023 7:35 AM	NOx ppm	Low	Span CC470005	8.89 ppm	8.71 ppm	12/1/2025	-0.18 ppm	±0.5 ppm	10 ppm	Unit online; Passed
						B32022		CO, NO, NOX, BALN		
3/14/2023 7:35 AM	75-NOx ppm	High	Zero CC415330	0.0 ppm	0.0 ppm	12/13/2030	0 ppm	±10 ppm	200 ppm	Unit online; Passed
						B72022		CO, O2, BALN		
3/14/2023 7:35 AM	75-NOx ppm	High	Span CC269367	182.1 ppm	180.9 ppm	11/28/2030	-1.2 ppm	±10 ppm	200 ppm	Unit online; Passed
						B32022		NO, NOX, BALN		
3/14/2023 7:35 AM	75-NOx ppm	Low	Zero CC415330	0.00 ppm	0.05 ppm	12/13/2030	0.05 ppm	±5 ppm	10 ppm	Unit online; Passed
						B72022		CO, O2, BALN		
3/14/2023 7:35 AM	75-NOx ppm	Low	Span CC470005	8.89 ppm	8.71 ppm	12/1/2025	-0.18 ppm	±5 ppm	10 ppm	Unit online; Passed
						B32022		CO, NO, NOX, BALN		
3/14/2023 7:35 AM	CO ppm	High	Zero CC269367	0.0 ppm	0.0 ppm	11/28/2030	0 ppm	±20 ppm	200 ppm	Unit online; Passed
						B32022		NO, NOX, BALN		
3/14/2023 7:35 AM	CO ppm	High	Span CC415330	182.0 ppm	180.3 ppm	12/13/2030	-1.7 ppm	±20 ppm	200 ppm	Unit online; Passed
						B72022		CO, O2, BALN		
3/14/2023 7:35 AM	CO ppm	Low	Zero CC269367	0.00 ppm	0.00 ppm	11/28/2030	0 ppm	±1 ppm	10 ppm	Unit online; Passed
						B32022		NO, NOX, BALN		
3/14/2023 7:35 AM	CO ppm	Low	Span CC470005	8.95 ppm	8.77 ppm	12/1/2025	-0.18 ppm	±1 ppm	10 ppm	Unit online; Passed
						B32022		CO, NO, NOX, BALN		
3/14/2023 7:35 AM	NH3/NOx ppm	High	Zero CC415330	0.0 ppm	0.0 ppm	12/13/2030	0 ppm	±10 ppm	200 ppm	Unit online; Passed
						B72022		CO, O2, BALN		
3/14/2023 7:35 AM	NH3/NOx ppm	High	Span CC269367	182.1 ppm	177.7 ppm	11/28/2030	-4.4 ppm	±10 ppm	200 ppm	Unit online; Passed
						B32022		NO, NOX, BALN		
3/14/2023 7:35 AM	NH3/NOx ppm	Low	Zero CC415330	0.00 ppm	0.15 ppm	12/13/2030	0.15 ppm	±0.5 ppm	10 ppm	Unit online; Passed
						B72022		CO, O2, BALN		
3/14/2023 7:35 AM	NH3/NOx ppm	Low	Span CC470005	8.89 ppm	8.82 ppm	12/1/2025	-0.07 ppm	±0.5 ppm	10 ppm	Unit online; Passed
						B32022		CO, NO, NOX, BALN		
3/14/2023 7:35 AM	O2 %	Single	Zero CC269367	0.00%	0.00%	11/28/2030	0%	±1%	25%	Unit online; Passed
						B32022		NO, NOX, BALN		
3/14/2023 7:35 AM	O2 %	Single	Span CC415330	22.53%	22.30%	12/13/2030	-0.23%	±1%	25%	Unit online; Passed
						B72022		CO, O2, BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
			Cylinder ID	Value	Value	EPA Vendor ID	EPA Gas Type Codes			
3/14/2023 7:35 AM	75-O2 %	Single	Zero	0.00%	0.00%	B32022	0%	±1%	25%	Unit online; Passed
			CC269367	11/28/2030		B32022		NO, NOX, BALN		
3/14/2023 7:35 AM	75-O2 %	Single	Span	22.53%	22.30%	B72022	-0.23%	±1%	25%	Unit online; Passed
			CC415330	12/13/2030		B72022		CO, O2, BALN		
3/28/2023 7:36 AM	NOx ppm	High	Zero	0.0 ppm	0.0 ppm	B72023	0 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0155775	1/23/2031		B72023		CO, O2, BALN		
3/28/2023 7:36 AM	NOx ppm	High	Span	178.4 ppm	177.8 ppm	B32022	-0.6 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC1622	12/28/2030		B32022		NO, NOX, BALN		
3/28/2023 7:36 AM	NOx ppm	Low	Zero	0.00 ppm	0.05 ppm	B72023	0.05 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			EB0155775	1/23/2031		B72023		CO, O2, BALN		
3/28/2023 7:36 AM	NOx ppm	Low	Span	9.05 ppm	8.95 ppm	B32022	-0.1 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC304196	12/29/2025		B32022		CO, NO, NOX, BALN		
3/28/2023 7:36 AM	75-NOx ppm	High	Zero	0.0 ppm	0.0 ppm	B72023	0 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0155775	1/23/2031		B72023		CO, O2, BALN		
3/28/2023 7:36 AM	75-NOx ppm	High	Span	178.4 ppm	177.8 ppm	B32022	-0.6 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC1622	12/28/2030		B32022		NO, NOX, BALN		
3/28/2023 7:36 AM	75-NOx ppm	Low	Zero	0.00 ppm	0.05 ppm	B72023	0.05 ppm	±5 ppm	10 ppm	Unit online; Passed
			EB0155775	1/23/2031		B72023		CO, O2, BALN		
3/28/2023 7:36 AM	75-NOx ppm	Low	Span	9.05 ppm	8.95 ppm	B32022	-0.1 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC304196	12/29/2025		B32022		CO, NO, NOX, BALN		
3/28/2023 7:36 AM	CO ppm	High	Zero	0.0 ppm	0.0 ppm	B32022	0 ppm	±20 ppm	200 ppm	Unit online; Passed
			CC1622	12/28/2030		B32022		NO, NOX, BALN		
3/28/2023 7:36 AM	CO ppm	High	Span	182.8 ppm	173.8 ppm	B72023	-9 ppm	±20 ppm	200 ppm	Unit online; Passed
			EB0155775	1/23/2031		B72023		CO, O2, BALN		
3/28/2023 7:36 AM	CO ppm	Low	Zero	0.00 ppm	0.00 ppm	B32022	0 ppm	±1 ppm	10 ppm	Unit online; Passed
			CC1622	12/28/2030		B32022		NO, NOX, BALN		
3/28/2023 7:36 AM	CO ppm	Low	Span	9.04 ppm	8.39 ppm	B32022	-0.65 ppm	±1 ppm	10 ppm	Unit online; Passed
			CC304196	12/29/2025		B32022		CO, NO, NOX, BALN		
3/28/2023 7:36 AM	NH3/NOx ppm	High	Zero	0.0 ppm	0.0 ppm	B72023	0 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0155775	1/23/2031		B72023		CO, O2, BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	Allowable Drift	Instrument Span	Results
3/28/2023 7:36 AM	NH3/NOx ppm	High	Span 178.4 ppm	12/28/2030	173.9 ppm	-4.5 ppm	±10 ppm	NO, NOX, BALN	200 ppm	Unit online; Passed
			CC1622			B32022				
3/28/2023 7:36 AM	NH3/NOx ppm	Low	Zero 0.00 ppm	1/23/2031	0.15 ppm	0.15 ppm	±0.5 ppm	CO, O2, BALN	10 ppm	Unit online; Passed
			EB0155775			B72023				
3/28/2023 7:36 AM	NH3/NOx ppm	Low	Span 9.05 ppm	12/29/2025	9.04 ppm	-0.01 ppm	±0.5 ppm	CO, NO, NOX, BALN	10 ppm	Unit online; Passed
			CC304196			B32022				
3/28/2023 7:36 AM	O2 %	Single	Zero 0.00%	12/28/2030	0.00%	0%	±1%	NO, NOX, BALN	25%	Unit online; Passed
			CC1622			B32022				
3/28/2023 7:36 AM	O2 %	Single	Span 22.45%	1/23/2031	22.29%	-0.16%	±1%	CO, O2, BALN	25%	Unit online; Passed
			EB0155775			B72023				
3/28/2023 7:36 AM	75-O2 %	Single	Zero 0.00%	12/28/2030	0.00%	0%	±1%	NO, NOX, BALN	25%	Unit online; Passed
			CC1622			B32022				
3/28/2023 7:36 AM	75-O2 %	Single	Span 22.45%	1/23/2031	22.29%	-0.16%	±1%	CO, O2, BALN	25%	Unit online; Passed
			EB0155775			B72023				
4/4/2023 7:36 PM	NOx ppm	High	Zero 0.0 ppm	1/23/2031	0.0 ppm	0 ppm	±10 ppm	CO, O2, BALN	200 ppm	Unit online; Passed
			EB0155775			B72023				
4/4/2023 7:36 PM	NOx ppm	High	Span 178.4 ppm	12/28/2030	177.2 ppm	-1.2 ppm	±10 ppm	NO, NOX, BALN	200 ppm	Unit online; Passed
			CC1622			B32022				
4/4/2023 7:36 PM	NOx ppm	Low	Zero 0.00 ppm	1/23/2031	0.05 ppm	0.05 ppm	±0.5 ppm	CO, O2, BALN	10 ppm	Unit online; Passed
			EB0155775			B72023				
4/4/2023 7:36 PM	NOx ppm	Low	Span 9.05 ppm	12/29/2025	8.89 ppm	-0.16 ppm	±0.5 ppm	CO, NO, NOX, BALN	10 ppm	Unit online; Passed
			CC304196			B32022				
4/4/2023 7:36 PM	75-NOx ppm	High	Zero 0.0 ppm	1/23/2031	0.0 ppm	0 ppm	±10 ppm	CO, O2, BALN	200 ppm	Unit online; Passed
			EB0155775			B72023				
4/4/2023 7:36 PM	75-NOx ppm	High	Span 178.4 ppm	12/28/2030	177.2 ppm	-1.2 ppm	±10 ppm	NO, NOX, BALN	200 ppm	Unit online; Passed
			CC1622			B32022				
4/4/2023 7:36 PM	75-NOx ppm	Low	Zero 0.00 ppm	1/23/2031	0.05 ppm	0.05 ppm	±5 ppm	CO, O2, BALN	10 ppm	Unit online; Passed
			EB0155775			B72023				
4/4/2023 7:36 PM	75-NOx ppm	Low	Span 9.05 ppm	12/29/2025	8.89 ppm	-0.16 ppm	±5 ppm	CO, NO, NOX, BALN	10 ppm	Unit online; Passed
			CC304196			B32022				

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
		Cylinder ID				EPA Vendor ID	EPA Gas Type Codes			
4/4/2023 7:36 PM	CO ppm	High	Zero	0.0 ppm	0.0 ppm	12/28/2030	0 ppm	±20 ppm	200 ppm	Unit online; Passed
		CC1622				B32022		NO, NOX, BALN		
4/4/2023 7:36 PM	CO ppm	High	Span	182.8 ppm	180.2 ppm	1/23/2031	-2.6 ppm	±20 ppm	200 ppm	Unit online; Passed
		EB0155775				B72023		CO, O2, BALN		
4/4/2023 7:36 PM	CO ppm	Low	Zero	0.00 ppm	0.00 ppm	12/28/2030	0 ppm	±1 ppm	10 ppm	Unit online; Passed
		CC1622				B32022		NO, NOX, BALN		
4/4/2023 7:36 PM	CO ppm	Low	Span	9.04 ppm	9.03 ppm	12/29/2025	-0.01 ppm	±1 ppm	10 ppm	Unit online; Passed
		CC304196				B32022		CO, NO, NOX, BALN		
4/4/2023 7:36 PM	NH3/NOx ppm	High	Zero	0.0 ppm	0.0 ppm	1/23/2031	0 ppm	±10 ppm	200 ppm	Unit online; Passed
		EB0155775				B72023		CO, O2, BALN		
4/4/2023 7:36 PM	NH3/NOx ppm	High	Span	178.4 ppm	174.6 ppm	12/28/2030	-3.8 ppm	±10 ppm	200 ppm	Unit online; Passed
		CC1622				B32022		NO, NOX, BALN		
4/4/2023 7:36 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	0.15 ppm	1/23/2031	0.15 ppm	±0.5 ppm	10 ppm	Unit online; Passed
		EB0155775				B72023		CO, O2, BALN		
4/4/2023 7:36 PM	NH3/NOx ppm	Low	Span	9.05 ppm	8.86 ppm	12/29/2025	-0.19 ppm	±0.5 ppm	10 ppm	Unit online; Passed
		CC304196				B32022		CO, NO, NOX, BALN		
4/4/2023 7:36 PM	O2 %	Single	Zero	0.00%	0.00%	12/28/2030	0%	±1%	25%	Unit online; Passed
		CC1622				B32022		NO, NOX, BALN		
4/4/2023 7:36 PM	O2 %	Single	Span	22.45%	22.30%	1/23/2031	-0.15%	±1%	25%	Unit online; Passed
		EB0155775				B72023		CO, O2, BALN		
4/4/2023 7:36 PM	75-O2 %	Single	Zero	0.00%	0.00%	12/28/2030	0%	±1%	25%	Unit online; Passed
		CC1622				B32022		NO, NOX, BALN		
4/4/2023 7:36 PM	75-O2 %	Single	Span	22.45%	22.30%	1/23/2031	-0.15%	±1%	25%	Unit online; Passed
		EB0155775				B72023		CO, O2, BALN		
4/5/2023 4:51 AM	NOx ppm	High	Zero	0.0 ppm	0.0 ppm	1/23/2031	0 ppm	±10 ppm	200 ppm	Unit online; Passed
		EB0155775				B72023		CO, O2, BALN		
4/5/2023 4:51 AM	NOx ppm	High	Span	178.4 ppm	177.2 ppm	12/28/2030	-1.2 ppm	±10 ppm	200 ppm	Unit online; Passed
		CC1622				B32022		NO, NOX, BALN		
4/5/2023 4:51 AM	NOx ppm	Low	Zero	0.00 ppm	0.03 ppm	1/23/2031	0.03 ppm	±0.5 ppm	10 ppm	Unit online; Passed
		EB0155775				B72023		CO, O2, BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	Allowable Drift	Instrument Span	Results
		Cylinder ID					EPA Vendor ID	EPA Gas Type Codes		
4/5/2023 4:51 AM	NOx ppm	Low	Span	9.05 ppm	12/29/2025	8.96 ppm	-0.09 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC304196				B32022	CO, NO, NOX, BALN		
4/5/2023 4:51 AM	75-NOx ppm	High	Zero	0.0 ppm	1/23/2031	0.0 ppm	0 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0155775				B72023	CO, O2, BALN		
4/5/2023 4:51 AM	75-NOx ppm	High	Span	178.4 ppm	12/28/2030	177.2 ppm	-1.2 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC1622				B32022	NO, NOX, BALN		
4/5/2023 4:51 AM	75-NOx ppm	Low	Zero	0.00 ppm	1/23/2031	0.03 ppm	0.03 ppm	±5 ppm	10 ppm	Unit online; Passed
			EB0155775				B72023	CO, O2, BALN		
4/5/2023 4:51 AM	75-NOx ppm	Low	Span	9.05 ppm	12/29/2025	8.96 ppm	-0.09 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC304196				B32022	CO, NO, NOX, BALN		
4/5/2023 4:51 AM	CO ppm	High	Zero	0.0 ppm	12/28/2030	0.0 ppm	0 ppm	±20 ppm	200 ppm	Unit online; Passed
			CC1622				B32022	NO, NOX, BALN		
4/5/2023 4:51 AM	CO ppm	High	Span	182.8 ppm	1/23/2031	180.2 ppm	-2.6 ppm	±20 ppm	200 ppm	Unit online; Passed
			EB0155775				B72023	CO, O2, BALN		
4/5/2023 4:51 AM	CO ppm	Low	Zero	0.00 ppm	12/28/2030	0.00 ppm	0 ppm	±1 ppm	10 ppm	Unit online; Passed
			CC1622				B32022	NO, NOX, BALN		
4/5/2023 4:51 AM	CO ppm	Low	Span	9.04 ppm	12/29/2025	8.93 ppm	-0.11 ppm	±1 ppm	10 ppm	Unit online; Passed
			CC304196				B32022	CO, NO, NOX, BALN		
4/5/2023 4:51 AM	NH3/NOx ppm	High	Zero	0.0 ppm	1/23/2031	0.0 ppm	0 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0155775				B72023	CO, O2, BALN		
4/5/2023 4:51 AM	NH3/NOx ppm	High	Span	178.4 ppm	12/28/2030	174.4 ppm	-4 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC1622				B32022	NO, NOX, BALN		
4/5/2023 4:51 AM	NH3/NOx ppm	Low	Zero	0.00 ppm	1/23/2031	0.10 ppm	0.1 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			EB0155775				B72023	CO, O2, BALN		
4/5/2023 4:51 AM	NH3/NOx ppm	Low	Span	9.05 ppm	12/29/2025	8.84 ppm	-0.21 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC304196				B32022	CO, NO, NOX, BALN		
4/5/2023 4:51 AM	O2 %	Single	Zero	0.00%	12/28/2030	0.00%	0%	±1%	25%	Unit online; Passed
			CC1622				B32022	NO, NOX, BALN		
4/5/2023 4:51 AM	O2 %	Single	Span	22.45%	1/23/2031	22.30%	-0.15%	±1%	25%	Unit online; Passed
			EB0155775				B72023	CO, O2, BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
		Cylinder ID		Value	Value	EPA Vendor ID	EPA Gas Type Codes			
4/5/2023 4:51 AM	75-O2 %	Single	Zero	0.00%	0.00%	B32022	0%	±1%	25%	Unit online; Passed
			CC1622	12/28/2030		B32022		NO, NOX, BALN		
4/5/2023 4:51 AM	75-O2 %	Single	Span	22.45%	22.30%	B72023	-0.15%	±1%	25%	Unit online; Passed
			EB0155775	1/23/2031		B72023		CO, O2, BALN		
4/20/2023 6:05 PM	NOx ppm	High	Zero	0.0 ppm	0.0 ppm	B72023	0 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0155775	1/23/2031		B72023		CO, O2, BALN		
4/20/2023 6:05 PM	NOx ppm	High	Span	178.4 ppm	176.7 ppm	B32022	-1.7 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC1622	12/28/2030		B32022		NO, NOX, BALN		
4/20/2023 6:05 PM	NOx ppm	Low	Zero	0.00 ppm	0.06 ppm	B72023	0.06 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			EB0155775	1/23/2031		B72023		CO, O2, BALN		
4/20/2023 6:05 PM	NOx ppm	Low	Span	9.05 ppm	8.93 ppm	B32022	-0.12 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC304196	12/29/2025		B32022		CO, NO, NOX, BALN		
4/20/2023 6:05 PM	75-NOx ppm	High	Zero	0.0 ppm	0.0 ppm	B72023	0 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0155775	1/23/2031		B72023		CO, O2, BALN		
4/20/2023 6:05 PM	75-NOx ppm	High	Span	178.4 ppm	176.7 ppm	B32022	-1.7 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC1622	12/28/2030		B32022		NO, NOX, BALN		
4/20/2023 6:05 PM	75-NOx ppm	Low	Zero	0.00 ppm	0.06 ppm	B72023	0.06 ppm	±5 ppm	10 ppm	Unit online; Passed
			EB0155775	1/23/2031		B72023		CO, O2, BALN		
4/20/2023 6:05 PM	75-NOx ppm	Low	Span	9.05 ppm	8.93 ppm	B32022	-0.12 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC304196	12/29/2025		B32022		CO, NO, NOX, BALN		
4/20/2023 6:05 PM	CO ppm	High	Zero	0.0 ppm	0.0 ppm	B32022	0 ppm	±20 ppm	200 ppm	Unit online; Passed
			CC1622	12/28/2030		B32022		NO, NOX, BALN		
4/20/2023 6:05 PM	CO ppm	High	Span	182.8 ppm	180.4 ppm	B72023	-2.4 ppm	±20 ppm	200 ppm	Unit online; Passed
			EB0155775	1/23/2031		B72023		CO, O2, BALN		
4/20/2023 6:05 PM	CO ppm	Low	Zero	0.00 ppm	0.00 ppm	B32022	0 ppm	±1 ppm	10 ppm	Unit online; Passed
			CC1622	12/28/2030		B32022		NO, NOX, BALN		
4/20/2023 6:05 PM	CO ppm	Low	Span	9.04 ppm	9.33 ppm	B32022	0.29 ppm	±1 ppm	10 ppm	Unit online; Passed
			CC304196	12/29/2025		B32022		CO, NO, NOX, BALN		
4/20/2023 6:05 PM	NH3/NOx ppm	High	Zero	0.0 ppm	0.0 ppm	B72023	0 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0155775	1/23/2031		B72023		CO, O2, BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	Allowable Drift	Instrument Span	Results
4/20/2023 6:05 PM	NH3/NOx ppm	High	Span 178.4 ppm	176.3 ppm	12/28/2030	-2.1 ppm	±10 ppm	200 ppm	Unit online; Passed	
			CC1622	B32022			NO, NOX, BALN			
4/20/2023 6:05 PM	NH3/NOx ppm	Low	Zero 0.00 ppm	0.15 ppm	1/23/2031	0.15 ppm	±0.5 ppm	10 ppm	Unit online; Passed	
			EB0155775	B72023			CO, O2, BALN			
4/20/2023 6:05 PM	NH3/NOx ppm	Low	Span 9.05 ppm	9.13 ppm	12/29/2025	0.08 ppm	±0.5 ppm	10 ppm	Unit online; Passed	
			CC304196	B32022			CO, NO, NOX, BALN			
4/20/2023 6:05 PM	O2 %	Single	Zero 0.00%	0.00%	12/28/2030	0%	±1%	25%	Unit online; Passed	
			CC1622	B32022			NO, NOX, BALN			
4/20/2023 6:05 PM	O2 %	Single	Span 22.45%	22.31%	1/23/2031	-0.14%	±1%	25%	Unit online; Passed	
			EB0155775	B72023			CO, O2, BALN			
4/20/2023 6:05 PM	75-O2 %	Single	Zero 0.00%	0.00%	12/28/2030	0%	±1%	25%	Unit online; Passed	
			CC1622	B32022			NO, NOX, BALN			
4/20/2023 6:05 PM	75-O2 %	Single	Span 22.45%	22.31%	1/23/2031	-0.14%	±1%	25%	Unit online; Passed	
			EB0155775	B72023			CO, O2, BALN			
4/24/2023 8:50 PM	NOx ppm	High	Zero 0.0 ppm	0.0 ppm	1/23/2031	0 ppm	±10 ppm	200 ppm	Unit online; Passed	
			EB0155775	B72023			CO, O2, BALN			
4/24/2023 8:50 PM	NOx ppm	High	Span 178.4 ppm	177.1 ppm	12/28/2030	-1.3 ppm	±10 ppm	200 ppm	Unit online; Passed	
			CC1622	B32022			NO, NOX, BALN			
4/24/2023 8:50 PM	NOx ppm	Low	Zero 0.00 ppm	0.05 ppm	1/23/2031	0.05 ppm	±0.5 ppm	10 ppm	Unit online; Passed	
			EB0155775	B72023			CO, O2, BALN			
4/24/2023 8:50 PM	NOx ppm	Low	Span 9.05 ppm	8.89 ppm	12/29/2025	-0.16 ppm	±0.5 ppm	10 ppm	Unit online; Passed	
			CC304196	B32022			CO, NO, NOX, BALN			
4/24/2023 8:50 PM	75-NOx ppm	High	Zero 0.0 ppm	0.0 ppm	1/23/2031	0 ppm	±10 ppm	200 ppm	Unit online; Passed	
			EB0155775	B72023			CO, O2, BALN			
4/24/2023 8:50 PM	75-NOx ppm	High	Span 178.4 ppm	177.1 ppm	12/28/2030	-1.3 ppm	±10 ppm	200 ppm	Unit online; Passed	
			CC1622	B32022			NO, NOX, BALN			
4/24/2023 8:50 PM	75-NOx ppm	Low	Zero 0.00 ppm	0.05 ppm	1/23/2031	0.05 ppm	±5 ppm	10 ppm	Unit online; Passed	
			EB0155775	B72023			CO, O2, BALN			
4/24/2023 8:50 PM	75-NOx ppm	Low	Span 9.05 ppm	8.89 ppm	12/29/2025	-0.16 ppm	±5 ppm	10 ppm	Unit online; Passed	
			CC304196	B32022			CO, NO, NOX, BALN			

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	Allowable Drift	Instrument Span	Results
4/24/2023 8:50 PM	CO ppm	High	Zero	0.0 ppm	12/28/2030	0.0 ppm	0 ppm	±20 ppm	200 ppm	Unit online; Passed
			CC1622				B32022	NO, NOX, BALN		
4/24/2023 8:50 PM	CO ppm	High	Span	182.8 ppm	180.0 ppm	-2.8 ppm	-2.8 ppm	±20 ppm	200 ppm	Unit online; Passed
			EB0155775				B72023	CO, O2, BALN		
4/24/2023 8:50 PM	CO ppm	Low	Zero	0.00 ppm	0.00 ppm	0 ppm	0 ppm	±1 ppm	10 ppm	Unit online; Passed
			CC1622				B32022	NO, NOX, BALN		
4/24/2023 8:50 PM	CO ppm	Low	Span	9.04 ppm	9.17 ppm	0.13 ppm	0.13 ppm	±1 ppm	10 ppm	Unit online; Passed
			CC304196				B32022	CO, NO, NOX, BALN		
4/24/2023 8:50 PM	NH3/NOx ppm	High	Zero	0.0 ppm	0.0 ppm	0 ppm	0 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0155775				B72023	CO, O2, BALN		
4/24/2023 8:50 PM	NH3/NOx ppm	High	Span	178.4 ppm	174.7 ppm	-3.7 ppm	-3.7 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC1622				B32022	NO, NOX, BALN		
4/24/2023 8:50 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	0.13 ppm	0.13 ppm	0.13 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			EB0155775				B72023	CO, O2, BALN		
4/24/2023 8:50 PM	NH3/NOx ppm	Low	Span	9.05 ppm	9.03 ppm	-0.02 ppm	-0.02 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC304196				B32022	CO, NO, NOX, BALN		
4/24/2023 8:50 PM	O2 %	Single	Zero	0.00%	0.00%	0%	0%	±1%	25%	Unit online; Passed
			CC1622				B32022	NO, NOX, BALN		
4/24/2023 8:50 PM	O2 %	Single	Span	22.45%	22.30%	-0.15%	-0.15%	±1%	25%	Unit online; Passed
			EB0155775				B72023	CO, O2, BALN		
4/24/2023 8:50 PM	75-O2 %	Single	Zero	0.00%	0.00%	0%	0%	±1%	25%	Unit online; Passed
			CC1622				B32022	NO, NOX, BALN		
4/24/2023 8:50 PM	75-O2 %	Single	Span	22.45%	22.30%	-0.15%	-0.15%	±1%	25%	Unit online; Passed
			EB0155775				B72023	CO, O2, BALN		
4/26/2023 6:50 PM	NOx ppm	High	Zero	0.0 ppm	0.0 ppm	0 ppm	0 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0155775				B72023	CO, O2, BALN		
4/26/2023 6:50 PM	NOx ppm	High	Span	178.4 ppm	177.0 ppm	-1.4 ppm	-1.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC1622				B32022	NO, NOX, BALN		
4/26/2023 6:50 PM	NOx ppm	Low	Zero	0.00 ppm	0.03 ppm	0.03 ppm	0.03 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			EB0155775				B72023	CO, O2, BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	EPA Vendor ID	Allowable Drift	Instrument Span	Results
4/26/2023 6:50 PM	NOx ppm	Low	Span	9.05 ppm	12/29/2025	8.82 ppm	-0.23 ppm	B32022	±0.5 ppm	10 ppm	Unit online; Passed
			CC304196						CO, NO, NOX, BALN		
4/26/2023 6:50 PM	75-NOx ppm	High	Zero	0.0 ppm	1/23/2031	0.0 ppm	0 ppm	B72023	±10 ppm	200 ppm	Unit online; Passed
			EB0155775						CO, O2, BALN		
4/26/2023 6:50 PM	75-NOx ppm	High	Span	178.4 ppm	12/28/2030	177.0 ppm	-1.4 ppm	B32022	±10 ppm	200 ppm	Unit online; Passed
			CC1622						NO, NOX, BALN		
4/26/2023 6:50 PM	75-NOx ppm	Low	Zero	0.00 ppm	1/23/2031	0.03 ppm	0.03 ppm	B72023	±5 ppm	10 ppm	Unit online; Passed
			EB0155775						CO, O2, BALN		
4/26/2023 6:50 PM	75-NOx ppm	Low	Span	9.05 ppm	12/29/2025	8.82 ppm	-0.23 ppm	B32022	±5 ppm	10 ppm	Unit online; Passed
			CC304196						CO, NO, NOX, BALN		
4/26/2023 6:50 PM	CO ppm	High	Zero	0.0 ppm	12/28/2030	0.0 ppm	0 ppm	B32022	±20 ppm	200 ppm	Unit online; Passed
			CC1622						NO, NOX, BALN		
4/26/2023 6:50 PM	CO ppm	High	Span	182.8 ppm	1/23/2031	181.1 ppm	-1.7 ppm	B72023	±20 ppm	200 ppm	Unit online; Passed
			EB0155775						CO, O2, BALN		
4/26/2023 6:50 PM	CO ppm	Low	Zero	0.00 ppm	12/28/2030	0.00 ppm	0 ppm	B32022	±1 ppm	10 ppm	Unit online; Passed
			CC1622						NO, NOX, BALN		
4/26/2023 6:50 PM	CO ppm	Low	Span	9.04 ppm	12/29/2025	9.13 ppm	0.09 ppm	B32022	±1 ppm	10 ppm	Unit online; Passed
			CC304196						CO, NO, NOX, BALN		
4/26/2023 6:50 PM	NH3/NOx ppm	High	Zero	0.0 ppm	1/23/2031	0.0 ppm	0 ppm	B72023	±10 ppm	200 ppm	Unit online; Passed
			EB0155775						CO, O2, BALN		
4/26/2023 6:50 PM	NH3/NOx ppm	High	Span	178.4 ppm	12/28/2030	174.0 ppm	-4.4 ppm	B32022	±10 ppm	200 ppm	Unit online; Passed
			CC1622						NO, NOX, BALN		
4/26/2023 6:50 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	1/23/2031	0.12 ppm	0.12 ppm	B72023	±0.5 ppm	10 ppm	Unit online; Passed
			EB0155775						CO, O2, BALN		
4/26/2023 6:50 PM	NH3/NOx ppm	Low	Span	9.05 ppm	12/29/2025	8.97 ppm	-0.08 ppm	B32022	±0.5 ppm	10 ppm	Unit online; Passed
			CC304196						CO, NO, NOX, BALN		
4/26/2023 6:50 PM	O2 %	Single	Zero	0.00%	12/28/2030	0.00%	0%	B32022	±1%	25%	Unit online; Passed
			CC1622						NO, NOX, BALN		
4/26/2023 6:50 PM	O2 %	Single	Span	22.45%	1/23/2031	22.31%	-0.14%	B72023	±1%	25%	Unit online; Passed
			EB0155775						CO, O2, BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	EPA Vendor ID	Allowable Drift	Instrument Span	Results
4/26/2023 6:50 PM	75-O2 %	Single	Zero	0.00%	12/28/2030	0.00%	0%	B32022	±1% NO, NOX, BALN	25%	Unit online; Passed
4/26/2023 6:50 PM	75-O2 %	Single	Span	22.45%	1/23/2031	22.31%	-0.14%	B72023	±1% CO, O2, BALN	25%	Unit online; Passed
5/1/2023 3:35 AM	NOx ppm	High	Zero	0.0 ppm	1/23/2031	0.0 ppm	0 ppm	B72023	±10 ppm CO, O2, BALN	200 ppm	Unit online; Passed
5/1/2023 3:35 AM	NOx ppm	High	Span	177.8 ppm	12/28/2030	176.8 ppm	-1 ppm	B32022	±10 ppm NO, NOX, BALN	200 ppm	Unit online; Passed
5/1/2023 3:35 AM	NOx ppm	Low	Zero	0.00 ppm	1/23/2031	0.05 ppm	0.05 ppm	B72023	±0.5 ppm CO, O2, BALN	10 ppm	Unit online; Passed
5/1/2023 3:35 AM	NOx ppm	Low	Span	9.07 ppm	12/29/2025	8.85 ppm	-0.22 ppm	B32022	±0.5 ppm CO, NO, NOX, BALN	10 ppm	Unit online; Passed
5/1/2023 3:35 AM	75-NOx ppm	High	Zero	0.0 ppm	1/23/2031	0.0 ppm	0 ppm	B72023	±10 ppm CO, O2, BALN	200 ppm	Unit online; Passed
5/1/2023 3:35 AM	75-NOx ppm	High	Span	177.8 ppm	12/28/2030	176.8 ppm	-1 ppm	B32022	±10 ppm NO, NOX, BALN	200 ppm	Unit online; Passed
5/1/2023 3:35 AM	75-NOx ppm	Low	Zero	0.00 ppm	1/23/2031	0.05 ppm	0.05 ppm	B72023	±5 ppm CO, O2, BALN	10 ppm	Unit online; Passed
5/1/2023 3:35 AM	75-NOx ppm	Low	Span	9.07 ppm	12/29/2025	8.85 ppm	-0.22 ppm	B32022	±5 ppm CO, NO, NOX, BALN	10 ppm	Unit online; Passed
5/1/2023 3:35 AM	CO ppm	High	Zero	0.0 ppm	12/28/2030	0.0 ppm	0 ppm	B32022	±20 ppm NO, NOX, BALN	200 ppm	Unit online; Passed
5/1/2023 3:35 AM	CO ppm	High	Span	182.4 ppm	1/23/2031	180.6 ppm	-1.8 ppm	B72023	±20 ppm CO, O2, BALN	200 ppm	Unit online; Passed
5/1/2023 3:35 AM	CO ppm	Low	Zero	0.00 ppm	12/28/2030	0.00 ppm	0 ppm	B32022	±1 ppm NO, NOX, BALN	10 ppm	Unit online; Passed
5/1/2023 3:35 AM	CO ppm	Low	Span	9.05 ppm	12/29/2025	9.15 ppm	0.1 ppm	B32022	±1 ppm CO, NO, NOX, BALN	10 ppm	Unit online; Passed
5/1/2023 3:35 AM	NH3/NOx ppm	High	Zero	0.0 ppm	1/23/2031	0.0 ppm	0 ppm	B72023	±10 ppm CO, O2, BALN	200 ppm	Unit online; Passed

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	Allowable Drift	Instrument Span	Results
5/1/2023 3:35 AM	NH3/NOx ppm	High	Span	177.8 ppm	12/28/2030	175.6 ppm	-2.2 ppm	±10 ppm	200 ppm	Unit online; Passed
			AAL7534				B32022	NO, NOX, BALN		
5/1/2023 3:35 AM	NH3/NOx ppm	Low	Zero	0.00 ppm	1/23/2031	0.13 ppm	0.13 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			EB0157271				B72023	CO, O2, BALN		
5/1/2023 3:35 AM	NH3/NOx ppm	Low	Span	9.07 ppm	12/29/2025	9.13 ppm	0.06 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC260235				B32022	CO, NO, NOX, BALN		
5/1/2023 3:35 AM	O2 %	Single	Zero	0.00%	12/28/2030	0.00%	0%	±1%	25%	Unit online; Passed
			AAL7534				B32022	NO, NOX, BALN		
5/1/2023 3:35 AM	O2 %	Single	Span	22.59%	1/23/2031	22.47%	-0.12%	±1%	25%	Unit online; Passed
			EB0157271				B72023	CO, O2, BALN		
5/1/2023 3:35 AM	75-O2 %	Single	Zero	0.00%	12/28/2030	0.00%	0%	±1%	25%	Unit online; Passed
			AAL7534				B32022	NO, NOX, BALN		
5/1/2023 3:35 AM	75-O2 %	Single	Span	22.59%	1/23/2031	22.47%	-0.12%	±1%	25%	Unit online; Passed
			EB0157271				B72023	CO, O2, BALN		
5/24/2023 9:28 AM	NOx ppm	High	Zero	0.0 ppm	2/20/2031	-0.4 ppm	-0.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC245072				B72023	CO, O2, BALN		
5/24/2023 9:28 AM	NOx ppm	High	Span	180.7 ppm	1/6/2031	181.6 ppm	0.9 ppm	±10 ppm	200 ppm	Unit online; Passed
			SG9139995BAL				B32023	NO, NOX, BALN		
5/24/2023 9:28 AM	NOx ppm	Low	Zero	0.00 ppm	2/20/2031	0.06 ppm	0.06 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC245072				B72023	CO, O2, BALN		
5/24/2023 9:28 AM	NOx ppm	Low	Span	9.18 ppm	1/25/2026	9.19 ppm	0.01 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC476222				B32023	CO, NO, NOX, BALN		
5/24/2023 9:28 AM	75-NOx ppm	High	Zero	0.0 ppm	2/20/2031	-0.4 ppm	-0.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC245072				B72023	CO, O2, BALN		
5/24/2023 9:28 AM	75-NOx ppm	High	Span	180.7 ppm	1/6/2031	181.6 ppm	0.9 ppm	±10 ppm	200 ppm	Unit online; Passed
			SG9139995BAL				B32023	NO, NOX, BALN		
5/24/2023 9:28 AM	75-NOx ppm	Low	Zero	0.00 ppm	2/20/2031	0.06 ppm	0.06 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC245072				B72023	CO, O2, BALN		
5/24/2023 9:28 AM	75-NOx ppm	Low	Span	9.18 ppm	1/25/2026	9.19 ppm	0.01 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC476222				B32023	CO, NO, NOX, BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
		Cylinder ID				EPA Vendor ID	EPA Gas Type Codes			
5/24/2023 9:28 AM	CO ppm	High	Zero	0.0 ppm	-0.6 ppm	B32023	-0.6 ppm	±10 ppm	200 ppm	Unit online; Passed
			SG9139995BAL	1/6/2031		B32023		NO,NOX,BALN		
5/24/2023 9:28 AM	CO ppm	High	Span	180.4 ppm	173.4 ppm	B72023	-7 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC245072	2/20/2031		B72023		CO,O2,BALN		
5/24/2023 9:28 AM	CO ppm	Low	Zero	0.00 ppm	-0.13 ppm	B32023	-0.13 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			SG9139995BAL	1/6/2031		B32023		NO,NOX,BALN		
5/24/2023 9:28 AM	CO ppm	Low	Span	9.16 ppm	9.11 ppm	B32023	-0.05 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC476222	1/25/2026		B32023		CO,NO,NOX,BALN		
5/24/2023 9:28 AM	NH3/NOx ppm	High	Zero	0.0 ppm	-0.3 ppm	B72023	-0.3 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC245072	2/20/2031		B72023		CO,O2,BALN		
5/24/2023 9:28 AM	NH3/NOx ppm	High	Span	180.7 ppm	179.1 ppm	B32023	-1.6 ppm	±10 ppm	200 ppm	Unit online; Passed
			SG9139995BAL	1/6/2031		B32023		NO,NOX,BALN		
5/24/2023 9:28 AM	NH3/NOx ppm	Low	Zero	0.00 ppm	0.22 ppm	B72023	0.22 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC245072	2/20/2031		B72023		CO,O2,BALN		
5/24/2023 9:28 AM	NH3/NOx ppm	Low	Span	9.18 ppm	9.34 ppm	B32023	0.16 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC476222	1/25/2026		B32023		CO,NO,NOX,BALN		
5/24/2023 9:28 AM	O2 %	Single	Zero	0.00%	-0.05%	B32023	-0.05%	±1%	25%	Unit online; Passed
			SG9139995BAL	1/6/2031		B32023		NO,NOX,BALN		
5/24/2023 9:28 AM	O2 %	Single	Span	22.50%	22.50%	B72023	0%	±1%	25%	Unit online; Passed
			CC245072	2/20/2031		B72023		CO,O2,BALN		
5/24/2023 9:28 AM	75-O2 %	Single	Zero	0.00%	-0.05%	B32023	-0.05%	±1%	25%	Unit online; Passed
			SG9139995BAL	1/6/2031		B32023		NO,NOX,BALN		
5/24/2023 9:28 AM	75-O2 %	Single	Span	22.50%	22.50%	B72023	0%	±1%	25%	Unit online; Passed
			CC245072	2/20/2031		B72023		CO,O2,BALN		
5/25/2023 2:24 PM	NOx ppm	High	Zero	0.0 ppm	-0.5 ppm	B72023	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC245072	2/20/2031		B72023		CO,O2,BALN		
5/25/2023 2:24 PM	NOx ppm	High	Span	180.7 ppm	181.0 ppm	B32023	0.3 ppm	±10 ppm	200 ppm	Unit online; Passed
			SG9139995BAL	1/6/2031		B32023		NO,NOX,BALN		
5/25/2023 2:24 PM	NOx ppm	Low	Zero	0.00 ppm	-0.01 ppm	B72023	-0.01 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC245072	2/20/2031		B72023		CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
			Cylinder ID	Value	Value		EPA Vendor ID	EPA Gas Type	Codes	
5/25/2023 2:24 PM	NOx ppm	Low	Span	9.18 ppm	9.09 ppm	1/25/2026	-0.09 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC476222				B32023	CO,NO,NOX,BALN		
5/25/2023 2:24 PM	75-NOx ppm	High	Zero	0.0 ppm	-0.5 ppm	2/20/2031	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC245072				B72023	CO,O2,BALN		
5/25/2023 2:24 PM	75-NOx ppm	High	Span	180.7 ppm	181.0 ppm	1/6/2031	0.3 ppm	±10 ppm	200 ppm	Unit online; Passed
			SG9139995BAL				B32023	NO,NOX,BALN		
5/25/2023 2:24 PM	75-NOx ppm	Low	Zero	0.00 ppm	-0.01 ppm	2/20/2031	-0.01 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC245072				B72023	CO,O2,BALN		
5/25/2023 2:24 PM	75-NOx ppm	Low	Span	9.18 ppm	9.09 ppm	1/25/2026	-0.09 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC476222				B32023	CO,NO,NOX,BALN		
5/25/2023 2:24 PM	CO ppm	High	Zero	0.0 ppm	-0.2 ppm	1/6/2031	-0.2 ppm	±10 ppm	200 ppm	Unit online; Passed
			SG9139995BAL				B32023	NO,NOX,BALN		
5/25/2023 2:24 PM	CO ppm	High	Span	180.4 ppm	190.2 ppm	2/20/2031	9.8 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC245072				B72023	CO,O2,BALN		
5/25/2023 2:24 PM	CO ppm	Low	Zero	0.00 ppm	-0.01 ppm	1/6/2031	-0.01 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			SG9139995BAL				B32023	NO,NOX,BALN		
5/25/2023 2:24 PM	CO ppm	Low	Span	9.16 ppm	8.92 ppm	1/25/2026	-0.24 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC476222				B32023	CO,NO,NOX,BALN		
5/25/2023 2:24 PM	NH3/NOx ppm	High	Zero	0.0 ppm	-0.5 ppm	2/20/2031	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC245072				B72023	CO,O2,BALN		
5/25/2023 2:24 PM	NH3/NOx ppm	High	Span	180.7 ppm	178.5 ppm	1/6/2031	-2.2 ppm	±10 ppm	200 ppm	Unit online; Passed
			SG9139995BAL				B32023	NO,NOX,BALN		
5/25/2023 2:24 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	0.06 ppm	2/20/2031	0.06 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC245072				B72023	CO,O2,BALN		
5/25/2023 2:24 PM	NH3/NOx ppm	Low	Span	9.18 ppm	9.25 ppm	1/25/2026	0.07 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC476222				B32023	CO,NO,NOX,BALN		
5/25/2023 2:24 PM	O2 %	Single	Zero	0.00%	0.01%	1/6/2031	0.01%	±1%	25%	Unit online; Passed
			SG9139995BAL				B32023	NO,NOX,BALN		
5/25/2023 2:24 PM	O2 %	Single	Span	22.50%	22.53%	2/20/2031	0.03%	±1%	25%	Unit online; Passed
			CC245072				B72023	CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
		Cylinder ID				EPA Vendor ID	EPA Gas Type Codes			
5/25/2023 2:24 PM	75-O2 %	Single	Zero	0.00%	0.01%	B32023	±1%	NO,NOX,BALN	25%	Unit online; Passed
			SG9139995BAL	1/6/2031						
5/25/2023 2:24 PM	75-O2 %	Single	Span	22.50%	22.53%	B72023	±1%	CO,O2,BALN	25%	Unit online; Passed
			CC245072	2/20/2031						
5/26/2023 7:38 AM	NOx ppm	High	Zero	0.0 ppm	-0.5 ppm	B72023	±10 ppm	CO,O2,BALN	200 ppm	Unit online; Passed
			CC245072	2/20/2031						
5/26/2023 7:38 AM	NOx ppm	High	Span	180.7 ppm	180.7 ppm	B32023	±10 ppm	NO,NOX,BALN	200 ppm	Unit online; Passed
			SG9139995BAL	1/6/2031						
5/26/2023 7:38 AM	NOx ppm	Low	Zero	0.00 ppm	-0.01 ppm	B72023	±0.5 ppm	CO,O2,BALN	10 ppm	Unit online; Passed
			CC245072	2/20/2031						
5/26/2023 7:38 AM	NOx ppm	Low	Span	9.18 ppm	9.08 ppm	B32023	±0.5 ppm	CO,NO,NOX,BALN	10 ppm	Unit online; Passed
			CC476222	1/25/2026						
5/26/2023 7:38 AM	75-NOx ppm	High	Zero	0.0 ppm	-0.5 ppm	B72023	±10 ppm	CO,O2,BALN	200 ppm	Unit online; Passed
			CC245072	2/20/2031						
5/26/2023 7:38 AM	75-NOx ppm	High	Span	180.7 ppm	180.7 ppm	B32023	±10 ppm	NO,NOX,BALN	200 ppm	Unit online; Passed
			SG9139995BAL	1/6/2031						
5/26/2023 7:38 AM	75-NOx ppm	Low	Zero	0.00 ppm	-0.01 ppm	B72023	±5 ppm	CO,O2,BALN	10 ppm	Unit online; Passed
			CC245072	2/20/2031						
5/26/2023 7:38 AM	75-NOx ppm	Low	Span	9.18 ppm	9.08 ppm	B32023	±5 ppm	CO,NO,NOX,BALN	10 ppm	Unit online; Passed
			CC476222	1/25/2026						
5/26/2023 7:38 AM	CO ppm	High	Zero	0.0 ppm	-0.2 ppm	B32023	±10 ppm	NO,NOX,BALN	200 ppm	Unit online; Passed
			SG9139995BAL	1/6/2031						
5/26/2023 7:38 AM	CO ppm	High	Span	180.4 ppm	180.2 ppm	B72023	±10 ppm	CO,O2,BALN	200 ppm	Unit online; Passed
			CC245072	2/20/2031						
5/26/2023 7:38 AM	CO ppm	Low	Zero	0.00 ppm	0.00 ppm	B32023	±0.5 ppm	NO,NOX,BALN	10 ppm	Unit online; Passed
			SG9139995BAL	1/6/2031						
5/26/2023 7:38 AM	CO ppm	Low	Span	9.16 ppm	9.26 ppm	B32023	±0.5 ppm	CO,NO,NOX,BALN	10 ppm	Unit online; Passed
			CC476222	1/25/2026						
5/26/2023 7:38 AM	NH3/NOx ppm	High	Zero	0.0 ppm	-0.4 ppm	B72023	±10 ppm	CO,O2,BALN	200 ppm	Unit online; Passed
			CC245072	2/20/2031						

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	Allowable Drift	Instrument Span	Results
5/26/2023 7:38 AM	NH3/NOx ppm	High	Span 180.7 ppm	178.1 ppm	1/6/2031	-2.6 ppm	±10 ppm	200 ppm	Unit online; Passed	
			SG9139995BAL	B32023			NO,NOX,BALN			
5/26/2023 7:38 AM	NH3/NOx ppm	Low	Zero 0.00 ppm	0.14 ppm	2/20/2031	0.14 ppm	±0.5 ppm	10 ppm	Unit online; Passed	
			CC245072	B72023			CO,O2,BALN			
5/26/2023 7:38 AM	NH3/NOx ppm	Low	Span 9.18 ppm	9.27 ppm	1/25/2026	0.09 ppm	±0.5 ppm	10 ppm	Unit online; Passed	
			CC476222	B32023			CO,NO,NOX,BALN			
5/26/2023 7:38 AM	O2 %	Single	Zero 0.00%	-0.08%	1/6/2031	-0.08%	±1%	25%	Unit online; Passed	
			SG9139995BAL	B32023			NO,NOX,BALN			
5/26/2023 7:38 AM	O2 %	Single	Span 22.50%	22.49%	2/20/2031	-0.01%	±1%	25%	Unit online; Passed	
			CC245072	B72023			CO,O2,BALN			
5/26/2023 7:38 AM	75-O2 %	Single	Zero 0.00%	-0.08%	1/6/2031	-0.08%	±1%	25%	Unit online; Passed	
			SG9139995BAL	B32023			NO,NOX,BALN			
5/26/2023 7:38 AM	75-O2 %	Single	Span 22.50%	22.49%	2/20/2031	-0.01%	±1%	25%	Unit online; Passed	
			CC245072	B72023			CO,O2,BALN			
5/27/2023 7:36 AM	NOx ppm	High	Zero 0.0 ppm	-0.5 ppm	2/20/2031	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed	
			CC245072	B72023			CO,O2,BALN			
5/27/2023 7:36 AM	NOx ppm	High	Span 180.7 ppm	180.9 ppm	1/6/2031	0.2 ppm	±10 ppm	200 ppm	Unit online; Passed	
			SG9139995BAL	B32023			NO,NOX,BALN			
5/27/2023 7:36 AM	NOx ppm	Low	Zero 0.00 ppm	0.00 ppm	2/20/2031	0 ppm	±0.5 ppm	10 ppm	Unit online; Passed	
			CC245072	B72023			CO,O2,BALN			
5/27/2023 7:36 AM	NOx ppm	Low	Span 9.18 ppm	9.09 ppm	1/25/2026	-0.09 ppm	±0.5 ppm	10 ppm	Unit online; Passed	
			CC476222	B32023			CO,NO,NOX,BALN			
5/27/2023 7:36 AM	75-NOx ppm	High	Zero 0.0 ppm	-0.5 ppm	2/20/2031	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed	
			CC245072	B72023			CO,O2,BALN			
5/27/2023 7:36 AM	75-NOx ppm	High	Span 180.7 ppm	180.9 ppm	1/6/2031	0.2 ppm	±10 ppm	200 ppm	Unit online; Passed	
			SG9139995BAL	B32023			NO,NOX,BALN			
5/27/2023 7:36 AM	75-NOx ppm	Low	Zero 0.00 ppm	0.00 ppm	2/20/2031	0 ppm	±5 ppm	10 ppm	Unit online; Passed	
			CC245072	B72023			CO,O2,BALN			
5/27/2023 7:36 AM	75-NOx ppm	Low	Span 9.18 ppm	9.09 ppm	1/25/2026	-0.09 ppm	±5 ppm	10 ppm	Unit online; Passed	
			CC476222	B32023			CO,NO,NOX,BALN			

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	EPA Vendor ID	Allowable Drift	Instrument Span	Results
5/27/2023 7:36 AM	CO ppm	High	Zero	0.0 ppm	1/6/2031	-0.3 ppm	-0.3 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
5/27/2023 7:36 AM	CO ppm	High	Span	180.4 ppm	2/20/2031	177.9 ppm	-2.5 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
5/27/2023 7:36 AM	CO ppm	Low	Zero	0.00 ppm	1/6/2031	0.01 ppm	0.01 ppm	B32023	±0.5 ppm NO,NOX,BALN	10 ppm	Unit online; Passed
5/27/2023 7:36 AM	CO ppm	Low	Span	9.16 ppm	1/25/2026	9.72 ppm	0.56 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Fail Above; Fail OOC
5/27/2023 7:36 AM	NH3/NOx ppm	High	Zero	0.0 ppm	2/20/2031	-0.4 ppm	-0.4 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
5/27/2023 7:36 AM	NH3/NOx ppm	High	Span	180.7 ppm	1/6/2031	177.8 ppm	-2.9 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
5/27/2023 7:36 AM	NH3/NOx ppm	Low	Zero	0.00 ppm	2/20/2031	0.12 ppm	0.12 ppm	B72023	±0.5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
5/27/2023 7:36 AM	NH3/NOx ppm	Low	Span	9.18 ppm	1/25/2026	9.25 ppm	0.07 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
5/27/2023 7:36 AM	O2 %	Single	Zero	0.00%	1/6/2031	-0.06%	-0.06%	B32023	±1% NO,NOX,BALN	25%	Unit online; Passed
5/27/2023 7:36 AM	O2 %	Single	Span	22.50%	2/20/2031	22.50%	0%	B72023	±1% CO,O2,BALN	25%	Unit online; Passed
5/27/2023 7:36 AM	75-O2 %	Single	Zero	0.00%	1/6/2031	-0.06%	-0.06%	B32023	±1% NO,NOX,BALN	25%	Unit online; Passed
5/27/2023 7:36 AM	75-O2 %	Single	Span	22.50%	2/20/2031	22.50%	0%	B72023	±1% CO,O2,BALN	25%	Unit online; Passed
5/27/2023 8:19 AM	NOx ppm	High	Zero	0.0 ppm	2/20/2031	-0.5 ppm	-0.5 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
5/27/2023 8:19 AM	NOx ppm	High	Span	180.7 ppm	1/6/2031	181.0 ppm	0.3 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
5/27/2023 8:19 AM	NOx ppm	Low	Zero	0.00 ppm	2/20/2031	-0.01 ppm	-0.01 ppm	B72023	±0.5 ppm CO,O2,BALN	10 ppm	Unit online; Passed

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	EPA Vendor ID	Allowable Drift	Instrument Span	Results
5/27/2023 8:19 AM	NOx ppm	Low	Span	9.18 ppm	1/25/2026	9.08 ppm	-0.1 ppm	B32023	±0.5 ppm	10 ppm	Unit online; Passed
			CC476222						CO,NO,NOX,BALN		
5/27/2023 8:19 AM	75-NOx ppm	High	Zero	0.0 ppm	2/20/2031	-0.5 ppm	-0.5 ppm	B72023	±10 ppm	200 ppm	Unit online; Passed
			CC245072						CO,O2,BALN		
5/27/2023 8:19 AM	75-NOx ppm	High	Span	180.7 ppm	1/6/2031	181.0 ppm	0.3 ppm	B32023	±10 ppm	200 ppm	Unit online; Passed
			SG9139995BAL						NO,NOX,BALN		
5/27/2023 8:19 AM	75-NOx ppm	Low	Zero	0.00 ppm	2/20/2031	-0.01 ppm	-0.01 ppm	B72023	±5 ppm	10 ppm	Unit online; Passed
			CC245072						CO,O2,BALN		
5/27/2023 8:19 AM	75-NOx ppm	Low	Span	9.18 ppm	1/25/2026	9.08 ppm	-0.1 ppm	B32023	±5 ppm	10 ppm	Unit online; Passed
			CC476222						CO,NO,NOX,BALN		
5/27/2023 8:19 AM	CO ppm	High	Zero	0.0 ppm	1/6/2031	-0.3 ppm	-0.3 ppm	B32023	±10 ppm	200 ppm	Unit online; Passed
			SG9139995BAL						NO,NOX,BALN		
5/27/2023 8:19 AM	CO ppm	High	Span	180.4 ppm	2/20/2031	178.9 ppm	-1.5 ppm	B72023	±10 ppm	200 ppm	Unit online; Passed
			CC245072						CO,O2,BALN		
5/27/2023 8:19 AM	CO ppm	Low	Zero	0.00 ppm	1/6/2031	0.02 ppm	0.02 ppm	B32023	±0.5 ppm	10 ppm	Unit online; Passed
			SG9139995BAL						NO,NOX,BALN		
5/27/2023 8:19 AM	CO ppm	Low	Span	9.16 ppm	1/25/2026	9.15 ppm	-0.01 ppm	B32023	±0.5 ppm	10 ppm	Unit online; Passed
			CC476222						CO,NO,NOX,BALN		
5/27/2023 8:19 AM	NH3/NOx ppm	High	Zero	0.0 ppm	2/20/2031	-0.4 ppm	-0.4 ppm	B72023	±10 ppm	200 ppm	Unit online; Passed
			CC245072						CO,O2,BALN		
5/27/2023 8:19 AM	NH3/NOx ppm	High	Span	180.7 ppm	1/6/2031	178.3 ppm	-2.4 ppm	B32023	±10 ppm	200 ppm	Unit online; Passed
			SG9139995BAL						NO,NOX,BALN		
5/27/2023 8:19 AM	NH3/NOx ppm	Low	Zero	0.00 ppm	2/20/2031	0.09 ppm	0.09 ppm	B72023	±0.5 ppm	10 ppm	Unit online; Passed
			CC245072						CO,O2,BALN		
5/27/2023 8:19 AM	NH3/NOx ppm	Low	Span	9.18 ppm	1/25/2026	9.22 ppm	0.04 ppm	B32023	±0.5 ppm	10 ppm	Unit online; Passed
			CC476222						CO,NO,NOX,BALN		
5/27/2023 8:19 AM	O2 %	Single	Zero	0.00%	1/6/2031	-0.06%	-0.06%	B32023	±1%	25%	Unit online; Passed
			SG9139995BAL						NO,NOX,BALN		
5/27/2023 8:19 AM	O2 %	Single	Span	22.50%	2/20/2031	22.50%	0%	B72023	±1%	25%	Unit online; Passed
			CC245072						CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	Allowable Drift	Instrument Span	Results
			Cylinder ID				EPA Vendor ID	EPA Gas Type	Codes	
5/27/2023 8:19 AM	75-O2 %	Single	Zero	0.00%	1/6/2031	-0.06%	-0.06%	±1%	25%	Unit online; Passed
			SG9139995BAL				B32023	NO,NOX,BALN		
5/27/2023 8:19 AM	75-O2 %	Single	Span	22.50%	2/20/2031	22.50%	0%	±1%	25%	Unit online; Passed
			CC245072				B72023	CO,O2,BALN		
6/5/2023 6:06 PM	NOx ppm	High	Zero	0.0 ppm	2/20/2031	-0.4 ppm	-0.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC245072				B72023	CO,O2,BALN		
6/5/2023 6:06 PM	NOx ppm	High	Span	180.7 ppm	1/6/2031	181.8 ppm	1.1 ppm	±10 ppm	200 ppm	Unit online; Passed
			SG9139995BAL				B32023	NO,NOX,BALN		
6/5/2023 6:06 PM	NOx ppm	Low	Zero	0.00 ppm	2/20/2031	0.05 ppm	0.05 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC245072				B72023	CO,O2,BALN		
6/5/2023 6:06 PM	NOx ppm	Low	Span	9.18 ppm	1/25/2026	9.18 ppm	0 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC476222				B32023	CO,NO,NOX,BALN		
6/5/2023 6:06 PM	75-NOx ppm	High	Zero	0.0 ppm	2/20/2031	-0.4 ppm	-0.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC245072				B72023	CO,O2,BALN		
6/5/2023 6:06 PM	75-NOx ppm	High	Span	180.7 ppm	1/6/2031	181.8 ppm	1.1 ppm	±10 ppm	200 ppm	Unit online; Passed
			SG9139995BAL				B32023	NO,NOX,BALN		
6/5/2023 6:06 PM	75-NOx ppm	Low	Zero	0.00 ppm	2/20/2031	0.05 ppm	0.05 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC245072				B72023	CO,O2,BALN		
6/5/2023 6:06 PM	75-NOx ppm	Low	Span	9.18 ppm	1/25/2026	9.18 ppm	0 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC476222				B32023	CO,NO,NOX,BALN		
6/5/2023 6:06 PM	CO ppm	High	Zero	0.0 ppm	1/6/2031	-0.2 ppm	-0.2 ppm	±10 ppm	200 ppm	Unit online; Passed
			SG9139995BAL				B32023	NO,NOX,BALN		
6/5/2023 6:06 PM	CO ppm	High	Span	180.4 ppm	2/20/2031	181.5 ppm	1.1 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC245072				B72023	CO,O2,BALN		
6/5/2023 6:06 PM	CO ppm	Low	Zero	0.00 ppm	1/6/2031	0.15 ppm	0.15 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			SG9139995BAL				B32023	NO,NOX,BALN		
6/5/2023 6:06 PM	CO ppm	Low	Span	9.16 ppm	1/25/2026	9.37 ppm	0.21 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC476222				B32023	CO,NO,NOX,BALN		
6/5/2023 6:06 PM	NH3/NOx ppm	High	Zero	0.0 ppm	2/20/2031	-0.4 ppm	-0.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC245072				B72023	CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
		Cylinder ID					EPA Vendor ID	EPA Gas Type	Codes	
6/5/2023 6:06 PM	NH3/NOx ppm	High	Span	180.7 ppm	179.3 ppm	1/6/2031	-1.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			SG9139995BAL				B32023	NO,NOX,BALN		
6/5/2023 6:06 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	0.12 ppm	2/20/2031	0.12 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC245072				B72023	CO,O2,BALN		
6/5/2023 6:06 PM	NH3/NOx ppm	Low	Span	9.18 ppm	9.07 ppm	1/25/2026	-0.11 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC476222				B32023	CO,NO,NOX,BALN		
6/5/2023 6:06 PM	O2 %	Single	Zero	0.00%	0.00%	1/6/2031	0%	±1%	25%	Unit online; Passed
			SG9139995BAL				B32023	NO,NOX,BALN		
6/5/2023 6:06 PM	O2 %	Single	Span	22.50%	22.57%	2/20/2031	0.07%	±1%	25%	Unit online; Passed
			CC245072				B72023	CO,O2,BALN		
6/5/2023 6:06 PM	75-O2 %	Single	Zero	0.00%	0.00%	1/6/2031	0%	±1%	25%	Unit online; Passed
			SG9139995BAL				B32023	NO,NOX,BALN		
6/5/2023 6:06 PM	75-O2 %	Single	Span	22.50%	22.57%	2/20/2031	0.07%	±1%	25%	Unit online; Passed
			CC245072				B72023	CO,O2,BALN		
6/21/2023 7:36 AM	NOx ppm	High	Zero	0.0 ppm	-0.4 ppm	4/14/2031	-0.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		
6/21/2023 7:36 AM	NOx ppm	High	Span	177.7 ppm	176.8 ppm	3/14/2031	-0.9 ppm	±10 ppm	200 ppm	Unit online; Passed
			AAL18904				B32023	NO,NOX,BALN		
6/21/2023 7:36 AM	NOx ppm	Low	Zero	0.00 ppm	0.06 ppm	4/14/2031	0.06 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		
6/21/2023 7:36 AM	NOx ppm	Low	Span	9.11 ppm	9.17 ppm	3/15/2026	0.06 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC403919				B32023	CO,NO,NOX,BALN		
6/21/2023 7:36 AM	75-NOx ppm	High	Zero	0.0 ppm	-0.4 ppm	4/14/2031	-0.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		
6/21/2023 7:36 AM	75-NOx ppm	High	Span	177.7 ppm	176.8 ppm	3/14/2031	-0.9 ppm	±10 ppm	200 ppm	Unit online; Passed
			AAL18904				B32023	NO,NOX,BALN		
6/21/2023 7:36 AM	75-NOx ppm	Low	Zero	0.00 ppm	0.06 ppm	4/14/2031	0.06 ppm	±5 ppm	10 ppm	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		
6/21/2023 7:36 AM	75-NOx ppm	Low	Span	9.11 ppm	9.17 ppm	3/15/2026	0.06 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC403919				B32023	CO,NO,NOX,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	Allowable Drift	Instrument Span	Results
6/21/2023 7:36 AM	CO ppm	High	Zero	0.0 ppm	3/14/2031	-0.2 ppm	-0.2 ppm	±10 ppm	200 ppm	Unit online; Passed
			AAL18904			B32023		NO,NOX,BALN		
6/21/2023 7:36 AM	CO ppm	High	Span	180.7 ppm	4/14/2031	180.5 ppm	-0.2 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0083068			B72023		CO,O2,BALN		
6/21/2023 7:36 AM	CO ppm	Low	Zero	0.00 ppm	3/14/2031	0.10 ppm	0.1 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			AAL18904			B32023		NO,NOX,BALN		
6/21/2023 7:36 AM	CO ppm	Low	Span	9.16 ppm	3/15/2026	9.05 ppm	-0.11 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC403919			B32023		CO,NO,NOX,BALN		
6/21/2023 7:36 AM	NH3/NOx ppm	High	Zero	0.0 ppm	4/14/2031	-0.3 ppm	-0.3 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0083068			B72023		CO,O2,BALN		
6/21/2023 7:36 AM	NH3/NOx ppm	High	Span	177.7 ppm	3/14/2031	176.0 ppm	-1.7 ppm	±10 ppm	200 ppm	Unit online; Passed
			AAL18904			B32023		NO,NOX,BALN		
6/21/2023 7:36 AM	NH3/NOx ppm	Low	Zero	0.00 ppm	4/14/2031	0.16 ppm	0.16 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			EB0083068			B72023		CO,O2,BALN		
6/21/2023 7:36 AM	NH3/NOx ppm	Low	Span	9.11 ppm	3/15/2026	9.26 ppm	0.15 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC403919			B32023		CO,NO,NOX,BALN		
6/21/2023 7:36 AM	O2 %	Single	Zero	0.00%	3/14/2031	-0.01%	-0.01%	±1%	25%	Unit online; Passed
			AAL18904			B32023		NO,NOX,BALN		
6/21/2023 7:36 AM	O2 %	Single	Span	22.51%	4/14/2031	22.55%	0.04%	±1%	25%	Unit online; Passed
			EB0083068			B72023		CO,O2,BALN		
6/21/2023 7:36 AM	75-O2 %	Single	Zero	0.00%	3/14/2031	-0.01%	-0.01%	±1%	25%	Unit online; Passed
			AAL18904			B32023		NO,NOX,BALN		
6/21/2023 7:36 AM	75-O2 %	Single	Span	22.51%	4/14/2031	22.55%	0.04%	±1%	25%	Unit online; Passed
			EB0083068			B72023		CO,O2,BALN		
6/24/2023 8:06 PM	NOx ppm	High	Zero	0.0 ppm	4/14/2031	-0.5 ppm	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0083068			B72023		CO,O2,BALN		
6/24/2023 8:06 PM	NOx ppm	High	Span	177.7 ppm	3/14/2031	176.3 ppm	-1.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			AAL18904			B32023		NO,NOX,BALN		
6/24/2023 8:06 PM	NOx ppm	Low	Zero	0.00 ppm	4/14/2031	0.02 ppm	0.02 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			EB0083068			B72023		CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
		Cylinder ID				EPA Vendor ID	EPA Gas Type Codes			
6/24/2023 8:06 PM	NOx ppm	Low	Span	9.11 ppm	9.09 ppm	3/15/2026	-0.02 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC403919			B32023		CO,NO,NOX,BALN		
6/24/2023 8:06 PM	75-NOx ppm	High	Zero	0.0 ppm	-0.5 ppm	4/14/2031	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0083068			B72023		CO,O2,BALN		
6/24/2023 8:06 PM	75-NOx ppm	High	Span	177.7 ppm	176.3 ppm	3/14/2031	-1.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			AAL18904			B32023		NO,NOX,BALN		
6/24/2023 8:06 PM	75-NOx ppm	Low	Zero	0.00 ppm	0.02 ppm	4/14/2031	0.02 ppm	±5 ppm	10 ppm	Unit online; Passed
			EB0083068			B72023		CO,O2,BALN		
6/24/2023 8:06 PM	75-NOx ppm	Low	Span	9.11 ppm	9.09 ppm	3/15/2026	-0.02 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC403919			B32023		CO,NO,NOX,BALN		
6/24/2023 8:06 PM	CO ppm	High	Zero	0.0 ppm	-0.3 ppm	3/14/2031	-0.3 ppm	±10 ppm	200 ppm	Unit online; Passed
			AAL18904			B32023		NO,NOX,BALN		
6/24/2023 8:06 PM	CO ppm	High	Span	180.7 ppm	179.8 ppm	4/14/2031	-0.9 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0083068			B72023		CO,O2,BALN		
6/24/2023 8:06 PM	CO ppm	Low	Zero	0.00 ppm	-0.01 ppm	3/14/2031	-0.01 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			AAL18904			B32023		NO,NOX,BALN		
6/24/2023 8:06 PM	CO ppm	Low	Span	9.16 ppm	9.07 ppm	3/15/2026	-0.09 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC403919			B32023		CO,NO,NOX,BALN		
6/24/2023 8:06 PM	NH3/NOx ppm	High	Zero	0.0 ppm	-0.4 ppm	4/14/2031	-0.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0083068			B72023		CO,O2,BALN		
6/24/2023 8:06 PM	NH3/NOx ppm	High	Span	177.7 ppm	175.2 ppm	3/14/2031	-2.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			AAL18904			B32023		NO,NOX,BALN		
6/24/2023 8:06 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	0.13 ppm	4/14/2031	0.13 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			EB0083068			B72023		CO,O2,BALN		
6/24/2023 8:06 PM	NH3/NOx ppm	Low	Span	9.11 ppm	9.18 ppm	3/15/2026	0.07 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC403919			B32023		CO,NO,NOX,BALN		
6/24/2023 8:06 PM	O2 %	Single	Zero	0.00%	0.00%	3/14/2031	0%	±1%	25%	Unit online; Passed
			AAL18904			B32023		NO,NOX,BALN		
6/24/2023 8:06 PM	O2 %	Single	Span	22.51%	22.55%	4/14/2031	0.04%	±1%	25%	Unit online; Passed
			EB0083068			B72023		CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	EPA Vendor ID	Allowable Drift	Instrument Span	Results
6/24/2023 8:06 PM	75-O2 %	Single	Zero	0.00%	3/14/2031	0.00%	0%	B32023	±1%	25%	Unit online; Passed
			AAL18904						NO,NOX,BALN		
6/24/2023 8:06 PM	75-O2 %	Single	Span	22.51%	4/14/2031	22.55%	0.04%	B72023	±1%	25%	Unit online; Passed
			EB0083068						CO,O2,BALN		
7/10/2023 6:35 PM	NOx ppm	High	Zero	0.0 ppm	4/14/2031	-0.4 ppm	-0.4 ppm	B72023	±10 ppm	200 ppm	Unit online; Passed
			EB0083068						CO,O2,BALN		
7/10/2023 6:35 PM	NOx ppm	High	Span	177.7 ppm	3/14/2031	177.4 ppm	-0.3 ppm	B32023	±10 ppm	200 ppm	Unit online; Passed
			AAL18904						NO,NOX,BALN		
7/10/2023 6:35 PM	NOx ppm	Low	Zero	0.00 ppm	4/14/2031	0.04 ppm	0.04 ppm	B72023	±0.5 ppm	10 ppm	Unit online; Passed
			EB0083068						CO,O2,BALN		
7/10/2023 6:35 PM	NOx ppm	Low	Span	9.11 ppm	3/15/2026	9.15 ppm	0.04 ppm	B32023	±0.5 ppm	10 ppm	Unit online; Passed
			CC403919						CO,NO,NOX,BALN		
7/10/2023 6:35 PM	75-NOx ppm	High	Zero	0.0 ppm	4/14/2031	-0.4 ppm	-0.4 ppm	B72023	±10 ppm	200 ppm	Unit online; Passed
			EB0083068						CO,O2,BALN		
7/10/2023 6:35 PM	75-NOx ppm	High	Span	177.7 ppm	3/14/2031	177.4 ppm	-0.3 ppm	B32023	±10 ppm	200 ppm	Unit online; Passed
			AAL18904						NO,NOX,BALN		
7/10/2023 6:35 PM	75-NOx ppm	Low	Zero	0.00 ppm	4/14/2031	0.04 ppm	0.04 ppm	B72023	±5 ppm	10 ppm	Unit online; Passed
			EB0083068						CO,O2,BALN		
7/10/2023 6:35 PM	75-NOx ppm	Low	Span	9.11 ppm	3/15/2026	9.15 ppm	0.04 ppm	B32023	±5 ppm	10 ppm	Unit online; Passed
			CC403919						CO,NO,NOX,BALN		
7/10/2023 6:35 PM	CO ppm	High	Zero	0.0 ppm	3/14/2031	-0.4 ppm	-0.4 ppm	B32023	±10 ppm	200 ppm	Unit online; Passed
			AAL18904						NO,NOX,BALN		
7/10/2023 6:35 PM	CO ppm	High	Span	180.7 ppm	4/14/2031	181.8 ppm	1.1 ppm	B72023	±10 ppm	200 ppm	Unit online; Passed
			EB0083068						CO,O2,BALN		
7/10/2023 6:35 PM	CO ppm	Low	Zero	0.00 ppm	3/14/2031	-0.10 ppm	-0.1 ppm	B32023	±0.5 ppm	10 ppm	Unit online; Passed
			AAL18904						NO,NOX,BALN		
7/10/2023 6:35 PM	CO ppm	Low	Span	9.16 ppm	3/15/2026	9.15 ppm	-0.01 ppm	B32023	±0.5 ppm	10 ppm	Unit online; Passed
			CC403919						CO,NO,NOX,BALN		
7/10/2023 6:35 PM	NH3/NOx ppm	High	Zero	0.0 ppm	4/14/2031	-0.4 ppm	-0.4 ppm	B72023	±10 ppm	200 ppm	Unit online; Passed
			EB0083068						CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
		Cylinder ID				EPA Vendor ID	EPA Gas Type Codes			
7/10/2023 6:35 PM	NH3/NOx ppm	High	Span	177.7 ppm	176.5 ppm	3/14/2031	-1.2 ppm	±10 ppm	200 ppm	Unit online; Passed
			AAL18904			B32023		NO,NOX,BALN		
7/10/2023 6:35 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	0.19 ppm	4/14/2031	0.19 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			EB0083068			B72023		CO,O2,BALN		
7/10/2023 6:35 PM	NH3/NOx ppm	Low	Span	9.11 ppm	9.29 ppm	3/15/2026	0.18 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC403919			B32023		CO,NO,NOX,BALN		
7/10/2023 6:35 PM	O2 %	Single	Zero	0.00%	-0.02%	3/14/2031	-0.02%	±1%	25%	Unit online; Passed
			AAL18904			B32023		NO,NOX,BALN		
7/10/2023 6:35 PM	O2 %	Single	Span	22.51%	22.54%	4/14/2031	0.03%	±1%	25%	Unit online; Passed
			EB0083068			B72023		CO,O2,BALN		
7/10/2023 6:35 PM	75-O2 %	Single	Zero	0.00%	-0.02%	3/14/2031	-0.02%	±1%	25%	Unit online; Passed
			AAL18904			B32023		NO,NOX,BALN		
7/10/2023 6:35 PM	75-O2 %	Single	Span	22.51%	22.54%	4/14/2031	0.03%	±1%	25%	Unit online; Passed
			EB0083068			B72023		CO,O2,BALN		
7/11/2023 7:20 PM	NOx ppm	High	Zero	0.0 ppm	-0.5 ppm	4/14/2031	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0083068			B72023		CO,O2,BALN		
7/11/2023 7:20 PM	NOx ppm	High	Span	177.7 ppm	176.5 ppm	3/14/2031	-1.2 ppm	±10 ppm	200 ppm	Unit online; Passed
			AAL18904			B32023		NO,NOX,BALN		
7/11/2023 7:20 PM	NOx ppm	Low	Zero	0.00 ppm	0.01 ppm	4/14/2031	0.01 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			EB0083068			B72023		CO,O2,BALN		
7/11/2023 7:20 PM	NOx ppm	Low	Span	9.11 ppm	9.08 ppm	3/15/2026	-0.03 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC403919			B32023		CO,NO,NOX,BALN		
7/11/2023 7:20 PM	75-NOx ppm	High	Zero	0.0 ppm	-0.5 ppm	4/14/2031	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0083068			B72023		CO,O2,BALN		
7/11/2023 7:20 PM	75-NOx ppm	High	Span	177.7 ppm	176.5 ppm	3/14/2031	-1.2 ppm	±10 ppm	200 ppm	Unit online; Passed
			AAL18904			B32023		NO,NOX,BALN		
7/11/2023 7:20 PM	75-NOx ppm	Low	Zero	0.00 ppm	0.01 ppm	4/14/2031	0.01 ppm	±5 ppm	10 ppm	Unit online; Passed
			EB0083068			B72023		CO,O2,BALN		
7/11/2023 7:20 PM	75-NOx ppm	Low	Span	9.11 ppm	9.08 ppm	3/15/2026	-0.03 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC403919			B32023		CO,NO,NOX,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
			Cylinder ID				EPA Vendor ID	EPA Gas Type Codes		
7/11/2023 7:20 PM	CO ppm	High	Zero	0.0 ppm	-0.5 ppm	3/14/2031	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			AAL18904				B32023	NO,NOX,BALN		
7/11/2023 7:20 PM	CO ppm	High	Span	180.7 ppm	180.5 ppm	4/14/2031	-0.2 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		
7/11/2023 7:20 PM	CO ppm	Low	Zero	0.00 ppm	-0.19 ppm	3/14/2031	-0.19 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			AAL18904				B32023	NO,NOX,BALN		
7/11/2023 7:20 PM	CO ppm	Low	Span	9.16 ppm	9.32 ppm	3/15/2026	0.16 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC403919				B32023	CO,NO,NOX,BALN		
7/11/2023 7:20 PM	NH3/NOx ppm	High	Zero	0.0 ppm	-0.4 ppm	4/14/2031	-0.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		
7/11/2023 7:20 PM	NH3/NOx ppm	High	Span	177.7 ppm	175.6 ppm	3/14/2031	-2.1 ppm	±10 ppm	200 ppm	Unit online; Passed
			AAL18904				B32023	NO,NOX,BALN		
7/11/2023 7:20 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	0.13 ppm	4/14/2031	0.13 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		
7/11/2023 7:20 PM	NH3/NOx ppm	Low	Span	9.11 ppm	9.23 ppm	3/15/2026	0.12 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC403919				B32023	CO,NO,NOX,BALN		
7/11/2023 7:20 PM	O2 %	Single	Zero	0.00%	-0.02%	3/14/2031	-0.02%	±1%	25%	Unit online; Passed
			AAL18904				B32023	NO,NOX,BALN		
7/11/2023 7:20 PM	O2 %	Single	Span	22.51%	22.53%	4/14/2031	0.02%	±1%	25%	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		
7/11/2023 7:20 PM	75-O2 %	Single	Zero	0.00%	-0.02%	3/14/2031	-0.02%	±1%	25%	Unit online; Passed
			AAL18904				B32023	NO,NOX,BALN		
7/11/2023 7:20 PM	75-O2 %	Single	Span	22.51%	22.53%	4/14/2031	0.02%	±1%	25%	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		
7/11/2023 6:05 PM	NOx ppm	High	Zero	0.0 ppm	-0.5 ppm	4/14/2031	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		
7/11/2023 6:05 PM	NOx ppm	High	Span	177.7 ppm	176.9 ppm	3/14/2031	-0.8 ppm	±10 ppm	200 ppm	Unit online; Passed
			AAL18904				B32023	NO,NOX,BALN		
7/11/2023 6:05 PM	NOx ppm	Low	Zero	0.00 ppm	0.01 ppm	4/14/2031	0.01 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
		Cylinder ID		Value	Value		EPA Vendor ID	EPA Gas Type Codes		
7/13/2023 6:05 PM	NOx ppm	Low	Span	9.11 ppm	9.10 ppm	3/15/2026	-0.01 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC403919				B32023	CO,NO,NOX,BALN		
7/13/2023 6:05 PM	75-NOx ppm	High	Zero	0.0 ppm	-0.5 ppm	4/14/2031	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		
7/13/2023 6:05 PM	75-NOx ppm	High	Span	177.7 ppm	176.9 ppm	3/14/2031	-0.8 ppm	±10 ppm	200 ppm	Unit online; Passed
			AAL18904				B32023	NO,NOX,BALN		
7/13/2023 6:05 PM	75-NOx ppm	Low	Zero	0.00 ppm	0.01 ppm	4/14/2031	0.01 ppm	±5 ppm	10 ppm	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		
7/13/2023 6:05 PM	75-NOx ppm	Low	Span	9.11 ppm	9.10 ppm	3/15/2026	-0.01 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC403919				B32023	CO,NO,NOX,BALN		
7/13/2023 6:05 PM	CO ppm	High	Zero	0.0 ppm	-0.3 ppm	3/14/2031	-0.3 ppm	±10 ppm	200 ppm	Unit online; Passed
			AAL18904				B32023	NO,NOX,BALN		
7/13/2023 6:05 PM	CO ppm	High	Span	180.7 ppm	181.0 ppm	4/14/2031	0.3 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		
7/13/2023 6:05 PM	CO ppm	Low	Zero	0.00 ppm	-0.07 ppm	3/14/2031	-0.07 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			AAL18904				B32023	NO,NOX,BALN		
7/13/2023 6:05 PM	CO ppm	Low	Span	9.16 ppm	9.18 ppm	3/15/2026	0.02 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC403919				B32023	CO,NO,NOX,BALN		
7/13/2023 6:05 PM	NH3/NOx ppm	High	Zero	0.0 ppm	-0.4 ppm	4/14/2031	-0.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		
7/13/2023 6:05 PM	NH3/NOx ppm	High	Span	177.7 ppm	175.2 ppm	3/14/2031	-2.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			AAL18904				B32023	NO,NOX,BALN		
7/13/2023 6:05 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	0.14 ppm	4/14/2031	0.14 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		
7/13/2023 6:05 PM	NH3/NOx ppm	Low	Span	9.11 ppm	9.25 ppm	3/15/2026	0.14 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC403919				B32023	CO,NO,NOX,BALN		
7/13/2023 6:05 PM	O2 %	Single	Zero	0.00%	-0.02%	3/14/2031	-0.02%	±1%	25%	Unit online; Passed
			AAL18904				B32023	NO,NOX,BALN		
7/13/2023 6:05 PM	O2 %	Single	Span	22.51%	22.54%	4/14/2031	0.03%	±1%	25%	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
		Cylinder ID				EPA Vendor ID	EPA Gas Type Codes			
7/13/2023 6:05 PM	75-O2 %	Single	Zero	0.00%	-0.02%	B32023	-0.02%	±1%	25%	Unit online; Passed
			AAL18904	3/14/2031		B32023		NO,NOX,BALN		
7/13/2023 6:05 PM	75-O2 %	Single	Span	22.51%	22.54%	B72023	0.03%	±1%	25%	Unit online; Passed
			EB0083068	4/14/2031		B72023		CO,O2,BALN		
7/14/2023 5:36 PM	NOx ppm	High	Zero	0.0 ppm	-0.5 ppm	B72023	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0083068	4/14/2031		B72023		CO,O2,BALN		
7/14/2023 5:36 PM	NOx ppm	High	Span	177.7 ppm	177.2 ppm	B32023	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			AAL18904	3/14/2031		B32023		NO,NOX,BALN		
7/14/2023 5:36 PM	NOx ppm	Low	Zero	0.00 ppm	0.00 ppm	B72023	0 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			EB0083068	4/14/2031		B72023		CO,O2,BALN		
7/14/2023 5:36 PM	NOx ppm	Low	Span	9.11 ppm	9.07 ppm	B32023	-0.04 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC403919	3/15/2026		B32023		CO,NO,NOX,BALN		
7/14/2023 5:36 PM	75-NOx ppm	High	Zero	0.0 ppm	-0.5 ppm	B72023	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0083068	4/14/2031		B72023		CO,O2,BALN		
7/14/2023 5:36 PM	75-NOx ppm	High	Span	177.7 ppm	177.2 ppm	B32023	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			AAL18904	3/14/2031		B32023		NO,NOX,BALN		
7/14/2023 5:36 PM	75-NOx ppm	Low	Zero	0.00 ppm	0.00 ppm	B72023	0 ppm	±5 ppm	10 ppm	Unit online; Passed
			EB0083068	4/14/2031		B72023		CO,O2,BALN		
7/14/2023 5:36 PM	75-NOx ppm	Low	Span	9.11 ppm	9.07 ppm	B32023	-0.04 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC403919	3/15/2026		B32023		CO,NO,NOX,BALN		
7/14/2023 5:36 PM	CO ppm	High	Zero	0.0 ppm	-0.4 ppm	B32023	-0.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			AAL18904	3/14/2031		B32023		NO,NOX,BALN		
7/14/2023 5:36 PM	CO ppm	High	Span	180.7 ppm	180.8 ppm	B72023	0.1 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0083068	4/14/2031		B72023		CO,O2,BALN		
7/14/2023 5:36 PM	CO ppm	Low	Zero	0.00 ppm	-0.11 ppm	B32023	-0.11 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			AAL18904	3/14/2031		B32023		NO,NOX,BALN		
7/14/2023 5:36 PM	CO ppm	Low	Span	9.16 ppm	9.18 ppm	B32023	0.02 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC403919	3/15/2026		B32023		CO,NO,NOX,BALN		
7/14/2023 5:36 PM	NH3/NOx ppm	High	Zero	0.0 ppm	-0.4 ppm	B72023	-0.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0083068	4/14/2031		B72023		CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
		Cylinder ID				EPA Vendor ID	EPA Gas Type Codes			
7/14/2023 5:36 PM	NH3/NOx ppm	High	Span	177.7 ppm	175.5 ppm	3/14/2031	-2.2 ppm	±10 ppm	200 ppm	Unit online; Passed
		AAL18904				B32023		NO,NOX,BALN		
7/14/2023 5:36 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	0.14 ppm	4/14/2031	0.14 ppm	±0.5 ppm	10 ppm	Unit online; Passed
		EB0083068				B72023		CO,O2,BALN		
7/14/2023 5:36 PM	NH3/NOx ppm	Low	Span	9.11 ppm	9.23 ppm	3/15/2026	0.12 ppm	±0.5 ppm	10 ppm	Unit online; Passed
		CC403919				B32023		CO,NO,NOX,BALN		
7/14/2023 5:36 PM	O2 %	Single	Zero	0.00%	-0.02%	3/14/2031	-0.02%	±1%	25%	Unit online; Passed
		AAL18904				B32023		NO,NOX,BALN		
7/14/2023 5:36 PM	O2 %	Single	Span	22.51%	22.55%	4/14/2031	0.04%	±1%	25%	Unit online; Passed
		EB0083068				B72023		CO,O2,BALN		
7/14/2023 5:36 PM	75-O2 %	Single	Zero	0.00%	-0.02%	3/14/2031	-0.02%	±1%	25%	Unit online; Passed
		AAL18904				B32023		NO,NOX,BALN		
7/14/2023 5:36 PM	75-O2 %	Single	Span	22.51%	22.55%	4/14/2031	0.04%	±1%	25%	Unit online; Passed
		EB0083068				B72023		CO,O2,BALN		
7/18/2023 5:51 PM	NOx ppm	High	Zero	0.0 ppm	-0.5 ppm	4/14/2031	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
		EB0083068				B72023		CO,O2,BALN		
7/18/2023 5:51 PM	NOx ppm	High	Span	177.7 ppm	176.8 ppm	3/14/2031	-0.9 ppm	±10 ppm	200 ppm	Unit online; Passed
		AAL18904				B32023		NO,NOX,BALN		
7/18/2023 5:51 PM	NOx ppm	Low	Zero	0.00 ppm	0.01 ppm	4/14/2031	0.01 ppm	±0.5 ppm	10 ppm	Unit online; Passed
		EB0083068				B72023		CO,O2,BALN		
7/18/2023 5:51 PM	NOx ppm	Low	Span	9.11 ppm	9.08 ppm	3/15/2026	-0.03 ppm	±0.5 ppm	10 ppm	Unit online; Passed
		CC403919				B32023		CO,NO,NOX,BALN		
7/18/2023 5:51 PM	75-NOx ppm	High	Zero	0.0 ppm	-0.5 ppm	4/14/2031	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
		EB0083068				B72023		CO,O2,BALN		
7/18/2023 5:51 PM	75-NOx ppm	High	Span	177.7 ppm	176.8 ppm	3/14/2031	-0.9 ppm	±10 ppm	200 ppm	Unit online; Passed
		AAL18904				B32023		NO,NOX,BALN		
7/18/2023 5:51 PM	75-NOx ppm	Low	Zero	0.00 ppm	0.01 ppm	4/14/2031	0.01 ppm	±5 ppm	10 ppm	Unit online; Passed
		EB0083068				B72023		CO,O2,BALN		
7/18/2023 5:51 PM	75-NOx ppm	Low	Span	9.11 ppm	9.08 ppm	3/15/2026	-0.03 ppm	±5 ppm	10 ppm	Unit online; Passed
		CC403919				B32023		CO,NO,NOX,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	EPA Vendor ID	Allowable Drift	Instrument Span	Results
7/13/2023 5:51 PM	CO ppm	High	Zero	0.0 ppm	3/14/2031	-0.3 ppm	-0.3 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
7/13/2023 5:51 PM	CO ppm	High	Span	180.7 ppm	4/14/2031	181.8 ppm	1.1 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
7/13/2023 5:51 PM	CO ppm	Low	Zero	0.00 ppm	3/14/2031	-0.02 ppm	-0.02 ppm	B32023	±0.5 ppm NO,NOX,BALN	10 ppm	Unit online; Passed
7/13/2023 5:51 PM	CO ppm	Low	Span	9.16 ppm	3/15/2026	9.18 ppm	0.02 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
7/18/2023 5:51 PM	NH3/NOx ppm	High	Zero	0.0 ppm	4/14/2031	-0.3 ppm	-0.3 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
7/18/2023 5:51 PM	NH3/NOx ppm	High	Span	177.7 ppm	3/14/2031	177.5 ppm	-0.2 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
7/18/2023 5:51 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	4/14/2031	0.21 ppm	0.21 ppm	B72023	±0.5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
7/18/2023 5:51 PM	NH3/NOx ppm	Low	Span	9.11 ppm	3/15/2026	9.24 ppm	0.13 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
7/18/2023 5:51 PM	O2 %	Single	Zero	0.00%	3/14/2031	0.00%	0%	B32023	±1% NO,NOX,BALN	25%	Unit online; Passed
7/18/2023 5:51 PM	O2 %	Single	Span	22.51%	4/14/2031	22.55%	0.04%	B72023	±1% CO,O2,BALN	25%	Unit online; Passed
7/18/2023 5:51 PM	75-O2 %	Single	Zero	0.00%	3/14/2031	0.00%	0%	B32023	±1% NO,NOX,BALN	25%	Unit online; Passed
7/18/2023 5:51 PM	75-O2 %	Single	Span	22.51%	4/14/2031	22.55%	0.04%	B72023	±1% CO,O2,BALN	25%	Unit online; Passed
7/19/2023 5:36 PM	NOx ppm	High	Zero	0.0 ppm	4/14/2031	-0.5 ppm	-0.5 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
7/19/2023 5:36 PM	NOx ppm	High	Span	177.7 ppm	3/14/2031	176.6 ppm	-1.1 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
7/19/2023 5:36 PM	NOx ppm	Low	Zero	0.00 ppm	4/14/2031	0.00 ppm	0 ppm	B72023	±0.5 ppm CO,O2,BALN	10 ppm	Unit online; Passed

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	Allowable Drift	Instrument Span	Results
7/19/2023 5:36 PM	NOx ppm	Low	Span 9.11 ppm	9.04 ppm	3/15/2026	9.04 ppm	-0.07 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC403919				B32023	CO,NO,NOX,BALN		
7/19/2023 5:36 PM	75-NOx ppm	High	Zero 0.0 ppm	-0.5 ppm	4/14/2031	-0.5 ppm	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		
7/19/2023 5:36 PM	75-NOx ppm	High	Span 177.7 ppm	176.6 ppm	3/14/2031	176.6 ppm	-1.1 ppm	±10 ppm	200 ppm	Unit online; Passed
			AAL18904				B32023	NO,NOX,BALN		
7/19/2023 5:36 PM	75-NOx ppm	Low	Zero 0.00 ppm	0.00 ppm	4/14/2031	0.00 ppm	0 ppm	±5 ppm	10 ppm	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		
7/19/2023 5:36 PM	75-NOx ppm	Low	Span 9.11 ppm	9.04 ppm	3/15/2026	9.04 ppm	-0.07 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC403919				B32023	CO,NO,NOX,BALN		
7/19/2023 5:36 PM	CO ppm	High	Zero 0.0 ppm	-0.3 ppm	3/14/2031	-0.3 ppm	-0.3 ppm	±10 ppm	200 ppm	Unit online; Passed
			AAL18904				B32023	NO,NOX,BALN		
7/19/2023 5:36 PM	CO ppm	High	Span 180.7 ppm	181.8 ppm	4/14/2031	181.8 ppm	1.1 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		
7/19/2023 5:36 PM	CO ppm	Low	Zero 0.00 ppm	-0.03 ppm	3/14/2031	-0.03 ppm	-0.03 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			AAL18904				B32023	NO,NOX,BALN		
7/19/2023 5:36 PM	CO ppm	Low	Span 9.16 ppm	9.12 ppm	3/15/2026	9.12 ppm	-0.04 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC403919				B32023	CO,NO,NOX,BALN		
7/19/2023 5:36 PM	NH3/NOx ppm	High	Zero 0.0 ppm	-0.3 ppm	4/14/2031	-0.3 ppm	-0.3 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		
7/19/2023 5:36 PM	NH3/NOx ppm	High	Span 177.7 ppm	177.7 ppm	3/14/2031	177.7 ppm	0 ppm	±10 ppm	200 ppm	Unit online; Passed
			AAL18904				B32023	NO,NOX,BALN		
7/19/2023 5:36 PM	NH3/NOx ppm	Low	Zero 0.00 ppm	0.15 ppm	4/14/2031	0.15 ppm	0.15 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		
7/19/2023 5:36 PM	NH3/NOx ppm	Low	Span 9.11 ppm	9.18 ppm	3/15/2026	9.18 ppm	0.07 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC403919				B32023	CO,NO,NOX,BALN		
7/19/2023 5:36 PM	O2 %	Single	Zero 0.00%	0.01%	3/14/2031	0.01%	0.01%	±1%	25%	Unit online; Passed
			AAL18904				B32023	NO,NOX,BALN		
7/19/2023 5:36 PM	O2 %	Single	Span 22.51%	22.57%	4/14/2031	22.57%	0.06%	±1%	25%	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	EPA Vendor ID	Allowable Drift	Instrument Span	Results
7/19/2023 5:36 PM	75-O2 %	Single	Zero	0.00%	0.01%	3/14/2031	0.01%	B32023	±1%	25%	Unit online; Passed
			AAL18904						NO,NOX,BALN		
7/19/2023 5:36 PM	75-O2 %	Single	Span	22.51%	22.57%	4/14/2031	0.06%	B72023	±1%	25%	Unit online; Passed
			EB0083068						CO,O2,BALN		
7/22/2023 6:06 PM	NOx ppm	High	Zero	0.0 ppm	-0.5 ppm	4/14/2031	-0.5 ppm	B72023	±10 ppm	200 ppm	Unit online; Passed
			EB0083068						CO,O2,BALN		
7/22/2023 6:06 PM	NOx ppm	High	Span	177.7 ppm	176.9 ppm	3/14/2031	-0.8 ppm	B32023	±10 ppm	200 ppm	Unit online; Passed
			AAL18904						NO,NOX,BALN		
7/22/2023 6:06 PM	NOx ppm	Low	Zero	0.00 ppm	0.01 ppm	4/14/2031	0.01 ppm	B72023	±0.5 ppm	10 ppm	Unit online; Passed
			EB0083068						CO,O2,BALN		
7/22/2023 6:06 PM	NOx ppm	Low	Span	9.11 ppm	9.09 ppm	3/15/2026	-0.02 ppm	B32023	±0.5 ppm	10 ppm	Unit online; Passed
			CC403919						CO,NO,NOX,BALN		
7/22/2023 6:06 PM	75-NOx ppm	High	Zero	0.0 ppm	-0.5 ppm	4/14/2031	-0.5 ppm	B72023	±10 ppm	200 ppm	Unit online; Passed
			EB0083068						CO,O2,BALN		
7/22/2023 6:06 PM	75-NOx ppm	High	Span	177.7 ppm	176.9 ppm	3/14/2031	-0.8 ppm	B32023	±10 ppm	200 ppm	Unit online; Passed
			AAL18904						NO,NOX,BALN		
7/22/2023 6:06 PM	75-NOx ppm	Low	Zero	0.00 ppm	0.01 ppm	4/14/2031	0.01 ppm	B72023	±5 ppm	10 ppm	Unit online; Passed
			EB0083068						CO,O2,BALN		
7/22/2023 6:06 PM	75-NOx ppm	Low	Span	9.11 ppm	9.09 ppm	3/15/2026	-0.02 ppm	B32023	±5 ppm	10 ppm	Unit online; Passed
			CC403919						CO,NO,NOX,BALN		
7/22/2023 6:06 PM	CO ppm	High	Zero	0.0 ppm	-0.3 ppm	3/14/2031	-0.3 ppm	B32023	±10 ppm	200 ppm	Unit online; Passed
			AAL18904						NO,NOX,BALN		
7/22/2023 6:06 PM	CO ppm	High	Span	180.7 ppm	182.6 ppm	4/14/2031	1.9 ppm	B72023	±10 ppm	200 ppm	Unit online; Passed
			EB0083068						CO,O2,BALN		
7/22/2023 6:06 PM	CO ppm	Low	Zero	0.00 ppm	0.00 ppm	3/14/2031	0 ppm	B32023	±0.5 ppm	10 ppm	Unit online; Passed
			AAL18904						NO,NOX,BALN		
7/22/2023 6:06 PM	CO ppm	Low	Span	9.16 ppm	9.39 ppm	3/15/2026	0.23 ppm	B32023	±0.5 ppm	10 ppm	Unit online; Passed
			CC403919						CO,NO,NOX,BALN		
7/22/2023 6:06 PM	NH3/NOx ppm	High	Zero	0.0 ppm	-0.3 ppm	4/14/2031	-0.3 ppm	B72023	±10 ppm	200 ppm	Unit online; Passed
			EB0083068						CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
7/22/2023 6:06 PM	NH3/NOx ppm	High	Span	177.7 ppm	177.4 ppm	3/14/2031	-0.3 ppm	±10 ppm	200 ppm	Unit online; Passed
			AAL18904				B32023	NO,NOX,BALN		
7/22/2023 6:06 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	0.14 ppm	4/14/2031	0.14 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		
7/22/2023 6:06 PM	NH3/NOx ppm	Low	Span	9.11 ppm	9.18 ppm	3/15/2026	0.07 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC403919				B32023	CO,NO,NOX,BALN		
7/22/2023 6:06 PM	O2 %	Single	Zero	0.00%	0.02%	3/14/2031	0.02%	±1%	25%	Unit online; Passed
			AAL18904				B32023	NO,NOX,BALN		
7/22/2023 6:06 PM	O2 %	Single	Span	22.51%	22.57%	4/14/2031	0.06%	±1%	25%	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		
7/22/2023 6:06 PM	75-O2 %	Single	Zero	0.00%	0.02%	3/14/2031	0.02%	±1%	25%	Unit online; Passed
			AAL18904				B32023	NO,NOX,BALN		
7/22/2023 6:06 PM	75-O2 %	Single	Span	22.51%	22.57%	4/14/2031	0.06%	±1%	25%	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		
7/24/2023 5:36 PM	NOx ppm	High	Zero	0.0 ppm	-0.5 ppm	4/14/2031	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		
7/24/2023 5:36 PM	NOx ppm	High	Span	177.7 ppm	176.7 ppm	3/14/2031	-1 ppm	±10 ppm	200 ppm	Unit online; Passed
			AAL18904				B32023	NO,NOX,BALN		
7/24/2023 5:36 PM	NOx ppm	Low	Zero	0.00 ppm	0.01 ppm	4/14/2031	0.01 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		
7/24/2023 5:36 PM	NOx ppm	Low	Span	9.11 ppm	9.05 ppm	3/15/2026	-0.06 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC403919				B32023	CO,NO,NOX,BALN		
7/24/2023 5:36 PM	75-NOx ppm	High	Zero	0.0 ppm	-0.5 ppm	4/14/2031	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		
7/24/2023 5:36 PM	75-NOx ppm	High	Span	177.7 ppm	176.7 ppm	3/14/2031	-1 ppm	±10 ppm	200 ppm	Unit online; Passed
			AAL18904				B32023	NO,NOX,BALN		
7/24/2023 5:36 PM	75-NOx ppm	Low	Zero	0.00 ppm	0.01 ppm	4/14/2031	0.01 ppm	±5 ppm	10 ppm	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		
7/24/2023 5:36 PM	75-NOx ppm	Low	Span	9.11 ppm	9.05 ppm	3/15/2026	-0.06 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC403919				B32023	CO,NO,NOX,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
7/24/2023 5:36 PM	CO ppm	High	Zero	0.0 ppm	-0.2 ppm	3/14/2031	-0.2 ppm	±10 ppm	200 ppm	Unit online; Passed
			AAL18904				B32023	NO,NOX,BALN		
7/24/2023 5:36 PM	CO ppm	High	Span	180.7 ppm	182.2 ppm	4/14/2031	1.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		
7/24/2023 5:36 PM	CO ppm	Low	Zero	0.00 ppm	0.10 ppm	3/14/2031	0.1 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			AAL18904				B32023	NO,NOX,BALN		
7/24/2023 5:36 PM	CO ppm	Low	Span	9.16 ppm	9.28 ppm	3/15/2026	0.12 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC403919				B32023	CO,NO,NOX,BALN		
7/24/2023 5:36 PM	NH3/NOx ppm	High	Zero	0.0 ppm	-0.3 ppm	4/14/2031	-0.3 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		
7/24/2023 5:36 PM	NH3/NOx ppm	High	Span	177.7 ppm	176.9 ppm	3/14/2031	-0.8 ppm	±10 ppm	200 ppm	Unit online; Passed
			AAL18904				B32023	NO,NOX,BALN		
7/24/2023 5:36 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	0.16 ppm	4/14/2031	0.16 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		
7/24/2023 5:36 PM	NH3/NOx ppm	Low	Span	9.11 ppm	9.18 ppm	3/15/2026	0.07 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC403919				B32023	CO,NO,NOX,BALN		
7/24/2023 5:36 PM	O2 %	Single	Zero	0.00%	0.01%	3/14/2031	0.01%	±1%	25%	Unit online; Passed
			AAL18904				B32023	NO,NOX,BALN		
7/24/2023 5:36 PM	O2 %	Single	Span	22.51%	22.58%	4/14/2031	0.07%	±1%	25%	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		
7/24/2023 5:36 PM	75-O2 %	Single	Zero	0.00%	0.01%	3/14/2031	0.01%	±1%	25%	Unit online; Passed
			AAL18904				B32023	NO,NOX,BALN		
7/24/2023 5:36 PM	75-O2 %	Single	Span	22.51%	22.58%	4/14/2031	0.07%	±1%	25%	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		
7/25/2023 6:36 PM	NOx ppm	High	Zero	0.0 ppm	-0.5 ppm	4/14/2031	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		
7/25/2023 6:36 PM	NOx ppm	High	Span	177.7 ppm	176.5 ppm	3/14/2031	-1.2 ppm	±10 ppm	200 ppm	Unit online; Passed
			AAL18904				B32023	NO,NOX,BALN		
7/25/2023 6:36 PM	NOx ppm	Low	Zero	0.00 ppm	0.00 ppm	4/14/2031	0 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	Allowable Drift	Instrument Span	Results
7/25/2023 6:36 PM	NOx ppm	Low	Span	9.11 ppm	3/15/2026	9.07 ppm	-0.04 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC403919				B32023	CO,NO,NOX,BALN		
7/25/2023 6:36 PM	75-NOx ppm	High	Zero	0.0 ppm	4/14/2031	-0.5 ppm	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		
7/25/2023 6:36 PM	75-NOx ppm	High	Span	177.7 ppm	3/14/2031	176.5 ppm	-1.2 ppm	±10 ppm	200 ppm	Unit online; Passed
			AAL18904				B32023	NO,NOX,BALN		
7/25/2023 6:36 PM	75-NOx ppm	Low	Zero	0.00 ppm	4/14/2031	0.00 ppm	0 ppm	±5 ppm	10 ppm	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		
7/25/2023 6:36 PM	75-NOx ppm	Low	Span	9.11 ppm	3/15/2026	9.07 ppm	-0.04 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC403919				B32023	CO,NO,NOX,BALN		
7/25/2023 6:36 PM	CO ppm	High	Zero	0.0 ppm	3/14/2031	-0.2 ppm	-0.2 ppm	±10 ppm	200 ppm	Unit online; Passed
			AAL18904				B32023	NO,NOX,BALN		
7/25/2023 6:36 PM	CO ppm	High	Span	180.7 ppm	4/14/2031	182.1 ppm	1.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		
7/25/2023 6:36 PM	CO ppm	Low	Zero	0.00 ppm	3/14/2031	0.12 ppm	0.12 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			AAL18904				B32023	NO,NOX,BALN		
7/25/2023 6:36 PM	CO ppm	Low	Span	9.16 ppm	3/15/2026	9.42 ppm	0.26 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC403919				B32023	CO,NO,NOX,BALN		
7/25/2023 6:36 PM	NH3/NOx ppm	High	Zero	0.0 ppm	4/14/2031	-0.3 ppm	-0.3 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		
7/25/2023 6:36 PM	NH3/NOx ppm	High	Span	177.7 ppm	3/14/2031	176.5 ppm	-1.2 ppm	±10 ppm	200 ppm	Unit online; Passed
			AAL18904				B32023	NO,NOX,BALN		
7/25/2023 6:36 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	4/14/2031	0.16 ppm	0.16 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		
7/25/2023 6:36 PM	NH3/NOx ppm	Low	Span	9.11 ppm	3/15/2026	9.17 ppm	0.06 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC403919				B32023	CO,NO,NOX,BALN		
7/25/2023 6:36 PM	O2 %	Single	Zero	0.00%	3/14/2031	0.02%	0.02%	±1%	25%	Unit online; Passed
			AAL18904				B32023	NO,NOX,BALN		
7/25/2023 6:36 PM	O2 %	Single	Span	22.51%	4/14/2031	22.58%	0.07%	±1%	25%	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	Allowable Drift	Instrument Span	Results
7/25/2023 6:36 PM	75-O2 %	Single	Zero	0.00%	3/14/2031	0.02%	0.02%	±1%	25%	Unit online; Passed
			AAL18904				B32023	NO,NOX,BALN		
7/25/2023 6:36 PM	75-O2 %	Single	Span	22.51%	22.58%	0.07%	0.07%	±1%	25%	Unit online; Passed
			EB0083068				B72023	CO,O2,BALN		
7/27/2023 5:51 PM	NOx ppm	High	Zero	0.0 ppm	6/9/2031	-0.5 ppm	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0141960				B72023	CO,O2,BALN		
7/27/2023 5:51 PM	NOx ppm	High	Span	177.4 ppm	178.6 ppm	1.2 ppm	1.2 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC314637				B32023	NO,NOX,BALN		
7/27/2023 5:51 PM	NOx ppm	Low	Zero	0.00 ppm	6/9/2031	0.00 ppm	0 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			EB0141960				B72023	CO,O2,BALN		
7/27/2023 5:51 PM	NOx ppm	Low	Span	9.13 ppm	9.06 ppm	-0.07 ppm	-0.07 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC408913				B32023	CO,NO,NOX,BALN		
7/27/2023 5:51 PM	75-NOx ppm	High	Zero	0.0 ppm	6/9/2031	-0.5 ppm	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0141960				B72023	CO,O2,BALN		
7/27/2023 5:51 PM	75-NOx ppm	High	Span	177.4 ppm	178.6 ppm	1.2 ppm	1.2 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC314637				B32023	NO,NOX,BALN		
7/27/2023 5:51 PM	75-NOx ppm	Low	Zero	0.00 ppm	6/9/2031	0.00 ppm	0 ppm	±5 ppm	10 ppm	Unit online; Passed
			EB0141960				B72023	CO,O2,BALN		
7/27/2023 5:51 PM	75-NOx ppm	Low	Span	9.13 ppm	9.06 ppm	-0.07 ppm	-0.07 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC408913				B32023	CO,NO,NOX,BALN		
7/27/2023 5:51 PM	CO ppm	High	Zero	0.0 ppm	6/13/2031	-0.2 ppm	-0.2 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC314637				B32023	NO,NOX,BALN		
7/27/2023 5:51 PM	CO ppm	High	Span	181.8 ppm	183.0 ppm	1.2 ppm	1.2 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0141960				B72023	CO,O2,BALN		
7/27/2023 5:51 PM	CO ppm	Low	Zero	0.00 ppm	6/13/2031	0.09 ppm	0.09 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC314637				B32023	NO,NOX,BALN		
7/27/2023 5:51 PM	CO ppm	Low	Span	9.03 ppm	9.28 ppm	0.25 ppm	0.25 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC408913				B32023	CO,NO,NOX,BALN		
7/27/2023 5:51 PM	NH3/NOx ppm	High	Zero	0.0 ppm	6/9/2031	-0.4 ppm	-0.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0141960				B72023	CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	EPA Vendor ID	Allowable Drift	Instrument Span	Results
7/27/2023 5:51 PM	NH3/NOx ppm	High	Span	177.4 ppm	6/13/2031	178.2 ppm	0.8 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
7/27/2023 5:51 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	6/9/2031	0.13 ppm	0.13 ppm	B72023	±0.5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
7/27/2023 5:51 PM	NH3/NOx ppm	Low	Span	9.13 ppm	6/13/2026	9.32 ppm	0.19 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
7/27/2023 5:51 PM	O2 %	Single	Zero	0.00%	6/13/2031	0.02%	0.02%	B32023	±1% NO,NOX,BALN	25%	Unit online; Passed
7/27/2023 5:51 PM	O2 %	Single	Span	22.54%	6/9/2031	22.59%	0.05%	B72023	±1% CO,O2,BALN	25%	Unit online; Passed
7/27/2023 5:51 PM	75-O2 %	Single	Zero	0.00%	6/13/2031	0.02%	0.02%	B32023	±1% NO,NOX,BALN	25%	Unit online; Passed
7/27/2023 5:51 PM	75-O2 %	Single	Span	22.54%	6/9/2031	22.59%	0.05%	B72023	±1% CO,O2,BALN	25%	Unit online; Passed
7/28/2023 6:05 PM	NOx ppm	High	Zero	0.0 ppm	6/9/2031	-0.5 ppm	-0.5 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
7/28/2023 6:05 PM	NOx ppm	High	Span	177.4 ppm	6/13/2031	178.4 ppm	1 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
7/28/2023 6:05 PM	NOx ppm	Low	Zero	0.00 ppm	6/9/2031	0.00 ppm	0 ppm	B72023	±0.5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
7/28/2023 6:05 PM	NOx ppm	Low	Span	9.13 ppm	6/13/2026	9.06 ppm	-0.07 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
7/28/2023 6:05 PM	75-NOx ppm	High	Zero	0.0 ppm	6/9/2031	-0.5 ppm	-0.5 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
7/28/2023 6:05 PM	75-NOx ppm	High	Span	177.4 ppm	6/13/2031	178.4 ppm	1 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
7/28/2023 6:05 PM	75-NOx ppm	Low	Zero	0.00 ppm	6/9/2031	0.00 ppm	0 ppm	B72023	±5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
7/28/2023 6:05 PM	75-NOx ppm	Low	Span	9.13 ppm	6/13/2026	9.06 ppm	-0.07 ppm	B32023	±5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	EPA Vendor ID	Allowable Drift	Instrument Span	Results
7/23/2023 6:05 PM	CO ppm	High	Zero	0.0 ppm	6/13/2031	-0.2 ppm	-0.2 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
7/23/2023 6:05 PM	CO ppm	High	Span	181.8 ppm	6/9/2031	181.8 ppm	0 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
7/23/2023 6:05 PM	CO ppm	Low	Zero	0.00 ppm	6/13/2031	0.14 ppm	0.14 ppm	B32023	±0.5 ppm NO,NOX,BALN	10 ppm	Unit online; Passed
7/28/2023 6:05 PM	CO ppm	Low	Span	9.03 ppm	6/13/2026	9.24 ppm	0.21 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
7/28/2023 6:05 PM	NH3/NOx ppm	High	Zero	0.0 ppm	6/9/2031	-0.4 ppm	-0.4 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
7/28/2023 6:05 PM	NH3/NOx ppm	High	Span	177.4 ppm	6/13/2031	177.8 ppm	0.4 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
7/28/2023 6:05 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	6/9/2031	0.14 ppm	0.14 ppm	B72023	±0.5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
7/28/2023 6:05 PM	NH3/NOx ppm	Low	Span	9.13 ppm	6/13/2026	9.30 ppm	0.17 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
7/28/2023 6:05 PM	O2 %	Single	Zero	0.00%	6/13/2031	0.02%	0.02%	B32023	±1% NO,NOX,BALN	25%	Unit online; Passed
7/28/2023 6:05 PM	O2 %	Single	Span	22.54%	6/9/2031	22.59%	0.05%	B72023	±1% CO,O2,BALN	25%	Unit online; Passed
7/28/2023 6:05 PM	75-O2 %	Single	Zero	0.00%	6/13/2031	0.02%	0.02%	B32023	±1% NO,NOX,BALN	25%	Unit online; Passed
7/28/2023 6:05 PM	75-O2 %	Single	Span	22.54%	6/9/2031	22.59%	0.05%	B72023	±1% CO,O2,BALN	25%	Unit online; Passed
7/29/2023 6:05 PM	NOx ppm	High	Zero	0.0 ppm	6/9/2031	-0.5 ppm	-0.5 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
7/29/2023 6:05 PM	NOx ppm	High	Span	177.4 ppm	6/13/2031	178.0 ppm	0.6 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
7/29/2023 6:05 PM	NOx ppm	Low	Zero	0.00 ppm	6/9/2031	0.00 ppm	0 ppm	B72023	±0.5 ppm CO,O2,BALN	10 ppm	Unit online; Passed

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
		Cylinder ID		Value	Value		EPA Vendor ID	EPA Gas Type	Codes	
7/29/2023 6:05 PM	NOx ppm	Low	Span	9.13 ppm	9.04 ppm	6/13/2026	-0.09 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC408913				B32023	CO,NO,NOX,BALN		
7/29/2023 6:05 PM	75-NOx ppm	High	Zero	0.0 ppm	-0.5 ppm	6/9/2031	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0141960				B72023	CO,O2,BALN		
7/29/2023 6:05 PM	75-NOx ppm	High	Span	177.4 ppm	178.0 ppm	6/13/2031	0.6 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC314637				B32023	NO,NOX,BALN		
7/29/2023 6:05 PM	75-NOx ppm	Low	Zero	0.00 ppm	0.00 ppm	6/9/2031	0 ppm	±5 ppm	10 ppm	Unit online; Passed
			EB0141960				B72023	CO,O2,BALN		
7/29/2023 6:05 PM	75-NOx ppm	Low	Span	9.13 ppm	9.04 ppm	6/13/2026	-0.09 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC408913				B32023	CO,NO,NOX,BALN		
7/29/2023 6:05 PM	CO ppm	High	Zero	0.0 ppm	-0.4 ppm	6/13/2031	-0.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC314637				B32023	NO,NOX,BALN		
7/29/2023 6:05 PM	CO ppm	High	Span	181.8 ppm	182.6 ppm	6/9/2031	0.8 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0141960				B72023	CO,O2,BALN		
7/29/2023 6:05 PM	CO ppm	Low	Zero	0.00 ppm	-0.04 ppm	6/13/2031	-0.04 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC314637				B32023	NO,NOX,BALN		
7/29/2023 6:05 PM	CO ppm	Low	Span	9.03 ppm	9.40 ppm	6/13/2026	0.37 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC408913				B32023	CO,NO,NOX,BALN		
7/29/2023 6:05 PM	NH3/NOx ppm	High	Zero	0.0 ppm	-0.4 ppm	6/9/2031	-0.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0141960				B72023	CO,O2,BALN		
7/29/2023 6:05 PM	NH3/NOx ppm	High	Span	177.4 ppm	177.5 ppm	6/13/2031	0.1 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC314637				B32023	NO,NOX,BALN		
7/29/2023 6:05 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	0.11 ppm	6/9/2031	0.11 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			EB0141960				B72023	CO,O2,BALN		
7/29/2023 6:05 PM	NH3/NOx ppm	Low	Span	9.13 ppm	9.23 ppm	6/13/2026	0.1 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC408913				B32023	CO,NO,NOX,BALN		
7/29/2023 6:05 PM	O2 %	Single	Zero	0.00%	0.02%	6/13/2031	0.02%	±1% 25%	25%	Unit online; Passed
			CC314637				B32023	NO,NOX,BALN		
7/29/2023 6:05 PM	O2 %	Single	Span	22.54%	22.58%	6/9/2031	0.04%	±1% 25%	25%	Unit online; Passed
			EB0141960				B72023	CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
		Cylinder ID		Value	Value		EPA Vendor ID	EPA Gas Type	Codes	
7/29/2023 6:05 PM	75-O2 %	Single	Zero	0.00%	0.02%	6/13/2031	0.02%	±1%	25%	Unit online; Passed
			CC314637				B32023	NO,NOX,BALN		
7/29/2023 6:05 PM	75-O2 %	Single	Span	22.54%	22.58%	6/9/2031	0.04%	±1%	25%	Unit online; Passed
			EB0141960				B72023	CO,O2,BALN		
8/1/2023 6:05 PM	NOx ppm	High	Zero	0.0 ppm	-0.5 ppm	6/9/2031	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0141960				B72023	CO,O2,BALN		
8/1/2023 6:05 PM	NOx ppm	High	Span	177.4 ppm	178.1 ppm	6/13/2031	0.7 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC314637				B32023	NO,NOX,BALN		
8/1/2023 6:05 PM	NOx ppm	Low	Zero	0.00 ppm	0.01 ppm	6/9/2031	0.01 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			EB0141960				B72023	CO,O2,BALN		
8/1/2023 6:05 PM	NOx ppm	Low	Span	9.13 ppm	9.03 ppm	6/13/2026	-0.1 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC408913				B32023	CO,NO,NOX,BALN		
8/1/2023 6:05 PM	75-NOx ppm	High	Zero	0.0 ppm	-0.5 ppm	6/9/2031	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0141960				B72023	CO,O2,BALN		
8/1/2023 6:05 PM	75-NOx ppm	High	Span	177.4 ppm	178.1 ppm	6/13/2031	0.7 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC314637				B32023	NO,NOX,BALN		
8/1/2023 6:05 PM	75-NOx ppm	Low	Zero	0.00 ppm	0.01 ppm	6/9/2031	0.01 ppm	±5 ppm	10 ppm	Unit online; Passed
			EB0141960				B72023	CO,O2,BALN		
8/1/2023 6:05 PM	75-NOx ppm	Low	Span	9.13 ppm	9.03 ppm	6/13/2026	-0.1 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC408913				B32023	CO,NO,NOX,BALN		
8/1/2023 6:05 PM	CO ppm	High	Zero	0.0 ppm	-0.2 ppm	6/13/2031	-0.2 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC314637				B32023	NO,NOX,BALN		
8/1/2023 6:05 PM	CO ppm	High	Span	181.8 ppm	183.5 ppm	6/9/2031	1.7 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0141960				B72023	CO,O2,BALN		
8/1/2023 6:05 PM	CO ppm	Low	Zero	0.00 ppm	0.09 ppm	6/13/2031	0.09 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC314637				B32023	NO,NOX,BALN		
8/1/2023 6:05 PM	CO ppm	Low	Span	9.03 ppm	9.30 ppm	6/13/2026	0.27 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC408913				B32023	CO,NO,NOX,BALN		
8/1/2023 6:05 PM	NH3/NOx ppm	High	Zero	0.0 ppm	-0.3 ppm	6/9/2031	-0.3 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0141960				B72023	CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	EPA Vendor ID	Allowable Drift	Instrument Span	Results
8/1/2023 6:05 PM	NH3/NOx ppm	High	Span	177.4 ppm	6/13/2031	177.2 ppm	-0.2 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
8/1/2023 6:05 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	6/9/2031	0.15 ppm	0.15 ppm	B72023	±0.5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
8/1/2023 6:05 PM	NH3/NOx ppm	Low	Span	9.13 ppm	6/13/2026	9.24 ppm	0.11 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
8/1/2023 6:05 PM	O2 %	Single	Zero	0.00%	6/13/2031	0.03%	0.03%	B32023	±1% NO,NOX,BALN	25%	Unit online; Passed
8/1/2023 6:05 PM	O2 %	Single	Span	22.54%	6/9/2031	22.60%	0.06%	B72023	±1% CO,O2,BALN	25%	Unit online; Passed
8/1/2023 6:05 PM	75-O2 %	Single	Zero	0.00%	6/13/2031	0.03%	0.03%	B32023	±1% NO,NOX,BALN	25%	Unit online; Passed
8/1/2023 6:05 PM	75-O2 %	Single	Span	22.54%	6/9/2031	22.60%	0.06%	B72023	±1% CO,O2,BALN	25%	Unit online; Passed
8/2/2023 6:05 PM	NOx ppm	High	Zero	0.0 ppm	6/9/2031	-0.5 ppm	-0.5 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
8/2/2023 6:05 PM	NOx ppm	High	Span	177.4 ppm	6/13/2031	177.8 ppm	0.4 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
8/2/2023 6:05 PM	NOx ppm	Low	Zero	0.00 ppm	6/9/2031	0.00 ppm	0 ppm	B72023	±0.5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
8/2/2023 6:05 PM	NOx ppm	Low	Span	9.13 ppm	6/13/2026	9.03 ppm	-0.1 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
8/2/2023 6:05 PM	75-NOx ppm	High	Zero	0.0 ppm	6/9/2031	-0.5 ppm	-0.5 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
8/2/2023 6:05 PM	75-NOx ppm	High	Span	177.4 ppm	6/13/2031	177.8 ppm	0.4 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
8/2/2023 6:05 PM	75-NOx ppm	Low	Zero	0.00 ppm	6/9/2031	0.00 ppm	0 ppm	B72023	±5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
8/2/2023 6:05 PM	75-NOx ppm	Low	Span	9.13 ppm	6/13/2026	9.03 ppm	-0.1 ppm	B32023	±5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	EPA Vendor ID	Allowable Drift	Instrument Span	Results
8/2/2023 6:05 PM	CO ppm	High	Zero	0.0 ppm	6/13/2031	-0.2 ppm	-0.2 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
8/2/2023 6:05 PM	CO ppm	High	Span	181.8 ppm	6/9/2031	183.2 ppm	1.4 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
8/2/2023 6:05 PM	CO ppm	Low	Zero	0.00 ppm	6/13/2031	0.06 ppm	0.06 ppm	B32023	±0.5 ppm NO,NOX,BALN	10 ppm	Unit online; Passed
8/2/2023 6:05 PM	CO ppm	Low	Span	9.03 ppm	6/13/2026	9.48 ppm	0.45 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
8/2/2023 6:05 PM	NH3/NOx ppm	High	Zero	0.0 ppm	6/9/2031	-0.4 ppm	-0.4 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
8/2/2023 6:05 PM	NH3/NOx ppm	High	Span	177.4 ppm	6/13/2031	176.9 ppm	-0.5 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
8/2/2023 6:05 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	6/9/2031	0.12 ppm	0.12 ppm	B72023	±0.5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
8/2/2023 6:05 PM	NH3/NOx ppm	Low	Span	9.13 ppm	6/13/2026	9.20 ppm	0.07 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
8/2/2023 6:05 PM	O2 %	Single	Zero	0.00%	6/13/2031	0.03%	0.03%	B32023	±1% NO,NOX,BALN	25%	Unit online; Passed
8/2/2023 6:05 PM	O2 %	Single	Span	22.54%	6/9/2031	22.60%	0.06%	B72023	±1% CO,O2,BALN	25%	Unit online; Passed
8/2/2023 6:05 PM	75-O2 %	Single	Zero	0.00%	6/13/2031	0.03%	0.03%	B32023	±1% NO,NOX,BALN	25%	Unit online; Passed
8/2/2023 6:05 PM	75-O2 %	Single	Span	22.54%	6/9/2031	22.60%	0.06%	B72023	±1% CO,O2,BALN	25%	Unit online; Passed
8/9/2023 6:06 PM	NOx ppm	High	Zero	0.0 ppm	6/9/2031	-0.6 ppm	-0.6 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
8/9/2023 6:06 PM	NOx ppm	High	Span	177.4 ppm	6/13/2031	178.2 ppm	0.8 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
8/9/2023 6:06 PM	NOx ppm	Low	Zero	0.00 ppm	6/9/2031	0.02 ppm	0.02 ppm	B72023	±0.5 ppm CO,O2,BALN	10 ppm	Unit online; Passed

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
		Cylinder ID				EPA Vendor ID	EPA Gas Type Codes			
8/9/2023 6:06 PM	NOx ppm	Low	Span	9.13 ppm	9.10 ppm	6/13/2026	-0.03 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC408913			B32023		CO,NO,NOX,BALN		
8/9/2023 6:06 PM	75-NOx ppm	High	Zero	0.0 ppm	-0.6 ppm	6/9/2031	-0.6 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0141960			B72023		CO,O2,BALN		
8/9/2023 6:06 PM	75-NOx ppm	High	Span	177.4 ppm	178.2 ppm	6/13/2031	0.8 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC314637			B32023		NO,NOX,BALN		
8/9/2023 6:06 PM	75-NOx ppm	Low	Zero	0.00 ppm	0.02 ppm	6/9/2031	0.02 ppm	±5 ppm	10 ppm	Unit online; Passed
			EB0141960			B72023		CO,O2,BALN		
8/9/2023 6:06 PM	75-NOx ppm	Low	Span	9.13 ppm	9.10 ppm	6/13/2026	-0.03 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC408913			B32023		CO,NO,NOX,BALN		
8/9/2023 6:06 PM	CO ppm	High	Zero	0.0 ppm	-0.1 ppm	6/13/2031	-0.1 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC314637			B32023		NO,NOX,BALN		
8/9/2023 6:06 PM	CO ppm	High	Span	181.8 ppm	182.7 ppm	6/9/2031	0.9 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0141960			B72023		CO,O2,BALN		
8/9/2023 6:06 PM	CO ppm	Low	Zero	0.00 ppm	0.02 ppm	6/13/2031	0.02 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC314637			B32023		NO,NOX,BALN		
8/9/2023 6:06 PM	CO ppm	Low	Span	9.03 ppm	9.44 ppm	6/13/2026	0.41 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC408913			B32023		CO,NO,NOX,BALN		
8/9/2023 6:06 PM	NH3/NOx ppm	High	Zero	0.0 ppm	-0.3 ppm	6/9/2031	-0.3 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0141960			B72023		CO,O2,BALN		
8/9/2023 6:06 PM	NH3/NOx ppm	High	Span	177.4 ppm	176.0 ppm	6/13/2031	-1.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC314637			B32023		NO,NOX,BALN		
8/9/2023 6:06 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	0.18 ppm	6/9/2031	0.18 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			EB0141960			B72023		CO,O2,BALN		
8/9/2023 6:06 PM	NH3/NOx ppm	Low	Span	9.13 ppm	9.23 ppm	6/13/2026	0.1 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC408913			B32023		CO,NO,NOX,BALN		
8/9/2023 6:06 PM	O2 %	Single	Zero	0.00%	0.04%	6/13/2031	0.04%	±1%	25%	Unit online; Passed
			CC314637			B32023		NO,NOX,BALN		
8/9/2023 6:06 PM	O2 %	Single	Span	22.54%	22.60%	6/9/2031	0.06%	±1%	25%	Unit online; Passed
			EB0141960			B72023		CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	EPA Vendor ID	Allowable Drift	Instrument Span	Results
8/9/2023 6:06 PM	75-O2 %	Single	Zero	0.00%	0.04%	6/13/2031	0.04%	B32023	±1% NO,NOX,BALN	25%	Unit online; Passed
8/9/2023 6:06 PM	75-O2 %	Single	Span	22.54%	22.60%	6/9/2031	0.06%	B72023	±1% CO,O2,BALN	25%	Unit online; Passed
8/13/2023 5:07 PM	NOx ppm	High	Zero	0.0 ppm	-0.6 ppm	6/9/2031	-0.6 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
8/13/2023 5:07 PM	NOx ppm	High	Span	177.4 ppm	178.0 ppm	6/13/2031	0.6 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
8/13/2023 5:07 PM	NOx ppm	Low	Zero	0.00 ppm	0.01 ppm	6/9/2031	0.01 ppm	B72023	±0.5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
8/13/2023 5:07 PM	NOx ppm	Low	Span	9.13 ppm	9.06 ppm	6/13/2026	-0.07 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
8/13/2023 5:07 PM	75-NOx ppm	High	Zero	0.0 ppm	-0.6 ppm	6/9/2031	-0.6 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
8/13/2023 5:07 PM	75-NOx ppm	High	Span	177.4 ppm	178.0 ppm	6/13/2031	0.6 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
8/13/2023 5:07 PM	75-NOx ppm	Low	Zero	0.00 ppm	0.01 ppm	6/9/2031	0.01 ppm	B72023	±5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
8/13/2023 5:07 PM	75-NOx ppm	Low	Span	9.13 ppm	9.06 ppm	6/13/2026	-0.07 ppm	B32023	±5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
8/13/2023 5:07 PM	CO ppm	High	Zero	0.0 ppm	-0.4 ppm	6/13/2031	-0.4 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
8/13/2023 5:07 PM	CO ppm	High	Span	181.8 ppm	182.8 ppm	6/9/2031	1 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
8/13/2023 5:07 PM	CO ppm	Low	Zero	0.00 ppm	-0.24 ppm	6/13/2031	-0.24 ppm	B32023	±0.5 ppm NO,NOX,BALN	10 ppm	Unit online; Passed
8/13/2023 5:07 PM	CO ppm	Low	Span	9.03 ppm	9.16 ppm	6/13/2026	0.13 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
8/13/2023 5:07 PM	NH3/NOx ppm	High	Zero	0.0 ppm	-0.3 ppm	6/9/2031	-0.3 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
		Cylinder ID				EPA Vendor ID	EPA Gas Type Codes			
8/13/2023 5:07 PM	NH3/NOx ppm	High	Span	177.4 ppm	176.0 ppm	6/13/2031	-1.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC314637			B32023		NO,NOX,BALN		
8/13/2023 5:07 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	0.14 ppm	6/9/2031	0.14 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			EB0141960			B72023		CO,O2,BALN		
8/13/2023 5:07 PM	NH3/NOx ppm	Low	Span	9.13 ppm	9.18 ppm	6/13/2026	0.05 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC408913			B32023		CO,NO,NOX,BALN		
8/13/2023 5:07 PM	O2 %	Single	Zero	0.00%	0.05%	6/13/2031	0.05%	±1%	25%	Unit online; Passed
			CC314637			B32023		NO,NOX,BALN		
8/13/2023 5:07 PM	O2 %	Single	Span	22.54%	22.61%	6/9/2031	0.07%	±1%	25%	Unit online; Passed
			EB0141960			B72023		CO,O2,BALN		
8/13/2023 5:07 PM	75-O2 %	Single	Zero	0.00%	0.05%	6/13/2031	0.05%	±1%	25%	Unit online; Passed
			CC314637			B32023		NO,NOX,BALN		
8/13/2023 5:07 PM	75-O2 %	Single	Span	22.54%	22.61%	6/9/2031	0.07%	±1%	25%	Unit online; Passed
			EB0141960			B72023		CO,O2,BALN		
8/14/2023 5:07 PM	NOx ppm	High	Zero	0.0 ppm	-0.6 ppm	6/9/2031	-0.6 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0141960			B72023		CO,O2,BALN		
8/14/2023 5:07 PM	NOx ppm	High	Span	177.4 ppm	177.8 ppm	6/13/2031	0.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC314637			B32023		NO,NOX,BALN		
8/14/2023 5:07 PM	NOx ppm	Low	Zero	0.00 ppm	0.00 ppm	6/9/2031	0 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			EB0141960			B72023		CO,O2,BALN		
8/14/2023 5:07 PM	NOx ppm	Low	Span	9.13 ppm	9.04 ppm	6/13/2026	-0.09 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC408913			B32023		CO,NO,NOX,BALN		
8/14/2023 5:07 PM	75-NOx ppm	High	Zero	0.0 ppm	-0.6 ppm	6/9/2031	-0.6 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0141960			B72023		CO,O2,BALN		
8/14/2023 5:07 PM	75-NOx ppm	High	Span	177.4 ppm	177.8 ppm	6/13/2031	0.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC314637			B32023		NO,NOX,BALN		
8/14/2023 5:07 PM	75-NOx ppm	Low	Zero	0.00 ppm	0.00 ppm	6/9/2031	0 ppm	±5 ppm	10 ppm	Unit online; Passed
			EB0141960			B72023		CO,O2,BALN		
8/14/2023 5:07 PM	75-NOx ppm	Low	Span	9.13 ppm	9.04 ppm	6/13/2026	-0.09 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC408913			B32023		CO,NO,NOX,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	EPA Vendor ID	Allowable Drift	Instrument Span	Results
8/14/2023 5:07 PM	CO ppm	High	Zero	0.0 ppm	6/13/2031	-0.3 ppm	-0.3 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
8/14/2023 5:07 PM	CO ppm	High	Span	181.8 ppm	6/9/2031	182.8 ppm	1 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
8/14/2023 5:07 PM	CO ppm	Low	Zero	0.00 ppm	6/13/2031	-0.18 ppm	-0.18 ppm	B32023	±0.5 ppm NO,NOX,BALN	10 ppm	Unit online; Passed
8/14/2023 5:07 PM	CO ppm	Low	Span	9.03 ppm	6/13/2026	9.20 ppm	0.17 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
8/14/2023 5:07 PM	NH3/NOx ppm	High	Zero	0.0 ppm	6/9/2031	-0.4 ppm	-0.4 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
8/14/2023 5:07 PM	NH3/NOx ppm	High	Span	177.4 ppm	6/13/2031	175.8 ppm	-1.6 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
8/14/2023 5:07 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	6/9/2031	0.14 ppm	0.14 ppm	B72023	±0.5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
8/14/2023 5:07 PM	NH3/NOx ppm	Low	Span	9.13 ppm	6/13/2026	9.16 ppm	0.03 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
8/14/2023 5:07 PM	O2 %	Single	Zero	0.00%	6/13/2031	0.04%	0.04%	B32023	±1% NO,NOX,BALN	25%	Unit online; Passed
8/14/2023 5:07 PM	O2 %	Single	Span	22.54%	6/9/2031	22.62%	0.08%	B72023	±1% CO,O2,BALN	25%	Unit online; Passed
8/14/2023 5:07 PM	75-O2 %	Single	Zero	0.00%	6/13/2031	0.04%	0.04%	B32023	±1% NO,NOX,BALN	25%	Unit online; Passed
8/14/2023 5:07 PM	75-O2 %	Single	Span	22.54%	6/9/2031	22.62%	0.08%	B72023	±1% CO,O2,BALN	25%	Unit online; Passed
8/15/2023 4:52 PM	NOx ppm	High	Zero	0.0 ppm	6/9/2031	-0.6 ppm	-0.6 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
8/15/2023 4:52 PM	NOx ppm	High	Span	177.4 ppm	6/13/2031	178.1 ppm	0.7 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
8/15/2023 4:52 PM	NOx ppm	Low	Zero	0.00 ppm	6/9/2031	0.00 ppm	0 ppm	B72023	±0.5 ppm CO,O2,BALN	10 ppm	Unit online; Passed

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	Allowable Drift	Instrument Span	Results
8/15/2023 4:52 PM	NOx ppm	Low	Span 9.13 ppm	9.01 ppm	6/13/2026	-0.12 ppm	B32023	±0.5 ppm	10 ppm	Unit online; Passed
			CC408913					CO,NO,NOX,BALN		
8/15/2023 4:52 PM	75-NOx ppm	High	Zero 0.0 ppm	-0.6 ppm	6/9/2031	-0.6 ppm	B72023	±10 ppm	200 ppm	Unit online; Passed
			EB0141960					CO,O2,BALN		
8/15/2023 4:52 PM	75-NOx ppm	High	Span 177.4 ppm	178.1 ppm	6/13/2031	0.7 ppm	B32023	±10 ppm	200 ppm	Unit online; Passed
			CC314637					NO,NOX,BALN		
8/15/2023 4:52 PM	75-NOx ppm	Low	Zero 0.00 ppm	0.00 ppm	6/9/2031	0 ppm	B72023	±5 ppm	10 ppm	Unit online; Passed
			EB0141960					CO,O2,BALN		
8/15/2023 4:52 PM	75-NOx ppm	Low	Span 9.13 ppm	9.01 ppm	6/13/2026	-0.12 ppm	B32023	±5 ppm	10 ppm	Unit online; Passed
			CC408913					CO,NO,NOX,BALN		
8/15/2023 4:52 PM	CO ppm	High	Zero 0.0 ppm	-0.3 ppm	6/13/2031	-0.3 ppm	B32023	±10 ppm	200 ppm	Unit online; Passed
			CC314637					NO,NOX,BALN		
8/15/2023 4:52 PM	CO ppm	High	Span 181.8 ppm	182.6 ppm	6/9/2031	0.8 ppm	B72023	±10 ppm	200 ppm	Unit online; Passed
			EB0141960					CO,O2,BALN		
8/15/2023 4:52 PM	CO ppm	Low	Zero 0.00 ppm	-0.12 ppm	6/13/2031	-0.12 ppm	B32023	±0.5 ppm	10 ppm	Unit online; Passed
			CC314637					NO,NOX,BALN		
8/15/2023 4:52 PM	CO ppm	Low	Span 9.03 ppm	8.93 ppm	6/13/2026	-0.1 ppm	B32023	±0.5 ppm	10 ppm	Unit online; Passed
			CC408913					CO,NO,NOX,BALN		
8/15/2023 4:52 PM	NH3/NOx ppm	High	Zero 0.0 ppm	-0.4 ppm	6/9/2031	-0.4 ppm	B72023	±10 ppm	200 ppm	Unit online; Passed
			EB0141960					CO,O2,BALN		
8/15/2023 4:52 PM	NH3/NOx ppm	High	Span 177.4 ppm	175.8 ppm	6/13/2031	-1.6 ppm	B32023	±10 ppm	200 ppm	Unit online; Passed
			CC314637					NO,NOX,BALN		
8/15/2023 4:52 PM	NH3/NOx ppm	Low	Zero 0.00 ppm	0.10 ppm	6/9/2031	0.1 ppm	B72023	±0.5 ppm	10 ppm	Unit online; Passed
			EB0141960					CO,O2,BALN		
8/15/2023 4:52 PM	NH3/NOx ppm	Low	Span 9.13 ppm	9.13 ppm	6/13/2026	0 ppm	B32023	±0.5 ppm	10 ppm	Unit online; Passed
			CC408913					CO,NO,NOX,BALN		
8/15/2023 4:52 PM	O2 %	Single	Zero 0.00%	0.04%	6/13/2031	0.04%	B32023	±1%	25%	Unit online; Passed
			CC314637					NO,NOX,BALN		
8/15/2023 4:52 PM	O2 %	Single	Span 22.54%	22.61%	6/9/2031	0.07%	B72023	±1%	25%	Unit online; Passed
			EB0141960					CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	EPA Vendor ID	Allowable Drift	Instrument Span	Results
8/15/2023 4:52 PM	75-O2 %	Single	Zero	0.00%	6/13/2031	0.04%	0.04%	B32023	±1% NO,NOX,BALN	25%	Unit online; Passed
8/15/2023 4:52 PM	75-O2 %	Single	Span	22.54%	6/9/2031	22.61%	0.07%	B72023	±1% CO,O2,BALN	25%	Unit online; Passed
8/16/2023 2:36 PM	NOx ppm	High	Zero	0.0 ppm	6/9/2031	-0.6 ppm	-0.6 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
8/16/2023 2:36 PM	NOx ppm	High	Span	177.4 ppm	6/13/2031	178.5 ppm	1.1 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
8/16/2023 2:36 PM	NOx ppm	Low	Zero	0.00 ppm	6/9/2031	-0.01 ppm	-0.01 ppm	B72023	±0.5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
8/16/2023 2:36 PM	NOx ppm	Low	Span	9.13 ppm	6/13/2026	9.05 ppm	-0.08 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
8/16/2023 2:36 PM	75-NOx ppm	High	Zero	0.0 ppm	6/9/2031	-0.6 ppm	-0.6 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
8/16/2023 2:36 PM	75-NOx ppm	High	Span	177.4 ppm	6/13/2031	178.5 ppm	1.1 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
8/16/2023 2:36 PM	75-NOx ppm	Low	Zero	0.00 ppm	6/9/2031	-0.01 ppm	-0.01 ppm	B72023	±5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
8/16/2023 2:36 PM	75-NOx ppm	Low	Span	9.13 ppm	6/13/2026	9.05 ppm	-0.08 ppm	B32023	±5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
8/16/2023 2:36 PM	CO ppm	High	Zero	0.0 ppm	6/13/2031	-0.3 ppm	-0.3 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
8/16/2023 2:36 PM	CO ppm	High	Span	181.8 ppm	6/9/2031	181.7 ppm	-0.1 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
8/16/2023 2:36 PM	CO ppm	Low	Zero	0.00 ppm	6/13/2031	-0.08 ppm	-0.08 ppm	B32023	±0.5 ppm NO,NOX,BALN	10 ppm	Unit online; Passed
8/16/2023 2:36 PM	CO ppm	Low	Span	9.03 ppm	6/13/2026	9.06 ppm	0.03 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
8/16/2023 2:36 PM	NH3/NOx ppm	High	Zero	0.0 ppm	6/9/2031	-0.4 ppm	-0.4 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	EPA Vendor ID	Allowable Drift	Instrument Span	Results
8/16/2023 2:36 PM	NH3/NOx ppm	High	Span	177.4 ppm	6/13/2031	175.9 ppm	-1.5 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
8/16/2023 2:36 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	6/9/2031	0.12 ppm	0.12 ppm	B72023	±0.5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
8/16/2023 2:36 PM	NH3/NOx ppm	Low	Span	9.13 ppm	6/13/2026	9.16 ppm	0.03 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
8/16/2023 2:36 PM	O2 %	Single	Zero	0.00%	6/13/2031	0.04%	0.04%	B32023	±1% NO,NOX,BALN	25%	Unit online; Passed
8/16/2023 2:36 PM	O2 %	Single	Span	22.54%	6/9/2031	22.62%	0.08%	B72023	±1% CO,O2,BALN	25%	Unit online; Passed
8/16/2023 2:36 PM	75-O2 %	Single	Zero	0.00%	6/13/2031	0.04%	0.04%	B32023	±1% NO,NOX,BALN	25%	Unit online; Passed
8/16/2023 2:36 PM	75-O2 %	Single	Span	22.54%	6/9/2031	22.62%	0.08%	B72023	±1% CO,O2,BALN	25%	Unit online; Passed
8/22/2023 6:36 AM	NOx ppm	High	Zero	0.0 ppm	6/9/2031	-0.5 ppm	-0.5 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
8/22/2023 6:36 AM	NOx ppm	High	Span	177.4 ppm	6/13/2031	178.2 ppm	0.8 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
8/22/2023 6:36 AM	NOx ppm	Low	Zero	0.00 ppm	6/9/2031	0.02 ppm	0.02 ppm	B72023	±0.5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
8/22/2023 6:36 AM	NOx ppm	Low	Span	9.13 ppm	6/13/2026	9.07 ppm	-0.06 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
8/22/2023 6:36 AM	75-NOx ppm	High	Zero	0.0 ppm	6/9/2031	-0.5 ppm	-0.5 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
8/22/2023 6:36 AM	75-NOx ppm	High	Span	177.4 ppm	6/13/2031	178.2 ppm	0.8 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
8/22/2023 6:36 AM	75-NOx ppm	Low	Zero	0.00 ppm	6/9/2031	0.02 ppm	0.02 ppm	B72023	±5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
8/22/2023 6:36 AM	75-NOx ppm	Low	Span	9.13 ppm	6/9/2031	9.07 ppm	-0.06 ppm	B32023	±5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
			Cylinder ID				EPA Vendor ID	EPA Gas Type Codes		
8/22/2023 6:36 AM	CO ppm	High	Zero	0.0 ppm	-0.5 ppm	6/13/2031	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC314637				B32023	NO,NOX,BALN		
8/22/2023 6:36 AM	CO ppm	High	Span	181.8 ppm	181.3 ppm	6/9/2031	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0141960				B72023	CO,O2,BALN		
8/22/2023 6:36 AM	CO ppm	Low	Zero	0.00 ppm	-0.36 ppm	6/13/2031	-0.36 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC314637				B32023	NO,NOX,BALN		
8/22/2023 6:36 AM	CO ppm	Low	Span	9.03 ppm	8.68 ppm	6/13/2026	-0.35 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC408913				B32023	CO,NO,NOX,BALN		
8/22/2023 6:36 AM	NH3/NOx ppm	High	Zero	0.0 ppm	-0.3 ppm	6/9/2031	-0.3 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0141960				B72023	CO,O2,BALN		
8/22/2023 6:36 AM	NH3/NOx ppm	High	Span	177.4 ppm	174.2 ppm	6/13/2031	-3.2 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC314637				B32023	NO,NOX,BALN		
8/22/2023 6:36 AM	NH3/NOx ppm	Low	Zero	0.00 ppm	0.14 ppm	6/9/2031	0.14 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			EB0141960				B72023	CO,O2,BALN		
8/22/2023 6:36 AM	NH3/NOx ppm	Low	Span	9.13 ppm	9.07 ppm	6/13/2026	-0.06 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC408913				B32023	CO,NO,NOX,BALN		
8/22/2023 6:36 AM	O2 %	Single	Zero	0.00%	0.05%	6/13/2031	0.05%	±1%	25%	Unit online; Passed
			CC314637				B32023	NO,NOX,BALN		
8/22/2023 6:36 AM	O2 %	Single	Span	22.54%	22.62%	6/9/2031	0.08%	±1%	25%	Unit online; Passed
			EB0141960				B72023	CO,O2,BALN		
8/22/2023 6:36 AM	75-O2 %	Single	Zero	0.00%	0.05%	6/13/2031	0.05%	±1%	25%	Unit online; Passed
			CC314637				B32023	NO,NOX,BALN		
8/22/2023 6:36 AM	75-O2 %	Single	Span	22.54%	22.62%	6/9/2031	0.08%	±1%	25%	Unit online; Passed
			EB0141960				B72023	CO,O2,BALN		
8/25/2023 6:06 PM	NOx ppm	High	Zero	0.0 ppm	-0.6 ppm	6/9/2031	-0.6 ppm	±10 ppm	200 ppm	Unit online; Passed
			EB0141960				B72023	CO,O2,BALN		
8/25/2023 6:06 PM	NOx ppm	High	Span	177.4 ppm	178.5 ppm	6/13/2031	1.1 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC314637				B32023	NO,NOX,BALN		
8/25/2023 6:06 PM	NOx ppm	Low	Zero	0.00 ppm	0.00 ppm	6/9/2031	0 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			EB0141960				B72023	CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
			Cylinder ID	Value	Value	EPA Vendor ID	EPA Gas Type Codes			
8/25/2023 6:06 PM	NOx ppm	Low	Span CC408913	9.13 ppm 6/13/2026	9.06 ppm B32023	-0.07 ppm	±0.5 ppm	10 ppm	CO,NO,NOX,BALN	Unit online; Passed
8/25/2023 6:06 PM	75-NOx ppm	High	Zero EB0141960	0.0 ppm 6/9/2031	-0.6 ppm B72023	-0.6 ppm	±10 ppm	200 ppm	CO,O2,BALN	Unit online; Passed
8/25/2023 6:06 PM	75-NOx ppm	High	Span CC314637	177.4 ppm 6/13/2031	178.5 ppm B32023	1.1 ppm	±10 ppm	200 ppm	NO,NOX,BALN	Unit online; Passed
8/25/2023 6:06 PM	75-NOx ppm	Low	Zero EB0141960	0.00 ppm 6/9/2031	0.00 ppm B72023	0 ppm	±5 ppm	10 ppm	CO,O2,BALN	Unit online; Passed
8/25/2023 6:06 PM	75-NOx ppm	Low	Span CC408913	9.13 ppm 6/13/2026	9.06 ppm B32023	-0.07 ppm	±5 ppm	10 ppm	CO,NO,NOX,BALN	Unit online; Passed
8/25/2023 6:06 PM	CO ppm	High	Zero CC314637	0.0 ppm 6/13/2031	-0.4 ppm B32023	-0.4 ppm	±10 ppm	200 ppm	NO,NOX,BALN	Unit online; Passed
8/25/2023 6:06 PM	CO ppm	High	Span EB0141960	181.8 ppm 6/9/2031	180.1 ppm B72023	-1.7 ppm	±10 ppm	200 ppm	CO,O2,BALN	Unit online; Passed
8/25/2023 6:06 PM	CO ppm	Low	Zero CC314637	0.00 ppm 6/13/2031	0.10 ppm B32023	0.1 ppm	±0.5 ppm	10 ppm	NO,NOX,BALN	Unit online; Passed
8/25/2023 6:06 PM	CO ppm	Low	Span CC408913	9.03 ppm 6/13/2026	9.02 ppm B32023	-0.01 ppm	±0.5 ppm	10 ppm	CO,NO,NOX,BALN	Unit online; Passed
8/25/2023 6:06 PM	NH3/NOx ppm	High	Zero EB0141960	0.0 ppm 6/9/2031	-0.3 ppm B72023	-0.3 ppm	±10 ppm	200 ppm	CO,O2,BALN	Unit online; Passed
8/25/2023 6:06 PM	NH3/NOx ppm	High	Span CC314637	177.4 ppm 6/13/2031	176.9 ppm B32023	-0.5 ppm	±10 ppm	200 ppm	NO,NOX,BALN	Unit online; Passed
8/25/2023 6:06 PM	NH3/NOx ppm	Low	Zero EB0141960	0.00 ppm 6/9/2031	0.17 ppm B72023	0.17 ppm	±0.5 ppm	10 ppm	CO,O2,BALN	Unit online; Passed
8/25/2023 6:06 PM	NH3/NOx ppm	Low	Span CC408913	9.13 ppm 6/13/2026	9.06 ppm B32023	-0.07 ppm	±0.5 ppm	10 ppm	CO,NO,NOX,BALN	Unit online; Passed
8/25/2023 6:06 PM	O2 %	Single	Zero CC314637	0.00% 6/13/2031	0.03% B32023	0.03%	±1%	25%	NO,NOX,BALN	Unit online; Passed
8/25/2023 6:06 PM	O2 %	Single	Span EB0141960	22.54% 6/9/2031	22.61% B72023	0.07%	±1%	25%	CO,O2,BALN	Unit online; Passed

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	EPA Vendor ID	Allowable Drift	Instrument Span	Results
8/25/2023 6:06 PM	75-O2 %	Single	Zero	0.00%	6/13/2031	0.03%	0.03%	B32023	±1%	25%	Unit online; Passed
			CC314637						NO,NOX,BALN		
8/25/2023 6:06 PM	75-O2 %	Single	Span	22.54%	6/9/2031	22.61%	0.07%	B72023	±1%	25%	Unit online; Passed
			EB0141960						CO,O2,BALN		
8/28/2023 5:36 PM	NOx ppm	High	Zero	0.0 ppm	6/9/2031	-0.6 ppm	-0.6 ppm	B72023	±10 ppm	200 ppm	Unit online; Passed
			EB0141960						CO,O2,BALN		
8/28/2023 5:36 PM	NOx ppm	High	Span	177.4 ppm	6/13/2031	178.3 ppm	0.9 ppm	B32023	±10 ppm	200 ppm	Unit online; Passed
			CC314637						NO,NOX,BALN		
8/28/2023 5:36 PM	NOx ppm	Low	Zero	0.00 ppm	6/9/2031	0.00 ppm	0 ppm	B72023	±0.5 ppm	10 ppm	Unit online; Passed
			EB0141960						CO,O2,BALN		
8/28/2023 5:36 PM	NOx ppm	Low	Span	9.13 ppm	6/13/2026	9.05 ppm	-0.08 ppm	B32023	±0.5 ppm	10 ppm	Unit online; Passed
			CC408913						CO,NO,NOX,BALN		
8/28/2023 5:36 PM	75-NOx ppm	High	Zero	0.0 ppm	6/9/2031	-0.6 ppm	-0.6 ppm	B72023	±10 ppm	200 ppm	Unit online; Passed
			EB0141960						CO,O2,BALN		
8/28/2023 5:36 PM	75-NOx ppm	High	Span	177.4 ppm	6/13/2031	178.3 ppm	0.9 ppm	B32023	±10 ppm	200 ppm	Unit online; Passed
			CC314637						NO,NOX,BALN		
8/28/2023 5:36 PM	75-NOx ppm	Low	Zero	0.00 ppm	6/9/2031	0.00 ppm	0 ppm	B72023	±5 ppm	10 ppm	Unit online; Passed
			EB0141960						CO,O2,BALN		
8/28/2023 5:36 PM	75-NOx ppm	Low	Span	9.13 ppm	6/13/2026	9.05 ppm	-0.08 ppm	B32023	±5 ppm	10 ppm	Unit online; Passed
			CC408913						CO,NO,NOX,BALN		
8/28/2023 5:36 PM	CO ppm	High	Zero	0.0 ppm	6/13/2031	-0.5 ppm	-0.5 ppm	B32023	±10 ppm	200 ppm	Unit online; Passed
			CC314637						NO,NOX,BALN		
8/28/2023 5:36 PM	CO ppm	High	Span	181.8 ppm	6/9/2031	181.8 ppm	0 ppm	B72023	±10 ppm	200 ppm	Unit online; Passed
			EB0141960						CO,O2,BALN		
8/28/2023 5:36 PM	CO ppm	Low	Zero	0.00 ppm	6/13/2031	-0.03 ppm	-0.03 ppm	B32023	±0.5 ppm	10 ppm	Unit online; Passed
			CC314637						NO,NOX,BALN		
8/28/2023 5:36 PM	CO ppm	Low	Span	9.03 ppm	6/13/2026	8.79 ppm	-0.24 ppm	B32023	±0.5 ppm	10 ppm	Unit online; Passed
			CC408913						CO,NO,NOX,BALN		
8/28/2023 5:36 PM	NH3/NOx ppm	High	Zero	0.0 ppm	6/9/2031	-0.3 ppm	-0.3 ppm	B72023	±10 ppm	200 ppm	Unit online; Passed
			EB0141960						CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	Allowable Drift	Instrument Span	Results
8/28/2023 5:36 PM	NH3/NOx ppm	High	Span	177.4 ppm	6/13/2031	176.5 ppm	-0.9 ppm	±10 ppm	200 ppm	Unit online; Passed
							B32023	NO,NOX,BALN		
8/28/2023 5:36 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	6/9/2031	0.17 ppm	0.17 ppm	±0.5 ppm	10 ppm	Unit online; Passed
							B72023	CO,O2,BALN		
8/28/2023 5:36 PM	NH3/NOx ppm	Low	Span	9.13 ppm	6/13/2026	9.02 ppm	-0.11 ppm	±0.5 ppm	10 ppm	Unit online; Passed
							B32023	CO,NO,NOX,BALN		
8/28/2023 5:36 PM	O2 %	Single	Zero	0.00%	6/13/2031	0.04%	0.04%	±1%	25%	Unit online; Passed
							B32023	NO,NOX,BALN		
8/28/2023 5:36 PM	O2 %	Single	Span	22.54%	6/9/2031	22.61%	0.07%	±1%	25%	Unit online; Passed
							B72023	CO,O2,BALN		
8/28/2023 5:36 PM	75-O2 %	Single	Zero	0.00%	6/13/2031	0.04%	0.04%	±1%	25%	Unit online; Passed
							B32023	NO,NOX,BALN		
8/28/2023 5:36 PM	75-O2 %	Single	Span	22.54%	6/9/2031	22.61%	0.07%	±1%	25%	Unit online; Passed
							B72023	CO,O2,BALN		
8/30/2023 5:51 PM	NOx ppm	High	Zero	0.0 ppm	6/26/2031	-0.5 ppm	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
							B72023	CO,O2,BALN		
8/30/2023 5:51 PM	NOx ppm	High	Span	181.2 ppm	7/24/2031	179.8 ppm	-1.4 ppm	±10 ppm	200 ppm	Unit online; Passed
							B32023	NO,NOX,BALN		
8/30/2023 5:51 PM	NOx ppm	Low	Zero	0.00 ppm	6/26/2031	0.01 ppm	0.01 ppm	±0.5 ppm	10 ppm	Unit online; Passed
							B72023	CO,O2,BALN		
8/30/2023 5:51 PM	NOx ppm	Low	Span	9.09 ppm	6/13/2026	9.01 ppm	-0.08 ppm	±0.5 ppm	10 ppm	Unit online; Passed
							B32023	CO,NO,NOX,BALN		
8/30/2023 5:51 PM	75-NOx ppm	High	Zero	0.0 ppm	6/26/2031	-0.5 ppm	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
							B72023	CO,O2,BALN		
8/30/2023 5:51 PM	75-NOx ppm	High	Span	181.2 ppm	7/24/2031	179.8 ppm	-1.4 ppm	±10 ppm	200 ppm	Unit online; Passed
							B32023	NO,NOX,BALN		
8/30/2023 5:51 PM	75-NOx ppm	Low	Zero	0.00 ppm	6/26/2031	0.01 ppm	0.01 ppm	±5 ppm	10 ppm	Unit online; Passed
							B72023	CO,O2,BALN		
8/30/2023 5:51 PM	75-NOx ppm	Low	Span	9.09 ppm	6/13/2026	9.01 ppm	-0.08 ppm	±5 ppm	10 ppm	Unit online; Passed
							B32023	CO,NO,NOX,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
		Cylinder ID				EPA Vendor ID	EPA Gas Type Codes			
8/30/2023 5:51 PM	CO ppm	High	Zero	0.0 ppm	-0.6 ppm	7/24/2031	-0.6 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC316597			B32023		NO,NOX,BALN		
8/30/2023 5:51 PM	CO ppm	High	Span	183.8 ppm	182.7 ppm	6/26/2031	-1.1 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM030038			B72023		CO,O2,BALN		
8/30/2023 5:51 PM	CO ppm	Low	Zero	0.00 ppm	-0.16 ppm	7/24/2031	-0.16 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC316597			B32023		NO,NOX,BALN		
8/30/2023 5:51 PM	CO ppm	Low	Span	9.11 ppm	9.31 ppm	6/13/2026	0.2 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC255556			B32023		CO,NO,NOX,BALN		
8/30/2023 5:51 PM	NH3/NOx ppm	High	Zero	0.0 ppm	-0.3 ppm	6/26/2031	-0.3 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM030038			B72023		CO,O2,BALN		
8/30/2023 5:51 PM	NH3/NOx ppm	High	Span	181.2 ppm	179.0 ppm	7/24/2031	-2.2 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC316597			B32023		NO,NOX,BALN		
8/30/2023 5:51 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	0.25 ppm	6/26/2031	0.25 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			ALM030038			B72023		CO,O2,BALN		
8/30/2023 5:51 PM	NH3/NOx ppm	Low	Span	9.09 ppm	9.32 ppm	6/13/2026	0.23 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC255556			B32023		CO,NO,NOX,BALN		
8/30/2023 5:51 PM	O2 %	Single	Zero	0.00%	0.04%	7/24/2031	0.04%	±1%	25%	Unit online; Passed
			CC316597			B32023		NO,NOX,BALN		
8/30/2023 5:51 PM	O2 %	Single	Span	22.50%	22.60%	6/26/2031	0.1%	±1%	25%	Unit online; Passed
			ALM030038			B72023		CO,O2,BALN		
8/30/2023 5:51 PM	75-O2 %	Single	Zero	0.00%	0.04%	7/24/2031	0.04%	±1%	25%	Unit online; Passed
			CC316597			B32023		NO,NOX,BALN		
8/30/2023 5:51 PM	75-O2 %	Single	Span	22.50%	22.60%	6/26/2031	0.1%	±1%	25%	Unit online; Passed
			ALM030038			B72023		CO,O2,BALN		
9/6/2023 5:37 PM	NOx ppm	High	Zero	0.0 ppm	-0.6 ppm	6/26/2031	-0.6 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM030038			B72023		CO,O2,BALN		
9/6/2023 5:37 PM	NOx ppm	High	Span	181.2 ppm	181.5 ppm	7/24/2031	0.3 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC316597			B32023		NO,NOX,BALN		
9/6/2023 5:37 PM	NOx ppm	Low	Zero	0.00 ppm	0.02 ppm	6/26/2031	0.02 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			ALM030038			B72023		CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
		Cylinder ID		Value	Value		EPA Vendor ID	EPA Gas Type Codes		
9/6/2023 5:37 PM	NOx ppm	Low	Span	9.09 ppm	9.11 ppm	6/13/2026	0.02 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC255556				B32023	CO,NO,NOX,BALN		
9/6/2023 5:37 PM	75-NOx ppm	High	Zero	0.0 ppm	-0.6 ppm	6/26/2031	-0.6 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		
9/6/2023 5:37 PM	75-NOx ppm	High	Span	181.2 ppm	181.5 ppm	7/24/2031	0.3 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC316597				B32023	NO,NOX,BALN		
9/6/2023 5:37 PM	75-NOx ppm	Low	Zero	0.00 ppm	0.02 ppm	6/26/2031	0.02 ppm	±5 ppm	10 ppm	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		
9/6/2023 5:37 PM	75-NOx ppm	Low	Span	9.09 ppm	9.11 ppm	6/13/2026	0.02 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC255556				B32023	CO,NO,NOX,BALN		
9/6/2023 5:37 PM	CO ppm	High	Zero	0.0 ppm	-0.4 ppm	7/24/2031	-0.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC316597				B32023	NO,NOX,BALN		
9/6/2023 5:37 PM	CO ppm	High	Span	183.8 ppm	182.2 ppm	6/26/2031	-1.6 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		
9/6/2023 5:37 PM	CO ppm	Low	Zero	0.00 ppm	0.04 ppm	7/24/2031	0.04 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC316597				B32023	NO,NOX,BALN		
9/6/2023 5:37 PM	CO ppm	Low	Span	9.11 ppm	9.29 ppm	6/13/2026	0.18 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC255556				B32023	CO,NO,NOX,BALN		
9/6/2023 5:37 PM	NH3/NOx ppm	High	Zero	0.0 ppm	-0.3 ppm	6/26/2031	-0.3 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		
9/6/2023 5:37 PM	NH3/NOx ppm	High	Span	181.2 ppm	177.7 ppm	7/24/2031	-3.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC316597				B32023	NO,NOX,BALN		
9/6/2023 5:37 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	0.23 ppm	6/26/2031	0.23 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		
9/6/2023 5:37 PM	NH3/NOx ppm	Low	Span	9.09 ppm	9.27 ppm	6/13/2026	0.18 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC255556				B32023	CO,NO,NOX,BALN		
9/6/2023 5:37 PM	O2 %	Single	Zero	0.00%	0.04%	7/24/2031	0.04%	±1%	25%	Unit online; Passed
			CC316597				B32023	NO,NOX,BALN		
9/6/2023 5:37 PM	O2 %	Single	Span	22.50%	22.59%	6/26/2031	0.09%	±1%	25%	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	EPA Vendor ID	Allowable Drift	Instrument Span	Results
9/6/2023 5:37 PM	75-O2 %	Single	Zero	0.00%	7/24/2031	0.04%	0.04%	B32023	±1% NO,NOX,BALN	25%	Unit online; Passed
9/6/2023 5:37 PM	75-O2 %	Single	Span	22.50%	6/26/2031	22.59%	0.09%	B72023	±1% CO,O2,BALN	25%	Unit online; Passed
9/7/2023 5:37 PM	NOx ppm	High	Zero	0.0 ppm	6/26/2031	-0.6 ppm	-0.6 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
9/7/2023 5:37 PM	NOx ppm	High	Span	181.2 ppm	7/24/2031	181.7 ppm	0.5 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
9/7/2023 5:37 PM	NOx ppm	Low	Zero	0.00 ppm	6/26/2031	0.01 ppm	0.01 ppm	B72023	±0.5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
9/7/2023 5:37 PM	NOx ppm	Low	Span	9.09 ppm	6/13/2026	9.08 ppm	-0.01 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
9/7/2023 5:37 PM	75-NOx ppm	High	Zero	0.0 ppm	6/26/2031	-0.6 ppm	-0.6 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
9/7/2023 5:37 PM	75-NOx ppm	High	Span	181.2 ppm	7/24/2031	181.7 ppm	0.5 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
9/7/2023 5:37 PM	75-NOx ppm	Low	Zero	0.00 ppm	6/26/2031	0.01 ppm	0.01 ppm	B72023	±5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
9/7/2023 5:37 PM	75-NOx ppm	Low	Span	9.09 ppm	6/13/2026	9.08 ppm	-0.01 ppm	B32023	±5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
9/7/2023 5:37 PM	CO ppm	High	Zero	0.0 ppm	7/24/2031	-0.5 ppm	-0.5 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
9/7/2023 5:37 PM	CO ppm	High	Span	183.8 ppm	6/26/2031	180.7 ppm	-3.1 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
9/7/2023 5:37 PM	CO ppm	Low	Zero	0.00 ppm	7/24/2031	-0.07 ppm	-0.07 ppm	B32023	±0.5 ppm NO,NOX,BALN	10 ppm	Unit online; Passed
9/7/2023 5:37 PM	CO ppm	Low	Span	9.11 ppm	6/13/2026	9.10 ppm	-0.01 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
9/7/2023 5:37 PM	NH3/NOx ppm	High	Zero	0.0 ppm	6/26/2031	-0.3 ppm	-0.3 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
		Cylinder ID		Value	Value		EPA Vendor ID	EPA Gas Type Codes		
9/7/2023 5:37 PM	NH3/NOx ppm	High	Span	181.2 ppm	177.7 ppm	7/24/2031	-3.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC316597				B32023	NO,NOX,BALN		
9/7/2023 5:37 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	0.14 ppm	6/26/2031	0.14 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		
9/7/2023 5:37 PM	NH3/NOx ppm	Low	Span	9.09 ppm	9.21 ppm	6/13/2026	0.12 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC255556				B32023	CO,NO,NOX,BALN		
9/7/2023 5:37 PM	O2 %	Single	Zero	0.00%	0.03%	7/24/2031	0.03%	±1%	25%	Unit online; Passed
			CC316597				B32023	NO,NOX,BALN		
9/7/2023 5:37 PM	O2 %	Single	Span	22.50%	22.60%	6/26/2031	0.1%	±1%	25%	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		
9/7/2023 5:37 PM	75-O2 %	Single	Zero	0.00%	0.03%	7/24/2031	0.03%	±1%	25%	Unit online; Passed
			CC316597				B32023	NO,NOX,BALN		
9/7/2023 5:37 PM	75-O2 %	Single	Span	22.50%	22.60%	6/26/2031	0.1%	±1%	25%	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		
9/8/2023 5:37 PM	NOx ppm	High	Zero	0.0 ppm	-0.5 ppm	6/26/2031	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		
9/8/2023 5:37 PM	NOx ppm	High	Span	181.2 ppm	181.6 ppm	7/24/2031	0.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC316597				B32023	NO,NOX,BALN		
9/8/2023 5:37 PM	NOx ppm	Low	Zero	0.00 ppm	0.01 ppm	6/26/2031	0.01 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		
9/8/2023 5:37 PM	NOx ppm	Low	Span	9.09 ppm	9.08 ppm	6/13/2026	-0.01 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC255556				B32023	CO,NO,NOX,BALN		
9/8/2023 5:37 PM	75-NOx ppm	High	Zero	0.0 ppm	-0.5 ppm	6/26/2031	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		
9/8/2023 5:37 PM	75-NOx ppm	High	Span	181.2 ppm	181.6 ppm	7/24/2031	0.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC316597				B32023	NO,NOX,BALN		
9/8/2023 5:37 PM	75-NOx ppm	Low	Zero	0.00 ppm	0.01 ppm	6/26/2031	0.01 ppm	±5 ppm	10 ppm	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		
9/8/2023 5:37 PM	75-NOx ppm	Low	Span	9.09 ppm	9.08 ppm	6/13/2026	-0.01 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC255556				B32023	CO,NO,NOX,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	Allowable Drift	Instrument Span	Results
		Cylinder ID					EPA Vendor ID	EPA Gas Type Codes		
9/8/2023 5:37 PM	CO ppm	High	Zero	0.0 ppm	7/24/2031	-0.5 ppm	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC316597				B32023	NO,NOX,BALN		
9/8/2023 5:37 PM	CO ppm	High	Span	183.8 ppm	6/26/2031	180.9 ppm	-2.9 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		
9/8/2023 5:37 PM	CO ppm	Low	Zero	0.00 ppm	7/24/2031	-0.09 ppm	-0.09 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC316597				B32023	NO,NOX,BALN		
9/8/2023 5:37 PM	CO ppm	Low	Span	9.11 ppm	6/13/2026	9.12 ppm	0.01 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC255556				B32023	CO,NO,NOX,BALN		
9/8/2023 5:37 PM	NH3/NOx ppm	High	Zero	0.0 ppm	6/26/2031	-0.3 ppm	-0.3 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		
9/8/2023 5:37 PM	NH3/NOx ppm	High	Span	181.2 ppm	7/24/2031	177.4 ppm	-3.8 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC316597				B32023	NO,NOX,BALN		
9/8/2023 5:37 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	6/26/2031	0.16 ppm	0.16 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		
9/8/2023 5:37 PM	NH3/NOx ppm	Low	Span	9.09 ppm	6/13/2026	9.20 ppm	0.11 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC255556				B32023	CO,NO,NOX,BALN		
9/8/2023 5:37 PM	O2 %	Single	Zero	0.00%	7/24/2031	0.04%	0.04%	±1%	25%	Unit online; Passed
			CC316597				B32023	NO,NOX,BALN		
9/8/2023 5:37 PM	O2 %	Single	Span	22.50%	6/26/2031	22.60%	0.1%	±1%	25%	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		
9/8/2023 5:37 PM	75-O2 %	Single	Zero	0.00%	7/24/2031	0.04%	0.04%	±1%	25%	Unit online; Passed
			CC316597				B32023	NO,NOX,BALN		
9/8/2023 5:37 PM	75-O2 %	Single	Span	22.50%	6/26/2031	22.60%	0.1%	±1%	25%	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		
9/9/2023 5:52 PM	NOx ppm	High	Zero	0.0 ppm	6/26/2031	-0.5 ppm	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		
9/9/2023 5:52 PM	NOx ppm	High	Span	181.2 ppm	7/24/2031	181.3 ppm	0.1 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC316597				B32023	NO,NOX,BALN		
9/9/2023 5:52 PM	NOx ppm	Low	Zero	0.00 ppm	6/26/2031	0.02 ppm	0.02 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
		Cylinder ID				EPA Vendor ID	EPA Gas Type Codes			
9/9/2023 5:52 PM	NOx ppm	Low	Span	9.09 ppm	9.10 ppm	6/13/2026	0.01 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC255556			B32023		CO,NO,NOX,BALN		
9/9/2023 5:52 PM	75-NOx ppm	High	Zero	0.0 ppm	-0.5 ppm	6/26/2031	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM030038			B72023		CO,O2,BALN		
9/9/2023 5:52 PM	75-NOx ppm	High	Span	181.2 ppm	181.3 ppm	7/24/2031	0.1 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC316597			B32023		NO,NOX,BALN		
9/9/2023 5:52 PM	75-NOx ppm	Low	Zero	0.00 ppm	0.02 ppm	6/26/2031	0.02 ppm	±5 ppm	10 ppm	Unit online; Passed
			ALM030038			B72023		CO,O2,BALN		
9/9/2023 5:52 PM	75-NOx ppm	Low	Span	9.09 ppm	9.10 ppm	6/13/2026	0.01 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC255556			B32023		CO,NO,NOX,BALN		
9/9/2023 5:52 PM	CO ppm	High	Zero	0.0 ppm	-0.3 ppm	7/24/2031	-0.3 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC316597			B32023		NO,NOX,BALN		
9/9/2023 5:52 PM	CO ppm	High	Span	183.8 ppm	183.0 ppm	6/26/2031	-0.8 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM030038			B72023		CO,O2,BALN		
9/9/2023 5:52 PM	CO ppm	Low	Zero	0.00 ppm	0.12 ppm	7/24/2031	0.12 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC316597			B32023		NO,NOX,BALN		
9/9/2023 5:52 PM	CO ppm	Low	Span	9.11 ppm	9.37 ppm	6/13/2026	0.26 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC255556			B32023		CO,NO,NOX,BALN		
9/9/2023 5:52 PM	NH3/NOx ppm	High	Zero	0.0 ppm	-0.4 ppm	6/26/2031	-0.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM030038			B72023		CO,O2,BALN		
9/9/2023 5:52 PM	NH3/NOx ppm	High	Span	181.2 ppm	176.9 ppm	7/24/2031	-4.3 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC316597			B32023		NO,NOX,BALN		
9/9/2023 5:52 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	0.11 ppm	6/26/2031	0.11 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			ALM030038			B72023		CO,O2,BALN		
9/9/2023 5:52 PM	NH3/NOx ppm	Low	Span	9.09 ppm	9.17 ppm	6/13/2026	0.08 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC255556			B32023		CO,NO,NOX,BALN		
9/9/2023 5:52 PM	O2 %	Single	Zero	0.00%	0.05%	7/24/2031	0.05%	±1%	25%	Unit online; Passed
			CC316597			B32023		NO,NOX,BALN		
9/9/2023 5:52 PM	O2 %	Single	Span	22.50%	22.59%	6/26/2031	0.09%	±1%	25%	Unit online; Passed
			ALM030038			B72023		CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	EPA Vendor ID	Actual Drift	Allowable Drift	Instrument Span	Results
		Cylinder ID							EPA Gas Type Codes		
9/9/2023 5:52 PM	75-O2 %	Single	Zero	0.00%	0.05%	7/24/2031	B32023	0.05%	±1%	25%	Unit online; Passed
			CC316597						NO,NOX,BALN		
9/9/2023 5:52 PM	75-O2 %	Single	Span	22.50%	22.59%	6/26/2031	B72023	0.09%	±1%	25%	Unit online; Passed
			ALM030038						CO,O2,BALN		
9/10/2023 5:52 PM	NOx ppm	High	Zero	0.0 ppm	-0.5 ppm	6/26/2031	B72023	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM030038						CO,O2,BALN		
9/10/2023 5:52 PM	NOx ppm	High	Span	181.2 ppm	181.3 ppm	7/24/2031	B32023	0.1 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC316597						NO,NOX,BALN		
9/10/2023 5:52 PM	NOx ppm	Low	Zero	0.00 ppm	0.01 ppm	6/26/2031	B72023	0.01 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			ALM030038						CO,O2,BALN		
9/10/2023 5:52 PM	NOx ppm	Low	Span	9.09 ppm	9.08 ppm	6/13/2026	B32023	-0.01 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC255556						CO,NO,NOX,BALN		
9/10/2023 5:52 PM	75-NOx ppm	High	Zero	0.0 ppm	-0.5 ppm	6/26/2031	B72023	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM030038						CO,O2,BALN		
9/10/2023 5:52 PM	75-NOx ppm	High	Span	181.2 ppm	181.3 ppm	7/24/2031	B32023	0.1 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC316597						NO,NOX,BALN		
9/10/2023 5:52 PM	75-NOx ppm	Low	Zero	0.00 ppm	0.01 ppm	6/26/2031	B72023	0.01 ppm	±5 ppm	10 ppm	Unit online; Passed
			ALM030038						CO,O2,BALN		
9/10/2023 5:52 PM	75-NOx ppm	Low	Span	9.09 ppm	9.08 ppm	6/13/2026	B32023	-0.01 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC255556						CO,NO,NOX,BALN		
9/10/2023 5:52 PM	CO ppm	High	Zero	0.0 ppm	-0.4 ppm	7/24/2031	B32023	-0.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC316597						NO,NOX,BALN		
9/10/2023 5:52 PM	CO ppm	High	Span	183.8 ppm	181.3 ppm	6/26/2031	B72023	-2.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM030038						CO,O2,BALN		
9/10/2023 5:52 PM	CO ppm	Low	Zero	0.00 ppm	0.07 ppm	7/24/2031	B32023	0.07 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC316597						NO,NOX,BALN		
9/10/2023 5:52 PM	CO ppm	Low	Span	9.11 ppm	9.32 ppm	6/13/2026	B32023	0.21 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC255556						CO,NO,NOX,BALN		
9/10/2023 5:52 PM	NH3/NOx ppm	High	Zero	0.0 ppm	-0.3 ppm	6/26/2031	B72023	-0.3 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM030038						CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
9/10/2023 5:52 PM	NH3/NOx ppm	High	Span	181.2 ppm	176.8 ppm	7/24/2031	-4.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC316597			B32023		NO,NOX,BALN		
9/10/2023 5:52 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	0.19 ppm	6/26/2031	0.19 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			ALM030038			B72023		CO,O2,BALN		
9/10/2023 5:52 PM	NH3/NOx ppm	Low	Span	9.09 ppm	9.20 ppm	6/13/2026	0.11 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC255556			B32023		CO,NO,NOX,BALN		
9/10/2023 5:52 PM	O2 %	Single	Zero	0.00%	0.05%	7/24/2031	0.05%	±1%	25%	Unit online; Passed
			CC316597			B32023		NO,NOX,BALN		
9/10/2023 5:52 PM	O2 %	Single	Span	22.50%	22.60%	6/26/2031	0.1%	±1%	25%	Unit online; Passed
			ALM030038			B72023		CO,O2,BALN		
9/10/2023 5:52 PM	75-O2 %	Single	Zero	0.00%	0.05%	7/24/2031	0.05%	±1%	25%	Unit online; Passed
			CC316597			B32023		NO,NOX,BALN		
9/10/2023 5:52 PM	75-O2 %	Single	Span	22.50%	22.60%	6/26/2031	0.1%	±1%	25%	Unit online; Passed
			ALM030038			B72023		CO,O2,BALN		
9/14/2023 5:52 PM	NOx ppm	High	Zero	0.0 ppm	-0.4 ppm	6/26/2031	-0.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM030038			B72023		CO,O2,BALN		
9/14/2023 5:52 PM	NOx ppm	High	Span	181.2 ppm	181.3 ppm	7/24/2031	0.1 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC316597			B32023		NO,NOX,BALN		
9/14/2023 5:52 PM	NOx ppm	Low	Zero	0.00 ppm	0.03 ppm	6/26/2031	0.03 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			ALM030038			B72023		CO,O2,BALN		
9/14/2023 5:52 PM	NOx ppm	Low	Span	9.09 ppm	9.07 ppm	6/13/2026	-0.02 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC255556			B32023		CO,NO,NOX,BALN		
9/14/2023 5:52 PM	75-NOx ppm	High	Zero	0.0 ppm	-0.4 ppm	6/26/2031	-0.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM030038			B72023		CO,O2,BALN		
9/14/2023 5:52 PM	75-NOx ppm	High	Span	181.2 ppm	181.3 ppm	7/24/2031	0.1 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC316597			B32023		NO,NOX,BALN		
9/14/2023 5:52 PM	75-NOx ppm	Low	Zero	0.00 ppm	0.03 ppm	6/26/2031	0.03 ppm	±5 ppm	10 ppm	Unit online; Passed
			ALM030038			B72023		CO,O2,BALN		
9/14/2023 5:52 PM	75-NOx ppm	Low	Span	9.09 ppm	9.07 ppm	6/13/2026	-0.02 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC255556			B32023		CO,NO,NOX,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	EPA Vendor ID	Allowable Drift	Instrument Span	Results
9/14/2023 5:52 PM	CO ppm	High	Zero	0.0 ppm	-0.4 ppm	7/24/2031	-0.4 ppm	B32023	±10 ppm	200 ppm	Unit online; Passed
			CC316597						NO,NOX,BALN		
9/14/2023 5:52 PM	CO ppm	High	Span	183.8 ppm	183.6 ppm	6/26/2031	-0.2 ppm	B72023	±10 ppm	200 ppm	Unit online; Passed
			ALM030038						CO,O2,BALN		
9/14/2023 5:52 PM	CO ppm	Low	Zero	0.00 ppm	0.03 ppm	7/24/2031	0.03 ppm	B32023	±0.5 ppm	10 ppm	Unit online; Passed
			CC316597						NO,NOX,BALN		
9/14/2023 5:52 PM	CO ppm	Low	Span	9.11 ppm	9.26 ppm	6/13/2026	0.15 ppm	B32023	±0.5 ppm	10 ppm	Unit online; Passed
			CC255556						CO,NO,NOX,BALN		
9/14/2023 5:52 PM	NH3/NOx ppm	High	Zero	0.0 ppm	-0.3 ppm	6/26/2031	-0.3 ppm	B72023	±10 ppm	200 ppm	Unit online; Passed
			ALM030038						CO,O2,BALN		
9/14/2023 5:52 PM	NH3/NOx ppm	High	Span	181.2 ppm	180.1 ppm	7/24/2031	-1.1 ppm	B32023	±10 ppm	200 ppm	Unit online; Passed
			CC316597						NO,NOX,BALN		
9/14/2023 5:52 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	0.18 ppm	6/26/2031	0.18 ppm	B72023	±0.5 ppm	10 ppm	Unit online; Passed
			ALM030038						CO,O2,BALN		
9/14/2023 5:52 PM	NH3/NOx ppm	Low	Span	9.09 ppm	9.16 ppm	6/13/2026	0.07 ppm	B32023	±0.5 ppm	10 ppm	Unit online; Passed
			CC255556						CO,NO,NOX,BALN		
9/14/2023 5:52 PM	O2 %	Single	Zero	0.00%	0.05%	7/24/2031	0.05%	B32023	±1%	25%	Unit online; Passed
			CC316597						NO,NOX,BALN		
9/14/2023 5:52 PM	O2 %	Single	Span	22.50%	22.60%	6/26/2031	0.1%	B72023	±1%	25%	Unit online; Passed
			ALM030038						CO,O2,BALN		
9/14/2023 5:52 PM	75-O2 %	Single	Zero	0.00%	0.05%	7/24/2031	0.05%	B32023	±1%	25%	Unit online; Passed
			CC316597						NO,NOX,BALN		
9/14/2023 5:52 PM	75-O2 %	Single	Span	22.50%	22.60%	6/26/2031	0.1%	B72023	±1%	25%	Unit online; Passed
			ALM030038						CO,O2,BALN		
9/15/2023 5:07 PM	NOx ppm	High	Zero	0.0 ppm	-0.5 ppm	6/26/2031	-0.5 ppm	B72023	±10 ppm	200 ppm	Unit online; Passed
			ALM030038						CO,O2,BALN		
9/15/2023 5:07 PM	NOx ppm	High	Span	181.2 ppm	181.3 ppm	7/24/2031	0.1 ppm	B32023	±10 ppm	200 ppm	Unit online; Passed
			CC316597						NO,NOX,BALN		
9/15/2023 5:07 PM	NOx ppm	Low	Zero	0.00 ppm	0.02 ppm	6/26/2031	0.02 ppm	B72023	±0.5 ppm	10 ppm	Unit online; Passed
			ALM030038						CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
		Cylinder ID				EPA Vendor ID	EPA Gas Type Codes			
9/15/2023 5:07 PM	NOx ppm	Low	Span	9.09 ppm	9.09 ppm	6/13/2026	0 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC255556			B32023		CO,NO,NOX,BALN		
9/15/2023 5:07 PM	75-NOx ppm	High	Zero	0.0 ppm	-0.5 ppm	6/26/2031	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM030038			B72023		CO,O2,BALN		
9/15/2023 5:07 PM	75-NOx ppm	High	Span	181.2 ppm	181.3 ppm	7/24/2031	0.1 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC316597			B32023		NO,NOX,BALN		
9/15/2023 5:07 PM	75-NOx ppm	Low	Zero	0.00 ppm	0.02 ppm	6/26/2031	0.02 ppm	±5 ppm	10 ppm	Unit online; Passed
			ALM030038			B72023		CO,O2,BALN		
9/15/2023 5:07 PM	75-NOx ppm	Low	Span	9.09 ppm	9.09 ppm	6/13/2026	0 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC255556			B32023		CO,NO,NOX,BALN		
9/15/2023 5:07 PM	CO ppm	High	Zero	0.0 ppm	-0.4 ppm	7/24/2031	-0.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC316597			B32023		NO,NOX,BALN		
9/15/2023 5:07 PM	CO ppm	High	Span	183.8 ppm	183.0 ppm	6/26/2031	-0.8 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM030038			B72023		CO,O2,BALN		
9/15/2023 5:07 PM	CO ppm	Low	Zero	0.00 ppm	0.05 ppm	7/24/2031	0.05 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC316597			B32023		NO,NOX,BALN		
9/15/2023 5:07 PM	CO ppm	Low	Span	9.11 ppm	9.20 ppm	6/13/2026	0.09 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC255556			B32023		CO,NO,NOX,BALN		
9/15/2023 5:07 PM	NH3/NOx ppm	High	Zero	0.0 ppm	-0.4 ppm	6/26/2031	-0.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM030038			B72023		CO,O2,BALN		
9/15/2023 5:07 PM	NH3/NOx ppm	High	Span	181.2 ppm	179.9 ppm	7/24/2031	-1.3 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC316597			B32023		NO,NOX,BALN		
9/15/2023 5:07 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	0.11 ppm	6/26/2031	0.11 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			ALM030038			B72023		CO,O2,BALN		
9/15/2023 5:07 PM	NH3/NOx ppm	Low	Span	9.09 ppm	9.11 ppm	6/13/2026	0.02 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC255556			B32023		CO,NO,NOX,BALN		
9/15/2023 5:07 PM	O2 %	Single	Zero	0.00%	0.06%	7/24/2031	0.06%	±1%	25%	Unit online; Passed
			CC316597			B32023		NO,NOX,BALN		
9/15/2023 5:07 PM	O2 %	Single	Span	22.50%	22.60%	6/26/2031	0.1%	±1%	25%	Unit online; Passed
			ALM030038			B72023		CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	Allowable Drift	Instrument Span	Results
		Cylinder ID					EPA Vendor ID	EPA Gas Type Codes		
9/15/2023 5:07 PM	75-O2 %	Single	Zero	0.00%	7/24/2031	0.06%	0.06%	±1%	25%	Unit online; Passed
			CC316597				B32023	NO,NOX,BALN		
9/15/2023 5:07 PM	75-O2 %	Single	Span	22.50%	6/26/2031	22.60%	0.1%	±1%	25%	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		
9/18/2023 5:37 PM	NOx ppm	High	Zero	0.0 ppm	6/26/2031	-0.5 ppm	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		
9/18/2023 5:37 PM	NOx ppm	High	Span	181.2 ppm	7/24/2031	181.8 ppm	0.6 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC316597				B32023	NO,NOX,BALN		
9/18/2023 5:37 PM	NOx ppm	Low	Zero	0.00 ppm	6/26/2031	0.02 ppm	0.02 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		
9/18/2023 5:37 PM	NOx ppm	Low	Span	9.09 ppm	6/13/2026	9.07 ppm	-0.02 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC255556				B32023	CO,NO,NOX,BALN		
9/18/2023 5:37 PM	75-NOx ppm	High	Zero	0.0 ppm	6/26/2031	-0.5 ppm	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		
9/18/2023 5:37 PM	75-NOx ppm	High	Span	181.2 ppm	7/24/2031	181.8 ppm	0.6 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC316597				B32023	NO,NOX,BALN		
9/18/2023 5:37 PM	75-NOx ppm	Low	Zero	0.00 ppm	6/26/2031	0.02 ppm	0.02 ppm	±5 ppm	10 ppm	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		
9/18/2023 5:37 PM	75-NOx ppm	Low	Span	9.09 ppm	6/13/2026	9.07 ppm	-0.02 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC255556				B32023	CO,NO,NOX,BALN		
9/18/2023 5:37 PM	CO ppm	High	Zero	0.0 ppm	7/24/2031	-0.5 ppm	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC316597				B32023	NO,NOX,BALN		
9/18/2023 5:37 PM	CO ppm	High	Span	183.8 ppm	6/26/2031	183.2 ppm	-0.6 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		
9/18/2023 5:37 PM	CO ppm	Low	Zero	0.00 ppm	7/24/2031	-0.06 ppm	-0.06 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC316597				B32023	NO,NOX,BALN		
9/13/2023 5:37 PM	CO ppm	Low	Span	9.11 ppm	6/13/2026	9.02 ppm	-0.09 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC255556				B32023	CO,NO,NOX,BALN		
9/18/2023 5:37 PM	NH3/NOx ppm	High	Zero	0.0 ppm	6/26/2031	-0.4 ppm	-0.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	Allowable Drift	Instrument Span	Results
9/18/2023 5:37 PM	NH3/NOx ppm	High	Span	181.2 ppm	7/24/2031	180.1 ppm	-1.1 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC316597				B32023	NO,NOX,BALN		
9/18/2023 5:37 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	6/26/2031	0.17 ppm	0.17 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		
9/18/2023 5:37 PM	NH3/NOx ppm	Low	Span	9.09 ppm	6/13/2026	9.15 ppm	0.06 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC255556				B32023	CO,NO,NOX,BALN		
9/18/2023 5:37 PM	O2 %	Single	Zero	0.00%	7/24/2031	0.06%	0.06%	±1%	25%	Unit online; Passed
			CC316597				B32023	NO,NOX,BALN		
9/18/2023 5:37 PM	O2 %	Single	Span	22.50%	6/26/2031	22.62%	0.12%	±1%	25%	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		
9/18/2023 5:37 PM	75-O2 %	Single	Zero	0.00%	7/24/2031	0.06%	0.06%	±1%	25%	Unit online; Passed
			CC316597				B32023	NO,NOX,BALN		
9/18/2023 5:37 PM	75-O2 %	Single	Span	22.50%	6/26/2031	22.62%	0.12%	±1%	25%	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		
9/21/2023 6:52 PM	NOx ppm	High	Zero	0.0 ppm	6/26/2031	-0.4 ppm	-0.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		
9/21/2023 6:52 PM	NOx ppm	High	Span	181.2 ppm	7/24/2031	181.2 ppm	0 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC316597				B32023	NO,NOX,BALN		
9/21/2023 6:52 PM	NOx ppm	Low	Zero	0.00 ppm	6/26/2031	0.05 ppm	0.05 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		
9/21/2023 6:52 PM	NOx ppm	Low	Span	9.09 ppm	6/13/2026	9.14 ppm	0.05 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC255556				B32023	CO,NO,NOX,BALN		
9/21/2023 6:52 PM	75-NOx ppm	High	Zero	0.0 ppm	6/26/2031	-0.4 ppm	-0.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		
9/21/2023 6:52 PM	75-NOx ppm	High	Span	181.2 ppm	7/24/2031	181.2 ppm	0 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC316597				B32023	NO,NOX,BALN		
9/21/2023 6:52 PM	75-NOx ppm	Low	Zero	0.00 ppm	6/26/2031	0.05 ppm	0.05 ppm	±5 ppm	10 ppm	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		
9/21/2023 6:52 PM	75-NOx ppm	Low	Span	9.09 ppm	6/13/2026	9.14 ppm	0.05 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC255556				B32023	CO,NO,NOX,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
		Cylinder ID				EPA Vendor ID	EPA Gas Type Codes			
9/21/2023 6:52 PM	CO ppm	High	Zero	0.0 ppm	-0.6 ppm	7/24/2031	-0.6 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC316597			B32023		NO,NOX,BALN		
9/21/2023 6:52 PM	CO ppm	High	Span	183.8 ppm	181.9 ppm	6/26/2031	-1.9 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM030038			B72023		CO,O2,BALN		
9/21/2023 6:52 PM	CO ppm	Low	Zero	0.00 ppm	0.11 ppm	7/24/2031	0.11 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC316597			B32023		NO,NOX,BALN		
9/21/2023 6:52 PM	CO ppm	Low	Span	9.11 ppm	9.39 ppm	6/13/2026	0.28 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC255556			B32023		CO,NO,NOX,BALN		
9/21/2023 6:52 PM	NH3/NOx ppm	High	Zero	0.0 ppm	-0.4 ppm	6/26/2031	-0.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM030038			B72023		CO,O2,BALN		
9/21/2023 6:52 PM	NH3/NOx ppm	High	Span	181.2 ppm	179.1 ppm	7/24/2031	-2.1 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC316597			B32023		NO,NOX,BALN		
9/21/2023 6:52 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	0.11 ppm	6/26/2031	0.11 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			ALM030038			B72023		CO,O2,BALN		
9/21/2023 6:52 PM	NH3/NOx ppm	Low	Span	9.09 ppm	9.08 ppm	6/13/2026	-0.01 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC255556			B32023		CO,NO,NOX,BALN		
9/21/2023 6:52 PM	O2 %	Single	Zero	0.00%	0.06%	7/24/2031	0.06%	±1%	25%	Unit online; Passed
			CC316597			B32023		NO,NOX,BALN		
9/21/2023 6:52 PM	O2 %	Single	Span	22.50%	22.60%	6/26/2031	0.1%	±1%	25%	Unit online; Passed
			ALM030038			B72023		CO,O2,BALN		
9/21/2023 6:52 PM	75-O2 %	Single	Zero	0.00%	0.06%	7/24/2031	0.06%	±1%	25%	Unit online; Passed
			CC316597			B32023		NO,NOX,BALN		
9/21/2023 6:52 PM	75-O2 %	Single	Span	22.50%	22.60%	6/26/2031	0.1%	±1%	25%	Unit online; Passed
			ALM030038			B72023		CO,O2,BALN		
9/26/2023 8:36 AM	NOx ppm	High	Zero	0.0 ppm	-0.4 ppm	6/26/2031	-0.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM030038			B72023		CO,O2,BALN		
9/26/2023 8:36 AM	NOx ppm	High	Span	181.2 ppm	181.3 ppm	7/24/2031	0.1 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC316597			B32023		NO,NOX,BALN		
9/25/2023 8:36 AM	NOx ppm	Low	Zero	0.00 ppm	0.04 ppm	6/26/2031	0.04 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			ALM030038			B72023		CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	Allowable Drift	Instrument Span	Results
		Cylinder ID					EPA Vendor ID	EPA Gas Type Codes		
9/26/2023 8:36 AM	NOx ppm	Low	Span	9.09 ppm	6/13/2026	9.14 ppm	0.05 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC255556				B32023	CO,NO,NOX,BALN		
9/26/2023 8:36 AM	75-NOx ppm	High	Zero	0.0 ppm	6/26/2031	-0.4 ppm	-0.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		
9/26/2023 8:36 AM	75-NOx ppm	High	Span	181.2 ppm	7/24/2031	181.3 ppm	0.1 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC316597				B32023	NO,NOX,BALN		
9/26/2023 8:36 AM	75-NOx ppm	Low	Zero	0.00 ppm	6/26/2031	0.04 ppm	0.04 ppm	±5 ppm	10 ppm	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		
9/25/2023 8:36 AM	75-NOx ppm	Low	Span	9.09 ppm	6/13/2026	9.14 ppm	0.05 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC255556				B32023	CO,NO,NOX,BALN		
9/25/2023 8:36 AM	CO ppm	High	Zero	0.0 ppm	7/24/2031	-0.6 ppm	-0.6 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC316597				B32023	NO,NOX,BALN		
9/25/2023 8:36 AM	CO ppm	High	Span	183.8 ppm	6/26/2031	182.5 ppm	-1.3 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		
9/25/2023 8:36 AM	CO ppm	Low	Zero	0.00 ppm	7/24/2031	0.06 ppm	0.06 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC316597				B32023	NO,NOX,BALN		
9/26/2023 8:36 AM	CO ppm	Low	Span	9.11 ppm	6/13/2026	9.27 ppm	0.16 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC255556				B32023	CO,NO,NOX,BALN		
9/26/2023 8:36 AM	NH3/NOx ppm	High	Zero	0.0 ppm	6/26/2031	-0.4 ppm	-0.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		
9/26/2023 8:36 AM	NH3/NOx ppm	High	Span	181.2 ppm	7/24/2031	178.7 ppm	-2.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC316597				B32023	NO,NOX,BALN		
9/26/2023 8:36 AM	NH3/NOx ppm	Low	Zero	0.00 ppm	6/26/2031	0.14 ppm	0.14 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		
9/26/2023 8:36 AM	NH3/NOx ppm	Low	Span	9.09 ppm	6/13/2026	9.08 ppm	-0.01 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC255556				B32023	CO,NO,NOX,BALN		
9/26/2023 8:36 AM	O2 %	Single	Zero	0.00%	7/24/2031	0.06%	0.06%	±1%	25%	Unit online; Passed
			CC316597				B32023	NO,NOX,BALN		
9/26/2023 8:36 AM	O2 %	Single	Span	22.50%	6/26/2031	22.61%	0.11%	±1%	25%	Unit online; Passed
			ALM030038				B72023	CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	EPA Vendor ID	Allowable Drift	Instrument Span	Results
9/23/2023 8:36 AM	75-O2 %	Single	Zero	0.00%	7/24/2031	0.06%	0.06%	B32023	±1% NO,NOX,BALN	25%	Unit online; Passed
9/23/2023 8:36 AM	75-O2 %	Single	Span	22.50%	6/26/2031	22.61%	0.11%	B72023	±1% CO,O2,BALN	25%	Unit online; Passed
10/4/2023 5:25 PM	NOx ppm	High	Zero	0.0 ppm	7/5/2031	-0.4 ppm	-0.4 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
10/4/2023 5:25 PM	NOx ppm	High	Span	176.4 ppm	6/13/2031	178.5 ppm	2.1 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
10/4/2023 5:25 PM	NOx ppm	Low	Zero	0.00 ppm	7/5/2031	0.03 ppm	0.03 ppm	B72023	±0.5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
10/4/2023 5:25 PM	NOx ppm	Low	Span	9.11 ppm	7/13/2026	9.16 ppm	0.05 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
10/4/2023 5:25 PM	75-NOx ppm	High	Zero	0.0 ppm	7/5/2031	-0.4 ppm	-0.4 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
10/4/2023 5:25 PM	75-NOx ppm	High	Span	176.4 ppm	6/13/2031	178.5 ppm	2.1 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
10/4/2023 5:25 PM	75-NOx ppm	Low	Zero	0.00 ppm	7/5/2031	0.03 ppm	0.03 ppm	B72023	±5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
10/4/2023 5:25 PM	75-NOx ppm	Low	Span	9.11 ppm	7/13/2026	9.16 ppm	0.05 ppm	B32023	±5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
10/4/2023 5:25 PM	CO ppm	High	Zero	0.0 ppm	6/13/2031	-0.4 ppm	-0.4 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
10/4/2023 5:25 PM	CO ppm	High	Span	182.4 ppm	7/5/2031	179.6 ppm	-2.8 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
10/4/2023 5:25 PM	CO ppm	Low	Zero	0.00 ppm	6/13/2031	-0.13 ppm	-0.13 ppm	B32023	±0.5 ppm NO,NOX,BALN	10 ppm	Unit online; Passed
10/4/2023 5:25 PM	CO ppm	Low	Span	9.10 ppm	7/13/2026	8.91 ppm	-0.19 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
10/4/2023 5:25 PM	NH3/NOx ppm	High	Zero	0.0 ppm	7/5/2031	-0.3 ppm	-0.3 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	EPA Vendor ID	Allowable Drift	Instrument Span	Results
10/4/2023 5:25 PM	NH3/NOx ppm	High	Span	176.4 ppm	6/13/2031	178.0 ppm	1.6 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
10/4/2023 5:25 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	7/5/2031	0.17 ppm	0.17 ppm	B72023	±0.5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
10/4/2023 5:25 PM	NH3/NOx ppm	Low	Span	9.11 ppm	7/13/2026	9.24 ppm	0.13 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
10/4/2023 5:25 PM	O2 %	Single	Zero	0.00%	6/13/2031	-0.06%	-0.06%	B32023	±1% NO,NOX,BALN	25%	Unit online; Passed
10/4/2023 5:25 PM	O2 %	Single	Span	22.52%	7/5/2031	22.54%	0.02%	B72023	±1% CO,O2,BALN	25%	Unit online; Passed
10/4/2023 5:25 PM	75-O2 %	Single	Zero	0.00%	6/13/2031	-0.06%	-0.06%	B32023	±1% NO,NOX,BALN	25%	Unit online; Passed
10/4/2023 5:25 PM	75-O2 %	Single	Span	22.52%	7/5/2031	22.54%	0.02%	B72023	±1% CO,O2,BALN	25%	Unit online; Passed
10/5/2023 5:07 PM	NOx ppm	High	Zero	0.0 ppm	7/5/2031	-0.5 ppm	-0.5 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
10/5/2023 5:07 PM	NOx ppm	High	Span	176.4 ppm	6/13/2031	177.7 ppm	1.3 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
10/5/2023 5:07 PM	NOx ppm	Low	Zero	0.00 ppm	7/5/2031	0.00 ppm	0 ppm	B72023	±0.5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
10/5/2023 5:07 PM	NOx ppm	Low	Span	9.11 ppm	7/13/2026	9.04 ppm	-0.07 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
10/5/2023 5:07 PM	75-NOx ppm	High	Zero	0.0 ppm	7/5/2031	-0.5 ppm	-0.5 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
10/5/2023 5:07 PM	75-NOx ppm	High	Span	176.4 ppm	6/13/2031	177.7 ppm	1.3 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
10/5/2023 5:07 PM	75-NOx ppm	Low	Zero	0.00 ppm	7/5/2031	0.00 ppm	0 ppm	B72023	±5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
10/5/2023 5:07 PM	75-NOx ppm	Low	Span	9.11 ppm	7/13/2026	9.04 ppm	-0.07 ppm	B32023	±5 ppm CO,O2,BALN	10 ppm	Unit online; Passed

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
			Cylinder ID			EPA Vendor ID	EPA Gas Type Codes			
10/5/2023 5:07 PM	CO ppm	High	Zero	0.0 ppm	-0.4 ppm	6/13/2031	-0.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM-040085			B32023		NO,NOX,BALN		
10/5/2023 5:07 PM	CO ppm	High	Span	182.4 ppm	180.9 ppm	7/5/2031	-1.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC158302			B72023		CO,O2,BALN		
10/5/2023 5:07 PM	CO ppm	Low	Zero	0.00 ppm	-0.21 ppm	6/13/2031	-0.21 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			ALM-040085			B32023		NO,NOX,BALN		
10/5/2023 5:07 PM	CO ppm	Low	Span	9.10 ppm	8.92 ppm	7/13/2026	-0.18 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC285248			B32023		CO,NO,NOX,BALN		
10/5/2023 5:07 PM	NH3/NOx ppm	High	Zero	0.0 ppm	-0.3 ppm	7/5/2031	-0.3 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC158302			B72023		CO,O2,BALN		
10/5/2023 5:07 PM	NH3/NOx ppm	High	Span	176.4 ppm	177.8 ppm	6/13/2031	1.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM-040085			B32023		NO,NOX,BALN		
10/5/2023 5:07 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	0.22 ppm	7/5/2031	0.22 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC158302			B72023		CO,O2,BALN		
10/5/2023 5:07 PM	NH3/NOx ppm	Low	Span	9.11 ppm	9.26 ppm	7/13/2026	0.15 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC285248			B32023		CO,NO,NOX,BALN		
10/5/2023 5:07 PM	O2 %	Single	Zero	0.00%	-0.06%	6/13/2031	-0.06%	±1%	25%	Unit online; Passed
			ALM-040085			B32023		NO,NOX,BALN		
10/5/2023 5:07 PM	O2 %	Single	Span	22.52%	22.54%	7/5/2031	0.02%	±1%	25%	Unit online; Passed
			CC158302			B72023		CO,O2,BALN		
10/5/2023 5:07 PM	75-O2 %	Single	Zero	0.00%	-0.06%	6/13/2031	-0.06%	±1%	25%	Unit online; Passed
			ALM-040085			B32023		NO,NOX,BALN		
10/5/2023 5:07 PM	75-O2 %	Single	Span	22.52%	22.54%	7/5/2031	0.02%	±1%	25%	Unit online; Passed
			CC158302			B72023		CO,O2,BALN		
10/6/2023 5:07 PM	NOx ppm	High	Zero	0.0 ppm	-0.5 ppm	7/5/2031	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC158302			B72023		CO,O2,BALN		
10/6/2023 5:07 PM	NOx ppm	High	Span	176.4 ppm	177.9 ppm	6/13/2031	1.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM-040085			B32023		NO,NOX,BALN		
10/6/2023 5:07 PM	NOx ppm	Low	Zero	0.00 ppm	0.00 ppm	7/5/2031	0 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC158302			B72023		CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
		Cylinder ID		Value	Value		EPA Vendor ID	EPA Gas Type Codes		
10/6/2023 5:07 PM	NOx ppm	Low	Span	9.11 ppm	9.04 ppm	7/13/2026	-0.07 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC285248				B32023	CO,NO,NOX,BALN		
10/6/2023 5:07 PM	75-NOx ppm	High	Zero	0.0 ppm	-0.5 ppm		-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC158302			7/5/2031	B72023	CO,O2,BALN		
10/6/2023 5:07 PM	75-NOx ppm	High	Span	176.4 ppm	177.9 ppm		1.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM-040085			6/13/2031	B32023	NO,NOX,BALN		
10/6/2023 5:07 PM	75-NOx ppm	Low	Zero	0.00 ppm	0.00 ppm		0 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC158302			7/5/2031	B72023	CO,O2,BALN		
10/6/2023 5:07 PM	75-NOx ppm	Low	Span	9.11 ppm	9.04 ppm		-0.07 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC285248			7/13/2026	B32023	CO,NO,NOX,BALN		
10/6/2023 5:07 PM	CO ppm	High	Zero	0.0 ppm	-0.4 ppm		-0.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM-040085			6/13/2031	B32023	NO,NOX,BALN		
10/6/2023 5:07 PM	CO ppm	High	Span	182.4 ppm	180.8 ppm		-1.6 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC158302			7/5/2031	B72023	CO,O2,BALN		
10/6/2023 5:07 PM	CO ppm	Low	Zero	0.00 ppm	-0.21 ppm		-0.21 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			ALM-040085			6/13/2031	B32023	NO,NOX,BALN		
10/6/2023 5:07 PM	CO ppm	Low	Span	9.10 ppm	9.11 ppm		0.01 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC285248			7/13/2026	B32023	CO,NO,NOX,BALN		
10/6/2023 5:07 PM	NH3/NOx ppm	High	Zero	0.0 ppm	-0.3 ppm		-0.3 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC158302			7/5/2031	B72023	CO,O2,BALN		
10/6/2023 5:07 PM	NH3/NOx ppm	High	Span	176.4 ppm	177.7 ppm		1.3 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM-040085			6/13/2031	B32023	NO,NOX,BALN		
10/6/2023 5:07 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	0.18 ppm		0.18 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC158302			7/5/2031	B72023	CO,O2,BALN		
10/6/2023 5:07 PM	NH3/NOx ppm	Low	Span	9.11 ppm	9.21 ppm		0.1 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC285248			7/13/2026	B32023	CO,NO,NOX,BALN		
10/6/2023 5:07 PM	O2 %	Single	Zero	0.00%	-0.06%		-0.06%	±1%	25%	Unit online; Passed
			ALM-040085			6/13/2031	B32023	NO,NOX,BALN		
10/6/2023 5:07 PM	O2 %	Single	Span	22.52%	22.55%		0.03%	±1%	25%	Unit online; Passed
			CC158302			7/5/2031	B72023	CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	EPA Vendor ID	Allowable Drift	Instrument Span	Results
10/16/2023 5:07 PM	75-O2 %	Single	Zero	0.00%	6/13/2031	-0.06%	-0.06%	B32023	±1% NO,NOX,BALN	25%	Unit online; Passed
10/16/2023 5:07 PM	75-O2 %	Single	Span	22.52%	7/5/2031	22.55%	0.03%	B72023	±1% CO,O2,BALN	25%	Unit online; Passed
10/16/2023 5:07 PM	NOx ppm	High	Zero	0.0 ppm	7/5/2031	-0.5 ppm	-0.5 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
10/16/2023 5:07 PM	NOx ppm	High	Span	176.4 ppm	6/13/2031	178.4 ppm	2 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
10/16/2023 5:07 PM	NOx ppm	Low	Zero	0.00 ppm	7/5/2031	0.02 ppm	0.02 ppm	B72023	±0.5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
10/16/2023 5:07 PM	NOx ppm	Low	Span	9.11 ppm	7/13/2026	9.14 ppm	0.03 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
10/16/2023 5:07 PM	75-NOx ppm	High	Zero	0.0 ppm	7/5/2031	-0.5 ppm	-0.5 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
10/16/2023 5:07 PM	75-NOx ppm	High	Span	176.4 ppm	6/13/2031	178.4 ppm	2 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
10/16/2023 5:07 PM	75-NOx ppm	Low	Zero	0.00 ppm	7/5/2031	0.02 ppm	0.02 ppm	B72023	±5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
10/16/2023 5:07 PM	75-NOx ppm	Low	Span	9.11 ppm	7/13/2026	9.14 ppm	0.03 ppm	B32023	±5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
10/16/2023 5:07 PM	CO ppm	High	Zero	0.0 ppm	6/13/2031	-0.2 ppm	-0.2 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
10/16/2023 5:07 PM	CO ppm	High	Span	182.4 ppm	7/5/2031	179.3 ppm	-3.1 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
10/16/2023 5:07 PM	CO ppm	Low	Zero	0.00 ppm	6/13/2031	0.13 ppm	0.13 ppm	B32023	±0.5 ppm NO,NOX,BALN	10 ppm	Unit online; Passed
10/16/2023 5:07 PM	CO ppm	Low	Span	9.10 ppm	7/13/2026	9.25 ppm	0.15 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
10/16/2023 5:07 PM	NH3/NOx ppm	High	Zero	0.0 ppm	7/5/2031	-0.3 ppm	-0.3 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
			Cylinder ID				EPA Vendor ID	EPA Gas Type	Codes	
10/16/2023 5:07 PM	NH3/NOx ppm	High	Span	176.4 ppm	175.4 ppm	6/13/2031	-1 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM-040085				B32023	NO,NOX,BALN		
10/16/2023 5:07 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	0.24 ppm	7/5/2031	0.24 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC158302				B72023	CO,O2,BALN		
10/16/2023 5:07 PM	NH3/NOx ppm	Low	Span	9.11 ppm	9.16 ppm	7/13/2026	0.05 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC285248				B32023	CO,NO,NOX,BALN		
10/16/2023 5:07 PM	O2 %	Single	Zero	0.00%	-0.06%	6/13/2031	-0.06%	±1%	25%	Unit online; Passed
			ALM-040085				B32023	NO,NOX,BALN		
10/16/2023 5:07 PM	O2 %	Single	Span	22.52%	22.55%	7/5/2031	0.03%	±1%	25%	Unit online; Passed
			CC158302				B72023	CO,O2,BALN		
10/16/2023 5:07 PM	75-O2 %	Single	Zero	0.00%	-0.06%	6/13/2031	-0.06%	±1%	25%	Unit online; Passed
			ALM-040085				B32023	NO,NOX,BALN		
10/16/2023 5:07 PM	75-O2 %	Single	Span	22.52%	22.55%	7/5/2031	0.03%	±1%	25%	Unit online; Passed
			CC158302				B72023	CO,O2,BALN		
10/18/2023 5:07 PM	NOx ppm	High	Zero	0.0 ppm	-0.5 ppm	7/5/2031	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC158302				B72023	CO,O2,BALN		
10/18/2023 5:07 PM	NOx ppm	High	Span	176.4 ppm	178.1 ppm	6/13/2031	1.7 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM-040085				B32023	NO,NOX,BALN		
10/18/2023 5:07 PM	NOx ppm	Low	Zero	0.00 ppm	0.01 ppm	7/5/2031	0.01 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC158302				B72023	CO,O2,BALN		
10/18/2023 5:07 PM	NOx ppm	Low	Span	9.11 ppm	9.07 ppm	7/13/2026	-0.04 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC285248				B32023	CO,NO,NOX,BALN		
10/18/2023 5:07 PM	75-NOx ppm	High	Zero	0.0 ppm	-0.5 ppm	7/5/2031	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC158302				B72023	CO,O2,BALN		
10/18/2023 5:07 PM	75-NOx ppm	High	Span	176.4 ppm	178.1 ppm	6/13/2031	1.7 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM-040085				B32023	NO,NOX,BALN		
10/18/2023 5:07 PM	75-NOx ppm	Low	Zero	0.00 ppm	0.01 ppm	7/5/2031	0.01 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC158302				B72023	CO,O2,BALN		
10/18/2023 5:07 PM	75-NOx ppm	Low	Span	9.11 ppm	9.07 ppm	7/13/2026	-0.04 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC285248				B32023	CO,NO,NOX,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	EPA Vendor ID	Actual Drift	Allowable Drift	Instrument Span	Results
			Cylinder ID						EPA Gas Type Codes		
10/18/2023 5:07 PM	CO ppm	High	Zero	0.0 ppm	-0.2 ppm	6/13/2031	B32023	-0.2 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM-040085						NO,NOX,BALN		
10/18/2023 5:07 PM	CO ppm	High	Span	182.4 ppm	179.9 ppm	7/5/2031	B72023	-2.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC158302						CO,O2,BALN		
10/18/2023 5:07 PM	CO ppm	Low	Zero	0.00 ppm	0.08 ppm	6/13/2031	B32023	0.08 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			ALM-040085						NO,NOX,BALN		
10/18/2023 5:07 PM	CO ppm	Low	Span	9.10 ppm	9.28 ppm	7/13/2026	B32023	0.18 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC285248						CO,NO,NOX,BALN		
10/18/2023 5:07 PM	NH3/NOx ppm	High	Zero	0.0 ppm	-0.3 ppm	7/5/2031	B72023	-0.3 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC158302						CO,O2,BALN		
10/18/2023 5:07 PM	NH3/NOx ppm	High	Span	176.4 ppm	175.0 ppm	6/13/2031	B32023	-1.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM-040085						NO,NOX,BALN		
10/18/2023 5:07 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	0.21 ppm	7/5/2031	B72023	0.21 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC158302						CO,O2,BALN		
10/18/2023 5:07 PM	NH3/NOx ppm	Low	Span	9.11 ppm	9.11 ppm	7/13/2026	B32023	0 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC285248						CO,NO,NOX,BALN		
10/18/2023 5:07 PM	O2 %	Single	Zero	0.00%	-0.06%	6/13/2031	B32023	-0.06%	±1%	25%	Unit online; Passed
			ALM-040085						NO,NOX,BALN		
10/18/2023 5:07 PM	O2 %	Single	Span	22.52%	22.55%	7/5/2031	B72023	0.03%	±1%	25%	Unit online; Passed
			CC158302						CO,O2,BALN		
10/18/2023 5:07 PM	75-O2 %	Single	Zero	0.00%	-0.06%	6/13/2031	B32023	-0.06%	±1%	25%	Unit online; Passed
			ALM-040085						NO,NOX,BALN		
10/18/2023 5:07 PM	75-O2 %	Single	Span	22.52%	22.55%	7/5/2031	B72023	0.03%	±1%	25%	Unit online; Passed
			CC158302						CO,O2,BALN		
10/19/2023 5:08 PM	NOx ppm	High	Zero	0.0 ppm	-0.5 ppm	7/5/2031	B72023	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC158302						CO,O2,BALN		
10/19/2023 5:08 PM	NOx ppm	High	Span	176.4 ppm	178.5 ppm	6/13/2031	B32023	2.1 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM-040085						NO,NOX,BALN		
10/19/2023 5:08 PM	NOx ppm	Low	Zero	0.00 ppm	0.01 ppm	7/5/2031	B72023	0.01 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC158302						CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
			Cylinder ID	Value	Value		EPA Vendor ID	EPA Gas Type	Codes	
10/19/2023 5:08 PM	NOx ppm	Low	Span	9.11 ppm	9.05 ppm	7/13/2026	-0.06 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC285248				B32023	CO,NO,NOX,BALN		
10/19/2023 5:08 PM	75-NOx ppm	High	Zero	0.0 ppm	-0.5 ppm	7/5/2031	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC158302				B72023	CO,O2,BALN		
10/19/2023 5:08 PM	75-NOx ppm	High	Span	176.4 ppm	178.5 ppm	6/13/2031	2.1 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM-040085				B32023	NO,NOX,BALN		
10/19/2023 5:08 PM	75-NOx ppm	Low	Zero	0.00 ppm	0.01 ppm	7/5/2031	0.01 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC158302				B72023	CO,O2,BALN		
10/19/2023 5:08 PM	75-NOx ppm	Low	Span	9.11 ppm	9.05 ppm	7/13/2026	-0.06 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC285248				B32023	CO,NO,NOX,BALN		
10/19/2023 5:08 PM	CO ppm	High	Zero	0.0 ppm	-0.3 ppm	6/13/2031	-0.3 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM-040085				B32023	NO,NOX,BALN		
10/19/2023 5:08 PM	CO ppm	High	Span	182.4 ppm	180.6 ppm	7/5/2031	-1.8 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC158302				B72023	CO,O2,BALN		
10/19/2023 5:08 PM	CO ppm	Low	Zero	0.00 ppm	-0.05 ppm	6/13/2031	-0.05 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			ALM-040085				B32023	NO,NOX,BALN		
10/19/2023 5:08 PM	CO ppm	Low	Span	9.10 ppm	9.05 ppm	7/13/2026	-0.05 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC285248				B32023	CO,NO,NOX,BALN		
10/19/2023 5:08 PM	NH3/NOx ppm	High	Zero	0.0 ppm	-0.3 ppm	7/5/2031	-0.3 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC158302				B72023	CO,O2,BALN		
10/19/2023 5:08 PM	NH3/NOx ppm	High	Span	176.4 ppm	175.5 ppm	6/13/2031	-0.9 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM-040085				B32023	NO,NOX,BALN		
10/19/2023 5:08 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	0.17 ppm	7/5/2031	0.17 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC158302				B72023	CO,O2,BALN		
10/19/2023 5:08 PM	NH3/NOx ppm	Low	Span	9.11 ppm	9.08 ppm	7/13/2026	-0.03 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC285248				B32023	CO,NO,NOX,BALN		
10/19/2023 5:08 PM	O2 %	Single	Zero	0.00%	-0.07%	6/13/2031	-0.07%	±1%	25%	Unit online; Passed
			ALM-040085				B32023	NO,NOX,BALN		
10/19/2023 5:08 PM	O2 %	Single	Span	22.52%	22.55%	7/5/2031	0.03%	±1%	25%	Unit online; Passed
			CC158302				B72023	CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	EPA Vendor ID	Allowable Drift	Instrument Span	Results
10/19/2023 5:08 PM	75-O2 %	Single	Zero	0.00%	6/13/2031	-0.07%	-0.07%	B32023	±1% NO,NOX,BALN	25%	Unit online; Passed
10/19/2023 5:08 PM	75-O2 %	Single	Span	22.52%	7/5/2031	22.55%	0.03%	B72023	±1% CO,O2,BALN	25%	Unit online; Passed
10/26/2023 6:23 PM	NOx ppm	High	Zero	0.0 ppm	7/5/2031	-0.4 ppm	-0.4 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
10/26/2023 6:23 PM	NOx ppm	High	Span	176.4 ppm	6/13/2031	177.9 ppm	1.5 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
10/26/2023 6:23 PM	NOx ppm	Low	Zero	0.00 ppm	7/5/2031	0.03 ppm	0.03 ppm	B72023	±0.5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
10/26/2023 6:23 PM	NOx ppm	Low	Span	9.11 ppm	7/13/2026	9.11 ppm	0 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
10/26/2023 6:23 PM	75-NOx ppm	High	Zero	0.0 ppm	7/5/2031	-0.4 ppm	-0.4 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
10/26/2023 6:23 PM	75-NOx ppm	High	Span	176.4 ppm	6/13/2031	177.9 ppm	1.5 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
10/26/2023 6:23 PM	75-NOx ppm	Low	Zero	0.00 ppm	7/5/2031	0.03 ppm	0.03 ppm	B72023	±5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
10/26/2023 6:23 PM	75-NOx ppm	Low	Span	9.11 ppm	7/13/2026	9.11 ppm	0 ppm	B32023	±5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
10/26/2023 6:23 PM	CO ppm	High	Zero	0.0 ppm	6/13/2031	-0.2 ppm	-0.2 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
10/26/2023 6:23 PM	CO ppm	High	Span	182.4 ppm	7/5/2031	181.5 ppm	-0.9 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
10/26/2023 6:23 PM	CO ppm	Low	Zero	0.00 ppm	6/13/2031	0.02 ppm	0.02 ppm	B32023	±0.5 ppm NO,NOX,BALN	10 ppm	Unit online; Passed
10/26/2023 6:23 PM	CO ppm	Low	Span	9.10 ppm	7/13/2026	9.13 ppm	0.03 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
10/26/2023 6:23 PM	NH3/NOx ppm	High	Zero	0.0 ppm	7/5/2031	-0.4 ppm	-0.4 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	Allowable Drift	Instrument Span	Results
		Cylinder ID					EPA Vendor ID	EPA Gas Type	Codes	
10/26/2023 6:23 PM	NH3/NOx ppm	High	Span	176.4 ppm	6/13/2031	174.5 ppm	-1.9 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM-040085				B32023	NO,NOX,BALN		
10/26/2023 6:23 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	7/5/2031	0.13 ppm	0.13 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC158302				B72023	CO,O2,BALN		
10/26/2023 6:23 PM	NH3/NOx ppm	Low	Span	9.11 ppm	7/13/2026	9.32 ppm	0.21 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC285248				B32023	CO,NO,NOX,BALN		
10/26/2023 6:23 PM	O2 %	Single	Zero	0.00%	6/13/2031	-0.06%	-0.06%	±1%	25%	Unit online; Passed
			ALM-040085				B32023	NO,NOX,BALN		
10/26/2023 6:23 PM	O2 %	Single	Span	22.52%	7/5/2031	22.54%	0.02%	±1%	25%	Unit online; Passed
			CC158302				B72023	CO,O2,BALN		
10/26/2023 6:23 PM	75-O2 %	Single	Zero	0.00%	6/13/2031	-0.06%	-0.06%	±1%	25%	Unit online; Passed
			ALM-040085				B32023	NO,NOX,BALN		
10/26/2023 6:23 PM	75-O2 %	Single	Span	22.52%	7/5/2031	22.54%	0.02%	±1%	25%	Unit online; Passed
			CC158302				B72023	CO,O2,BALN		
10/29/2023 6:22 PM	NOx ppm	High	Zero	0.0 ppm	7/5/2031	-0.5 ppm	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC158302				B72023	CO,O2,BALN		
10/29/2023 6:22 PM	NOx ppm	High	Span	176.4 ppm	6/13/2031	178.6 ppm	2.2 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM-040085				B32023	NO,NOX,BALN		
10/29/2023 6:22 PM	NOx ppm	Low	Zero	0.00 ppm	7/5/2031	0.01 ppm	0.01 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC158302				B72023	CO,O2,BALN		
10/29/2023 6:22 PM	NOx ppm	Low	Span	9.11 ppm	7/13/2026	9.12 ppm	0.01 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC285248				B32023	CO,NO,NOX,BALN		
10/29/2023 6:22 PM	75-NOx ppm	High	Zero	0.0 ppm	7/5/2031	-0.5 ppm	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC158302				B72023	CO,O2,BALN		
10/29/2023 6:22 PM	75-NOx ppm	High	Span	176.4 ppm	6/13/2031	178.6 ppm	2.2 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM-040085				B32023	NO,NOX,BALN		
10/29/2023 6:22 PM	75-NOx ppm	Low	Zero	0.00 ppm	7/5/2031	0.01 ppm	0.01 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC158302				B72023	CO,O2,BALN		
10/29/2023 6:22 PM	75-NOx ppm	Low	Span	9.11 ppm	7/13/2026	9.12 ppm	0.01 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC285248				B32023	CO,NO,NOX,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	EPA Vendor ID	Allowable Drift	Instrument Span	Results
10/29/2023 6:22 PM	CO ppm	High	Zero	0.0 ppm	6/13/2031	-0.1 ppm	-0.1 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
10/29/2023 6:22 PM	CO ppm	High	Span	182.4 ppm	7/5/2031	179.9 ppm	-2.5 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
10/29/2023 6:22 PM	CO ppm	Low	Zero	0.00 ppm	6/13/2031	0.15 ppm	0.15 ppm	B32023	±0.5 ppm NO,NOX,BALN	10 ppm	Unit online; Passed
10/29/2023 6:22 PM	CO ppm	Low	Span	9.10 ppm	7/13/2026	9.16 ppm	0.06 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
10/29/2023 6:22 PM	NH3/NOx ppm	High	Zero	0.0 ppm	7/5/2031	-0.4 ppm	-0.4 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
10/29/2023 6:22 PM	NH3/NOx ppm	High	Span	176.4 ppm	6/13/2031	174.9 ppm	-1.5 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
10/29/2023 6:22 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	7/5/2031	0.12 ppm	0.12 ppm	B72023	±0.5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
10/29/2023 6:22 PM	NH3/NOx ppm	Low	Span	9.11 ppm	7/13/2026	9.34 ppm	0.23 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
10/29/2023 6:22 PM	O2 %	Single	Zero	0.00%	6/13/2031	-0.09%	-0.09%	B32023	±1% NO,NOX,BALN	25%	Unit online; Passed
10/29/2023 6:22 PM	O2 %	Single	Span	22.52%	7/5/2031	22.53%	0.01%	B72023	±1% CO,O2,BALN	25%	Unit online; Passed
10/29/2023 6:22 PM	75-O2 %	Single	Zero	0.00%	6/13/2031	-0.09%	-0.09%	B32023	±1% NO,NOX,BALN	25%	Unit online; Passed
10/29/2023 6:22 PM	75-O2 %	Single	Span	22.52%	7/5/2031	22.53%	0.01%	B72023	±1% CO,O2,BALN	25%	Unit online; Passed
10/30/2023 6:07 AM	NOx ppm	High	Zero	0.0 ppm	7/5/2031	-0.4 ppm	-0.4 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
10/30/2023 6:07 AM	NOx ppm	High	Span	176.4 ppm	6/13/2031	179.3 ppm	2.9 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
10/30/2023 6:07 AM	NOx ppm	Low	Zero	0.00 ppm	7/5/2031	0.02 ppm	0.02 ppm	B72023	±0.5 ppm CO,O2,BALN	10 ppm	Unit online; Passed

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	EPA Vendor ID	Allowable Drift	Instrument Span	Results
10/30/2023 6:07 AM	NOx ppm	Low	Span	9.11 ppm	7/13/2026	9.15 ppm	0.04 ppm	B32023	±0.5 ppm	10 ppm	Unit online; Passed
			CC285248						CO,NO,NOX,BALN		
10/30/2023 6:07 AM	75-NOx ppm	High	Zero	0.0 ppm	7/5/2031	-0.4 ppm	-0.4 ppm	B72023	±10 ppm	200 ppm	Unit online; Passed
			CC158302						CO,O2,BALN		
10/30/2023 6:07 AM	75-NOx ppm	High	Span	176.4 ppm	6/13/2031	179.3 ppm	2.9 ppm	B32023	±10 ppm	200 ppm	Unit online; Passed
			ALM-040085						NO,NOX,BALN		
10/30/2023 6:07 AM	75-NOx ppm	Low	Zero	0.00 ppm	7/5/2031	0.02 ppm	0.02 ppm	B72023	±5 ppm	10 ppm	Unit online; Passed
			CC158302						CO,O2,BALN		
10/30/2023 6:07 AM	75-NOx ppm	Low	Span	9.11 ppm	7/13/2026	9.15 ppm	0.04 ppm	B32023	±5 ppm	10 ppm	Unit online; Passed
			CC285248						CO,NO,NOX,BALN		
10/30/2023 6:07 AM	CO ppm	High	Zero	0.0 ppm	6/13/2031	-0.1 ppm	-0.1 ppm	B32023	±10 ppm	200 ppm	Unit online; Passed
			ALM-040085						NO,NOX,BALN		
10/30/2023 6:07 AM	CO ppm	High	Span	182.4 ppm	7/5/2031	179.8 ppm	-2.6 ppm	B72023	±10 ppm	200 ppm	Unit online; Passed
			CC158302						CO,O2,BALN		
10/30/2023 6:07 AM	CO ppm	Low	Zero	0.00 ppm	6/13/2031	0.15 ppm	0.15 ppm	B32023	±0.5 ppm	10 ppm	Unit online; Passed
			ALM-040085						NO,NOX,BALN		
10/30/2023 6:07 AM	CO ppm	Low	Span	9.10 ppm	7/13/2026	9.18 ppm	0.08 ppm	B32023	±0.5 ppm	10 ppm	Unit online; Passed
			CC285248						CO,NO,NOX,BALN		
10/30/2023 6:07 AM	NH3/NOx ppm	High	Zero	0.0 ppm	7/5/2031	-0.5 ppm	-0.5 ppm	B72023	±10 ppm	200 ppm	Unit online; Passed
			CC158302						CO,O2,BALN		
10/30/2023 6:07 AM	NH3/NOx ppm	High	Span	176.4 ppm	6/13/2031	175.8 ppm	-0.6 ppm	B32023	±10 ppm	200 ppm	Unit online; Passed
			ALM-040085						NO,NOX,BALN		
10/30/2023 6:07 AM	NH3/NOx ppm	Low	Zero	0.00 ppm	7/5/2031	0.06 ppm	0.06 ppm	B72023	±0.5 ppm	10 ppm	Unit online; Passed
			CC158302						CO,O2,BALN		
10/30/2023 6:07 AM	NH3/NOx ppm	Low	Span	9.11 ppm	7/13/2026	9.37 ppm	0.26 ppm	B32023	±0.5 ppm	10 ppm	Unit online; Passed
			CC285248						CO,NO,NOX,BALN		
10/30/2023 6:07 AM	O2 %	Single	Zero	0.00%	6/13/2031	-0.09%	-0.09%	B32023	±1%	25%	Unit online; Passed
			ALM-040085						NO,NOX,BALN		
10/30/2023 6:07 AM	O2 %	Single	Span	22.52%	7/5/2031	22.53%	0.01%	B72023	±1%	25%	Unit online; Passed
			CC158302						CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	EPA Vendor ID	Actual Drift	Allowable Drift	Instrument Span	Results
			Cylinder ID						EPA Gas Type Codes		
10/30/2023 6:07 AM	75-O2 %	Single	Zero	0.00%	-0.09%	6/13/2031	B32023	-0.09%	±1%	25%	Unit online; Passed
			ALM-040085						NO,NOX,BALN		
10/30/2023 6:07 AM	75-O2 %	Single	Span	22.52%	22.53%	7/5/2031	B72023	0.01%	±1%	25%	Unit online; Passed
			CC158302						CO,O2,BALN		
11/1/2023 5:22 PM	NOx ppm	High	Zero	0.0 ppm	-0.6 ppm	7/5/2031	B72023	-0.6 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC158302						CO,O2,BALN		
11/1/2023 5:22 PM	NOx ppm	High	Span	176.4 ppm	175.8 ppm	6/13/2031	B32023	-0.6 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM-040085						NO,NOX,BALN		
11/1/2023 5:22 PM	NOx ppm	Low	Zero	0.00 ppm	0.00 ppm	7/5/2031	B72023	0 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC158302						CO,O2,BALN		
11/1/2023 5:22 PM	NOx ppm	Low	Span	9.11 ppm	9.03 ppm	7/13/2026	B32023	-0.08 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC285248						CO,NO,NOX,BALN		
11/1/2023 5:22 PM	75-NOx ppm	High	Zero	0.0 ppm	-0.6 ppm	7/5/2031	B72023	-0.6 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC158302						CO,O2,BALN		
11/1/2023 5:22 PM	75-NOx ppm	High	Span	176.4 ppm	175.8 ppm	6/13/2031	B32023	-0.6 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM-040085						NO,NOX,BALN		
11/1/2023 5:22 PM	75-NOx ppm	Low	Zero	0.00 ppm	0.00 ppm	7/5/2031	B72023	0 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC158302						CO,O2,BALN		
11/1/2023 5:22 PM	75-NOx ppm	Low	Span	9.11 ppm	9.03 ppm	7/13/2026	B32023	-0.08 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC285248						CO,NO,NOX,BALN		
11/1/2023 5:22 PM	CO ppm	High	Zero	0.0 ppm	0.0 ppm	6/13/2031	B32023	0 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM-040085						NO,NOX,BALN		
11/1/2023 5:22 PM	CO ppm	High	Span	182.4 ppm	181.3 ppm	7/5/2031	B72023	-1.1 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC158302						CO,O2,BALN		
11/1/2023 5:22 PM	CO ppm	Low	Zero	0.00 ppm	0.32 ppm	6/13/2031	B32023	0.32 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			ALM-040085						NO,NOX,BALN		
11/1/2023 5:22 PM	CO ppm	Low	Span	9.10 ppm	9.17 ppm	7/13/2026	B32023	0.07 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC285248						CO,NO,NOX,BALN		
11/1/2023 5:22 PM	NH3/NOx ppm	High	Zero	0.0 ppm	-0.4 ppm	7/5/2031	B72023	-0.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC158302						CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
		Cylinder ID		Value	Value		EPA Vendor ID	EPA Gas Type	Codes	
11/1/2023 5:22 PM	NH3/NOx ppm	High	Span	176.4 ppm	174.8 ppm	6/13/2031	-1.6 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM-040085				B32023	NO,NOX,BALN		
11/1/2023 5:22 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	0.14 ppm	7/5/2031	0.14 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC158302				B72023	CO,O2,BALN		
11/1/2023 5:22 PM	NH3/NOx ppm	Low	Span	9.11 ppm	9.39 ppm	7/13/2026	0.28 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC285248				B32023	CO,NO,NOX,BALN		
11/1/2023 5:22 PM	O2 %	Single	Zero	0.00%	-0.07%	6/13/2031	-0.07%	±1%	25%	Unit online; Passed
			ALM-040085				B32023	NO,NOX,BALN		
11/1/2023 5:22 PM	O2 %	Single	Span	22.52%	22.54%	7/5/2031	0.02%	±1%	25%	Unit online; Passed
			CC158302				B72023	CO,O2,BALN		
11/1/2023 5:22 PM	75-O2 %	Single	Zero	0.00%	-0.07%	6/13/2031	-0.07%	±1%	25%	Unit online; Passed
			ALM-040085				B32023	NO,NOX,BALN		
11/1/2023 5:22 PM	75-O2 %	Single	Span	22.52%	22.54%	7/5/2031	0.02%	±1%	25%	Unit online; Passed
			CC158302				B72023	CO,O2,BALN		
11/6/2023 5:07 PM	NOx ppm	High	Zero	0.0 ppm	-0.6 ppm	7/5/2031	-0.6 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC158302				B72023	CO,O2,BALN		
11/6/2023 5:07 PM	NOx ppm	High	Span	176.4 ppm	175.5 ppm	6/13/2031	-0.9 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM-040085				B32023	NO,NOX,BALN		
11/6/2023 5:07 PM	NOx ppm	Low	Zero	0.00 ppm	0.02 ppm	7/5/2031	0.02 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC158302				B72023	CO,O2,BALN		
11/6/2023 5:07 PM	NOx ppm	Low	Span	9.11 ppm	9.11 ppm	7/13/2026	0 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC285248				B32023	CO,NO,NOX,BALN		
11/6/2023 5:07 PM	75-NOx ppm	High	Zero	0.0 ppm	-0.6 ppm	7/5/2031	-0.6 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC158302				B72023	CO,O2,BALN		
11/6/2023 5:07 PM	75-NOx ppm	High	Span	176.4 ppm	175.5 ppm	6/13/2031	-0.9 ppm	±10 ppm	200 ppm	Unit online; Passed
			ALM-040085				B32023	NO,NOX,BALN		
11/6/2023 5:07 PM	75-NOx ppm	Low	Zero	0.00 ppm	0.02 ppm	7/5/2031	0.02 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC158302				B72023	CO,O2,BALN		
11/6/2023 5:07 PM	75-NOx ppm	Low	Span	9.11 ppm	9.11 ppm	7/13/2026	0 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC285248				B32023	CO,NO,NOX,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	EPA Vendor ID	Allowable Drift	Instrument Span	Results
11/6/2023 5:07 PM	CO ppm	High	Zero	0.0 ppm	6/13/2031	-0.1 ppm	-0.1 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
11/6/2023 5:07 PM	CO ppm	High	Span	182.4 ppm	7/5/2031	182.1 ppm	-0.3 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
11/6/2023 5:07 PM	CO ppm	Low	Zero	0.00 ppm	6/13/2031	0.15 ppm	0.15 ppm	B32023	±0.5 ppm NO,NOX,BALN	10 ppm	Unit online; Passed
11/6/2023 5:07 PM	CO ppm	Low	Span	9.10 ppm	7/13/2026	9.03 ppm	-0.07 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
11/6/2023 5:07 PM	NH3/NOx ppm	High	Zero	0.0 ppm	7/5/2031	-0.4 ppm	-0.4 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
11/6/2023 5:07 PM	NH3/NOx ppm	High	Span	176.4 ppm	6/13/2031	173.8 ppm	-2.6 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
11/6/2023 5:07 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	7/5/2031	0.12 ppm	0.12 ppm	B72023	±0.5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
11/6/2023 5:07 PM	NH3/NOx ppm	Low	Span	9.11 ppm	7/13/2026	9.30 ppm	0.19 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
11/6/2023 5:07 PM	O2 %	Single	Zero	0.00%	6/13/2031	-0.08%	-0.08%	B32023	±1% NO,NOX,BALN	25%	Unit online; Passed
11/6/2023 5:07 PM	O2 %	Single	Span	22.52%	7/5/2031	22.53%	0.01%	B72023	±1% CO,O2,BALN	25%	Unit online; Passed
11/6/2023 5:07 PM	75-O2 %	Single	Zero	0.00%	6/13/2031	-0.08%	-0.08%	B32023	±1% NO,NOX,BALN	25%	Unit online; Passed
11/6/2023 5:07 PM	75-O2 %	Single	Span	22.52%	7/5/2031	22.53%	0.01%	B72023	±1% CO,O2,BALN	25%	Unit online; Passed
11/13/2023 4:21 PM	NOx ppm	High	Zero	0.0 ppm	8/8/2031	-0.5 ppm	-0.5 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
11/13/2023 4:21 PM	NOx ppm	High	Span	181.1 ppm	7/24/2031	178.6 ppm	-2.5 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
11/13/2023 4:21 PM	NOx ppm	Low	Zero	0.00 ppm	8/8/2031	0.02 ppm	0.02 ppm	B72023	±0.5 ppm CO,O2,BALN	10 ppm	Unit online; Passed

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	EPA Vendor ID	Allowable Drift	Instrument Span	Results
11/13/2023 4:21 PM	NOx ppm	Low	Span	8.69 ppm	9/22/2026	8.68 ppm	-0.01 ppm	B32023	±0.5 ppm	10 ppm	Unit online; Passed
			CC283708						CO,NO,NOX,BALN		
11/13/2023 4:21 PM	75-NOx ppm	High	Zero	0.0 ppm	8/8/2031	-0.5 ppm	-0.5 ppm	B72023	±10 ppm	200 ppm	Unit online; Passed
			CC38198						CO,O2,BALN		
11/13/2023 4:21 PM	75-NOx ppm	High	Span	181.1 ppm	7/24/2031	178.6 ppm	-2.5 ppm	B32023	±10 ppm	200 ppm	Unit online; Passed
			CC59333						NO,NOX,BALN		
11/13/2023 4:21 PM	75-NOx ppm	Low	Zero	0.00 ppm	8/8/2031	0.02 ppm	0.02 ppm	B72023	±5 ppm	10 ppm	Unit online; Passed
			CC38198						CO,O2,BALN		
11/13/2023 4:21 PM	75-NOx ppm	Low	Span	8.69 ppm	9/22/2026	8.68 ppm	-0.01 ppm	B32023	±5 ppm	10 ppm	Unit online; Passed
			CC283708						CO,NO,NOX,BALN		
11/13/2023 4:21 PM	CO ppm	High	Zero	0.0 ppm	7/24/2031	0.0 ppm	0 ppm	B32023	±10 ppm	200 ppm	Unit online; Passed
			CC59333						NO,NOX,BALN		
11/13/2023 4:21 PM	CO ppm	High	Span	181.1 ppm	8/8/2031	181.1 ppm	0 ppm	B72023	±10 ppm	200 ppm	Unit online; Passed
			CC38198						CO,O2,BALN		
11/13/2023 4:21 PM	CO ppm	Low	Zero	0.00 ppm	7/24/2031	-0.23 ppm	-0.23 ppm	B32023	±0.5 ppm	10 ppm	Unit online; Passed
			CC59333						NO,NOX,BALN		
11/13/2023 4:21 PM	CO ppm	Low	Span	9.20 ppm	9/22/2026	9.21 ppm	0.01 ppm	B32023	±0.5 ppm	10 ppm	Unit online; Passed
			CC283708						CO,NO,NOX,BALN		
11/13/2023 4:21 PM	NH3/NOx ppm	High	Zero	0.0 ppm	8/8/2031	-0.4 ppm	-0.4 ppm	B72023	±10 ppm	200 ppm	Unit online; Passed
			CC38198						CO,O2,BALN		
11/13/2023 4:21 PM	NH3/NOx ppm	High	Span	181.1 ppm	7/24/2031	180.4 ppm	-0.7 ppm	B32023	±10 ppm	200 ppm	Unit online; Passed
			CC59333						NO,NOX,BALN		
11/13/2023 4:21 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	8/8/2031	0.13 ppm	0.13 ppm	B72023	±0.5 ppm	10 ppm	Unit online; Passed
			CC38198						CO,O2,BALN		
11/13/2023 4:21 PM	NH3/NOx ppm	Low	Span	8.69 ppm	9/22/2026	8.85 ppm	0.16 ppm	B32023	±0.5 ppm	10 ppm	Unit online; Passed
			CC283708						CO,NO,NOX,BALN		
11/13/2023 4:21 PM	O2 %	Single	Zero	0.00%	7/24/2031	-0.07%	-0.07%	B32023	±1%	25%	Unit online; Passed
			CC59333						NO,NOX,BALN		
11/13/2023 4:21 PM	O2 %	Single	Span	22.50%	8/8/2031	22.50%	0%	B72023	±1%	25%	Unit online; Passed
			CC38198						CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	EPA Vendor ID	Allowable Drift	Instrument Span	Results
11/13/2023 4:21 PM	75-O2 %	Single	Zero	0.00%	7/24/2031	-0.07%	-0.07%	B32023	±1% NO,NOX,BALN	25%	Unit online; Passed
11/13/2023 4:21 PM	75-O2 %	Single	Span	22.50%	8/8/2031	22.50%	0%	B72023	±1% CO,O2,BALN	25%	Unit online; Passed
11/14/2023 5:51 PM	NOx ppm	High	Zero	0.0 ppm	8/8/2031	-0.6 ppm	-0.6 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
11/14/2023 5:51 PM	NOx ppm	High	Span	181.1 ppm	7/24/2031	178.2 ppm	-2.9 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
11/14/2023 5:51 PM	NOx ppm	Low	Zero	0.00 ppm	8/8/2031	0.00 ppm	0 ppm	B72023	±0.5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
11/14/2023 5:51 PM	NOx ppm	Low	Span	8.69 ppm	9/22/2026	8.65 ppm	-0.04 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
11/14/2023 5:51 PM	75-NOx ppm	High	Zero	0.0 ppm	8/8/2031	-0.6 ppm	-0.6 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
11/14/2023 5:51 PM	75-NOx ppm	High	Span	181.1 ppm	7/24/2031	178.2 ppm	-2.9 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
11/14/2023 5:51 PM	75-NOx ppm	Low	Zero	0.00 ppm	8/8/2031	0.00 ppm	0 ppm	B72023	±5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
11/14/2023 5:51 PM	75-NOx ppm	Low	Span	8.69 ppm	9/22/2026	8.65 ppm	-0.04 ppm	B32023	±5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
11/14/2023 5:51 PM	CO ppm	High	Zero	0.0 ppm	7/24/2031	-0.1 ppm	-0.1 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
11/14/2023 5:51 PM	CO ppm	High	Span	181.1 ppm	8/8/2031	181.6 ppm	0.5 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
11/14/2023 5:51 PM	CO ppm	Low	Zero	0.00 ppm	7/24/2031	-0.27 ppm	-0.27 ppm	B32023	±0.5 ppm NO,NOX,BALN	10 ppm	Unit online; Passed
11/14/2023 5:51 PM	CO ppm	Low	Span	9.20 ppm	9/22/2026	9.21 ppm	0.01 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
11/14/2023 5:51 PM	NH3/NOx ppm	High	Zero	0.0 ppm	8/8/2031	-0.4 ppm	-0.4 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	EPA Vendor ID	Actual Drift	Allowable Drift	Instrument Span	Results
			Cylinder ID						EPA Gas Type Codes		
11/14/2023 5:51 PM	NH3/NOx ppm	High	Span	181.1 ppm	180.0 ppm	7/24/2031	B32023	-1.1 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC59333						NO,NOX,BALN		
11/14/2023 5:51 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	0.08 ppm	8/8/2031	B72023	0.08 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC38198						CO,O2,BALN		
11/14/2023 5:51 PM	NH3/NOx ppm	Low	Span	8.69 ppm	8.80 ppm	9/22/2026	B32023	0.11 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC283708						CO,NO,NOX,BALN		
11/14/2023 5:51 PM	O2 %	Single	Zero	0.00%	-0.06%	7/24/2031	B32023	-0.06%	±1%	25%	Unit online; Passed
			CC59333						NO,NOX,BALN		
11/14/2023 5:51 PM	O2 %	Single	Span	22.50%	22.51%	8/8/2031	B72023	0.01%	±1%	25%	Unit online; Passed
			CC38198						CO,O2,BALN		
11/14/2023 5:51 PM	75-O2 %	Single	Zero	0.00%	-0.06%	7/24/2031	B32023	-0.06%	±1%	25%	Unit online; Passed
			CC59333						NO,NOX,BALN		
11/14/2023 5:51 PM	75-O2 %	Single	Span	22.50%	22.51%	8/8/2031	B72023	0.01%	±1%	25%	Unit online; Passed
			CC38198						CO,O2,BALN		
11/15/2023 3:51 PM	NOx ppm	High	Zero	0.0 ppm	-0.6 ppm	8/8/2031	B72023	-0.6 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC38198						CO,O2,BALN		
11/15/2023 3:51 PM	NOx ppm	High	Span	181.1 ppm	178.4 ppm	7/24/2031	B32023	-2.7 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC59333						NO,NOX,BALN		
11/15/2023 3:51 PM	NOx ppm	Low	Zero	0.00 ppm	0.01 ppm	8/8/2031	B72023	0.01 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC38198						CO,O2,BALN		
11/15/2023 3:51 PM	NOx ppm	Low	Span	8.69 ppm	8.64 ppm	9/22/2026	B32023	-0.05 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC283708						CO,NO,NOX,BALN		
11/15/2023 3:51 PM	75-NOx ppm	High	Zero	0.0 ppm	-0.6 ppm	8/8/2031	B72023	-0.6 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC38198						CO,O2,BALN		
11/15/2023 3:51 PM	75-NOx ppm	High	Span	181.1 ppm	178.4 ppm	7/24/2031	B32023	-2.7 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC59333						NO,NOX,BALN		
11/15/2023 3:51 PM	75-NOx ppm	Low	Zero	0.00 ppm	0.01 ppm	8/8/2031	B72023	0.01 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC38198						CO,O2,BALN		
11/15/2023 3:51 PM	75-NOx ppm	Low	Span	8.69 ppm	8.64 ppm	9/22/2026	B32023	-0.05 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC283708						CO,NO,NOX,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Cylinder ID	Reference Value	Expiration Date	Measured Value	Actual Drift	EPA Vendor ID	Allowable Drift	Instrument Span	Results
11/15/2023 3:51 PM	CO ppm	High	Zero	CC59333	0.0 ppm	7/24/2031	-0.1 ppm	-0.1 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
11/15/2023 3:51 PM	CO ppm	High	Span	CC38198	181.1 ppm	8/8/2031	182.1 ppm	1 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
11/15/2023 3:51 PM	CO ppm	Low	Zero	CC59333	0.00 ppm	7/24/2031	-0.26 ppm	-0.26 ppm	B32023	±0.5 ppm NO,NOX,BALN	10 ppm	Unit online; Passed
11/15/2023 3:51 PM	CO ppm	Low	Span	CC283708	9.20 ppm	9/22/2026	9.23 ppm	0.03 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
11/15/2023 3:51 PM	NH3/NOx ppm	High	Zero	CC38198	0.0 ppm	8/8/2031	-0.4 ppm	-0.4 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
11/15/2023 3:51 PM	NH3/NOx ppm	High	Span	CC59333	181.1 ppm	7/24/2031	180.3 ppm	-0.8 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
11/15/2023 3:51 PM	NH3/NOx ppm	Low	Zero	CC38198	0.00 ppm	8/8/2031	0.08 ppm	0.08 ppm	B72023	±0.5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
11/15/2023 3:51 PM	NH3/NOx ppm	Low	Span	CC283708	8.69 ppm	9/22/2026	8.80 ppm	0.11 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
11/15/2023 3:51 PM	O2 %	Single	Zero	CC59333	0.00%	7/24/2031	-0.07%	-0.07%	B32023	±1% NO,NOX,BALN	25%	Unit online; Passed
11/15/2023 3:51 PM	O2 %	Single	Span	CC38198	22.50%	8/8/2031	22.50%	0%	B72023	±1% CO,O2,BALN	25%	Unit online; Passed
11/15/2023 3:51 PM	75-O2 %	Single	Zero	CC59333	0.00%	7/24/2031	-0.07%	-0.07%	B32023	±1% NO,NOX,BALN	25%	Unit online; Passed
11/15/2023 3:51 PM	75-O2 %	Single	Span	CC38198	22.50%	8/8/2031	22.50%	0%	B72023	±1% CO,O2,BALN	25%	Unit online; Passed
11/27/2023 5:07 PM	NOx ppm	High	Zero	CC38198	0.0 ppm	8/8/2031	-0.5 ppm	-0.5 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
11/27/2023 5:07 PM	NOx ppm	High	Span	CC59333	181.1 ppm	7/24/2031	178.3 ppm	-2.8 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
11/27/2023 5:07 PM	NOx ppm	Low	Zero	CC38198	0.00 ppm	8/8/2031	0.02 ppm	0.02 ppm	B72023	±0.5 ppm CO,O2,BALN	10 ppm	Unit online; Passed

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Actual Drift	Allowable Drift	Instrument Span	Results
		Cylinder ID	Expiration Date	EPA Vendor ID	EPA Gas Type Codes				
11/27/2023 5:07 PM	NOx ppm	Low	Span	8.69 ppm	8.66 ppm	-0.03 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC283708	9/22/2026	B32023		CO,NO,NOX,BALN		
11/27/2023 5:07 PM	75-NOx ppm	High	Zero	0.0 ppm	-0.5 ppm	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC38198	8/8/2031	B72023		CO,O2,BALN		
11/27/2023 5:07 PM	75-NOx ppm	High	Span	181.1 ppm	178.3 ppm	-2.8 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC59333	7/24/2031	B32023		NO,NOX,BALN		
11/27/2023 5:07 PM	75-NOx ppm	Low	Zero	0.00 ppm	0.02 ppm	0.02 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC38198	8/8/2031	B72023		CO,O2,BALN		
11/27/2023 5:07 PM	75-NOx ppm	Low	Span	8.69 ppm	8.66 ppm	-0.03 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC283708	9/22/2026	B32023		CO,NO,NOX,BALN		
11/27/2023 5:07 PM	CO ppm	High	Zero	0.0 ppm	0.0 ppm	0 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC59333	7/24/2031	B32023		NO,NOX,BALN		
11/27/2023 5:07 PM	CO ppm	High	Span	181.1 ppm	178.9 ppm	-2.2 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC38198	8/8/2031	B72023		CO,O2,BALN		
11/27/2023 5:07 PM	CO ppm	Low	Zero	0.00 ppm	-0.17 ppm	-0.17 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC59333	7/24/2031	B32023		NO,NOX,BALN		
11/27/2023 5:07 PM	CO ppm	Low	Span	9.20 ppm	9.24 ppm	0.04 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC283708	9/22/2026	B32023		CO,NO,NOX,BALN		
11/27/2023 5:07 PM	NH3/NOx ppm	High	Zero	0.0 ppm	-0.4 ppm	-0.4 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC38198	8/8/2031	B72023		CO,O2,BALN		
11/27/2023 5:07 PM	NH3/NOx ppm	High	Span	181.1 ppm	179.2 ppm	-1.9 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC59333	7/24/2031	B32023		NO,NOX,BALN		
11/27/2023 5:07 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	0.12 ppm	0.12 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC38198	8/8/2031	B72023		CO,O2,BALN		
11/27/2023 5:07 PM	NH3/NOx ppm	Low	Span	8.69 ppm	8.76 ppm	0.07 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC283708	9/22/2026	B32023		CO,NO,NOX,BALN		
11/27/2023 5:07 PM	O2 %	Single	Zero	0.00%	-0.08%	-0.08%	±1%	25%	Unit online; Passed
			CC59333	7/24/2031	B32023		NO,NOX,BALN		
11/27/2023 5:07 PM	O2 %	Single	Span	22.50%	22.49%	-0.01%	±1%	25%	Unit online; Passed
			CC38198	8/8/2031	B72023		CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	EPA Vendor ID	Allowable Drift	Instrument Span	Results
11/27/2023 5:07 PM	75-O2 %	Single	Zero	0.00%	7/24/2031	-0.08%	-0.08%	B32023	±1% NO,NOX,BALN	25%	Unit online; Passed
11/27/2023 5:07 PM	75-O2 %	Single	Span	22.50%	8/8/2031	22.49%	-0.01%	B72023	±1% CO,O2,BALN	25%	Unit online; Passed
11/30/2023 5:07 PM	NOx ppm	High	Zero	0.0 ppm	8/8/2031	-0.6 ppm	-0.6 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
11/30/2023 5:07 PM	NOx ppm	High	Span	181.1 ppm	7/24/2031	178.5 ppm	-2.6 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
11/30/2023 5:07 PM	NOx ppm	Low	Zero	0.00 ppm	8/8/2031	0.01 ppm	0.01 ppm	B72023	±0.5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
11/30/2023 5:07 PM	NOx ppm	Low	Span	8.69 ppm	9/22/2026	8.67 ppm	-0.02 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
11/30/2023 5:07 PM	75-NOx ppm	High	Zero	0.0 ppm	8/8/2031	-0.6 ppm	-0.6 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
11/30/2023 5:07 PM	75-NOx ppm	High	Span	181.1 ppm	7/24/2031	178.5 ppm	-2.6 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
11/30/2023 5:07 PM	75-NOx ppm	Low	Zero	0.00 ppm	8/8/2031	0.01 ppm	0.01 ppm	B72023	±5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
11/30/2023 5:07 PM	75-NOx ppm	Low	Span	8.69 ppm	9/22/2026	8.67 ppm	-0.02 ppm	B32023	±5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
11/30/2023 5:07 PM	CO ppm	High	Zero	0.0 ppm	7/24/2031	0.2 ppm	0.2 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
11/30/2023 5:07 PM	CO ppm	High	Span	181.1 ppm	8/8/2031	179.8 ppm	-1.3 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
11/30/2023 5:07 PM	CO ppm	Low	Zero	0.00 ppm	7/24/2031	0.00 ppm	0 ppm	B32023	±0.5 ppm NO,NOX,BALN	10 ppm	Unit online; Passed
11/30/2023 5:07 PM	CO ppm	Low	Span	9.20 ppm	9/22/2026	9.31 ppm	0.11 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
11/30/2023 5:07 PM	NH3/NOx ppm	High	Zero	0.0 ppm	8/8/2031	-0.4 ppm	-0.4 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed

Date/Time	Parameter	Analyzer Scale	Test Level	Cylinder ID	Reference Value	Expiration Date	Measured Value	Actual Drift	EPA Vendor ID	Allowable Drift	Instrument Span	Results
11/30/2023 5:07 PM	NH3/NOx ppm	High	Span	CC59333	181.1 ppm	7/24/2031	178.4 ppm	-2.7 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
11/30/2023 5:07 PM	NH3/NOx ppm	Low	Zero	CC38198	0.00 ppm	8/8/2031	0.13 ppm	0.13 ppm	B72023	±0.5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
11/30/2023 5:07 PM	NH3/NOx ppm	Low	Span	CC283708	8.69 ppm	9/22/2026	8.75 ppm	0.06 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
11/30/2023 5:07 PM	O2 %	Single	Zero	CC59333	0.00%	7/24/2031	-0.07%	-0.07%	B32023	±1% NO,NOX,BALN	25%	Unit online; Passed
11/30/2023 5:07 PM	O2 %	Single	Span	CC38198	22.50%	8/8/2031	22.49%	-0.01%	B72023	±1% CO,O2,BALN	25%	Unit online; Passed
11/30/2023 5:07 PM	75-O2 %	Single	Zero	CC59333	0.00%	7/24/2031	-0.07%	-0.07%	B32023	±1% NO,NOX,BALN	25%	Unit online; Passed
11/30/2023 5:07 PM	75-O2 %	Single	Span	CC38198	22.50%	8/8/2031	22.49%	-0.01%	B72023	±1% CO,O2,BALN	25%	Unit online; Passed
12/10/2023 6:06 PM	NOx ppm	High	Zero	CC38198	0.0 ppm	8/8/2031	-0.5 ppm	-0.5 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
12/10/2023 6:06 PM	NOx ppm	High	Span	CC59333	181.1 ppm	7/24/2031	179.2 ppm	-1.9 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
12/10/2023 6:06 PM	NOx ppm	Low	Zero	CC38198	0.00 ppm	8/8/2031	0.02 ppm	0.02 ppm	B72023	±0.5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
12/10/2023 6:06 PM	NOx ppm	Low	Span	CC283708	8.69 ppm	9/22/2026	8.63 ppm	-0.06 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
12/10/2023 6:06 PM	75-NOx ppm	High	Zero	CC38198	0.0 ppm	8/8/2031	-0.5 ppm	-0.5 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
12/10/2023 6:06 PM	75-NOx ppm	High	Span	CC59333	181.1 ppm	7/24/2031	179.2 ppm	-1.9 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
12/10/2023 6:06 PM	75-NOx ppm	Low	Zero	CC38198	0.00 ppm	8/8/2031	0.02 ppm	0.02 ppm	B72023	±5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
12/10/2023 6:06 PM	75-NOx ppm	Low	Span	CC283708	8.69 ppm	9/22/2026	8.63 ppm	-0.06 ppm	B32023	±5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed

Date/Time	Parameter	Analyzer Scale	Test Level	Cylinder ID	Reference Value	Expiration Date	Measured Value	Actual Drift	EPA Vendor ID	Allowable Drift	Instrument Span	Results
12/10/2023 6:06 PM	CO ppm	High	Zero	CC59333	0.0 ppm	7/24/2031	-0.1 ppm	-0.1 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
12/10/2023 6:06 PM	CO ppm	High	Span	CC38198	181.1 ppm	8/8/2031	179.7 ppm	-1.4 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
12/10/2023 6:06 PM	CO ppm	Low	Zero	CC59333	0.00 ppm	7/24/2031	-0.25 ppm	-0.25 ppm	B32023	±0.5 ppm NO,NOX,BALN	10 ppm	Unit online; Passed
12/10/2023 6:06 PM	CO ppm	Low	Span	CC283708	9.20 ppm	9/22/2026	9.18 ppm	-0.02 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
12/10/2023 6:06 PM	NH3/NOx ppm	High	Zero	CC38198	0.0 ppm	8/8/2031	-0.4 ppm	-0.4 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
12/10/2023 6:06 PM	NH3/NOx ppm	High	Span	CC59333	181.1 ppm	7/24/2031	179.3 ppm	-1.8 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
12/10/2023 6:06 PM	NH3/NOx ppm	Low	Zero	CC38198	0.00 ppm	8/8/2031	0.12 ppm	0.12 ppm	B72023	±0.5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
12/10/2023 6:06 PM	NH3/NOx ppm	Low	Span	CC283708	8.69 ppm	9/22/2026	8.70 ppm	0.01 ppm	B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
12/10/2023 6:06 PM	O2 %	Single	Zero	CC59333	0.00%	7/24/2031	-0.06%	-0.06%	B32023	±1% NO,NOX,BALN	25%	Unit online; Passed
12/10/2023 6:06 PM	O2 %	Single	Span	CC38198	22.50%	8/8/2031	22.50%	0%	B72023	±1% CO,O2,BALN	25%	Unit online; Passed
12/10/2023 6:06 PM	75-O2 %	Single	Zero	CC59333	0.00%	7/24/2031	-0.06%	-0.06%	B32023	±1% NO,NOX,BALN	25%	Unit online; Passed
12/10/2023 6:06 PM	75-O2 %	Single	Span	CC38198	22.50%	8/8/2031	22.50%	0%	B72023	±1% CO,O2,BALN	25%	Unit online; Passed
12/11/2023 5:06 PM	NOx ppm	High	Zero	CC38198	0.0 ppm	8/8/2031	-0.6 ppm	-0.6 ppm	B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
12/11/2023 5:06 PM	NOx ppm	High	Span	CC59333	181.1 ppm	7/24/2031	180.0 ppm	-1.1 ppm	B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
12/11/2023 5:06 PM	NOx ppm	Low	Zero	CC38198	0.00 ppm	8/8/2031	0.00 ppm	0 ppm	B72023	±0.5 ppm CO,O2,BALN	10 ppm	Unit online; Passed

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Expiration Date	Measured Value	Actual Drift	EPA Vendor ID	Allowable Drift	Instrument Span	Results
12/11/2023 5:06 PM	NOx ppm	Low	Span	8.69 ppm	9/22/2026	8.63 ppm	-0.06 ppm	B32023	±0.5 ppm	10 ppm	Unit online; Passed
			CC283708						CO,NO,NOX,BALN		
12/11/2023 5:06 PM	75-NOx ppm	High	Zero	0.0 ppm	8/8/2031	-0.6 ppm	-0.6 ppm	B72023	±10 ppm	200 ppm	Unit online; Passed
			CC38198						CO,O2,BALN		
12/11/2023 5:06 PM	75-NOx ppm	High	Span	181.1 ppm	7/24/2031	180.0 ppm	-1.1 ppm	B32023	±10 ppm	200 ppm	Unit online; Passed
			CC59333						NO,NOX,BALN		
12/11/2023 5:06 PM	75-NOx ppm	Low	Zero	0.00 ppm	8/8/2031	0.00 ppm	0 ppm	B72023	±5 ppm	10 ppm	Unit online; Passed
			CC38198						CO,O2,BALN		
12/11/2023 5:06 PM	75-NOx ppm	Low	Span	8.69 ppm	9/22/2026	8.63 ppm	-0.06 ppm	B32023	±5 ppm	10 ppm	Unit online; Passed
			CC283708						CO,NO,NOX,BALN		
12/11/2023 5:06 PM	CO ppm	High	Zero	0.0 ppm	7/24/2031	-0.1 ppm	-0.1 ppm	B32023	±10 ppm	200 ppm	Unit online; Passed
			CC59333						NO,NOX,BALN		
12/11/2023 5:06 PM	CO ppm	High	Span	181.1 ppm	8/8/2031	180.8 ppm	-0.3 ppm	B72023	±10 ppm	200 ppm	Unit online; Passed
			CC38198						CO,O2,BALN		
12/11/2023 5:06 PM	CO ppm	Low	Zero	0.00 ppm	7/24/2031	-0.24 ppm	-0.24 ppm	B32023	±0.5 ppm	10 ppm	Unit online; Passed
			CC59333						NO,NOX,BALN		
12/11/2023 5:06 PM	CO ppm	Low	Span	9.20 ppm	9/22/2026	9.18 ppm	-0.02 ppm	B32023	±0.5 ppm	10 ppm	Unit online; Passed
			CC283708						CO,NO,NOX,BALN		
12/11/2023 5:06 PM	NH3/NOx ppm	High	Zero	0.0 ppm	8/8/2031	-0.4 ppm	-0.4 ppm	B72023	±10 ppm	200 ppm	Unit online; Passed
			CC38198						CO,O2,BALN		
12/11/2023 5:06 PM	NH3/NOx ppm	High	Span	181.1 ppm	7/24/2031	179.7 ppm	-1.4 ppm	B32023	±10 ppm	200 ppm	Unit online; Passed
			CC59333						NO,NOX,BALN		
12/11/2023 5:06 PM	NH3/NOx ppm	Low	Zero	0.00 ppm	8/8/2031	0.11 ppm	0.11 ppm	B72023	±0.5 ppm	10 ppm	Unit online; Passed
			CC38198						CO,O2,BALN		
12/11/2023 5:06 PM	NH3/NOx ppm	Low	Span	8.69 ppm	9/22/2026	8.71 ppm	0.02 ppm	B32023	±0.5 ppm	10 ppm	Unit online; Passed
			CC283708						CO,NO,NOX,BALN		
12/11/2023 5:06 PM	O2 %	Single	Zero	0.00%	7/24/2031	-0.06%	-0.06%	B32023	±1%	25%	Unit online; Passed
			CC59333						NO,NOX,BALN		
12/11/2023 5:06 PM	O2 %	Single	Span	22.50%	8/8/2031	22.51%	0.01%	B72023	±1%	25%	Unit online; Passed
			CC38198						CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	EPA Vendor ID	Actual Drift	Allowable Drift	Instrument Span	Results
			Cylinder ID						EPA Gas Type Codes		
12/11/2023 5:06 PM	75-O2 %	Single	Zero CC59333	0.00%	-0.06%	7/24/2031	B32023	-0.06%	±1%	25%	Unit online; Passed
12/11/2023 5:06 PM	75-O2 %	Single	Span CC38198	22.50%	22.51%	8/8/2031	B72023	0.01%	±1%	25%	Unit online; Passed
12/12/2023 6:06 AM	NOx ppm	High	Zero CC38198	0.0 ppm	-0.6 ppm	8/8/2031	B72023	-0.6 ppm	±10 ppm	200 ppm	Unit online; Passed
12/12/2023 6:06 AM	NOx ppm	High	Span CC59333	181.1 ppm	180.4 ppm	7/24/2031	B32023	-0.7 ppm	±10 ppm	200 ppm	Unit online; Passed
12/12/2023 6:06 AM	NOx ppm	Low	Zero CC38198	0.00 ppm	0.02 ppm	8/8/2031	B72023	0.02 ppm	±0.5 ppm	10 ppm	Unit online; Passed
12/12/2023 6:06 AM	NOx ppm	Low	Span CC283708	8.69 ppm	8.70 ppm	9/22/2026	B32023	0.01 ppm	±0.5 ppm	10 ppm	Unit online; Passed
12/12/2023 6:06 AM	75-NOx ppm	High	Zero CC38198	0.0 ppm	-0.6 ppm	8/8/2031	B72023	-0.6 ppm	±10 ppm	200 ppm	Unit online; Passed
12/12/2023 6:06 AM	75-NOx ppm	High	Span CC59333	181.1 ppm	180.4 ppm	7/24/2031	B32023	-0.7 ppm	±10 ppm	200 ppm	Unit online; Passed
12/12/2023 6:06 AM	75-NOx ppm	Low	Zero CC38198	0.00 ppm	0.02 ppm	8/8/2031	B72023	0.02 ppm	±5 ppm	10 ppm	Unit online; Passed
12/12/2023 6:06 AM	75-NOx ppm	Low	Span CC283708	8.69 ppm	8.70 ppm	9/22/2026	B32023	0.01 ppm	±5 ppm	10 ppm	Unit online; Passed
12/12/2023 6:06 AM	CO ppm	High	Zero CC59333	0.0 ppm	-0.1 ppm	7/24/2031	B32023	-0.1 ppm	±10 ppm	200 ppm	Unit online; Passed
12/12/2023 6:06 AM	CO ppm	High	Span CC38198	181.1 ppm	183.3 ppm	8/8/2031	B72023	2.2 ppm	±10 ppm	200 ppm	Unit online; Passed
12/12/2023 6:06 AM	CO ppm	Low	Zero CC59333	0.00 ppm	-0.25 ppm	7/24/2031	B32023	-0.25 ppm	±0.5 ppm	10 ppm	Unit online; Passed
12/12/2023 6:06 AM	CO ppm	Low	Span CC283708	9.20 ppm	9.21 ppm	9/22/2026	B32023	0.01 ppm	±0.5 ppm	10 ppm	Unit online; Passed
12/12/2023 6:06 AM	NH3/NOx ppm	High	Zero CC38198	0.0 ppm	-0.5 ppm	8/8/2031	B72023	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
			Cylinder ID	Value	Value		EPA Vendor ID	EPA Gas Type Codes		
12/12/2023 6:06 AM	NH3/NOx ppm	High	Span	181.1 ppm	180.3 ppm	7/24/2031	-0.8 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC59333				B32023	NO,NOX,BALN		
12/12/2023 6:06 AM	NH3/NOx ppm	Low	Zero	0.00 ppm	0.05 ppm	8/8/2031	0.05 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC38198				B72023	CO,O2,BALN		
12/12/2023 6:06 AM	NH3/NOx ppm	Low	Span	8.69 ppm	8.70 ppm	9/22/2026	0.01 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC283708				B32023	CO,NO,NOX,BALN		
12/12/2023 6:06 AM	O2 %	Single	Zero	0.00%	-0.06%	7/24/2031	-0.06%	±1%	25%	Unit online; Passed
			CC59333				B32023	NO,NOX,BALN		
12/12/2023 6:06 AM	O2 %	Single	Span	22.50%	22.51%	8/8/2031	0.01%	±1%	25%	Unit online; Passed
			CC38198				B72023	CO,O2,BALN		
12/12/2023 6:06 AM	75-O2 %	Single	Zero	0.00%	-0.06%	7/24/2031	-0.06%	±1%	25%	Unit online; Passed
			CC59333				B32023	NO,NOX,BALN		
12/12/2023 6:06 AM	75-O2 %	Single	Span	22.50%	22.51%	8/8/2031	0.01%	±1%	25%	Unit online; Passed
			CC38198				B72023	CO,O2,BALN		
12/13/2023 10:00 AM	NOx ppm	High	Zero	0.0 ppm	-0.6 ppm	8/8/2031	-0.6 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC38198				B72023	CO,O2,BALN		
12/13/2023 10:00 AM	NOx ppm	High	Span	181.1 ppm	179.2 ppm	7/24/2031	-1.9 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC59333				B32023	NO,NOX,BALN		
12/13/2023 10:00 AM	NOx ppm	Low	Zero	0.00 ppm	0.00 ppm	8/8/2031	0 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC38198				B72023	CO,O2,BALN		
12/13/2023 10:00 AM	NOx ppm	Low	Span	8.69 ppm	8.63 ppm	9/22/2026	-0.06 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC283708				B32023	CO,NO,NOX,BALN		
12/13/2023 10:00 AM	75-NOx ppm	High	Zero	0.0 ppm	-0.6 ppm	8/8/2031	-0.6 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC38198				B72023	CO,O2,BALN		
12/13/2023 10:00 AM	75-NOx ppm	High	Span	181.1 ppm	179.2 ppm	7/24/2031	-1.9 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC59333				B32023	NO,NOX,BALN		
12/13/2023 10:00 AM	75-NOx ppm	Low	Zero	0.00 ppm	0.00 ppm	8/8/2031	0 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC38198				B72023	CO,O2,BALN		
12/13/2023 10:00 AM	75-NOx ppm	Low	Span	8.69 ppm	8.63 ppm	9/22/2026	-0.06 ppm	±5 ppm	10 ppm	Unit online; Passed
			CC283708				B32023	CO,NO,NOX,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
			Cylinder ID	Value	Value		EPA Vendor ID	EPA Gas Type	Codes	
12/13/2023 10:00 AM	CO ppm	High	Zero	0.0 ppm	-0.1 ppm	7/24/2031	-0.1 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC59333				B32023	NO,NOX,BALN		
12/13/2023 10:00 AM	CO ppm	High	Span	181.1 ppm	180.0 ppm	8/8/2031	-1.1 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC38198				B72023	CO,O2,BALN		
12/13/2023 10:00 AM	CO ppm	Low	Zero	0.00 ppm	-0.27 ppm	7/24/2031	-0.27 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC59333				B32023	NO,NOX,BALN		
12/13/2023 10:00 AM	CO ppm	Low	Span	9.20 ppm	9.52 ppm	9/22/2026	0.32 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC283708				B32023	CO,NO,NOX,BALN		
12/13/2023 10:00 AM	NH3/NOx ppm	High	Zero	0.0 ppm	-0.5 ppm	8/8/2031	-0.5 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC38198				B72023	CO,O2,BALN		
12/13/2023 10:00 AM	NH3/NOx ppm	High	Span	181.1 ppm	179.0 ppm	7/24/2031	-2.1 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC59333				B32023	NO,NOX,BALN		
12/13/2023 10:00 AM	NH3/NOx ppm	Low	Zero	0.00 ppm	0.04 ppm	8/8/2031	0.04 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC38198				B72023	CO,O2,BALN		
12/13/2023 10:00 AM	NH3/NOx ppm	Low	Span	8.69 ppm	8.64 ppm	9/22/2026	-0.05 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC283708				B32023	CO,NO,NOX,BALN		
12/13/2023 10:00 AM	O2 %	Single	Zero	0.00%	-0.07%	7/24/2031	-0.07%	±1%	25%	Unit online; Passed
			CC59333				B32023	NO,NOX,BALN		
12/13/2023 10:00 AM	O2 %	Single	Span	22.50%	22.50%	8/8/2031	0%	±1%	25%	Unit online; Passed
			CC38198				B72023	CO,O2,BALN		
12/13/2023 10:00 AM	75-O2 %	Single	Zero	0.00%	-0.07%	7/24/2031	-0.07%	±1%	25%	Unit online; Passed
			CC59333				B32023	NO,NOX,BALN		
12/13/2023 10:00 AM	75-O2 %	Single	Span	22.50%	22.50%	8/8/2031	0%	±1%	25%	Unit online; Passed
			CC38198				B72023	CO,O2,BALN		
12/16/2023 5:05 PM	NOx ppm	High	Zero	0.0 ppm	-0.6 ppm	8/8/2031	-0.6 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC38198				B72023	CO,O2,BALN		
12/16/2023 5:05 PM	NOx ppm	High	Span	181.1 ppm	178.8 ppm	7/24/2031	-2.3 ppm	±10 ppm	200 ppm	Unit online; Passed
			CC59333				B32023	NO,NOX,BALN		
12/16/2023 5:05 PM	NOx ppm	Low	Zero	0.00 ppm	0.00 ppm	8/8/2031	0 ppm	±0.5 ppm	10 ppm	Unit online; Passed
			CC38198				B72023	CO,O2,BALN		

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Expiration Date	Actual Drift	Allowable Drift	Instrument Span	Results
			Cylinder ID	Value	Value		EPA Vendor ID	EPA Gas Type Codes		
12/16/2023 5:05 PM	NOx ppm	Low	Span CC283708	8.69 ppm	8.61 ppm	9/22/2026	-0.08 ppm B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
12/16/2023 5:05 PM	75-NOx ppm	High	Zero CC38198	0.0 ppm	-0.6 ppm	8/8/2031	-0.6 ppm B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
12/16/2023 5:05 PM	75-NOx ppm	High	Span CC59333	181.1 ppm	178.8 ppm	7/24/2031	-2.3 ppm B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
12/16/2023 5:05 PM	75-NOx ppm	Low	Zero CC38198	0.00 ppm	0.00 ppm	8/8/2031	0 ppm B72023	±5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
12/16/2023 5:05 PM	75-NOx ppm	Low	Span CC283708	8.69 ppm	8.61 ppm	9/22/2026	-0.08 ppm B32023	±5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
12/16/2023 5:05 PM	CO ppm	High	Zero CC59333	0.0 ppm	0.0 ppm	7/24/2031	0 ppm B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
12/16/2023 5:05 PM	CO ppm	High	Span CC38198	181.1 ppm	180.3 ppm	8/8/2031	-0.8 ppm B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
12/16/2023 5:05 PM	CO ppm	Low	Zero CC59333	0.00 ppm	-0.01 ppm	7/24/2031	-0.01 ppm B32023	±0.5 ppm NO,NOX,BALN	10 ppm	Unit online; Passed
12/16/2023 5:05 PM	CO ppm	Low	Span CC283708	9.20 ppm	9.33 ppm	9/22/2026	0.13 ppm B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
12/16/2023 5:05 PM	NH3/NOx ppm	High	Zero CC38198	0.0 ppm	-0.4 ppm	8/8/2031	-0.4 ppm B72023	±10 ppm CO,O2,BALN	200 ppm	Unit online; Passed
12/16/2023 5:05 PM	NH3/NOx ppm	High	Span CC59333	181.1 ppm	180.3 ppm	7/24/2031	-0.8 ppm B32023	±10 ppm NO,NOX,BALN	200 ppm	Unit online; Passed
12/16/2023 5:05 PM	NH3/NOx ppm	Low	Zero CC38198	0.00 ppm	0.12 ppm	8/8/2031	0.12 ppm B72023	±0.5 ppm CO,O2,BALN	10 ppm	Unit online; Passed
12/16/2023 5:05 PM	NH3/NOx ppm	Low	Span CC283708	8.69 ppm	8.65 ppm	9/22/2026	-0.04 ppm B32023	±0.5 ppm CO,NO,NOX,BALN	10 ppm	Unit online; Passed
12/16/2023 5:05 PM	O2 %	Single	Zero CC59333	0.00%	-0.06%	7/24/2031	-0.06% B32023	±1% NO,NOX,BALN	25%	Unit online; Passed
12/16/2023 5:05 PM	O2 %	Single	Span CC38198	22.50%	22.50%	8/8/2031	0% B72023	±1% CO,O2,BALN	25%	Unit online; Passed

Date/Time	Parameter	Analyzer Scale	Test Level	Reference Value	Measured Value	Actual Drift	Allowable Drift	Instrument Span	Results
			Cylinder ID	Expiration Date	EPA Vendor ID	EPA Gas Type Codes			
12/16/2023 5:05 PM	75-O2 %	Single	Zero CC59333	0.00% 7/24/2031	-0.06% B32023	-0.06%	±1% NO,NOX,BALN	25%	Unit online; Passed
12/16/2023 5:05 PM	75-O2 %	Single	Span CC38198	22.50% 8/8/2031	22.50%	0%	±1% CO,O2,BALN	25%	Unit online; Passed

Attachment 4

SCR and CO Catalyst Temperature and Pressure Devices Calibrations Records



Instrumentation Calibration Record McGrath Peaker

73

Instrument Number | TE-403A

Instrument Name | McGrath Scr Catalyst Inlet Temp Element (B-255-TE-403A)

Zero | _____ Span | _____ Units | _____

Alarm 1 Setpoint | _____ Alarm 2 Setpoint | _____

Alarm 1 Inc/Dec | _____ Alarm 2 Inc/Dec | _____

Span Calibration- Found	0%	25	50%	75%	100%
	64°F	251	501	750	1003

Span Calibration- Left	0%	25	50%	75%	100%

Switch Calibration Setpoint Found | _____ Setpoint Left | _____

Function | Emissions Reduction

Location | SCR Catalyst

Technician | Gonzalez / Fisher | Completion Date | 5/5/23

Manufacturer	Fluke	Manufacturer	
Model	9144	Model	
Serial Number	A98356	Serial Number	
Calibration Due Date	2/9/25	Calibration Due Date	



Instrumentation Calibration Record McGrath Peaker

T5

Instrument Number | TE-403C

Instrument Name | McGrath Scr Catalyst Inlet Temp Element (B-255-TE-403C)

Zero | _____ Span | _____ Units | _____

Alarm 1 Setpoint | _____ Alarm 2 Setpoint | _____

Alarm 1 Inc/Dec | _____ Alarm 2 Inc/Dec | _____

	0% ^{64°F}	25	^{250°F}	50%	^{500°F}	75%	^{750°F}	100%	^{1000°F}
Span Calibration- Found	66		252		502		753		1004

Span Calibration- Left	0%	25	50%	75%	100%
------------------------	----	----	-----	-----	------

Switch Calibration Setpoint Found | _____ Setpoint Left | _____

Com | Emmissions Reduction

Location | SCR Catalyst

Technician | Gonzalez / Fisher Completion Date | 5/5/23

Manufacturer	<u>Fulke</u>	Manufacturer	_____
Model	<u>9144</u>	Model	_____
Serial Number	<u>A98356</u>	Serial Number	_____
Calibration Due Date	<u>2/9/25</u>	Calibration Due Date	_____

Attachment 5

Gas Fuel and Ammonia Flow Meters Calibrations Records

Certificate of Calibration

Calibration Date:	4/10/2023
Calibration Due Date:	4/2024
Customer:	Southern California Edison Peaker Unit

Certificate Lot Number:	230410
Technician:	Brandon Davis
CA Weights & Measures ID:	1918-04871

Calibrated Instrument Data

Equipment ID:	427111 (Sensor), 427110 (Flow Converter)	Manufacturer:	Yokogawa		
Location:	Unknown-Device shipped to MMCI headquarters for calibration.	Model Number:	DY050S1-NBLBA4-2N/KF1/SCT (Sensor), DYAS1-D2N/KF1/SCT (Flow Converter)		
Calibration Description:	Calibration with direct totalizer comparison from flow meter to calibration standard indicated below. Flow meter is a 2" vortex style with remote flow converter.	Serial Number:	S5F505478 620 (Sensor), S5F505492 620 (Flow Converter)		
Notes:	NA	As Found		As Left	
		Adjustment K	1.0000		1.0000
		mA Output 1:	0 to 345 cf/min		0 to 345 cf/min
		K-Factor:	8.833 Pulses/liter		8.833 Pulses/liter

Calibration Performance

Status:	Found (Hydraulic calibration of flow meter with direct totalizer comparison to MMCI standard)			Units:	US Gallons
				Tolerance:	± 0.80%
Test Rate/min	Sensor Total	Standard	Error Sensor/Standard		Pass/Fail
50	204.00	202.98	0.50%		Pass
100	307.00	305.99	0.33%		Pass
150	305.00	303.82	0.39%		Pass

Status:	Left (Hydraulic calibration of flow meter with direct totalizer comparison to MMCI standard)			Units:	US Gallons
				Tolerance:	± 0.80%
Test Rate/min	Sensor Total	Standard	Error Sensor/Standard		Pass/Fail

Calibration Standards

Standard 1:	Gravimetric Prover, MOD: 520, SN: 1329400009, Due: 5/2023, NIST: 191003002
Standard 2:	

Standard 3:	
Standard 4:	

Technician Signature 

This device has been calibrated using standards traceable to the National Institute of Standards & Technology (NIST). This certificate shall not be reproduced in any form, except in full, without the expressed written consent of MMCI. Please refer to MMCI's contact information above regarding any questions associated with this certificate.

Certificate of Calibration

Calibration Date:	4/10/2023
Calibration Due Date:	4/2024
Customer:	Southern California Edison Peaker Unit

Certificate Lot Number:	230410
Technician:	Brandon Davis
CA Weights & Measures ID:	1918-04871

Calibrated Instrument Data

Equipment ID:	427111 (Sensor), 427110 (Flow Converter)	Manufacturer:	Yokogawa		
Location:	Unknown-Device shipped to MMCI headquarters for calibration.	Model Number:	DY050S1-NBLBA4-2N/KF1/SCT (Sensor), DYAS1-D2N/KF1/SCT (Flow Converter)		
Calibration Description:	mA calibration of flow transmitter mA output #1	Serial Number:	S5F505478 620 (Sensor), S5F505492 620 (Flow Converter)		
Notes:	NA	As Found		As Left	
		Adjustment K	1.0000		1.0000
		mA Output 1:	0 to 345 cf/min		0 to 345 cf/min
		K-Factor:	8.833 Pulses/liter		8.833 Pulses/liter

Calibration Performance

Status:	Found (mA calibration of flow transmitter mA output with direct comparison to MMCI standard)			Units:	mA
				Tolerance:	± 0.20%
Sensor Simulated mA	Standard Measured	Error Sensor/Standard			Pass/Fail
4.00	4.000	0.00%			Pass
8.00	7.999	0.01%			Pass
12.00	11.997	0.03%			Pass
16.00	15.996	0.03%			Pass
20.00	19.995	0.03%			Pass

Status:	Left (mA calibration of flow transmitter mA output with direct comparison to MMCI standard)			Units:	mA
				Tolerance:	± 0.20%
Sensor Simulated mA	Standard Measured	Error Sensor/Standard			Pass/Fail

Calibration Standards

Standard 1:	Druck DPI620, SN: 233790, Due: 6/23/2024, NIST: 19-D9L7X-20-1
Standard 2:	

Standard 3:	
Standard 4:	

Technician Signature



This device has been calibrated using standards traceable to the National Institute of Standards & Technology (NIST). This certificate shall not be reproduced in any form, except in full, without the expressed written consent of MMCI. Please refer to MMCI's contact information above regarding any questions associated with this certificate.

Certificate of Calibration

Calibration Date:	4/11/2023
Calibration Due Date:	4/2024
Customer:	Southern California Edison Peaker Unit

Certificate Lot Number:	230411
Technician:	Brandon Davis
CA Weights & Measures ID:	1918-04871

Calibrated Instrument Data

Equipment ID:	441371	Manufacturer:	Micro Motion		
Location:	Unknown	Model Number:	CMF025M313NQBUEZZZ (Sensor), 1700R11ABUEZZZ (Xmtr)		
Calibration Description:	Gravimetric calibration with direct totalizer comparison from flow meter to calibration standard indicated below. Flow meter is a 0.25" coriolis with remote transmitter.	Serial Number:	14197004 (Sensor), 3157263 (Xmtr)		
Notes:	NA	As Found		As Left	
		Flo Cal Factor:	4.8692		4.8692
		mA Output 1:	0 to 150 lb/hr		0 to 150 lb/hr
		Pulse Out:	1500 pulses/lb		1500 pulses/lb

Calibration Performance

Status:	Found (Gravimetric calibration of flow meter with direct totalizer comparison to MMCI standard)			Units:	Pounds
				Tolerance:	± 0.50%
Test Rate/Hour	Sensor Total	Standard	Error Sensor/Standard		Pass/Fail
75	4.027	4.030	-0.07%		Pass
100	5.050	5.045	0.10%		Pass
150	8.615	8.605	0.12%		Pass

Status:	Left (Gravimetric calibration of flow meter with direct totalizer comparison to MMCI standard)			Units:	Pounds
				Tolerance:	± 0.50%
Test Rate/Hour	Sensor Total	Standard	Error Sensor/Standard		Pass/Fail

Calibration Standards

Standard 1:	Torrey Scale, Mod: L-EQ 10/20, SN: 116-006363, Due: 5/2023, NIST: 200204001
Standard 2:	

Standard 3:	
Standard 4:	

Technician Signature



This device has been calibrated using standards traceable to the National Institute of Standards & Technology (NIST). This certificate shall not be reproduced in any form, except in full, without the expressed written consent of MMCI. Please refer to MMCI's contact information above regarding any questions associated with this certificate.

Certificate of Calibration

Calibration Date:	4/11/2023
Calibration Due Date:	4/2024
Customer:	Southern California Edison Peaker Unit

Certificate Lot Number:	230411
Technician:	Brandon Davis
CA Weights & Measures ID:	1918-04871

Calibrated Instrument Data

Equipment ID:	441371	Manufacturer:	Micro Motion		
Location:	Unknown	Model Number:	CMF025M313NQBUEZZZ (Sensor), 1700R11ABUEZZZ (Xmtr)		
Calibration Description:	Calibration of transmitter mA output	Serial Number:	14197004 (Sensor), 3157263 (Xmtr)		
Notes:	NA	As Found		As Left	
		Flo Cal Factor:	4.8692		4.8692
		mA Output 1:	0 to 150 lb/hr		0 to 150 lb/hr
		Pulse Out:	1500 pulses/lb		1500 pulses/lb

Calibration Performance

Status:	Found (mA calibration of flow transmitter mA output with direct comparison to MMCI standard)			Units:	mA	
				Tolerance:	± 0.20%	
Sensor Simulated mA	Standard Measured	Error Sensor/Standard				Pass/Fail
4.00	4.000	0.00%				Pass
8.00	8.000	0.00%				Pass
12.00	11.998	0.02%				Pass
16.00	15.997	0.02%				Pass
20.00	19.997	0.02%				Pass

Status:	Left (mA calibration of flow transmitter mA output with direct comparison to MMCI standard)			Units:	mA	
				Tolerance:	± 0.20%	
Sensor Simulated mA	Standard Measured	Error Sensor/Standard				Pass/Fail

Calibration Standards

Standard 1:	Druck DPI620, SN: 233790, Due: 6/23/2024, NIST: 19-D9L7X-20-1
Standard 2:	

Standard 3:	
Standard 4:	

This device has been calibrated using standards traceable to the National Institute of Standards & Technology (NIST). This certificate shall not be reproduced in any form, except in full, without the expressed written consent of MMCI. Please refer to MMCI's contact information above regarding any questions associated with this certificate.

Technician Signature 

Attachment 6
Generator and Engine Specifications

Waukesha VGF Series



A New Look At
Reliable, Compact,
Low Emission
Gas Engines.

Whether you're designing an engine package for a new installation or retrofitting an engine room, space is always an issue. Your customers want more power squeezed into smaller spaces. So, how do you reconcile the need for more power and limited space? Simple, size up the Waukesha VGF family of gas engines. This series of compact, fuel efficient, low emission gas engines comes in 6, 8, 12 and 16 cylinder outputs with both

inline and vee configurations to meet almost any installation requirements.


Driving Down the Cost of Horsepower.

This compact engine is purposely designed from the ground up for stationary, spark ignited, gaseous fuel applications. The VGF is not limited to weight-sensitive mobile equipment requirements. Its high weight-to-power ratio provides a solid foundation for years of

consistent and dependable operation. For the size and price of other high speed engines, the VGF provides full power around the clock, with substantially reduced maintenance and operating costs. You get more engine, more reliability, more performance, more uptime, and more value – at less cost.

Because of their size to power ratio and high speed continuous-duty capability, VGF engines are remarkably cost-effective to

A Family of Compact Engines



their "clean" design requires less piping.

Strong Resemblance.

The VGF Series reflects Waukesha's usual robust structural strength, along with sophisticated engineering features. Derived from the design and manufacturing integrity of the classic Waukesha VHP, these engines have inherently long component life. So overhaul intervals will be few and far between. The performance and durability of the VGF Series make this engine

line a design classic in itself.

Fast and Lean.

VGF engines lead a clean life, thanks to Waukesha's lean burn technology. Our patented combustion system allows the VGF engine to meet clean air standards throughout the world. (Consult your distributor for site specific requirements.) Both GL (Gas Lean Burn) and GLD (Gas Lean Burn Draw Thru) models are available as standard

engines. They are designed to run efficiently with high or low fuel pressure systems for maximum application flexibility.

The VGF is also available in a rich burn naturally aspirated version (G) for applications where less horsepower is required or a 3-way catalyst is desired.

The Waukesha VGF Series. Simple to buy.

Simple to install. Simple to run. Simple to

meet emissions requirements. Simple to

maintain. How much simpler can we make

your engine choice?


to Fit Your Requirements.



VGF

First of all, we need to drive home a simple point. This is not a warmed over automotive engine designed for occasional rpm spikes. The VGF is industrial strength continuous high rpm horsepower in a range from 160 BHP (120 KWb) to 1065 BHP (800 KWb).

How this Hard-Working



- The *GL (Lean Burn) fuel system and high turbulence combustion chamber* provide excellent fuel efficiency and optimum combustion stability.
- VGF's *optimum BMEP* results in better service life for key components and long maintenance intervals.

■ *Patented high turbulence combustion chamber* allows ability to burn lean air fuel mixtures with an open combustion chamber. This produces low emissions across engine load and speed range. It also provides lower fuel consumption, which is typically 91%




of life cycle costs.

- *Low fuel pressure GLD model* has draw-thru carburetion capability which increases application potential. It allows operation with

■ *Individual, four valve, water cooled*

cylinder heads ensure optimum performance and reliability. Cooler cylinder head operation results in longer valve train component life and lower maintenance cost.



Engine Works so Well.


nance cost.


■ Waukesha's *multi-fuel capability* means you have various options for primary fuel requirements – unlike typical dual fuel systems where one fuel is used for primary requirements and the second fuel as back-up. The VGF is adaptable to natural gas and propane as well as digester gas and even landfill gas.

■ Standard *Custom Engine Control*® (CEC) Ignition Module provides precise timing and control for reduced emissions and fuel costs. No wearing parts means consistent ignition performance and lower maintenance costs.

■ Because of the *tremendous torque* of VGF engines, RPMs can be turned down to reduce fuel consumption and still maintain constant torque capability.

■ When required by local restrictions, the VGF can be ordered in a *rich burn draw*






■ *High temperature cooling capability* increases heat recovery opportunities, making the VGF a better fit for chiller or cogeneration applications.

■ Virtually all *wearing components* of the VGF series are *interchangeable* across the VGF engine series. This includes major components – pistons, rings, sleeves, conrods, cylinder heads, bearings and valve train parts. So not only do you have proven component reliability, you have lower parts inventory requirement even if you're running different models of the VGF.




■ *A fully counterweighted crankshaft* helps prevent engine vibration for a smoother running package. It minimizes main and rod bearing loads and maximizes component life. VGF shafts are T-drilled for continuous lubrication of main bearings, keeping them cooler

Engine Performance Refle



Emission | Performance vs. % Load

1500 & 1800
@ 7.8% O₂ Exhaust
130°F ICW, 185°F JW





and cleaner for longer life.

■ *Flanged connection points* make installation and alignment easy. Compact packaging simplifies set-up and adjustments for reduced costs.

■ *Maintenance is simplified* because the size of the engine and the thoughtful location

of service points and controls places everything conveniently within the technician's reach. The VGF engine

series allows *complete in-place maintenance and in-frame overhaul*. All VGFs have camshaft and connrod/crankshaft inspection doors. Vee engines have a base-type oil pan with inspection doors.


All these features point to the fact that

the VGF is intelligently designed and quality built to provide you with a compact, high speed gas engine in the Waukesha tradition of rugged reliability. This is an engine that will work long and hard for you. And it will have overhaul intervals so far apart you'll forget the meaning of downtime.

When it comes to life cycle numbers, the


Waukesha VGF adds up to be the right choice.

cts Company Performance.



Emission | Performance vs. % Load


1500 & 1800
@ 8.2% O₂ Exhaust (T.A. LUFT)
130° F ICW, 165° F JW



From design and development to manufacturing to total aftermarket support, the VGF family of engines symbolizes Waukesha's commitment to its customers. We build engines worthy of your installation

and your investment. And we stand behind them with a worldwide distributor network. **The Best in the Business.** The VGF line is backed by the best warranty of any industrial engine manufacturer. On

Why You Should Specify



VGF PERFORMANCE

Configuration	Engine Model	Intercooler Water Temperature °F (°C)
V16	P48GL/GLD	130° (54°)
V12	L36GL/GLD	130° (54°)
Inline 8	H24GL/GLD	130° (54°)
	H24G	
Inline 6	F18GL/GLD	130° (54°)
	F18G	

VGF STANDBY RATINGS

Configuration	Engine Model
V16	P48GL/GLD
V12	L36GL/GLD
Inline 8	H24GL/GLD
Inline 6	F18GL/GLD

VGF STATS

All data are based on standard conditions of

29.54 inches Hg. (100 kPa) barometric pressure,
77°F (25°C) ambient and induction air temperature,
30% relative humidity at 0.3 inches Hg. (1 kPa)
water vapor pressure, 185° F (85° C) engine jacket
water outlet temperature. Rating fuel standard:
Commercial quality dry natural gas.

Configuration	Engine Model	Bore x Stroke in (mm)
V16	P48GL/GLD	5.98 x 6.5 (152 x 165)
V12	L36GL/GLD	5.98 x 6.5 (152 x 165)
Inline 8	H24GL/GLD	5.98 x 6.5 (152 x 165)
	H24G	5.98 x 6.5 (152 x 165)
Inline 6	F18GL/GLD	5.98 x 6.5 (152 x 165)
	F18G	5.98 x 6.5 (152 x 165)

new engines, all parts and labor are covered for one full year. Five years on major forgings and castings. New service parts are warranted for one year along with the labor costs to replace them.

We built and backed this engine so you can be sure. The Waukesha VGF is the easy choice. The smart choice. The right choice.

the Waukesha VGF.


Continuous Brake Horsepower (kW/b) at Indicated RPM

1200 rpm		1500 rpm		1800 rpm	
710	(530)	885	(660)	1065	(800)
530	(400)	665	(500)	800	(600)
355	(265)	445	(330)	530	(400)
215	(160)	265	(200)	320	(240)
265	(200)	330	(250)	400	(300)
160	(120)	200	(150)	240	(180)

Brake Horsepower (kW/b) at Indicated RPM

1500 rpm (50 Hz)		1800 rpm (60 Hz)	
1025	(765)	1230	(918)
770	(574)	920	(686)
510	(380)	615	(460)
385	(285)	460	(345)

Displacement cu in (liters)	Weight		Height		Length		Width	
	lbs	(Kgs)	in	(mm)	in	(mm)	in	(mm)
2924 (48)	14,900	(6,759)	75.5	(1918)	106.5	(2705)	62.0	(1574)
2193 (36)	11,525	(5,228)	75.5	(1918)	88.0	(2235)	62.0	(1574)
1462 (24)	7,200	(3,265)	68.0	(1727)	94.5	(2395)	50.0	(1264)
1462 (24)	7,200	(3,265)	68.0	(1727)	94.5	(2395)	50.0	(1264)
1096 (18)	5,500	(2,495)	68.0	(1727)	80.5	(2043)	50.0	(1264)
1096 (18)	5,500	(2,495)	68.0	(1727)	80.5	(2043)	50.0	(1264)





*University of Illinois
in Chicago installed an
L36GL standby set.*

Strong Work Ethic.

Waukesha has nearly a century of stationary gas engine experience in rugged applications all over the world. Simply put, Waukesha engines are built to work. The VGF

continues that heritage.

The VGF is the perfect blend of compact size, brute strength and the right technology. No wasted weight, space or hardware. The result is an engine of timely design



*Waukesha VGF F18C
drives 90 kW induction
generator in parallel with
utility for Canonsburg,
Pennsylvania, wastewater
treatment plant.*

*Recovered heat is used to
maintain temperature in
digesters.*

The Proof is in



*A landfill at Vlagheide in Schijndel,
The Netherlands, has five H24GLDs
efficiently generating electricity from low
BTU landfill gas.*



**This F18GL engine
compresses gas for
Oceot Energy, Inc.
in south central
Alberta, Canada**



*Elkerliek Hospital in The Netherlands uses a P48GLD
and an H24GLD for standby power and cogeneration.*

and enduring reliability. It is economical to own and operate. And it lends itself to extraordinary installation versatility.

The VGF is manufactured in the USA at Waukesha Engine Division in

Wisconsin and in Europe at Waukesha Engine Division, Appingedam, The Netherlands.

Nearly one thousand VGF engines have been installed since product introduction in 1987.



A 700 kW VGF P48GL cogen system generates electricity and warehouse heat for this grain drying operation in Dungannon, Ontario.

the Performance.



CNG Transmission Corp. in Utica, NY, has a VGF H24GL driving a 375 kW generator for standby electrical power at a compressor station.



This greenhouse in Antwerpen,


Holland, obtains

electricity and heat from an F18GLD cogeneration system.




A VGF H24GL genset rated at 350 kW is used for peak shaving at Lethbridge

Regional Hospital in Alberta, Canada. Exhaust heat is converted to steam via a waste heat boiler.



Greensburg Sewer Authority, Greensburg, PA, uses a digester gas powered VGF F18G for aeration in wastewater treatment. Recovered engine heat maintains temperature in digesters.



The Stegeman Meat Factory in Deventer, The Netherlands, has an L36GLD cogen system supplying electricity as well as hot water for absorption cooling.



Waukesha

Waukesha Engine

Dresser, Inc.

1000 West St. Paul Avenue

Waukesha, WI 53188-4999

Fax: (262) 549-2795

(262) 547-3311

waukeshaengine.dresser.com

Waukesha Engine

Dresser Industrial Products, b.v.

Farmsumerweg 43, Postbus 330

9900 AH Appingedam, The Netherlands

Fax: (31) 596-628111

(31) 596-652222

Printed on Recycled Paper
Bulletin No. 1323
5M 07/01



101 Industrial Blvd.
Mansfield, TX 76063-3611
Phone (817) 840-5544
Fax (817) 453-0219
Toll Free 1-800-888-5557

Equipment Proposed:

Kohler 625 KW Standby Generator
Model Number 600RZW

- Engine Waukesha VGF L36GLD Lean-Burn
- Natural Gas Fueled
- Isochronous Governor
- Heavy duty dry type air filters
- Lube oil filters and drain connections to skid
- Unit mounted radiator with shell and fan guard
- 24-volt starter and alternator
- Battery with cables and mounting tray
- Battery charger 10 amp with charge rate meter
- Muffler critical type with flex connector

Engine block heater thermostatically controlled
Generator: 1800 RPM, PMG type excitation, 60 Hz, 3 phase, 480 volt, .25% constant voltage regulator, Class H insulation, direct flex disc drive, 10 or 12 lead reconnectable

Decision-Maker 550 Digital Generator Controller
Detailed Specs Included at end of this document

- All weather outdoor sound attenuated generator enclosure rated 85 dba @ 1 meter with interior lighting
- 1000 amp 100% rated main line circuit breaker mounted in connection box with GFI
- Spring isolators
- Dry contacts
- Alternator heater
- Woodward speed controller
- Marathon DVR2000EC voltage regulator
- Natural gas regulator 8" W.C.-2.5psig
- Natural gas solenoid valve

Project Management Services:

- Start-up, three up to 8-hour days including travel
- Training of owner's personnel at time of start up
- Operation of equipment at Wedlake Manufacturing in Tulsa, OK
- Factory load bank testing
- Submittal drawings
 - Dimensional drawings, Electrical drawings, Product specifications
- Production testing
- Technical assistance
- Operation and maintenance manuals for engine, generator and controls
- One (1) year warranty

Decision-Maker 550 Controller Details

Standards:

- NFPA 99
- NFPA 110, Level 1
- UL-508 (pending)

Hardware Features

- Alarm horn
- Battery circuits are fuse protected
- Controller mounts in four orientations locally or remotely up to a distance of 12 m (40 ft.)
- Five LED status indicating lights
- Latch-type emergency stop switch
- Vacuum fluorescent display
- Terminals for remote annunciator
- Three-position (run, off/reset, auto) selector switch

Shutdown Functions

Engine functions:

CLIFFORD POWER SYSTEMS, INC. IS COMMITTED TO BE THE PREFERRED LEADER IN THE POWER GENERATION INDUSTRY. WE WILL FULFILL THIS MISSION BY PROVIDING OUR CUSTOMERS WITH SERVICE ABOVE AND BEYOND THEIR EXPECTATIONS.

PAGE 002 OF 0

- Air damper fault, if equipped
- High coolant temperature
- High oil temperature
- Low coolant level
- Low oil pressure
- Overcrank
- Overspeed
- General functions:
 - Auxiliary—(up to 7 analog inputs each with a high and low programmable shutdown level). NOTE: Non-ECM models have 5 programmable shutdowns)
 - Auxiliary—Digital (up to 21 programmable shutdowns)
 - ECM communications loss (ECM models only)
 - Emergency stop
 - Internal fault
 - Master switch in off/reset position
 - Master switch error
 - Master switch open
 - NFPA 110 fault

Generator functions:

- Alternator protection against overload and short circuits
- Locked rotor (failed to crank)
- Over AC output voltage
- Overfrequency
- Under AC output voltage
- Underfrequency

Warning Functions

Engine functions:

- Coolant temperature signal loss
- High battery voltage
- High coolant temperature
- Low battery voltage
- Low coolant temperature
- Low fuel (level or pressure)*
- Low oil pressure
- Oil pressure signal loss
- Speed sensor fault
- Starting aid fault
- Weak battery

General functions:

- Auxiliary—(up to 7 analog inputs each with a high and low programmable warning level). NOTE: Non-ECM models have 5 programmable warnings)
- Auxiliary—Digital (up to 21 programmable warnings)
- Battery charger fault*
- Emergency power system (EPS) supplying load
- Engine cooldown delay
- Engine start delay
- Load shed kW overload
- Load shed underfrequency
- Master switch not in auto
- NFPA-110 fault
- System ready

Generator functions:

- AC sensing loss
- Generator running
- Ground fault*
- Overcurrent
- Underfrequency
- Requires optional input sensors.

User-Defined Common Faults

The user customizes outputs through a menu of shutdowns and warnings.

User defines up to 21 relay driver outputs (RDOs), (relays not included) from the following list of functions:

Engine functions:

- Air damper fault, if equipped
- Coolant temperature signal loss
- High battery voltage
- High coolant temperature shutdown
- High coolant temperature warning
- High oil temperature (ECM models only)
- Low battery voltage

CLIFFORD POWER SYSTEMS, INC. IS COMMITTED TO BE THE PREFERRED LEADER IN THE POWER GENERATION INDUSTRY. WE WILL FULFILL THIS MISSION BY PROVIDING OUR CUSTOMERS WITH SERVICE ABOVE AND BEYOND THEIR EXPECTATIONS.

- Low coolant level
- Low coolant temperature
- Low fuel (level or pressure)*
- Low oil pressure shutdown
- Low oil pressure warning
- Oil pressure signal loss
- Overcrank
- Overspeed
- Speed sensor fault
- Starting aid
- System ready
- Weak battery

General functions:

- Battery charger fault*
- ECM communications loss (ECM models only)
- EEPROM write failure
- Emergency stop
- Engine cooldown delay
- Engine start delay
- EPS supplying load
- Internal fault
- Load shed kW overload
- Load shed underfrequency
- Master switch error
- Master switch not in auto
- Master switch to off
- NFPA 110 common alarm fault

Generator functions:

- AC sensing loss
- Alternator protection against overload and short circuits
- Generator running
- Ground fault
- Locked rotor (failed to crank)
- Overcurrent
- Overfrequency
- Overvoltage
- Underfrequency
- Undervoltage

* Requires optional input sensors.

NFPA-110 Common Alarms

Additional annunciated alarms including NFPA 110 alarms.

Engine functions:

- High battery voltage
- High coolant temperature shutdown
- High coolant temperature warning
- Low battery voltage
- Low coolant temperature warning
- Low fuel (level or pressure)*
- Low oil pressure shutdown
- Low oil pressure warning
- Overcrank
- Overspeed

General functions:

- Battery charger fault*
- Master switch not in auto
- NFPA 110 common alarm

* Requires optional input sensors.

Monitoring

Standard Equipment and Features

Alarm horn

Indicators:

- Not in auto (yellow)
- Program mode (yellow)
- System ready (green)
- System shutdown (red)
- System warning (yellow)
- Switches and standard features:
- Keypad, 16-button multi-function soft-membrane environmentally sealed
- Lamp test

CLIFFORD POWER SYSTEMS, INC. IS COMMITTED TO BE THE PREFERRED LEADER IN THE POWER GENERATION INDUSTRY. WE WILL FULFILL THIS MISSION BY PROVIDING OUR CUSTOMERS WITH SERVICE ABOVE AND BEYOND THEIR EXPECTATIONS.

- Switch, auto/off-reset/run (engine start)
- Switch, emergency stop (normally closed contacts)
- Vacuum fluorescent display with two lines of 20 characters

Displays

Some engine displays are dependent upon enhanced electronic engine control availability.

Engine monitoring (metric or English units):

- Ambient temperature (ECM models only)
- Battery voltage
- Coolant—pressure, temperature, and level (ECM models only)
- Engine start countdown
- Fuel—pressure, temperature, fuel rate, amount of fuel used during last run (ECM models only)
- Oil—pressure, temperature, level, and crankcase pressure
- RPM

Generator monitoring:

- Current (L1, L2, L3), _ 0.25% accuracy
- Frequency, _ 0.5% accuracy
- Kilowatts, total per phase (L1, L2, L3), _ 0.5% accuracy
- KVA, total per phase (L1, L2, L3), _ 0.5% accuracy
- KVAR, total absorbing/generating per phase (L1, L2, L3), _ 0.5% accuracy
- Percent alternator duty level (actual load kW/standby kW rating)
- Power factor per phase, leading/absorbing
- Voltage (line-to-line, line-to-neutral for all phases), _ 0.25% accuracy

Operational records:

- Event history (stores up to 100 system events)
- Last start date
- Number of starts
- Number of starts since last maintenance
- Operating days since last maintenance
- Operating mode—standby or prime power
- Run time (total, loaded and unloaded hours, and total kW hours)
- Run time since maintenance (total, loaded, and unloaded hours and total kW hours)
- System shutdowns
- System warnings
- Time, date, and day of week

Time delays:

- Crank cycles for on/pause
- Crank cycles for overcrank shutdown
- Engine cooldown
- Engine start
- Load shed
- Voltage, over- and under-
- Starting aid

System parameters:

- Current, rated
- ECM serial number (ECM models only)
- Engine model number
- Engine serial number
- Frequency
- Generator set model number
- Generator set serial number
- Generator set spec number
- kW Rating
- Phase, single and three
- Unit number
- Voltage
- Voltage configuration, wye or delta

Inputs

Customer and remote inputs:

- Analog inputs 0-5 VDC (up to 7 user-defined analog inputs with multiple shutdown and warning levels). NOTE: Non-ECM models have 4 programmable shutdowns)
- Digital contact (up to 21 user-defined digital inputs with shutdown or warning levels).
- Ground fault detector*
- Remote emergency stop
- Remote reset
- Remote 2-wire start

Digital inputs:

- Air damper fault, if equipped
- Battery charger fault*
- Emergency stop

CLIFFORD POWER SYSTEMS, INC. IS COMMITTED TO BE THE PREFERRED LEADER IN THE POWER GENERATION INDUSTRY. WE WILL FULFILL THIS MISSION BY PROVIDING OUR CUSTOMERS WITH SERVICE ABOVE AND BEYOND THEIR EXPECTATIONS.

PAK 11000

KOHLER POWER SYSTEMS

Model: 600RZW

190-600 V

Gas



Ratings Range


	Lean-Burn GLD Engine		Rich-Burn GSID Engine	
	60 Hz	50 Hz	60 Hz	50 Hz
Standby: kW	620-625	524-528	600	500-504
kVA	775-781	665-660	750	625-630
Prime: kW	510-600	428-504	510-540	428-456
kVA	638-750	535-630	638-675	535-570

Generator Set Ratings

Alternator Voltage	Ph	Hz	Lean-Burn GLD Engine			Rich-Burn GSID Engine			
			130°C Rise Standby	105°C Rise Prime	80°C Rise Prime	130°C Rise Standby	105°C Rise Prime	80°C Rise Prime	
			KW/KVA	KW/KVA	KW/KVA	KW/KVA	KW/KVA	KW/KVA	
5M4032	120/208	3	60	620/775	590/738	570/713	600/750	535/669	535/669
	127/220	3	60	620/775	590/738	590/738	600/750	535/669	535/669
	139/240	3	60	625/781	595/744	570/713	600/750	540/675	540/675
	240/416	3	60	620/775	590/738	570/713	600/750	535/669	535/669
	277/480	3	60	625/781	595/744	570/713	600/750	540/675	540/675
	110/190	3	50	524/655	500/625	488/610	500/625	458/570	458/570
	115/200	3	50	524/655	500/625	472/590	500/625	458/570	458/570
	120/208	3	50	524/655	500/625	440/550	500/625	458/570	440/550
	220/380	3	50	524/655	500/625	488/610	500/625	458/570	458/570
	230/400	3	50	524/655	500/625	472/590	500/625	458/570	458/570
240/416	3	50	524/655	500/625	440/550	500/625	458/570	440/550	
5M4034	120/208	3	60	625/781	595/744	590/738	600/750	540/675	540/675
	127/220	3	60	625/781	595/744	595/744	600/750	540/675	540/675
	139/240	3	60	625/781	595/744	595/731	600/750	540/675	540/675
	240/416	3	60	625/781	595/744	590/738	600/750	540/675	540/675
	277/480	3	60	625/781	595/744	585/731	600/750	540/675	540/675
	110/190	3	50	528/660	504/630	500/625	504/630	458/570	458/570
	115/200	3	50	528/660	504/630	472/590	504/630	458/570	458/570
	120/208	3	50	528/660	504/630	428/535	504/630	458/570	428/535
	220/380	3	50	528/660	504/630	500/625	504/630	458/570	458/570
	230/400	3	50	528/660	504/630	472/590	504/630	458/570	458/570
240/416	3	50	528/660	504/630	428/535	504/630	458/570	428/535	
5M4036	120/208	3	60	625/781	595/744	600/750	600/750	540/675	540/675
	127/220	3	60	625/781	595/744	570/713	600/750	540/675	535/669
	139/240	3	60	625/781	595/744	595/744	600/750	540/675	540/675
	240/416	3	60	625/781	595/744	595/744	600/750	540/675	540/675
	277/480	3	60	625/781	595/744	595/744	600/750	540/675	540/675
	110/190	3	50	528/660	504/630	504/630	504/630	458/570	458/570
	115/200	3	50	528/660	504/630	504/630	504/630	458/570	458/570
	120/208	3	50	528/660	504/630	480/600	504/630	458/570	458/570
	220/380	3	50	528/660	504/630	504/630	504/630	458/570	458/570
	230/400	3	50	528/660	504/630	504/630	504/630	458/570	458/570
240/416	3	50	528/660	504/630	480/600	504/630	458/570	458/570	
5M4164	220/380	3	60	625/781	595/744	570/713	600/750	540/675	510/638
5M4166	220/380	3	60	625/781	595/744	595/744	600/750	540/675	540/675
5M4274	347/600	3	60	625/781	595/744	530/683	600/750	535/669	530/663
5M4276	347/600	3	60	625/781	595/744	585/706	600/750	540/675	540/675

Standard Features

- Kohler Co. provides one-source responsibility for the generating system and accessories.
- The generator set and its components are prototype-tested, factory-built, and production-tested.
- The generator set complies with ISO 8528-3, Class G4 requirements for transient performance.*
- A one-year limited warranty covers all systems and components. Two-, five-, and ten-year extended warranties are also available.
- Alternator features:
 - The brushless, rotating-field alternator has broadrange reconfigurability.
 - The pilot-excited, permanent-magnet (PM) alternator provides superior short-circuit capability.
- Other features:
 - The low coolant level shutdown prevents overheating (standard on radiator models only).
 - The generator set is direct-mounted to the skid.
 - An electronic, isochronous governor delivers precise frequency regulation.
 - Electronic engine controls manage the engine.
 - Lean-burn natural gas technology provides maximum power and fuel efficiency.
 - Rich-burn natural gas technology reduces harmful exhaust emissions when used with a catalytic converter.
- * This generator set does not meet NFPA 110 requirements for the one-step load acceptance and the 10-second start sequence.



RATINGS: All three-phase units are rated at 0.8 power factor. Standby Ratings: Standby ratings apply to installations served by a reliable utility source. The standby rating is applicable to varying loads for the duration of a power outage. There is no overload capability for this rating. Ratings are in accordance with ISO-3046/1, BS 6614, AS 2769, and DIN 6271. Prime Power Ratings: Prime power ratings apply to installations where utility power is unavailable or unreliable. At varying load, the number of generator set operating hours is unlimited. Rich Burn: A 10% overload capacity is available for one hour in twelve. Lean Burn: A 5% overload capacity is available for one hour in twelve. Ratings are in accordance with ISO-3046/1, BS 6614, AS 2769, and DIN 6271. For limited running time and base load ratings, consult the factory. Obtain the technical information bulletin (TIB-101) on ratings guidelines for the complete ratings definitions. The generator set manufacturer reserves the right to change the design or specifications without notice and without any obligation or liability whatsoever. GENERAL GUIDELINES FOR DEPARTION: Altitude: Deduct 1% for each 150 m (492 ft.) elevation above 500 m (1640 ft.). Temperature: Deduct 2% for each 10°C (18°F) temperature above 38°C (100°F).

04-91 (600RZW) 1/03

Alternator Specifications

Specifications		Alternator
Type		4-Pole, Rotating-Field
Exciter type		Brushless, Permanent-Magnet, Pilot Exciter
Voltage regulator		Solid State, Volts/Hz
Insulation:		NEMA MG1
Material		Class H, Synthetic, Nonhygroscopic
Temperature rise		130°C, 150°C, Standby
Bearing: quantity, type		1, Sealed
Coupling		Flexible Disc
Amortisseur windings		Full
Rotor balancing		125% (60 Hz), 150% (50 Hz)
Voltage regulation, no-load to full-load (with <0.5% drift due to temp. variation)		3-phase, ±0.25%
Unbalanced load capability		100% of Rated Standby Current
Peak motor starting kVA:		(35% dip for voltages below)
480 V/380 V	5M4032 (10 lead)	2200 (60Hz), 1375 (50Hz)
480 V/380 V	5M4034 (10 lead)	2600 (60Hz), 1750 (50Hz)
490 V/380 V	5M4036 (10 lead)	3150 (60Hz), 2100 (50Hz)
380 V	5M4164 (4 lead)	2250 (60Hz)
380 V	5M4166 (4 lead)	2750 (60Hz)
600 V	5M4274 (4 lead)	1545 (60Hz)
600 V	5M4276 (4 lead)	2800 (60Hz)

- NEMA MG1, IEEE, and ANSI standards compliance for temperature rise and motor starting.
- Sustained short-circuit current of up to 300% of the rated current for up to 10 seconds.
- Sustained short-circuit current enabling downstream circuit breakers to trip without collapsing the alternator field.
- Self-ventilated and dripproof construction.
- Superior voltage waveform from two-thirds pitch windings and skewed stator.
- Digital solid-state, volts-per-hertz voltage regulator with ±0.25% no-load to full-load regulation.
- Brushless alternator with brushless pilot exciter for excellent load response.

Application Data

Engine

Engine Specifications	60 Hz	50 Hz
Manufacturer	Waukesha Engine	
Engine model		
Lean-Burn GLD Engine	VGF L36GLD, 4-Cycle	
Rich-Burn GSID Engine	VGF L36GSID, 4-Cycle	
Engine type	Turbocharged, Intercooled	
Cylinder arrangement	12 V	
Displacement, L (cu. in.)	36 (2196)	
Bore and stroke, mm (in.)	152 x 165 (5.98 x 6.5)	
Compression ratio		
Lean-Burn GLD Engine	11:1	
Rich-Burn GSID Engine	8.7:1	
Piston speed, m/min. (ft./min.)	594 (1950)	495 (1625)
Main bearings: quantity, type	7, Half-Shell	
Rated rpm	1800	1500
Max. power at rated rpm, kWm (BHP)		
Lean-Burn GLD Engine	690 (925)	574 (770)
Rich-Burn GSID Engine	655 (880)	548 (735)
Cylinder head material	Cast Iron	
Piston: type, material	Aluminum Alloy	
Crankshaft material	Forged Steel	
Valve material, intake/exhaust:	Hard-Faced Steel	
Governor: type, make/model	Electronic	
Frequency regulation, no-load to full-load	Isochronous	
Frequency regulation, steady state	±0.50%	
Frequency	Field-Convertible	
Air cleaner type, all models	Dry	

Fuel

Fuel System	60 Hz	50 Hz
Fuel type	Natural Gas	
Fuel supply line inlet, mm (in.)	50.8 (2) ANSI 125 lb. Flange	
Natural gas fuel supply pressure, measured at the generator set fuel inlet after any fuel system equipment accessories, kPa (oz./in. ²)	2-34 (4.8-80)	
Particulate filter requirement, mm (in.)	0.005 (0.0002)	

Exhaust

Exhaust System	60 Hz	50 Hz
Exhaust flow at rated kW, m ³ /min. (cfm)		
Lean-Burn GLD Engine	135 (4765)	108 (3816)
Rich-Burn GSID Engine	105 (3755)	84 (2952)
Exhaust temperature at rated kW, dry exhaust, °C (°F)		
Lean-Burn GLD Engine	450 (843)	427 (800)
Rich-Burn GSID Engine	500 (1114)	579 (1074)
Maximum allowable back pressure, kPa (in. Hg)	9.73 (1.1)	
Engine exhaust outlet size, mm (in.)	See ADV Drawing	

Engine Electrical

Engine Electrical System	60 Hz	50 Hz
Ignition system	Electronic	
Battery charging, min.	Requires Float/Equalizer Battery Charger, 24 V, 10 A	
Starter motor rated voltage (DC)	24	
Battery, recommended cold cranking amps (CCA):		
Qty., CCA rating	2, 1150	
Battery voltage (DC)	12	

Lubrication

Lubricating System	60 Hz	50 Hz
Type	Full Pressure	
Oil pan capacity, L (qt.)	163 (43)	
Oil pan capacity with filter, L (gal.)	2, Cartridge	
Oil filter: quantity, type	Water-Cooled	
Oil cooler	SAE40 Allowable Sulfated Ash Content by Weight	
Oil requirements:	0.5-1.0% (GLD) 0.35-0.5% (GSID)	

G4-01 (800PZM) 1/98

Application Data

Cooling

Radiator System	60 Hz	50 Hz
Ambient temperature, °C (°F)	38 (100)	
Engine jacket water capacity, L (gal.)	166 (44)	
Engine auxiliary water capacity, L (gal.)	57 (15)	
Radiator jacket water capacity, including engine, L (gal.)	357 (94)	
Radiator auxiliary water capacity, including engine, L (gal.)	224 (59)	
Minimum engine jacket water flow, Lpm (gpm)		
Lean-Burn GLD Engine	625 (218)	697 (184)
Rich-Burn GSID Engine	997 (263)	841 (222)
Minimum engine auxiliary water flow, Lpm (gpm)	235 (62)	197 (52)
Heat rejected to cooling water at standby rated kW, wet exhaust, kW (Btu/min.)		
Lean-Burn GLD Engine	648 (26920)	401 (22780)
Rich-Burn GSID Engine	574 (26770)	476 (27080)
Heat rejected to auxiliary cooling water at standby rated kW, wet exhaust, kW (Btu/min.)		
Lean-Burn GLD Engine	182 (10370)	128 (7300)
Rich-Burn GSID Engine	131 (7470)	102 (5800)
Water pump type	Centrifugal	
Fan diameter, including blades, mm (in.)	1829 (72)	
Fan, kWm (HP)	31 (42)	19 (25)
Max. restriction of cooling air, intake and discharge side of radiator, kPa (in. H ₂ O)	0.125 (0.5)	

Operation Requirements

Air Requirements	60 Hz	50 Hz
Radiator-cooled cooling air, m ³ /min. (scfm)*	1690 (5970)	1410 (4990)
Combustion air, m ³ /min. (scfm)		
Lean-Burn GLD Engine	52 (1820)	42 (1485)
Rich-Burn GSID Engine	33 (1180)	27 (965)
Heat rejected to ambient air, kW (Btu/min.):		
Lean-Burn GLD Engine	34 (1933)	31 (1783)
Rich-Burn GSID Engine	42 (2383)	39 (2233)
Alternator	42 (2383)	35 (1960)

* Air density = 1.20 kg/m³ (0.075 lbm/ft³)

Fuel Consumption†	60 Hz	50 Hz
Natural Gas, m ³ /hr. (cfh) at % load	Lean-Burn Standby Rating	
100%	202 (7153)	164 (5794)
75%	159 (5603)	128 (4533)
50%	115 (4054)	93 (3271)
25%	71 (2504)	57 (2010)
Natural Gas, m ³ /hr. (cfh) at % load	Lean-Burn Prime Rating	
100%	184 (6551)	157 (5549)
75%	152 (5377)	123 (4348)
50%	111 (3903)	89 (3148)
25%	69 (2429)	55 (1948)
Natural Gas, m ³ /hr. (cfh) at % load	Rich-Burn Standby Rating	
100%	201 (7105)	165 (5819)
75%	159 (5602)	129 (4569)
50%	116 (4098)	94 (3319)
25%	73 (2594)	59 (2070)
Natural Gas, m ³ /hr. (cfh) at % load	Rich-Burn Prime Rating	
100%	186 (6559)	162 (5364)
75%	147 (5192)	120 (4228)
50%	108 (3825)	88 (3092)
25%	70 (2468)	55 (1956)

† Fuel energy content = 35.38 MJ/m³ (900 Btu/scf) saturated lower heating value.

Controller



Decision-Maker™ 550 Controller

Audiovisual annunciation.

Programmable microprocessor logic and digital display features.

Alternator safeguard circuit protection.

24-volt engine electrical system capability.

Remote start, remote annunciation, and remote communication options.

Refer to G6-46 for additional controller features and accessories.

Standard Features and Accessories

Standard Features

- Air Cleaner, Heavy Duty
- Air Cleaner Restriction Indicator
- Alternator Protection
- Oil Drain Extension
- Operation and Installation Literature
- Radiator Duct Flange

Accessories

- Enclosed Unit**
 - Sound Enclosure (with enclosed critical silencer)
 - Weather Enclosure (with enclosed critical silencer)
- Open Unit**
 - Exhaust Silencer, Critical, Lean-Burn GLD Engine:
60 Hz kit: PA-354896; 50 Hz kit: PA-354894
 - Exhaust Silencer, Critical, Rich-Burn GSID Engine:
Kit: PA-354894
 - Exhaust Silencer, Residential, Kit: PA-354892
 - Flexible Exhaust Connector, Stainless Steel
- Cooling System**
 - Block Heater
 - Remote Radiator Cooling
- Fuel System**
 - Air/Fuel Ratio Controller
 - Gas Regulator
 - Natural Gas Filter
 - Gas Solenoid Valve
- Electrical System**
 - Battery
 - Battery Charger, Equalize/Float Type
 - Battery Heater
 - Battery Rack and Cables
- Engine and Alternator**
 - Bus Bar Kits
 - CSA Certification
 - Alternator Strip Heater
 - Line Circuit Breaker (NEMA1 enclosure) *100% rated*
 - Line Circuit Breaker with Shunt Trip (NEMA1 enclosure)
 - Optional Alternators
 - Pre-Lube Pumps
 - Pre-Lube Pumps with Heaters
 - Rated Power Factor Testing
 - Remote Voltage Adjust Control
 - Spring Isolators

Maintenance and Literature

- General Maintenance Literature Kit
- Maintenance Kit (Includes air, oil, and fuel filters)
- Overhaul Literature Kit
- Production Literature Kit

Controller


- Common Failure Relay Kit
- Communications Products and PC Software
- Customer Connection Kit
- Dry Contact Kit (isolated alarm)
- Remote Annunciator Panel
- Remote Audiovisual Alarm Panel
- Remote Emergency Stop Kit
- Remote Mounting Cable
- Run Relay Kit

Miscellaneous Accessories

- Woodsward Controller*
- Manometer NVD 2000E*
-
-
-
-
-
-

Dimensions and Weights

Overall Size, L x W x H, mm (in.): 4824 x 2416 x 3111
 (193.8 x 95.1 x 122.5)
 Weight (radiator model), wet, kg (lb.): 10750 (23700)



NOTE: This drawing is provided for reference only and should not be used for planning installation. Contact your local distributor for more detailed information.

DISTRIBUTED BY:

Attachment 7
Generator Engine Operating Log

Attachment 8

Ammonia Tank PSV Calibrations Records

Basin Valve Company

Safety Valve Service Report

Customer : Southern California Edison

PO No :

Site : Peaker Generating Stations

Rev : 10

SO No : 220208-4

Date Required : 4/27/2023

4:00 PM

Location :

Repair Nameplate Data		By : EU		Original Nameplate Data		By : EU	
SAP Number :	Pretest Date :	Set Pressure : 50	Cold Set : 50	Back Press :			
Tag Number : PSV-201-A-C	Job Number : 13DA	Capacity : 383	% Over Pressure : 10	Units : GPM			
Manufacturer : CROSBY	Orifice : K	Temp Corr :		Code Stamp : LIV			
Model No : JLT-JOS-E-15-OR-J	Set Press : 50	Model Number : JLT-JOS-E-15-OR-J		Code Case :			
Serial No : HS06-55271	Cold Set : 50	Serial Number : HS06-55271					
Shop No : 82353544E	Back Press : 0	Last Repair Nameplate Data		By : EU			
Interval : 0	Capacity : 383	Set Pressure : 50	Cold Set : 50	Back Pressure :			
Units : GPM	Temp Corr : 0	Capacity : 383	Units : GPM	Unique ID : 188305			
Inter : 3 / 150 / RF	Outlet : 4 / 150 / RF	Repair Company : Basin Valve Company					
		Last Repair Date : 4/17/2018		VR : <input checked="" type="checkbox"/>			
		Field Location : MCGRATH					

Work Order :

Pre-Test Information

By : N/A

Pop PSIG : 51.82	New Valve :	No Pop :	Faulted :	Leaked At :
Probable Cause of Failure : Valve performed properly				

Conditions After Dismantling

By : N/A

Cleanliness Condition : Normal condition Mechanical Condition : Passed pretest-No internal inspection made

Parts	Pretest Conditions	Work Performed	Note
Bonnet Assembly			
Spring			
Upper Spring Washer			
Lower Spring Washer			
Compression Screw			
Compression Jam Nut			
Body			
Inlet Flange			
Outlet flange			
Guide			
Disc			
Disc Holder			
Nozzle			
Spindle			
		Qty	PO Number
		Part Number	Description
Overlap Adjustment			
Retaining Ring			
Floating Washer			
Spindle Nut			
Lock Nut			
Cap			
Top lever & Pin			
Lower Lever & Pin			
Upper Adj Ring & Pin			
Lower Adj Ring & Pin			
Bellovs			
Gaskets			

Critical Dimensions

Ring Adjustment / Spring Info

Disc : Min / Max :	After Repair :	Material Left :	Upper Ring : As Found :	After Repair :
Disc Holder : Min / Max :	After Repair :	Material Left :	Lower Ring : As Found :	After Repair :
Guide : Min / Max :	After Repair :	Material Left :	Compression Screw : As Found :	After Repair :
Nozzle : Min / Max : 4.654	After Repair :	Material Left :	Spindle / Stem : As Found :	After Repair :
			Spring Number : X36139	Spring Range : 39 - 60

Special Instructions :

Back Pressure Tested at PSIG <input type="text" value="30"/>	<input type="checkbox"/> O2 Cleaning Required	<input type="checkbox"/> Repair as Required	<input type="checkbox"/> Replace Soft Goods	Cap / Lever
Seat Leakage Tested at <input type="text" value="90"/>	<input type="checkbox"/> % of Set Pressure	<input checked="" type="checkbox"/> Final Test Only	<input type="checkbox"/> Pull from SP	<input checked="" type="checkbox"/> Seals Intact
	<input type="checkbox"/> Replace Next Shut Down	<input type="checkbox"/> Repair Return to SP	<input type="checkbox"/> Build From SP / Tag No	<input type="checkbox"/> Seals Broken
	<input type="checkbox"/> Add Face Seal Conn	<input type="checkbox"/> PreTest Return to SP	<input type="text" value=""/>	<input type="checkbox"/> Seals Missing
		<input type="checkbox"/> Change Set Pressure		<input checked="" type="checkbox"/> Open
				<input type="checkbox"/> Packed
				<input checked="" type="checkbox"/> Screwed

Valve Repair Note :

Assembled By : N/A Mach/Lapp By : N/A Final Test Information Cleaned By : N/A Inspected By : N/A

Test Specifications :	Test Medium : Liquid	Test Type : Seat	Set Pressure Definition : First Steady Stream	Code Stamp Applied : None
Set Pressure : 50	Average Test : 51.82	Test 1 : 51.82	Test 2 : 51.82	Test 3 : 51.82
Gauges Used :	Primary : CPU	Secondary : D-5		Hold Time :
Misc Items :	ID Tag : <input checked="" type="checkbox"/>	Paint :	Seals : <input checked="" type="checkbox"/>	Flange Protector : <input checked="" type="checkbox"/>
Final Test By : JR	Sig :	Final Test Date : 4/21/2023	Company : Basin Valve Company	
QC Witnessed By : Chris Bordewich	Sig :	QC Date : 4/24/2023	Company :	
VR Stamp Number/s : 20	Install Verified By :	Date :		

Basin Valve Company

Safety Valve Service Report

Customer : Southern California Edison

PO No :

Site : Peaker Generating Stations

Rev : 10

SO No : 220208-5

Date Required : 4/27/2023

4:00 PM

Location :

Repair Nameplate Data		By : EU		Original Nameplate Data		By : EU	
SAP Number :	Pretest Date :	Set Pressure : 50	Cold Set : 50	Back Press :			
Tag Number : P5V-201B	Job Number : 12FD	Capacity : 383	% Over Pressure : 10	Units : GPM			
Manufacturer : CROSBY	Office : K	Temp Corr :	Code Stamp : UV				
Model No : JLT-JOS-E-15-OR-J	Set Press : 50	Model Number : JLT-JOS-E-15-OR-J		Code Case :			
Serial No : HS06155268	Cold Set : 50	Serial Number : HS06155268					
Shop No : 82353644E	Back Press : 0	Last Repair Nameplate Data					
Interval : 0	Capacity : 383	Set Pressure : 0		Cold Set :		Back Pressure :	
Units : GPM	Temp Corr : 0	Capacity : 0		Units :		Unique ID :	
Inlet : 3 / 150 / RF	Outlet : 4 / 150 / RF	Repair Company :					
Last Repair Date :						VR :	
Field Location : MCGRATH							

Work Order : Pre-Test Information By : N/A

Pop PSIG : 51.04 New Valve : No Pop : Fouled : Leaked At :

Probable Cause of Failure : Valve performed properly

Cleanliness Condition : Normal condition Conditions After Dismantling By : N/A

Mechanical Condition : Passed pretest-No internal inspection made

Parts	Pretest Conditions	Work Performed	Note		
Bonnet Assembly					
Spring					
Upper Spring Washer					
Lower Spring Washer					
Compression Screw					
Compression Jam Nut					
Body					
Inlet Flange					
Outlet flange					
Guide					
Disc					
Disc Holder					
Nozzle					
Spindle		Qty	PO Number	Part Number	Description
Overlap Adjustment					
Retaining Ring					
Floating Washer					
Spindle Nut					
Lock Nut					
Cap					
Top lever & Pin					
Lower Lever & Pin					
Upper Adj. Ring & Pin					
Lower Adj. Ring & Pin					
Bellows					
Gaskets					

Critical Dimensions				Ring Adjustment / Spring Info			
Disc :	Min / Max :	After Repair :	Material Left :	Upper Ring :	As Found :	After Repair :	
Disc Holder :	Min / Max :	After Repair :	Material Left :	Lower Ring :	As Found :	After Repair :	
Guide :	Min / Max :	After Repair :	Material Left :	Compression Screw :	As Found :	After Repair :	
Nozzle :	Min / Max :	After Repair :	Material Left :	Spindle / Stem :	As Found :	After Repair :	
Special Instructions :				Spring Number :	Spring Range :		

Back Pressure Tested at PSIG O2 Cleaning Required Repair as Required Replace Soft Goods Cap / Lower

Seat Leakage Tested at % of Set Pressure Final Test Only P/I from SP Seals intact Open

Replace Next Shut Down Repair Return to SP Build From SP / Tag No Seals Broken Packed

Add Face Seal Conn Pre-Test Return to SP Seals Missing Screwed

Change Set Pressure

Valve Repair Note :

Assembled By : N/A	Mach/Lapp By : N/A	Final Test Information	Cleaned By : N/A	Inspected By : N/A
Test Specifications :	Test Medium : Liquid	Test Type : Seat	Set Pressure Definition : First Steady Stream	Code Stamp Applied : None
Set Pressure : 50	Average Test : 51.04	Test 1 : 51.04	Test 2 : 51.04	Test 3 : 51.04
Gauges Used :	Primary : CPU	Secondary : D-5	Hold Time :	
Misc Items :	ID Tag : <input checked="" type="checkbox"/>	Paint : <input type="checkbox"/>	Seals : <input checked="" type="checkbox"/>	Flange Protector : <input checked="" type="checkbox"/>
Final Test By : JR	Sig : _____	Final Test Date : 4/24/2023	Company : Basin Valve Company	
QC Witnessed By : Chris Bordewich	Sig : _____	QC Date : 4/24/2023	Company :	
VR Stamp Number : 20	Install Verified By :	Date :		

Attachment 9

VOC Emissions Log of Coating, Solvent, and Aerosol Spray

COATING, ADHESIVE, and SOLVENT USAGE CHART

COMPANY NAME: So Cal Edison McGrath Peaker PERMIT NUMBER: 7891
ADDRESS: 251 N Harbor Blvd Oxnard CA 93035 PREPARED BY (Print Name): Ali Aleshaikeer
TELEPHONE NUMBER: (805)-673-7228 SIGNATURE: _____
PERIOD: FROM January THROUGH December TITLE: Environmental Science Advisor
Maintain daily logs and submit copies monthly to Ali Aleshaikeer (ali.aleshaikeer@sce.com). Call 909.353.9609 for assistance.

COATING/AEROSOL/SOLVENT USAGE LOG


DATE	NAME, NUMBER, COLOR	COATING CATEGORY	MANUFACTURER	SIZE	QUANTITY USED	ROG CONTENT (lb/gal)	ROG (lb)
COATINGS							
5/1/2023	Loctite LB 8150 - Silver Grade Anti-Seize (<3% VOC)	Industrial Maintenance Coating	Loctite	PT	0.25	0.31291	0.00978
5/19/2023	Thomas & Betts Ocal Coating - Dark Grey Compound	Industrial Maintenance Coating	Thomas & Betts	PT	1	0.24000	0.03000
SOLVENT							
5/1/2023	Isopropyl Alcohol 99%	Cleaning high precision optics	Pharmco-Aeper	GAL	0.125	6.88134	0.86017
AEROSOL SPRAY							
5/1/2023	Rust-Oleum Zinc-Rich Coating - Cold Galvanizing Compound		Rust-Oleum	20 oz	0.8	8.10754	1.01344
5/6/2023	Rust-Oleum Gloss Enamel - Black Enamel		Rust-Oleum	12 oz	2	4.87519	0.91410
5/12/2023	LPS All Purpose Lubricant		LPS	12 oz	1.25	0.02704	0.00317
01/01-12/31	AeroKroil (<25% VOC)		Kano Laboratories	10 oz	1	2.29530	0.17932
01/01-12/31	WD-40		WD-40 Company	11 oz	1	1.64901	0.14171
Total VOC Emissions (Pounds)							3.15169

Attachment 10


Opacity Survey for Turbine Engine and Blackstart Generator Engine

FIGURE 9-1. RECORD OF VISUAL DETERMINATION OF OPACITY

Company: SCE
 Location: M^cGRATH
 Test No.: 1, 2, 3
 Date: 8-22-23
 Type Facility: TURBINE
 Control Device: SCR/CO CAT
 Hours of Observation: 641-659
 Observer: MART M^cCONN
 Observer Certification Date: 8-3-23
 Point of Emissions: STACK
 Observer Affiliation: Mowroast
 Height of Discharge Point: ~ 70'



KEY

X = Observer
 = Sun
 WD = Wind Direction
 O = Stack

CLOCK TIME	Initial			Final
Distance to Discharge	~250'			~250'
Direction from Discharge	E			E
Height of Observation Point	GROUND			GROUND
BACKGROUND DESCRIPTION	BLUE SKY			BLUE SKY
WEATHER CONDITIONS				
Wind Direction	N			N
Wind Speed	<5			<5
Ambient Temperature	65			65
SKY CONDITIONS (clear, over-cast, % clouds, etc.)	CLR			CLR
PLUME DESCRIPTION				
Color	N/A			N/A
Distance Visible	N/A			N/A
OTHER INFORMATION				

SUMMARY OF AVERAGE OPACITY

Set Number	Time	Opacity	
		Sum	Average
1	641-647	0	0
2	647-653	0	0
3	653-659	0	0

Readings ranged from 0 to 0 % opacity.
 The source was not in compliance with 20% at the time the evaluation was made.

Figure 9-2. Observation record.



Company SCE Observer M. McCune
 Location McGrath Type facility GAS TURBINE
 Test Number 1, 2, 3 - VCE Point of emissions STACK


HR.	MIN.	SECONDS				STEAM PLUME (check if applicable)		Comments
		0	15	30	45	ATTACHED	DETACHED	
	0							
	1							
	2							
	3							
	4							
	5							
	6							
	7							
	8							
	9							
	10							
	11							
	12							
	13							
	14							
	15							
	16							
	17							
	18							
	19							
	20							
	21							
	22							
	23							
	24							
	25							
	26							
	27							
	28							
	29							
	30							
	31							
	32							
	33							
	34							
	35							
	36							
	37							
	38							
	39							
	40							
	41							
	42							
	43							
	44							
	45							
	46							
	47							
	48							
	49							
	50							
	51							
	52							
	53							
	54							
	55							
	56							
	57							
	58							
	59							

"CEMS TIME"
PST


1
1
2
1
3
1

FIGURE 9-1. RECORD OF VISUAL DETERMINATION OF OPACITY

Company: SCF
 Location: MCGRAW
 Test No.: 1, 2, 3
 Date: 8-22-23
 Type Facility: BLACK START GENERATOR
 Control Device: N/A
 Hours of Observation: 7:09 - 7:27
 Observer: M. McCune
 Observer Certification Date: 8-3-23
 Point of Emissions: EXHAUST
 Observer Affiliation: MONTROSE
 Height of Discharge Point: EXHAUST ~ 20'



KEY

X = Observer
 = Sun
 WD = Wind Direction
 O = Stack

CLOCK TIME	Initial	Final
Distance to Discharge	~30'	~30'
Direction from Discharge	SE	SE
Height of Observation Point	GROUND	GROUND
BACKGROUND DESCRIPTION	WHITE TANK	WHITE TANK
WEATHER CONDITIONS		
Wind Direction	N	N
Wind Speed	<5	<5
Ambient Temperature	66	67
SKY CONDITIONS (clear, over-cast, % clouds, etc.)	CLR	CLR
PLUME DESCRIPTION		
Color	N/A	N/A
Distance Visible	N/A	N/A
OTHER INFORMATION		

SUMMARY OF AVERAGE OPACITY

Set Number	Time Start - End	Opacity	
		Sum	Average
1	709 - 715	0	0
2	715 - 721	0	0
3	721 - 727	0	0

Readings ranged from 0 to 0 % opacity.
 The source was not in compliance with 20% at the time the evaluation was made.

Figure 9-2. Observation record.



Company SCE Observer M. McCune
 Location MCGRAH Type facility BLACK START GENERATOR
 Test Number 1, 2, 3 Point of emissions EXHAUST

"Cons Time"
 PST
 1
 2
 3

HR.	MIN.	SECONDS				STEAM PLUME (check if applicable)		Comments
		0	15	30	45	ATTACHED	DETACHED	
	0							
	1							
	2							
	3							
	4							
	5							
	6							
	7							
	8							
	9							
	10							
	11							
	12							
	13							
	14							
	15							
	16							
	17							
	18							
	19							
	20							
	21							
	22							
	23							
	24							
	25							
	26							
	27							
	28							
	29							
	30							
	31							
	32							
	33							
	34							
	35							
	36							
	37							
	38							
	39							
	40							
	41							
	42							
	43							
	44							
	45							
	46							
	47							
	48							
	49							
	50							
	51							
	52							
	53							
	54							
	55							
	56							
	57							
	58							
	59							



Air Quality Training Program

Awards This Certificate To

Matt McCune

For Completion Of

MM106 - Visible Emissions Evaluation: Day Certification

In
Long Beach

On
Thursday, August 3, 2023

This certificate expires six months after the evaluation completion date.

A handwritten signature in blue ink that reads 'Heather Quiros'.

Heather Quiros, Acting Chief